

# INTERNATIONAL STANDARD

## NORME INTERNATIONALE



**Medical electrical equipment –**  
**Part 2-76: Particular requirements for the basic safety and essential performance**  
**of low energy ionized gas haemostasis equipment**

**Appareils électromédicaux –** IEC 60601-2-76:2018  
**Partie 2-76: Exigences particulières pour la sécurité de base et les performances**  
**essentiels des appareils d'hémostase à gaz ionisé à faible pouvoir calorifique**



## THIS PUBLICATION IS COPYRIGHT PROTECTED

Copyright © 2023 IEC, Geneva, Switzerland

All rights reserved. Unless otherwise specified, no part of this publication may be reproduced or utilized in any form or by any means, electronic or mechanical, including photocopying and microfilm, without permission in writing from either IEC or IEC's member National Committee in the country of the requester. If you have any questions about IEC copyright or have an enquiry about obtaining additional rights to this publication, please contact the address below or your local IEC member National Committee for further information.

Droits de reproduction réservés. Sauf indication contraire, aucune partie de cette publication ne peut être reproduite ni utilisée sous quelque forme que ce soit et par aucun procédé, électronique ou mécanique, y compris la photocopie et les microfilms, sans l'accord écrit de l'IEC ou du Comité national de l'IEC du pays du demandeur. Si vous avez des questions sur le copyright de l'IEC ou si vous désirez obtenir des droits supplémentaires sur cette publication, utilisez les coordonnées ci-après ou contactez le Comité national de l'IEC de votre pays de résidence.

IEC Secretariat  
3, rue de Varembe  
CH-1211 Geneva 20  
Switzerland

Tel.: +41 22 919 02 11  
[info@iec.ch](mailto:info@iec.ch)  
[www.iec.ch](http://www.iec.ch)

### About the IEC

The International Electrotechnical Commission (IEC) is the leading global organization that prepares and publishes International Standards for all electrical, electronic and related technologies.

### About IEC publications

The technical content of IEC publications is kept under constant review by the IEC. Please make sure that you have the latest edition, a corrigendum or an amendment might have been published.

#### IEC publications search - [webstore.iec.ch/advsearchform](http://webstore.iec.ch/advsearchform)

The advanced search enables to find IEC publications by a variety of criteria (reference number, text, technical committee, ...). It also gives information on projects, replaced and withdrawn publications.

#### IEC Just Published - [webstore.iec.ch/justpublished](http://webstore.iec.ch/justpublished)

Stay up to date on all new IEC publications. Just Published details all new publications released. Available online and once a month by email.

#### IEC Customer Service Centre - [webstore.iec.ch/csc](http://webstore.iec.ch/csc)

If you wish to give us your feedback on this publication or need further assistance, please contact the Customer Service Centre: [sales@iec.ch](mailto:sales@iec.ch).

#### IEC Products & Services Portal - [products.iec.ch](http://products.iec.ch)

Discover our powerful search engine and read freely all the publications previews. With a subscription you will always have access to up to date content tailored to your needs.

#### Electropedia - [www.electropedia.org](http://www.electropedia.org)

The world's leading online dictionary on electrotechnology, containing more than 22 300 terminological entries in English and French, with equivalent terms in 19 additional languages. Also known as the International Electrotechnical Vocabulary (IEV) online.

### A propos de l'IEC

La Commission Electrotechnique Internationale (IEC) est la première organisation mondiale qui élabore et publie des Normes internationales pour tout ce qui a trait à l'électricité, à l'électronique et aux technologies apparentées.

### A propos des publications IEC

Le contenu technique des publications IEC est constamment revu. Veuillez vous assurer que vous possédez l'édition la plus récente, un corrigendum ou amendement peut avoir été publié.

#### Recherche de publications IEC -

[webstore.iec.ch/advsearchform](http://webstore.iec.ch/advsearchform)

La recherche avancée permet de trouver des publications IEC en utilisant différents critères (numéro de référence, texte, comité d'études, ...). Elle donne aussi des informations sur les projets et les publications remplacées ou retirées.

#### IEC Just Published - [webstore.iec.ch/justpublished](http://webstore.iec.ch/justpublished)

Restez informé sur les nouvelles publications IEC. Just Published détaille les nouvelles publications parues. Disponible en ligne et une fois par mois par email.

#### Service Clients - [webstore.iec.ch/csc](http://webstore.iec.ch/csc)

Si vous désirez nous donner des commentaires sur cette publication ou si vous avez des questions contactez-nous: [sales@iec.ch](mailto:sales@iec.ch).

#### IEC Products & Services Portal - [products.iec.ch](http://products.iec.ch)

Découvrez notre puissant moteur de recherche et consultez gratuitement tous les aperçus des publications. Avec un abonnement, vous aurez toujours accès à un contenu à jour adapté à vos besoins.

#### Electropedia - [www.electropedia.org](http://www.electropedia.org)

Le premier dictionnaire d'électrotechnologie en ligne au monde, avec plus de 22 300 articles terminologiques en anglais et en français, ainsi que les termes équivalents dans 19 langues additionnelles. Egalement appelé Vocabulaire Electrotechnique International (IEV) en ligne.



IEC 60601-2-76

Edition 1.1 2023-08  
CONSOLIDATED VERSION

# INTERNATIONAL STANDARD

## NORME INTERNATIONALE



**Medical electrical equipment –**

**Part 2-76: Particular requirements for the basic safety and essential performance  
of low energy ionized gas haemostasis equipment**

**Appareils électromédicaux –** IEC 60601-2-76:2018

**Partie 2-76: Exigences particulières pour la sécurité de base et les performances  
essentielles des appareils d'hémostase à gaz ionisé à faible pouvoir calorifique**

INTERNATIONAL  
ELECTROTECHNICAL  
COMMISSION

COMMISSION  
ELECTROTECHNIQUE  
INTERNATIONALE

ICS 11.040

ISBN 978-2-8322-7502-3

**Warning! Make sure that you obtained this publication from an authorized distributor.**

**Attention! Veuillez vous assurer que vous avez obtenu cette publication via un distributeur agréé.**



## REDLINE VERSION

## VERSION REDLINE



**Medical electrical equipment –**  
**Part 2-76: Particular requirements for the basic safety and essential performance**  
**of low energy ionized gas haemostasis equipment**

**Appareils électromédicaux –** IEC 60601-2-76:2018  
**Partie 2-76: Exigences particulières pour la sécurité de base et les performances**  
**essentiels des appareils d'hémostase à gaz ionisé à faible pouvoir calorifique**

## CONTENTS

FOREWORD.....	3
INTRODUCTION.....	6
INTRODUCTION to Amendment 1 .....	6
201.1 Scope, object and related standards .....	7
201.2 Normative references.....	8
201.3 Terms and definitions.....	9
201.4 * General requirements .....	10
201.5 General requirements for testing ME EQUIPMENT .....	10
201.6 Classification of ME EQUIPMENT and ME SYSTEMS.....	10
201.7 ME EQUIPMENT identification, marking and documents .....	10
201.8 Protection against electrical HAZARDS from ME EQUIPMENT .....	11
201.9 Protection against mechanical HAZARDS of ME EQUIPMENT and ME SYSTEMS .....	19
201.10 Protection against unwanted and excessive radiation HAZARDS .....	19
201.11 Protection against excessive temperatures and other HAZARDS .....	19
201.12 Accuracy of controls and instruments and protection against hazardous outputs .....	20
201.13 HAZARDOUS SITUATIONS and fault conditions for ME EQUIPMENT .....	21
201.14 Programmable electrical medical systems (PEMS).....	21
201.15 Construction of ME EQUIPMENT.....	21
201.16 ME SYSTEMS .....	22
201.17 Electromagnetic compatibility of ME EQUIPMENT and ME SYSTEMS .....	22
Annex AA (informative) Particular guidance and rationale .....	23
Bibliography.....	26
Index of defined terms used in this particular standard.....	27
Figure 201.101 – Measurement of LEAKAGE CURRENT from the PLASMA FLARE.....	12
Figure 201.102 – Measurement of low frequency PATIENT LEAKAGE CURRENT .....	13
Figure 201.103 – Measurement of IONIZED GAS ACCESSORY CABLE LEAKAGE CURRENT.....	14
Figure 201.104 – Test apparatus for anchorages of IONIZED GAS ACCESSORY CABLES .....	19
Figure 201.105 – Measurement of temperature from the PLASMA FLARE .....	20

## INTERNATIONAL ELECTROTECHNICAL COMMISSION

### MEDICAL ELECTRICAL EQUIPMENT –

#### **Part 2-76: Particular requirements for the basic safety and essential performance of low energy ionized gas haemostasis equipment**

#### FOREWORD

- 1) The International Electrotechnical Commission (IEC) is a worldwide organization for standardization comprising all national electrotechnical committees (IEC National Committees). The object of IEC is to promote international co-operation on all questions concerning standardization in the electrical and electronic fields. To this end and in addition to other activities, IEC publishes International Standards, Technical Specifications, Technical Reports, Publicly Available Specifications (PAS) and Guides (hereafter referred to as "IEC Publication(s)"). Their preparation is entrusted to technical committees; any IEC National Committee interested in the subject dealt with may participate in this preparatory work. International, governmental and non-governmental organizations liaising with the IEC also participate in this preparation. IEC collaborates closely with the International Organization for Standardization (ISO) in accordance with conditions determined by agreement between the two organizations.
- 2) The formal decisions or agreements of IEC on technical matters express, as nearly as possible, an international consensus of opinion on the relevant subjects since each technical committee has representation from all interested IEC National Committees.
- 3) IEC Publications have the form of recommendations for international use and are accepted by IEC National Committees in that sense. While all reasonable efforts are made to ensure that the technical content of IEC Publications is accurate, IEC cannot be held responsible for the way in which they are used or for any misinterpretation by any end user.
- 4) In order to promote international uniformity, IEC National Committees undertake to apply IEC Publications transparently to the maximum extent possible in their national and regional publications. Any divergence between any IEC Publication and the corresponding national or regional publication shall be clearly indicated in the latter.
- 5) IEC itself does not provide any attestation of conformity. Independent certification bodies provide conformity assessment services and, in some areas, access to IEC marks of conformity. IEC is not responsible for any services carried out by independent certification bodies.
- 6) All users should ensure that they have the latest edition of this publication.
- 7) No liability shall attach to IEC or its directors, employees, servants or agents including individual experts and members of its technical committees and IEC National Committees for any personal injury, property damage or other damage of any nature whatsoever, whether direct or indirect, or for costs (including legal fees) and expenses arising out of the publication, use of, or reliance upon, this IEC Publication or any other IEC Publications.
- 8) Attention is drawn to the Normative references cited in this publication. Use of the referenced publications is indispensable for the correct application of this publication.
- 9) Attention is drawn to the possibility that some of the elements of this IEC Publication may be the subject of patent rights. IEC shall not be held responsible for identifying any or all such patent rights.

**This consolidated version of the official IEC Standard and its amendment has been prepared for user convenience.**

**IEC 60601-2-76 edition 1.1 contains the first edition (2018-04) [documents 62D/1554/FDIS and 62D/1573/RVD] and its amendment 1 (2023-08) [documents 62D/2047/FDIS and 62D/2076/RVD].**

**In this Redline version, a vertical line in the margin shows where the technical content is modified by amendment 1. Additions are in green text, deletions are in strikethrough red text. A separate Final version with all changes accepted is available in this publication.**

International standard IEC 60601-2-76 has been prepared by IEC subcommittee 62D: Electromedical equipment, of IEC technical committee 62: Electrical equipment in medical practice.

This document has been drafted in accordance with the ISO/IEC Directives, Part 2.

In this standard, the following print types are used:

- requirements and definitions: roman type;
- *test specifications: italic type;*
- informative material appearing outside of tables, such as notes, examples and references: in smaller type. Normative text of tables is also in a smaller type;
- TERMS DEFINED IN CLAUSE 3 OF THE GENERAL STANDARD, IN THIS PARTICULAR STANDARD OR AS NOTED: SMALL CAPITALS.

In referring to the structure of this standard, the term

- “clause” means one of the seventeen numbered divisions within the table of contents, inclusive of all subdivisions (e.g. Clause 7 includes subclauses 7.1, 7.2, etc.);
- “subclause” means a numbered subdivision of a clause (e.g. 7.1, 7.2 and 7.2.1 are all subclauses of Clause 7).

References to clauses within this standard are preceded by the term “Clause” followed by the clause number. References to subclauses within this particular standard are by number only.

In this standard, the conjunctive “or” is used as an “inclusive or” so a statement is true if any combination of the conditions is true.

The verbal forms used in this standard conform to usage described in Clause 7 of the ISO/IEC Directives, Part 2. For the purposes of this standard, the auxiliary verb:

- “shall” means that compliance with a requirement or a test is mandatory for compliance with this standard;
- “should” means that compliance with a requirement or a test is recommended but is not mandatory for compliance with this standard;
- “may” is used to describe a permissible way to achieve compliance with a requirement or test.

An asterisk (\*) as the first character of a title or at the beginning of a paragraph or table title indicates that there is guidance or rationale related to that item in Annex AA.

A list of all parts of the IEC 60601 series, published under the general title *Medical electrical equipment*, can be found on the IEC website.



The committee has decided that the contents of this document and its amendment will remain unchanged until the stability date indicated on the IEC website under [webstore.iec.ch](http://webstore.iec.ch) in the data related to the specific document. At this date, the document will be

- reconfirmed,
- withdrawn,
- replaced by a revised edition, or
- amended.

**IMPORTANT – The '*colour inside*' logo on the cover page of this publication indicates that it contains colours which are considered to be useful for the correct understanding of its contents. Users should therefore print this document using a colour printer.**

iTeh STANDARD PREVIEW  
(standards.iteh.ai)

IEC 60601-2-76:2018

<https://standards.iteh.ai/catalog/standards/sist/e87051db-4f66-46f9-ba10-ba1dffa493/iec-60601-2-76-2018>

## INTRODUCTION

The minimum safety requirements specified in this particular standard are considered to provide for a practical degree of safety in the operation of LOW ENERGY IONIZED GAS HAEMOSTASIS EQUIPMENT.

This particular standard amends and supplements IEC 60601-1:2005 and IEC 60601-1:2005/AMD1:2012 *Medical electrical equipment – Part 1: General requirements for basic safety and essential performance*, hereinafter referred to as the general standard (see 201.1.4).

The requirements are followed by specifications for the relevant tests.

A "Particular guidance and rationale" section giving some explanatory notes, where appropriate, about the more important requirements is included in Annex AA.

Clauses or subclauses for which there are explanatory notes in Annex AA are marked with an asterisk (\*).

It is considered that knowledge of the reasons for these requirements will not only facilitate the proper application of the standard but will, in due course, expedite any revision necessitated by changes in clinical practice or as a result of developments in technology. However, this annex does not form part of the requirements of this document.

## INTRODUCTION to Amendment 1

At the October 2019 meeting of IEC SC 62D in Shanghai, China, the subcommittee discussed the need for administrative/technical changes to most 62D standards after completion of the 29 amendment projects within the IEC 60601-1 series. Those projects were all completed and the amendments published in 2020.

The full list of IEC SC 62D documents that will be amended or revised can be found within the IEC document 62D/1792/DC. The results and comments on the DC can be found within 62D/1808/INF. The review report for this amendment is 62D/1811A/RR.

## MEDICAL ELECTRICAL EQUIPMENT –

### Part 2-76: Particular requirements for the basic safety and essential performance of low energy ionized gas haemostasis equipment

#### 201.1 Scope, object and related standards

Clause 1 of the general standard<sup>1</sup> applies, except as follows:

##### 201.1.1 Scope

*Replacement:*

This part of IEC 60601 applies to the BASIC SAFETY and ESSENTIAL PERFORMANCE of LOW ENERGY IONIZED GAS HAEMOSTASIS EQUIPMENT hereafter referred to as ME EQUIPMENT.

HAZARDS inherent in the intended physiological function of ME EQUIPMENT or ME SYSTEMS within the scope of this document are not covered by specific requirements in this document except in 7.2.13 and 8.4.1 of the general standard.

NOTE See also 4.2 of the general standard.

##### 201.1.2 Object

*Replacement:*

The object of this particular standard is to establish particular requirements for BASIC SAFETY and ESSENTIAL PERFORMANCE of LOW ENERGY IONIZED GAS HAEMOSTASIS EQUIPMENT as defined in 201.3.207.

##### 201.1.3 Collateral standards

*Addition:*

This particular standard refers to those applicable collateral standards that are listed in Clause 2 of the general standard and Clause 201.2 of this particular standard.

##### 201.1.4 Particular standards

*Replacement:*

In the IEC 60601 series, particular standards may modify, replace or delete requirements contained in the general standard and collateral standards as appropriate for the particular ME EQUIPMENT under consideration, and may add other BASIC SAFETY and ESSENTIAL PERFORMANCE requirements.

A requirement of a particular standard takes priority over the general standard.

---

<sup>1</sup> The general standard is IEC 60601-1:2005, IEC 60601-1:2005/AMD1:2012 and IEC 60601-1:2005/AMD2:2020, *Medical electrical equipment – Part 1: General requirements for basic safety and essential performance*.

For brevity, IEC 60601-1:2005, IEC 60601-1:2005/AMD1:2012 and IEC 60601-1:2005/AMD2:2020 are referred to in this particular standard as the general standard. Collateral standards are referred to by their document numbers.

The numbering of clauses and subclauses of this particular standard corresponds to that of the general standard with the prefix “201” (e.g. 201.1 in this document addresses the content of Clause 1 of the general standard) or applicable collateral standard with the prefix “20x” where x is the final digit(s) of the collateral standard document number (e.g. 202.4 in this particular standard addresses the content of Clause 4 of the IEC 60601-1-2 collateral standard, 203.4 in this particular standard addresses the content of Clause 4 of the 60601-1-3 collateral standard, etc.). The changes to the text of the general standard are specified by the use of the following words:

**"Replacement"** means that the clause or subclause of the general standard or applicable collateral standard is replaced completely by the text of this particular standard.

**"Addition"** means that the text of this particular standard is additional to the requirements of the general standard or applicable collateral standard.

**"Amendment"** means that the clause or subclause of the general standard or applicable collateral standard is amended as indicated by the text of this particular standard.

Subclauses, figures or tables which are additional to those of the general standard are numbered starting from 201.101, however because definitions in the general standard are numbered 3.1 through 3.147 additional definitions in this document are numbered beginning from 201.3.201. Additional annexes are lettered AA, BB, etc., and additional items aa), bb), etc.

Subclauses, figures or tables which are additional to those of a collateral standard are numbered starting from 20x, where “x” is the number of the collateral standard, e.g. 202 for IEC 60601-1-2, 203 for IEC 6060-1-3, etc.

The term “this document” is used to make reference to the general standard, any applicable collateral standards and this particular standard taken together.

Where there is no corresponding clause or subclause in this particular standard, the clause or subclause of the general standard or applicable collateral standard, although possibly not relevant, applies without modification; where it is intended that any part of the general standard or applicable collateral standard, although possibly relevant, is not to be applied, a statement to that effect is given in this particular standard.

## 201.2 Normative references

NOTE Informative references are listed in the bibliography.

Clause 2 of the general standard applies, except as follows:

*Addition:*

IEC 60601-1:2005, *Medical electrical equipment – Part 1: General requirements for basic safety and essential performance*  
IEC 60601-1:2005/AMD1:2012  
IEC 60601-1:2005/AMD2:2020

### 201.3 Terms and definitions

For the purposes of this document, the terms and definitions given in IEC 60601-1:2005, IEC 60601-1:2005/AMD1:2012, IEC 60601-1:2005/AMD2:2020 and the following apply.

ISO and IEC maintain terminological databases for use in standardization at the following addresses:

- IEC Electropedia: available at <http://www.electropedia.org/>
- ISO Online browsing platform: available at <http://www.iso.org/obp>

*Replacement of NOTE 1 by the following:*

NOTE 1 Where the terms "voltage" and "current" are used in this document, they mean the RMS values of a voltage or current unless stated otherwise. When a direct voltage and current is measured, the voltage or current is averaged over 1 s with a sampling rate of at least  $10^4$  sample/s. When an alternating, ~~direct~~ or composite voltage or current ~~averaged over 1 s unless stated otherwise~~ is measured, a minimum of one second is measured.

*Addition:*

#### 201.3.201

##### FINGERSWITCH

device generally included with an IONIZED GAS ACCESSORY which, when activated by the OPERATOR, enables IONIZED GAS output to be produced and, when released, disables the IONIZED GAS output

#### 201.3.202

##### \* HAEMOSTASIS

control or prevention of bleeding from capillaries and similar small blood vessels in biological tissues using IONIZED GAS

#### 201.3.203

##### IONIZED GAS

gas composed in part of charged particles and excited atoms or molecules

Note 1 to entry: The frequency used to create the IONIZED GAS is called the ionization frequency.

#### 201.3.204

##### IONIZED GAS ACCESSORY

ACCESSORY intended to direct IONIZED GAS to the PATIENT

#### 201.3.205

##### IONIZED GAS ACCESSORY CABLE

portion of the IONIZED GAS ACCESSORY between the IONIZED GAS CONNECTOR and the part held by the OPERATOR

#### 201.3.206

##### IONIZED GAS CONNECTOR

part of a IONIZED GAS ACCESSORY intended for connection to the output TERMINAL DEVICE of LOW ENERGY IONIZED GAS HAEMOSTASIS EQUIPMENT

#### 201.3.207

##### \* LOW ENERGY IONIZED GAS HAEMOSTASIS EQUIPMENT

ME EQUIPMENT including an IONIZED GAS ACCESSORY and any other associated ACCESSORIES which uses IONIZED GAS for the purpose of HAEMOSTASIS in biological tissue and limits the ionization frequency LEAKAGE CURRENT to 20 mA or less

### **201.3.208**

#### **MAXIMUM IONIZATION VOLTAGE**

maximum peak voltage used to create the IONIZED GAS under NORMAL CONDITION

### **201.3.209**

#### **PLASMA FLARE**

visible IONIZED GAS emitted from the IONIZED GAS ACCESSORY

## **201.4 \* General requirements**

Clause 4 of the general standard applies.

NOTE See Annex AA for additional information.

## **201.5 General requirements for testing ME EQUIPMENT**

Clause 5 of the general standard applies.

## **201.6 Classification of ME EQUIPMENT and ME SYSTEMS**

Clause 6 of the general standard applies.

## **201.7 ME EQUIPMENT identification, marking and documents**

Clause 7 of the general standard applies, except as follows:

### **201.7.9.2.1 \* General**

*IEC 60601-2-76:2018  
[https://standards.iteh.ai/catalog/standards/sist/e87051db-4f66-46f9-ba10-ba1dffa493/iec-](https://standards.iteh.ai/catalog/standards/sist/e87051db-4f66-46f9-ba10-ba1dffa493/iec-60601-2-76-2018-amd1-2023)  
Replace the third dashed item by the following dashed items:*

- advice for the OPERATOR to avoid LOW ENERGY IONIZED GAS HAEMOSTASIS EQUIPMENT output settings that may exceed the rating(s) of the IONIZED GAS ACCESSORY;
- any known contraindication(s) to the use of the ME EQUIPMENT. For LOW ENERGY IONIZED GAS HAEMOSTASIS EQUIPMENT this shall include, at minimum, a statement that the equipment is not intended for use where blood flow is at arterial pressure; and

*Additional subclause:*

#### **201.7.9.2.2.101 Additional warnings**

The instructions for use shall include the following warnings:

- 1) the IONIZED GAS ACCESSORY CABLE should be positioned in such a way that contact with the PATIENT or other leads is avoided;
- 2) for PATIENTS with cardiac pacemakers or other active implants, a possible HAZARD exists because interference with the action of the pacemaker may occur, or the pacemaker may be damaged. In case of doubt, approved qualified advice should be obtained.

#### **201.7.9.2.14 ACCESSORIES, supplementary equipment, used material**

*Addition:*

*The instructions for use shall include:*

Advice for the OPERATOR to avoid using the IONIZED GAS ACCESSORY with LOW ENERGY IONIZED GAS HAEMOSTASIS EQUIPMENT output settings that may exceed the rating(s) of the ACCESSORY.

### 201.7.9.3 Technical description

#### 201.7.9.3.1 General

*Addition:*

- Voltage output data. The MAXIMUM IONIZATION VOLTAGE.

## 201.8 Protection against electrical HAZARDS from ME EQUIPMENT

Clause 8 of the general standard applies, except as follows:

### 201.8.3 \* Classification of APPLIED PARTS

*Addition:*

- aa) IONIZED GAS ACCESSORIES shall be considered APPLIED PARTS, and shall be TYPE BF or TYPE CF because the PLASMA FLARE provides a conductive path between the IONIZED GAS ACCESSORY and the PATIENT.

*Additional subclause:*

#### 201.8.4.101 Excitation current monitoring circuits

Circuits monitoring the flow of current used to generate the IONIZED GAS shall be isolated from earth by at least two MEANS OF PATIENT PROTECTION.

### 201.8.5.1.2 \* MEANS OF PATIENT PROTECTION (MOPP)

*Amendment:*

IEC 60601-2-76:2018

For IONIZED GAS ACCESSORIES including IONIZED GAS ACCESSORY CABLES and IONIZED GAS CONNECTORS, separation between the circuits used to generate the IONIZED GAS and the ENCLOSURE including SIGNAL INPUT PARTS and SIGNAL OUTPUT PARTS need not be subjected to the dielectric strength test of 201.8.8.3.

The CREEPAGE DISTANCES and AIR CLEARANCES of the equipment other than IONIZED GAS ACCESSORIES tested according to 201.8.8.3.101 and that are intended to prevent contact with circuits used to generate the IONIZED GAS shall be as follows:

- For an ionization frequency less than 1 kHz, the general standard applies.
- For an ionization frequency between 1 kHz and 200 kHz, the greater value of the following calculation is chosen:

$$\text{CREEPAGE DISTANCE: } 16 - \frac{13f}{200} \text{ mm/kV, or } 21 - \frac{17,3f}{200} \text{ mm}$$

$$\text{AIR CLEARANCE: } 9,14 - \frac{6,14f}{200} \text{ mm/kV, or } 12,2 - \frac{8,2f}{200} \text{ mm}$$

where

$f$  is the frequency in kHz.

- For an ionization frequency higher than 200 kHz, the CREEPAGE DISTANCE and AIR CLEARANCE is at least 3 mm/kV or 4 mm, whichever is the greater.

The reference voltage shall be the maximum peak voltage.