

INTERNATIONAL STANDARD

NORME INTERNATIONALE

Medical electrical equipment – **STANDARD PREVIEW**
Part 2-46: Particular requirements for the basic safety and essential performance
of operating tables
(standards.iteh.ai)

Appareils électromédicaux – [IEC 60601-2-46:2016](https://standards.iteh.ai/catalog/standards/sist/b2b180e0-14f5-45b2-b191-100000000000/iec-60601-2-46-2016)
Partie 2-46: Exigences particulières pour la sécurité de base et les performances
essentielles des tables d'opération





THIS PUBLICATION IS COPYRIGHT PROTECTED

Copyright © 2016 IEC, Geneva, Switzerland

All rights reserved. Unless otherwise specified, no part of this publication may be reproduced or utilized in any form or by any means, electronic or mechanical, including photocopying and microfilm, without permission in writing from either IEC or IEC's member National Committee in the country of the requester. If you have any questions about IEC copyright or have an enquiry about obtaining additional rights to this publication, please contact the address below or your local IEC member National Committee for further information.

Droits de reproduction réservés. Sauf indication contraire, aucune partie de cette publication ne peut être reproduite ni utilisée sous quelque forme que ce soit et par aucun procédé, électronique ou mécanique, y compris la photocopie et les microfilms, sans l'accord écrit de l'IEC ou du Comité national de l'IEC du pays du demandeur. Si vous avez des questions sur le copyright de l'IEC ou si vous désirez obtenir des droits supplémentaires sur cette publication, utilisez les coordonnées ci-après ou contactez le Comité national de l'IEC de votre pays de résidence.

IEC Central Office
3, rue de Varembe
CH-1211 Geneva 20
Switzerland

Tel.: +41 22 919 02 11
Fax: +41 22 919 03 00
info@iec.ch
www.iec.ch

About the IEC

The International Electrotechnical Commission (IEC) is the leading global organization that prepares and publishes International Standards for all electrical, electronic and related technologies.

About IEC publications

The technical content of IEC publications is kept under constant review by the IEC. Please make sure that you have the latest edition, a corrigenda or an amendment might have been published.

IEC Catalogue - webstore.iec.ch/catalogue

The stand-alone application for consulting the entire bibliographical information on IEC International Standards, Technical Specifications, Technical Reports and other documents. Available for PC, Mac OS, Android Tablets and iPad.

IEC publications search - www.iec.ch/searchpub

The advanced search enables to find IEC publications by a variety of criteria (reference number, text, technical committee,...). It also gives information on projects, replaced and withdrawn publications.

IEC Just Published - webstore.iec.ch/justpublished

Stay up to date on all new IEC publications. Just Published details all new publications released. Available online and also once a month by email.

Electropedia - www.electropedia.org

The world's leading online dictionary of electronic and electrical terms containing 20 000 terms and definitions in English and French, with equivalent terms in 15 additional languages. Also known as the International Electrotechnical Vocabulary (IEV) online.

IEC Glossary - std.iec.ch/glossary

65 000 electrotechnical terminology entries in English and French extracted from the Terms and Definitions clause of IEC publications issued since 2002. Some entries have been collected from earlier publications of IEC TC 37, 77, 86 and CISPR.

IEC Customer Service Centre - webstore.iec.ch/csc

If you wish to give us your feedback on this publication or need further assistance, please contact the Customer Service Centre: csc@iec.ch.

A propos de l'IEC

La Commission Electrotechnique Internationale (IEC) est la première organisation mondiale qui élabore et publie des Normes internationales pour tout ce qui a trait à l'électricité, à l'électronique et aux technologies apparentées.

A propos des publications IEC

Le contenu technique des publications IEC est constamment revu. Veuillez vous assurer que vous possédez l'édition la plus récente, un corrigendum ou amendement peut avoir été publié.

Catalogue IEC - webstore.iec.ch/catalogue

Application autonome pour consulter tous les renseignements bibliographiques sur les Normes internationales, Spécifications techniques, Rapports techniques et autres documents de l'IEC. Disponible pour PC, Mac OS, tablettes Android et iPad.

Recherche de publications IEC - www.iec.ch/searchpub

La recherche avancée permet de trouver des publications IEC en utilisant différents critères (numéro de référence, texte, comité d'études,...). Elle donne aussi des informations sur les projets et les publications remplacées ou retirées.

IEC Just Published - webstore.iec.ch/justpublished

Restez informé sur les nouvelles publications IEC. Just Published détaille les nouvelles publications parues. Disponible en ligne et aussi une fois par mois par email.

Electropedia - www.electropedia.org

Le premier dictionnaire en ligne de termes électroniques et électriques. Il contient 20 000 termes et définitions en anglais et en français, ainsi que les termes équivalents dans 15 langues additionnelles. Egalement appelé Vocabulaire Electrotechnique International (IEV) en ligne.

Glossaire IEC - std.iec.ch/glossary

65 000 entrées terminologiques électrotechniques, en anglais et en français, extraites des articles Termes et Définitions des publications IEC parues depuis 2002. Plus certaines entrées antérieures extraites des publications des CE 37, 77, 86 et CISPR de l'IEC.

Service Clients - webstore.iec.ch/csc

Si vous désirez nous donner des commentaires sur cette publication ou si vous avez des questions contactez-nous: csc@iec.ch.

INTERNATIONAL STANDARD

NORME INTERNATIONALE

Medical electrical equipment –
Part 2-46: Particular requirements for the basic safety and essential performance
of operating tables

Appareils électromédicaux –
Partie 2-46: Exigences particulières pour la sécurité de base et les performances
essentiels des tables d'opération

INTERNATIONAL
ELECTROTECHNICAL
COMMISSION

COMMISSION
ELECTROTECHNIQUE
INTERNATIONALE

ICS 11.140

ISBN 978-2-8322-3565-2

Warning! Make sure that you obtained this publication from an authorized distributor.
Attention! Veuillez vous assurer que vous avez obtenu cette publication via un distributeur agréé.

CONTENTS

FOREWORD.....	3
INTRODUCTION.....	5
201.1 Scope, object and related standards.....	6
201.2 Normative references.....	8
201.3 Terms and definitions.....	8
201.4 General requirements.....	9
201.5 General requirements for testing OPERATING TABLES.....	9
201.6 Classification of OPERATING TABLES and ME SYSTEMS.....	9
201.7 OPERATING TABLES identification, marking and documents.....	9
201.8 Protection against electrical HAZARDS from OPERATING TABLES.....	10
201.9 Protection against MECHANICAL HAZARDS of OPERATING TABLES and ME SYSTEMS.....	10
201.10 Protection against unwanted and excessive radiation HAZARDS.....	13
201.11 Protection against excessive temperatures and other HAZARDS.....	13
201.12 Accuracy of controls and instruments and protection against hazardous outputs.....	13
201.13 HAZARDOUS SITUATIONS and fault conditions for OPERATING TABLES.....	14
201.14 PROGRAMMABLE ELECTRICAL MEDICAL SYSTEMS (PEMS).....	14
201.15 Construction of OPERATING TABLES.....	14
201.16 ME SYSTEMS.....	14
201.17 Electromagnetic compatibility of OPERATING TABLES and ME SYSTEMS.....	14
202 Electromagnetic disturbances – Requirements and tests.....	14
203 Radiation protection in diagnostic X-ray equipment.....	18
Annexes.....	19
Annex G (normative) Protection against HAZARDS of ignition of flammable anaesthetic mixtures.....	20
Annex AA (informative) Particular guidance and rationale.....	21
Bibliography.....	23
Index of defined terms used in this particular standard.....	24
Figure 202.101 – ENCLOSURE ad hoc test.....	16
Figure 202.102 – POWER SUPPLY CORD ad hoc test.....	17
Figure 202.103 – ACCESSORY cable ad hoc test.....	17
Figure AA.1 – Recommended distribution of mass in excess of 135 kg and examples of application.....	21
Table 201.101 – Determination of TENSILE SAFETY FACTOR.....	12
Table AA.1 – Recommended distribution of mass in excess of 135 kg and examples of application.....	22

INTERNATIONAL ELECTROTECHNICAL COMMISSION

MEDICAL ELECTRICAL EQUIPMENT –

**Part 2-46: Particular requirements for the basic safety
and essential performance of operating tables**

FOREWORD

- 1) The International Electrotechnical Commission (IEC) is a worldwide organization for standardization comprising all national electrotechnical committees (IEC National Committees). The object of IEC is to promote international co-operation on all questions concerning standardization in the electrical and electronic fields. To this end and in addition to other activities, IEC publishes International Standards, Technical Specifications, Technical Reports, Publicly Available Specifications (PAS) and Guides (hereafter referred to as “IEC Publication(s)”). Their preparation is entrusted to technical committees; any IEC National Committee interested in the subject dealt with may participate in this preparatory work. International, governmental and non-governmental organizations liaising with the IEC also participate in this preparation. IEC collaborates closely with the International Organization for Standardization (ISO) in accordance with conditions determined by agreement between the two organizations.
- 2) The formal decisions or agreements of IEC on technical matters express, as nearly as possible, an international consensus of opinion on the relevant subjects since each technical committee has representation from all interested IEC National Committees.
- 3) IEC Publications have the form of recommendations for international use and are accepted by IEC National Committees in that sense. While all reasonable efforts are made to ensure that the technical content of IEC Publications is accurate, IEC cannot be held responsible for the way in which they are used or for any misinterpretation by any end user.
- 4) In order to promote international uniformity, IEC National Committees undertake to apply IEC Publications transparently to the maximum extent possible in their national and regional publications. Any divergence between any IEC Publication and the corresponding national or regional publication shall be clearly indicated in the latter.
<https://standards.iteh.ai/catalog/standards/sist/b2b180e0-14f5-45b2-b191->
- 5) IEC itself does not provide any attestation of conformity. Independent certification bodies provide conformity assessment services and, in some areas, access to IEC marks of conformity. IEC is not responsible for any services carried out by independent certification bodies.
- 6) All users should ensure that they have the latest edition of this publication.
- 7) No liability shall attach to IEC or its directors, employees, servants or agents including individual experts and members of its technical committees and IEC National Committees for any personal injury, property damage or other damage of any nature whatsoever, whether direct or indirect, or for costs (including legal fees) and expenses arising out of the publication, use of, or reliance upon, this IEC Publication or any other IEC Publications.
- 8) Attention is drawn to the Normative references cited in this publication. Use of the referenced publications is indispensable for the correct application of this publication.
- 9) Attention is drawn to the possibility that some of the elements of this IEC Publication may be the subject of patent rights. IEC shall not be held responsible for identifying any or all such patent rights.

International standard IEC 60601-2-46 has been prepared by IEC subcommittee 62D Electromedical equipment, of IEC technical committee 62: Electrical equipment in medical practice.

This third edition cancels and replaces the second edition published in 2010 and constitutes a technical revision. This edition of IEC 60601-2-46 was revised to align structurally with the 2005 edition of IEC 60601-1 and with IEC 60601-1:2005/AMD1:2012.

The text of this standard is based on the following documents:

FDIS	Report on voting
62D/1365/FDIS	62D/1371/RVD

Full information on the voting for the approval of this standard can be found in the report on voting indicated in the above table.

This publication has been drafted in accordance with the ISO/IEC Directives, Part 2.

In this standard, the following print types are used:

- Requirements and definitions: roman type.
- *Test specifications: italic type.*
- Informative material appearing outside of tables, such as notes, examples and references: in smaller type. Normative text of tables is also in a smaller type.
- TERMS DEFINED IN CLAUSE 3 OF THE GENERAL STANDARD, IN THIS PARTICULAR STANDARD OR AS NOTED: SMALL CAPITALS.

In referring to the structure of this standard, the term

- “clause” means one of the numbered divisions within the table of contents, inclusive of all subdivisions (e.g. Clause 7 includes subclauses 7.1, 7.2, etc.);
- “subclause” means a numbered subdivision of a clause (e.g. 7.1, 7.2 and 7.2.1 are all subclauses of Clause 7).

References to clauses within this standard are preceded by the term “Clause” followed by the clause number. References to subclauses within this particular standard are by number only.

In this standard, the conjunctive “or” is used as an “inclusive or” so a statement is true if any combination of the conditions is true.

The verbal forms used in this standard conform to usage described in Annex H of the ISO/IEC Directives, Part 2. For the purposes of this standard, the auxiliary verb:

- “shall” means that compliance with a requirement or a test is mandatory for compliance with this standard;
- “should” means that compliance with a requirement or a test is recommended but is not mandatory for compliance with this standard;
- “may” is used to describe a permissible way to achieve compliance with a requirement or test.

An asterisk (*) as the first character of a title or at the beginning of a paragraph or table title indicates that there is guidance or rationale related to that item in Annex AA.

A list of all parts of the IEC 60601 series, published under the general title *Medical electrical equipment*, can be found on the IEC website.

The committee has decided that the contents of this publication will remain unchanged until the stability date indicated on the IEC website under "http://webstore.iec.ch" in the data related to the specific publication. At this date, the publication will be

- reconfirmed,
- withdrawn,
- replaced by a revised edition, or
- amended.

INTRODUCTION

This particular standard concerns the BASIC SAFETY and ESSENTIAL PERFORMANCE of OPERATING TABLES. It amends and supplements IEC 60601-1 (third edition, 2005) and its Amendment 1 (IEC 60601-1:2005/AMD1:2012), hereinafter referred to as the general standard.

The aim of this third edition is to bring this particular standard up to date with reference to the third edition of the general standard through reformatting and technical changes.

The requirements of this particular standard take priority over those of the general standard.

A “General guidance and rationale” for the more important requirements of this particular standard is included in Annex AA. It is considered that knowledge of the reasons for these requirements will not only facilitate the proper application of the standard but will, in due course, expedite any revision necessitated by changes in clinical practice or as a result of developments in technology. However, Annex AA does not form part of the requirements of this Standard.

iTeh STANDARD PREVIEW **(standards.iteh.ai)**

[IEC 60601-2-46:2016](https://standards.iteh.ai/catalog/standards/sist/b2b180e0-14f5-45b2-b191-2dcd24c961aa/iec-60601-2-46-2016)

<https://standards.iteh.ai/catalog/standards/sist/b2b180e0-14f5-45b2-b191-2dcd24c961aa/iec-60601-2-46-2016>

MEDICAL ELECTRICAL EQUIPMENT –

Part 2-46: Particular requirements for the basic safety and essential performance of operating tables

201.1 Scope, object and related standards

Clause 1 of the general standard¹⁾ applies, except as follows:

201.1.1 Scope

Replacement:

This particular standard specifies safety requirements for OPERATING TABLES, whether or not having electrical parts, including TRANSPORTERS, used for the transportation of the OPERATING TABLE top to or from the base or pedestal of an OPERATING TABLE with detachable OPERATING TABLE top.

NOTE See also 4.2 of the General Standard.

This particular standard does not apply to

- dental PATIENT chairs;
- examination chairs and couches;
- PATIENT-supporting systems of diagnostic and therapeutic devices; (see IEC 60601-2-43)
- OPERATING TABLE heating blankets; (see IEC 80601-2-35)
- PATIENT transfer equipment;
- delivery tables and beds;
- medical beds; (see IEC 60601-2-52)
- field tables.

If OPERATING TABLES will be used in combination with diagnostic and/or therapeutic devices the relevant requirements of each related particular standard are also applicable.

201.1.2 Object

Replacement:

The object of this particular standard is to establish particular BASIC SAFETY and ESSENTIAL PERFORMANCE requirements for OPERATING TABLES as defined in 201.3.201.

201.1.3 Collateral standards

Addition:

This particular standard refers to those applicable collateral standards that are listed in Clause 2 of the general standard and Clause 201.2 of this particular standard.

¹⁾ The general standard is IEC 60601-1:2005/AMD1:2012, *Medical electrical equipment – Part 1: General requirements for basic safety and essential performance.*

IEC 60601-1-2 and IEC 60601-1-3 apply as modified in Clauses 202 and 203 respectively. IEC 60601-1-8, IEC 60601-1-10, IEC 60601-1-11 and IEC 60601-1-12 do not apply. All other published collateral standards in the IEC 60601-1 series apply as published.

201.1.4 Particular standards

Replacement:

In the IEC 60601 series, particular standards may modify, replace or delete requirements contained in the general standard and collateral standards as appropriate for the particular ME EQUIPMENT under consideration, and may add other BASIC SAFETY and ESSENTIAL PERFORMANCE requirements.

A requirement of a particular standard takes priority over the general standard.

For brevity, IEC 60601-1 is referred to in this particular standard as the general standard. Collateral standards are referred to by their document number.

The numbering of clauses and subclauses of this particular standard corresponds to that of the general standard with the prefix "201" (e.g. 201.1 in this standard addresses the content of Clause 1 of the general standard) or applicable collateral standard with the prefix "20x" where x is the final digit(s) of the collateral standard document number (e.g. 202.4 in this particular standard addresses the content of Clause 4 of the IEC 60601-1-2 collateral standard, 203.4 in this particular standard addresses the content of Clause 4 of the IEC 60601-1-3 collateral standard, etc.). The changes to the text of the general standard are specified by the use of the following words:

"Replacement" means that the clause or subclause of the general standard or applicable collateral standard is replaced completely by the text of this particular standard.

"Addition" means that the text of this particular standard is additional to the requirements of the general standard or applicable collateral standard.

"Amendment" means that the clause or subclause of the general standard or applicable collateral standard is amended as indicated by the text of this particular standard.

Subclauses, figures or tables which are additional to those of the general standard are numbered starting from 201.101. However, due to the fact that definitions in the general standard are numbered 3.1 through 3.147, additional definitions in this standard are numbered beginning from 201.3.201. Additional annexes are lettered AA, BB, etc., and additional items aa), bb), etc.

Subclauses, figures or tables which are additional to those of a collateral standard are numbered starting from 20x, where "x" is the number of the collateral standard, e.g. 202 for IEC 60601-1-2, 203 for IEC 60601-1-3, etc.

The term "this standard" is used to make reference to the general standard, any applicable collateral standards and this particular standard taken together.

Where there is no corresponding clause or subclause in this particular standard, the clause or subclause of the general standard or applicable collateral standard, although possibly not relevant, applies without modification; where it is intended that any part of the general standard or applicable collateral standard, although possibly relevant, is not to be applied, a statement to that effect is given in this particular standard.

201.2 Normative references

Clause 2 of the general standard applies, except as follows:

Replacement:

IEC 60601-1-2:2014, *Medical electrical equipment – Part 1-2: General requirements for basic safety and essential performance – Collateral standard: Electromagnetic disturbances – Requirements and tests*

IEC 60601-1-3:2008, *Medical electrical equipment – Part 1-3: General requirements for basic safety and essential performance – Collateral standard: Radiation protection in diagnostic X-ray equipment*

IEC 60601-1-3:2008/AMD1:2013²

Addition:

IEC 60601-2-2, *Medical electrical equipment – Part 2-2: Particular requirements for the basic safety and essential performance of high frequency surgical equipment and high frequency surgical accessories*

IEC 60601-2-43, *Medical electrical equipment – Part 2-43: Particular requirements for the basic safety and essential performance of X-ray equipment for interventional procedures*

iTeh STANDARD PREVIEW
(standards.iteh.ai)

201.3 Terms and definitions

For the purposes of this document, the terms and definitions given in IEC 60601-1 apply, except as follows: <https://standards.iteh.ai/catalog/standards/sist/b2b180e0-14f5-45b2-b191-2dcd24c961aa/iec-60601-2-46-2016>

NOTE An index of defined terms is found beginning on page 24.

Addition:

201.3.201

MOBILE OPERATING TABLE

OPERATING TABLE intended to be relocated from one location to another while supported by its own wheels or equivalent means

201.3.202

NORMAL POSITION

position of the OPERATING TABLE top with all sections set in the horizontal position

201.3.203

OPERATING TABLE

device with the INTENDED USE of supporting and positioning a PATIENT during surgical procedures for not more than 24 h

Note 1 to entry: This includes pre- and post-operative phases in general, surgical/medical procedures under medical supervision.

² There exists a consolidated edition 2.1, which includes IEC 60601-1-3:2008 and its Amendment 1 (2013).

201.3.204**TRANSPORTER**

device intended for the transportation of an OPERATING TABLE top to or from the base or pedestal of an OPERATING TABLE, or the transportation of the OPERATING TABLE top complete with the base

Note 1 to entry: This definition does not include devices intended to simplify the transport of the PATIENT from one location to another without the transfer of parts associated with an OPERATING TABLE.

Note 2 to entry: The transportation can be done with or without a PATIENT in place.

201.3.205**TRENDELENBURG POSITION**

a supine PATIENT position where the body is in a single plane, with that plane inclined so that the head is lower than the pelvis

201.4 General requirements

Clause 4 of the general standard applies, except as follows.

201.4.3 Essential performance

Addition:

Besides the definition of the MANUFACTURER, the following ESSENTIAL PERFORMANCE is required from OPERATING TABLES:

- supporting a PATIENT without unwanted movement in a SINGLE FAULT CONDITION.

201.4.7 SINGLE FAULT CONDITION FOR OPERATING TABLES

Addition:

Additional SINGLE FAULT CONDITIONS to be regarded with OPERATING TABLES:

- flaw (impairment) in the transmission of commands from/to input devices.

The MANUFACTURER should provide means, where practical, to ensure that in a SINGLE FAULT CONDITION the PATIENT support platform of the OPERATING TABLE can return to a position for emergency treatment.

NOTE 101 Examples of positions for emergency treatment are TRENDELENBURG or positions for cardiopulmonary resuscitation (CPR), emergency back flattening.

201.5 General requirements for testing OPERATING TABLES

Clause 5 of the general standard applies.

201.6 Classification of OPERATING TABLES and ME SYSTEMS

Clause 6 of the general standard applies.

201.7 OPERATING TABLES identification, marking and documents

Clause 7 of the general standard applies, except as follows:

201.7.2 Marking on the outside of OPERATING TABLES or OPERATING TABLES parts

201.7.2.10 Applied parts

Amendment:

The APPLIED PART marking symbol according to Table D.1 (symbol 19, 20 or 21) shall be located in a prominent place. Compliance is checked by inspection.

201.7.2.21 Mass of MOBILE OPERATING TABLES

This subclause of the general standard does not apply.

201.7.9.2 Instructions for use

201.7.9.2.1 General

Addition:

Instructions for use shall include information, regarding potential HAZARDS related to cardiac defibrillators and cardiac defibrillator-monitors.

NOTE 101 Potential HAZARDS which have to be considered include but are not limited to: PATIENT burns, explosion HAZARDS or electrical shock of the PATIENT or OPERATOR.

201.8 Protection against electrical HAZARDS from OPERATING TABLES

Clause 8 of the general standard applies, except as follows:

201.8.6.7 POTENTIAL EQUALIZATION CONDUCTOR

Addition:

Where POTENTIAL EQUALIZATION is required, the APPLIED PARTS of OPERATING TABLES with ACCESSIBLE PARTS which are not PROTECTIVELY EARTHED shall be provided with a POTENTIAL EQUALIZATION terminal.

For OPERATING TABLES with a POTENTIAL EQUALIZATION terminal the impedance between the potential equalization terminal and any ACCESSIBLE PART shall not exceed 200 mΩ,

Compliance is checked by using the test method of 8.6.4 of the general standard.

201.9 Protection against MECHANICAL HAZARDS of OPERATING TABLES and ME SYSTEMS

Clause 9 of the general standard applies, except as follows:

201.9.2.3 Other MECHANICAL HAZARDS associated with moving parts

201.9.2.3.1 *Unintended movement

Addition:

Wireless remote control devices of OPERATING TABLES shall be clearly assigned by internal means to the individual items of OPERATING TABLES.

Compliance is checked by inspection.

201.9.4 Instability hazards

201.9.4.2.2 *Instability excluding transport position

Item a)

Addition:

OPERATING TABLES shall be subjected to SAFE WORKING LOAD.

NOTE See Figure AA.1 and Table AA.1 for guidance regarding mass distribution.

Additional requirement:

OPERATING TABLES with transferable OPERATING TABLE tops shall be designed and manufactured so as to minimize the RISK of physical injuries and of accidental separation of the OPERATING TABLE tops when being transferred.

Specifications concerning OPERATING TABLE top transfer operations shall indicate in the instructions for use the safety elements inherent in the transfer operation.

Compliance is checked by inspection and the following tests:

Having transferred the OPERATING TABLE top to the TRANSPORTER, the stability in NORMAL USE test of 9.4.2.2 of the general standard shall be carried out. The OPERATING TABLE top shall not disengage from the TRANSPORTER.

The test is then repeated with the OPERATING TABLE top being placed on the base and the stability test is carried out on the base immediately after transfer.

201.9.4.2.4.3 *Movement over a threshold

Addition:

If MOBILE OPERATING TABLES *and* TRANSPORTERS are not able to negotiate such obstacles safely, the MANUFACTURER shall include a warning in the instructions for use or determine which threshold can be negotiated safely and inform the OPERATOR accordingly.

201.9.4.3.1 Instability in transport position

Replacement of items b) and c) of the test procedure:

The MOBILE OPERATING TABLE or TRANSPORTER is placed with the SAFE WORKING LOAD in place, and the locking device (e.g. brakes) activated, on a plane covered with 2 mm to 4 mm thick vinyl flooring material and inclined at 6° from the horizontal plane on a concrete floor. Following initial elastic movement, initial creepage, and initial pivoting of castors, there shall be no movement of the MOBILE OPERATING TABLE or TRANSPORTER greater than 50 mm (in relation to the inclined plane). Any initial movement shall not result in an unacceptable RISK, taking into account the NORMAL USE of the MOBILE OPERATING TABLE or TRANSPORTER.

NOTE See Figure AA.1 and Table AA.1 for guidance regarding mass distribution.

201.9.8 MECHANICAL HAZARDS associated with support systems

201.9.8.1 General

Replacement of first dash:

- The construction of the support, suspension or actuation system shall be designed based upon Table 201.101 and the SAFE WORKING LOAD.

201.9.8.2 *TENSILE SAFETY FACTOR

Replacement:

Support systems shall maintain structural integrity during the EXPECTED SERVICE LIFE of the OPERATING TABLE or TRANSPORTER. TENSILE SAFETY FACTORS shall not be less than those shown in Table 201.101 unless an alternative method demonstrates structural integrity throughout the EXPECTED SERVICE LIFE of the OPERATING TABLE or TRANSPORTER.

Due to the fact that it is not always possible to determine in general whether a specific component or construction is impaired by wear, the decision shall be based on experience, tests and/or RISK MANAGEMENT and shall be documented accordingly. However, the MANUFACTURER is responsible for choosing the adequate TENSILE SAFETY FACTOR.

The OPERATING TABLE or TRANSPORTER shall be tested:

- with the SAFE WORKING LOAD (required PATIENT mass according to Figure AA.1 and Table AA.1) and a TENSILE SAFETY FACTOR according to Table 201.101:

Table 201.101 – Determination of TENSILE SAFETY FACTOR

Situation		Minimum TENSILE SAFETY FACTOR
No.	System part	Elongation
1	Support system not impaired by wear	Material having a specific elongation at break equal to or greater than 5 %
2	Support system not impaired by wear	Material having a specific elongation at break of less than 5 %
3	Support system impaired by wear	Material having a specific elongation at break equal to or greater than 5 %
4	Support system impaired by wear	Material having a specific elongation at break of less than 5 %

The material tensile strength and all external forces to be expected are quantifiable and known accurately.

Compliance with 201.9.8.1 and 201.9.8.2 is checked by inspection of the OPERATING TABLE or TRANSPORTER, the RISK MANAGEMENT FILE, the specifications of materials used and the processing specifications for these materials.

When test results are part of relevant information, testing consists of gradually applying a test load to the support assembly under test equal to the SAFE WORKING LOAD times the required TENSILE SAFETY FACTOR. The support assembly under test is to be in equilibrium after 1 min, or otherwise not result in an unacceptable RISK.

NOTE The 1 min time period might need to be longer for materials which might have creep type problems, such as plastics or other non-metallic materials.

201.9.8.3.2 *Static forces due to loading from persons

Replacement of item b):

- b) OPERATING TABLES and TRANSPORTERS shall be designed so that failure or permanent deformation shall not occur when subjected to 2,2 times SAFE WORKING LOAD.

NOTE See Figure AA.1 and Table AA.1 for guidance regarding mass distribution.

Compliance is checked by the following test:

- 1) *In NORMAL POSITION and at maximum height the OPERATING TABLES shall be statically loaded with 2,2 times SAFE WORKING LOAD. The deformation after 5 min is recorded. The OPERATING TABLES shall not be operated or moved during this part of the test.*

- 2) *The load is removed and replaced as soon as practical with SAFE WORKING LOAD.*
- 3) *After waiting 5 min in NORMAL POSITION and at maximum height the OPERATING TABLES shall be statically loaded with 2,2 times SAFE WORKING LOAD. The deformation after 5 min is recorded.*

The deflections are compared to the values measured under 1) and shall be within $\pm 2,5$ mm of the original readings.

- 4) *The load is removed and replaced with SAFE WORKING LOAD and the OPERATING TABLES shall operate over the full range of movements. The deformation/deflection shall be measured at the end of the head- and leg-section of the OPERATING TABLE. For accessories the measuring point shall be determined according the intended use.*

201.9.8.3.3 *Dynamic forces due to loading from persons

This subclause of the general standard does not apply.

201.10 Protection against unwanted and excessive radiation HAZARDS

Clause 10 of the general standard applies.

201.11 Protection against excessive temperatures and other HAZARDS

Clause 11 of the general standard applies, except as follows:

201.11.6 Overflow, spillage, leakage, ingress of water or particulate matter, cleaning, disinfection, sterilization and compatibility with substances used with the OPERATING TABLES

201.11.6.5 Ingress of water or particulate matter into OPERATING TABLES and ME SYSTEMS

Addition:

OPERATING TABLES shall be at least IPX4.

201.11.8 Interruption of the power supply/SUPPLY MAINS to OPERATING TABLES

Addition:

In the event of interruption of the SUPPLY MAINS, whether or not the SUPPLY MAINS is restored, the height and configuration of the OPERATING TABLE top shall not alter. Movement into NORMAL POSITION and/or TRENDLENBURG POSITION shall remain possible.

Compliance is checked as follows:

- a) *By test after interruption of the SUPPLY MAINS with the OPERATING TABLE top in any position, other than the NORMAL POSITION, midway between its maximum and minimum heights, subjected to SAFE WORKING LOAD with mass distributed according to Figure AA.1 and Table AA.1. Movement into and out of the NORMAL POSITION shall be obtainable using the methods described by the MANUFACTURER.*
- b) *By observation after restoration of the SUPPLY MAINS.*

201.12 Accuracy of controls and instruments and protection against hazardous outputs

Clause 12 of the general standard applies.