

# INTERNATIONAL STANDARD

# NORME INTERNATIONALE

AMENDMENT 1  
AMENDEMENT 1

Medical electrical equipment –  
Part 2-21: Particular requirements for the basic safety and essential performance  
of infant radiant warmers  
[\(standards.iteh.ai\)](https://standards.iteh.ai/)

Appareils électromédicaux –  
Partie 2-21: Exigences particulières pour la sécurité de base et les performances  
essentielles des incubateurs radiants pour nouveau-nés  
[IEC 60601-2-21:2009/AMD1:2016](https://standards.iteh.ai/catalog/standards/sist/8467265e-1e59-45dc-8e25-342111111111/iec-60601-2-21:2009/amd1:2016)





**THIS PUBLICATION IS COPYRIGHT PROTECTED**  
**Copyright © 2016 IEC, Geneva, Switzerland**

All rights reserved. Unless otherwise specified, no part of this publication may be reproduced or utilized in any form or by any means, electronic or mechanical, including photocopying and microfilm, without permission in writing from either IEC or IEC's member National Committee in the country of the requester. If you have any questions about IEC copyright or have an enquiry about obtaining additional rights to this publication, please contact the address below or your local IEC member National Committee for further information.

Droits de reproduction réservés. Sauf indication contraire, aucune partie de cette publication ne peut être reproduite ni utilisée sous quelque forme que ce soit et par aucun procédé, électronique ou mécanique, y compris la photocopie et les microfilms, sans l'accord écrit de l'IEC ou du Comité national de l'IEC du pays du demandeur. Si vous avez des questions sur le copyright de l'IEC ou si vous désirez obtenir des droits supplémentaires sur cette publication, utilisez les coordonnées ci-après ou contactez le Comité national de l'IEC de votre pays de résidence.

IEC Central Office  
3, rue de Varembe  
CH-1211 Geneva 20  
Switzerland

Tel.: +41 22 919 02 11  
Fax: +41 22 919 03 00  
[info@iec.ch](mailto:info@iec.ch)  
[www.iec.ch](http://www.iec.ch)

#### About the IEC

The International Electrotechnical Commission (IEC) is the leading global organization that prepares and publishes International Standards for all electrical, electronic and related technologies.

#### About IEC publications

The technical content of IEC publications is kept under constant review by the IEC. Please make sure that you have the latest edition, a corrigenda or an amendment might have been published.

#### IEC Catalogue - [webstore.iec.ch/catalogue](http://webstore.iec.ch/catalogue)

The stand-alone application for consulting the entire bibliographical information on IEC International Standards, Technical Specifications, Technical Reports and other documents. Available for PC, Mac OS, Android Tablets and iPad.

#### IEC publications search - [www.iec.ch/searchpub](http://www.iec.ch/searchpub)

The advanced search enables to find IEC publications by a variety of criteria (reference number, text, technical committee,...). It also gives information on projects, replaced and withdrawn publications.

#### IEC Just Published - [webstore.iec.ch/justpublished](http://webstore.iec.ch/justpublished)

Stay up to date on all new IEC publications. Just Published details all new publications released. Available online and also once a month by email.

#### Electropedia - [www.electropedia.org](http://www.electropedia.org)

The world's leading online dictionary of electronic and electrical terms containing 20 000 terms and definitions in English and French, with equivalent terms in 15 additional languages. Also known as the International Electrotechnical Vocabulary (IEV) online.

#### IEC Glossary - [std.iec.ch/glossary](http://std.iec.ch/glossary)

65 000 electrotechnical terminology entries in English and French extracted from the Terms and Definitions clause of IEC publications issued since 2002. Some entries have been collected from earlier publications of IEC TC 37, 77, 86 and CISPR.

#### IEC Customer Service Centre - [webstore.iec.ch/csc](http://webstore.iec.ch/csc)

If you wish to give us your feedback on this publication or need further assistance, please contact the Customer Service Centre: [csc@iec.ch](mailto:csc@iec.ch).

---

#### A propos de l'IEC

La Commission Electrotechnique Internationale (IEC) est la première organisation mondiale qui élabore et publie des Normes internationales pour tout ce qui a trait à l'électricité, à l'électronique et aux technologies apparentées.

#### A propos des publications IEC

Le contenu technique des publications IEC est constamment revu. Veuillez vous assurer que vous possédez l'édition la plus récente, un corrigendum ou amendement peut avoir été publié.

#### Catalogue IEC - [webstore.iec.ch/catalogue](http://webstore.iec.ch/catalogue)

Application autonome pour consulter tous les renseignements bibliographiques sur les Normes internationales, Spécifications techniques, Rapports techniques et autres documents de l'IEC. Disponible pour PC, Mac OS, tablettes Android et iPad.

#### Recherche de publications IEC - [www.iec.ch/searchpub](http://www.iec.ch/searchpub)

La recherche avancée permet de trouver des publications IEC en utilisant différents critères (numéro de référence, texte, comité d'études,...). Elle donne aussi des informations sur les projets et les publications remplacées ou retirées.

#### IEC Just Published - [webstore.iec.ch/justpublished](http://webstore.iec.ch/justpublished)

Restez informé sur les nouvelles publications IEC. Just Published détaille les nouvelles publications parues. Disponible en ligne et aussi une fois par mois par email.

#### Electropedia - [www.electropedia.org](http://www.electropedia.org)

Le premier dictionnaire en ligne de termes électroniques et électriques. Il contient 20 000 termes et définitions en anglais et en français, ainsi que les termes équivalents dans 15 langues additionnelles. Egalement appelé Vocabulaire Electrotechnique International (IEV) en ligne.

#### Glossaire IEC - [std.iec.ch/glossary](http://std.iec.ch/glossary)

65 000 entrées terminologiques électrotechniques, en anglais et en français, extraites des articles Termes et Définitions des publications IEC parues depuis 2002. Plus certaines entrées antérieures extraites des publications des CE 37, 77, 86 et CISPR de l'IEC.

#### Service Clients - [webstore.iec.ch/csc](http://webstore.iec.ch/csc)

Si vous désirez nous donner des commentaires sur cette publication ou si vous avez des questions contactez-nous: [csc@iec.ch](mailto:csc@iec.ch).

# INTERNATIONAL STANDARD

# NORME INTERNATIONALE

AMENDMENT 1  
AMENDEMENT 1

**Medical electrical equipment –**  
**Part 2-21: Particular requirements for the basic safety and essential performance**  
**of infant radiant warmers**

**Appareils électromédicaux –**  
**Partie 2-21: Exigences particulières pour la sécurité de base et les performances**  
**essentiels des incubateurs radiants pour nouveau-nés**

INTERNATIONAL  
ELECTROTECHNICAL  
COMMISSION

COMMISSION  
ELECTROTECHNIQUE  
INTERNATIONALE

ICS 11.040.10

ISBN 978-2-8322-3344-3

**Warning! Make sure that you obtained this publication from an authorized distributor.**  
**Attention! Veuillez vous assurer que vous avez obtenu cette publication via un distributeur agréé.**

## FOREWORD

This amendment has been prepared by subcommittee 62D: Electromedical equipment of IEC technical committee 62: Electrical equipment in medical practice.

The text of this amendment is based on the following documents:

FDIS	Report on voting
62D/1326/FDIS	62D/1347/RVD

Full information on the voting for the approval of this amendment can be found in the report on voting indicated in the above table.

The committee has decided that the contents of this amendment and the base publication will remain unchanged until the stability date indicated on the IEC website under "http://webstore.iec.ch" in the data related to the specific publication. At this date, the publication will be

- reconfirmed,
- withdrawn,
- replaced by a revised edition, or
- amended.

**iTeh STANDARD PREVIEW**  
**(standards.iteh.ai)**

[IEC 60601-2-21:2009/AMD1:2016](https://standards.iteh.ai/catalog/standards/sist/8467265e-1e59-45dc-8e25-de2e111ac59d/iec-60601-2-21-2009-amd1-2016)  
<https://standards.iteh.ai/catalog/standards/sist/8467265e-1e59-45dc-8e25-de2e111ac59d/iec-60601-2-21-2009-amd1-2016>

## INTRODUCTION

*Replace, in the second paragraph, "IEC 60601-1:2005" by "IEC 60601-1".*

**201.1 Scope, object and related standards**

*Replace, in footnote 1), "IEC 60601-1:2005" by "IEC 60601-1".*

**201.1.3 \* Collateral standards**

*Delete the asterisk (\*) from the title.*

*Replace the second paragraph by the following text:*

IEC 60601-1-2 applies as modified in Clause 202. IEC 60601-1-3 and IEC 60601-1-10 do not apply. All other published collateral standards in the IEC 60601-1 series apply as published.

**201.1.4 Particular standards**

*Add an asterisk at the beginning of the title, as follows:*

**201.1.4 \* Particular standards**

*Add the following paragraph at the end of this subclause:*

SKIN TEMPERATURE SENSORS which are applied to operate a BABY CONTROLLED RADIANT WARMER including the displayed value are considered to be not a CLINICAL THERMOMETER in the sense of the particular standard ISO 80601-2-56.

## 201.2 Normative references

*Replace* “IEC 60601-1-2:2007” by “IEC 60601-1-2”.

*Remove the reference to IEC 60601-1-10:2007.*

## 201.3 Terms and definitions

### 201.3.201

#### BABY CONTROLLED RADIANT WARMER

*Remove the note at the end of the entry.*

## 202 Electromagnetic compatibility – Requirements and tests

*Replace, in the first paragraph, “IEC 60601-1-2:2007” by “IEC 60601-1-2”.*

### 202.6.2.3 Radiated RF electromagnetic fields

*Replace the number, title and entire text by the following new subclause number, title and text:*

#### 202.8.9 IMMUNITY TEST LEVELS [IEC 60601-2-21:2009/AMD1:2016](https://standards.iteh.ai/catalog/standards/sist/8467265e-1e59-45dc-8e25-de2e111ac59d/iec-60601-2-21-2009-amd1-2016)

*Addition:* <https://standards.iteh.ai/catalog/standards/sist/8467265e-1e59-45dc-8e25-de2e111ac59d/iec-60601-2-21-2009-amd1-2016>

For radiated radio-frequency electromagnetic fields, the INFANT RADIANT WARMER and/or system shall

- continue to perform its intended function as specified by the MANUFACTURER at a level up to 3 V/m for the frequency range stated in the collateral standard for EMC.

NOTE An INFANT RADIANT WARMER is not considered to be used in a HOME HEALTHCARE ENVIRONMENT.

## 210 Process requirements for the development of physiologic closed-loop controllers

*Delete the entire Clause 210.*

## **Annex AA** (informative)

### **Particular guidance and rationale**

*Add, immediately after the annex title, the following new text and figure:*


#### **AA.1 Requirements and the safety concept of this standard**

Compliance with the minimum safety requirements specified in this particular standard is predominantly checked by measurement of physical quantities such as the temperature. In most cases the spatial location of the measuring site or the temporal development of the quantity is of interest. Therefore, the expert group of this standard considered it helpful to provide a synopsis of the requirements of this standard. Hence, Figure AA.1 illustrates the requirements and their schematic measuring sites or expected temporal development. The requirements as given by their clauses are set in brackets.

**iTeh STANDARD PREVIEW**  
**(standards.iteh.ai)**

[IEC 60601-2-21:2009/AMD1:2016](https://standards.iteh.ai/catalog/standards/sist/8467265e-1e59-45dc-8e25-de2e111ac59d/iec-60601-2-21-2009-amd1-2016)

<https://standards.iteh.ai/catalog/standards/sist/8467265e-1e59-45dc-8e25-de2e111ac59d/iec-60601-2-21-2009-amd1-2016>



Temperature alarm ±1 °C	(201.15.4.2.1)
Overtemperature alarm 40 °C	(201.15.4.2.1, Test 2 addition to item b)
Interruption of power supply alarm	(201.12.3.101)
Maximum surface temperature (normal condition) 40 °C (for metals) 43 °C (for other materials) Maximum surface temperature (single fault condition) 42 °C (for metals) 45 °C (for other materials)	(201.11.1.2.2)
Every 15 min alarm in Manual Mode for irradiance level > 10 mW/cm <sup>2</sup>	(201.12.2.103)

IEC

NOTE Number in brackets indicate the relevant subclauses

**Figure AA.1 – Illustration of the main requirements of this standard**

## AA.2 Particular guidance

### Subclause 201.1.3 – Collateral standards

*Delete the title and entire text.*

*Add the following new text:*

### Subclause 201.1.4 – Particular standards

It is the primary purpose of a BABY CONTROLLED RADIANT WARMER to maintain the temperature as measured by a SKIN TEMPERATURE SENSOR. Hence, SKIN TEMPERATURE SENSORS which are

applied to operate a BABY CONTROLLED RADIANT WARMER including the displayed value are considered to be not a CLINICAL THERMOMETER in the sense of the particular standard ISO 80601-2-56 unless they are specifically extended to measure the body temperature.

The term body temperature is used for all other temperatures of the human body except SKIN TEMPERATURE as defined in IEC 60601-2-19.

#### **Subclause 201.9.6.2.1.101 – Audible alarm sound level**

*Add, after the existing text, the following new text:*

Reflecting rooms represent the acoustic situation in an intensive care nursery more realistically than non-reflecting or semi-anechoic rooms that are very often used for sound pressure measurements. However, reflecting rooms are not well defined and deliver less reproducible values due to their variable size and geometry. The more idealized reverberation chambers deliver very reproducible results but are sometimes difficult to get for tests.

Henceforth, the test can alternatively be performed in a semi-anechoic chamber that is very often used to measure operating sound pressure level. Using a semi-anechoic chamber for the measurements, the thresholds are lowered. This takes into account that reverberation chambers when compared with semi-anechoic chambers obtain sound pressure levels that are reflected mainly at the ceiling which can be considered as low compared to the typical height of a device and to a minor extent by the lateral walls. For measurements in a semi-anechoic chamber and with a measurement distance of 3 m, the thresholds of 65 dB(A) and 50 dB(A) are lowered by 5 dB to 60 dB(A) and 45 dB(A), respectively.

Furthermore, if in the semi-anechoic chamber a distance of 3 m between the device and the microphone as required is not feasible, the distance can be decreased to no less than 2 m. The thresholds of 65 dB(A) and 50 dB(A) are then lowered by 1,5 dB to 63,5 dB(A) and 48,5 dB(A), respectively. This takes into account that the measured sound pressure level is increased by 3,5 dB, compared to a test with a 3-m distance (reciprocal distance  $1/r$  law).

#### **Subclause 202.6.2.3.1 – Requirements**

*Delete the title and entire text.*

#### **Subclause 210.5.1 – Instructions for use**

*Delete the title and entire text.*

#### **Subclause 210.6.3 – PCLS VARIABLE logging**

*Delete the title and entire text.*

#### **Subclause 210.8.2.2.6 – Responses of the PCLS**

*Delete the title and entire text.*



## Bibliography

Delete the following references:

- [31] ISO 3743-1:1994, *Acoustics – Determination of sound power levels of noise sources – Engineering methods for small, movable sources in reverberant fields – Part 1: Comparison method for hard-walled test rooms*
- [32] ISO 7767:1997, *Oxygen monitors for monitoring patient breathing mixtures – Safety requirements (withdrawn)*

Add the following reference:

- [31] ISO 80601-2-56, *Medical electrical equipment – Part 2-56: Particular requirements for basic safety and essential performance of clinical thermometers for body temperature measurement*

## Index of defined terms used in this particular standard

Replace all instances of “IEC 60601-1:2005” by “IEC 60601-1”.

Replace all instances of “IEC 60601-1-6:2006” by “IEC 60601-1-6”.

Replace all instances of “IEC 60601-1-10:2007” by “IEC 60601-1-10”.

Replace all instances of “IEC 60601-1-8:2006” by “IEC 60601-1-8”.

Replace all instances of “IEC 60601-2-19:2009” by “IEC 60601-2-19”.

Replace all instances of “IEC 60601-2-20:2009” by “IEC 60601-2-20”.

Replace all instances of “IEC 60601-2-50:2009” by “IEC 60601-2-50”.

Replace all instances of “IEC 80601-2-35:2009” by “IEC 80601-2-35”.

Replace the term

LIFE SUPPORTING EQUIPMENT ..... IEC 60601-1-2:2007, 3.18

by the following new term:

HOME HEALTHCARE ENVIRONMENT ..... IEC 60601-1-11, 3.2

---