



SLOVENSKI STANDARD
SIST EN ISO 5360:2008
01-april-2008

BUXca Yý U.
SIST EN 1280-1:2000
SIST EN 1280-1:2000/A1:2000

Anestezijski hlapilniki (vaporizatorji) - Sistemi za nalivanje posebnih hlapnih anestetikov (ISO 5360:2006)

Anaesthetic vaporizers - Agent-specific filling systems (ISO 5360:2006)

Anästhesiemittelverdampfer - Substanzspezifische Füllsysteme (ISO 5360:2006)

Evaporateurs d'anesthésie - Systèmes de remplissage spécifiques à l'agent (ISO 5360:2006)

[SIST EN ISO 5360:2008](https://standards.iteh.ai/catalog/standards/sist/a5e56b28-ac60-415e-a74d-4c0343d04991/sist-en-iso-5360-2008)

[https://standards.iteh.ai/catalog/standards/sist/a5e56b28-ac60-415e-a74d-](https://standards.iteh.ai/catalog/standards/sist/a5e56b28-ac60-415e-a74d-4c0343d04991/sist-en-iso-5360-2008)

[4c0343d04991/sist-en-iso-5360-2008](https://standards.iteh.ai/catalog/standards/sist/a5e56b28-ac60-415e-a74d-4c0343d04991/sist-en-iso-5360-2008)

Ta slovenski standard je istoveten z: EN ISO 5360:2007

ICS:

11.040.10

SIST EN ISO 5360:2008

en

iTeh STANDARD PREVIEW
(standards.iteh.ai)

SIST EN ISO 5360:2008

<https://standards.iteh.ai/catalog/standards/sist/a5e56b28-ac60-415e-a74d-4c0343d04991/sist-en-iso-5360-2008>

English Version

Anaesthetic vaporizers - Agent-specific filling systems (ISO
5360:2006)

Evaporateurs d'anesthésie - Systèmes de remplissage
spécifiques à l'agent (ISO 5360:2006)

Anästhesiemittelverdampfer - Substanzspezifische
Füllsysteme (ISO 5360:2006)

This European Standard was approved by CEN on 19 November 2007.

CEN members are bound to comply with the CEN/CENELEC Internal Regulations which stipulate the conditions for giving this European Standard the status of a national standard without any alteration. Up-to-date lists and bibliographical references concerning such national standards may be obtained on application to the CEN Management Centre or to any CEN member.

This European Standard exists in three official versions (English, French, German). A version in any other language made by translation under the responsibility of a CEN member into its own language and notified to the CEN Management Centre has the same status as the official versions.

CEN members are the national standards bodies of Austria, Belgium, Bulgaria, Cyprus, Czech Republic, Denmark, Estonia, Finland, France, Germany, Greece, Hungary, Iceland, Ireland, Italy, Latvia, Lithuania, Luxembourg, Malta, Netherlands, Norway, Poland, Portugal, Romania, Slovakia, Slovenia, Spain, Sweden, Switzerland and United Kingdom.

[SIST EN ISO 5360:2008](https://standards.iteh.ai/catalog/standards/sist/a5e56b28-ac60-415e-a74d-4c0343d04991/sist-en-iso-5360-2008)

<https://standards.iteh.ai/catalog/standards/sist/a5e56b28-ac60-415e-a74d-4c0343d04991/sist-en-iso-5360-2008>



EUROPEAN COMMITTEE FOR STANDARDIZATION
COMITÉ EUROPÉEN DE NORMALISATION
EUROPÄISCHES KOMITEE FÜR NORMUNG

Management Centre: rue de Stassart, 36 B-1050 Brussels

Contents

Page

Foreword.....	3
Annex ZA (informative) Relationship between this European Standard and the Essential Requirements of EU Directive 93/42/EEC Medical Devices.....	4

iTeh STANDARD PREVIEW (standards.iteh.ai)

[SIST EN ISO 5360:2008](https://standards.iteh.ai/catalog/standards/sist/a5e56b28-ac60-415e-a74d-4c0343d04991/sist-en-iso-5360-2008)

<https://standards.iteh.ai/catalog/standards/sist/a5e56b28-ac60-415e-a74d-4c0343d04991/sist-en-iso-5360-2008>

Foreword

The text of ISO 5360:2006 has been prepared by Technical Committee ISO/TC 121 "Anaesthetic and respiratory equipment" of the International Organization for Standardization (ISO) and has been taken over as EN ISO 5360:2007 by Technical Committee CEN/TC 215 "Respiratory and anaesthetic equipment", the secretariat of which is held by BSI.

This European Standard shall be given the status of a national standard, either by publication of an identical text or by endorsement, at the latest by June 2008, and conflicting national standards shall be withdrawn at the latest by June 2008.

This document supersedes EN 1280-1:1997.

This document has been prepared under a mandate given to CEN by the European Commission and the European Free Trade Association, and supports essential requirements of EU Directive(s).

For relationship with EU Directive(s), see informative Annex ZA, which is an integral part of this document.

According to the CEN/CENELEC Internal Regulations, the national standards organizations of the following countries are bound to implement this European Standard: Austria, Belgium, Bulgaria, Cyprus, Czech Republic, Denmark, Estonia, Finland, France, Germany, Greece, Hungary, Iceland, Ireland, Italy, Latvia, Lithuania, Luxembourg, Malta, Netherlands, Norway, Poland, Portugal, Romania, Slovakia, Slovenia, Spain, Sweden, Switzerland and the United Kingdom.

ITeH STANDARD PREVIEW
(standards.iteh.ai)
SIST EN ISO 5360:2008
<https://standards.iteh.ai/catalog/standards/sist/5e56b28-ac60-415e-a74d-4c0343d04991/sist-en-iso-5360-2008>
Endorsement notice

The text of ISO 5360:2006 has been approved by CEN as EN ISO 5360:2007 without any modifications.

Annexe ZA (informative)

Relationship between this European Standard and the Essential Requirements of EU Directive 93/42/EEC Medical Devices

This European Standard has been prepared under a mandate given to CEN by the European Commission and the European Free Trade Association to provide a means of conforming to Essential Requirements of the New Approach Directive 93/42/EEC Medical devices.

Once this standard is cited in the Official Journal of the European Communities under that Directive and has been implemented as a national standard in at least one Member State, compliance with the clauses of this standard given in table ZA confers, within the limits of the scope of this standard, a presumption of conformity with the corresponding Essential Requirements of that Directive and associated EFTA regulations.

Table ZA — Correspondence between this European Standard and Directive 93/42/EEC Medical devices

Clause(s)/sub-clause(s) of this EN	Essential Requirements (ERs) of Directive 93/42/EEC	Qualifying remarks/Notes
4	1, 2, 9.1	
5	1, 2, 9.1	
6	1, 2, 7.5, 9.1	
6.6	4	
7	1, 2, 9.1	
8	1, 2, 3	
9	1, 2, 3, 4, 5, 6, 7.5, 9.1	
10	1, 2, 3, 4, 9.1	
11	2, 9.1, 13.2	
12.1	13.5	
12.1 a)	13.3 a)	
12.1 c)	13.3 b)	
12.2.1 a)	13.3 a)	
12.2.1 b)	13.3 b)	
12.2.1 c)	13.3 b)	
12.2.1 d)	13.3 e)	
12.2.1 e)	13.3.f)	
12.2.1 f)	13.3 f), 13.3 i)	
12.2.2	6, 7.6, 9.1, 13.3 k)	
12.3 a)	13.6 a)	
12.3 b)	6, 7.6, 9.1, 13.3 k)	
12.3 c)	13.6 b), 13.6 d)	

Clause(s)/sub-clause(s) of this EN	Essential Requirements (ERs) of Directive 93/42/EEC	Qualifying remarks/Notes
12.3 d)	13.6 d)	
12.3 e)	13.3 k)	

WARNING — Other requirements and other EU Directives may be applicable to the product(s) falling within the scope of this standard.

iTeh STANDARD PREVIEW (standards.iteh.ai)

[SIST EN ISO 5360:2008](https://standards.iteh.ai/catalog/standards/sist/a5e56b28-ac60-415e-a74d-4c0343d04991/sist-en-iso-5360-2008)

<https://standards.iteh.ai/catalog/standards/sist/a5e56b28-ac60-415e-a74d-4c0343d04991/sist-en-iso-5360-2008>

iTeh STANDARD PREVIEW
(standards.iteh.ai)

SIST EN ISO 5360:2008

<https://standards.iteh.ai/catalog/standards/sist/a5e56b28-ac60-415e-a74d-4c0343d04991/sist-en-iso-5360-2008>

**Anaesthetic vaporizers — Agent-specific
filling systems**

*Évaporateurs d'anesthésie — Systèmes de remplissage spécifiques à
l'agent*

iTeh STANDARD PREVIEW
(standards.iteh.ai)

[SIST EN ISO 5360:2008](https://standards.iteh.ai/catalog/standards/sist/a5e56b28-ac60-415e-a74d-4c0343d04991/sist-en-iso-5360-2008)

<https://standards.iteh.ai/catalog/standards/sist/a5e56b28-ac60-415e-a74d-4c0343d04991/sist-en-iso-5360-2008>



PDF disclaimer

This PDF file may contain embedded typefaces. In accordance with Adobe's licensing policy, this file may be printed or viewed but shall not be edited unless the typefaces which are embedded are licensed to and installed on the computer performing the editing. In downloading this file, parties accept therein the responsibility of not infringing Adobe's licensing policy. The ISO Central Secretariat accepts no liability in this area.

Adobe is a trademark of Adobe Systems Incorporated.

Details of the software products used to create this PDF file can be found in the General Info relative to the file; the PDF-creation parameters were optimized for printing. Every care has been taken to ensure that the file is suitable for use by ISO member bodies. In the unlikely event that a problem relating to it is found, please inform the Central Secretariat at the address given below.

iTeh STANDARD PREVIEW
(standards.iteh.ai)

[SIST EN ISO 5360:2008](https://standards.iteh.ai/catalog/standards/sist/a5e56b28-ac60-415e-a74d-4c0343d04991/sist-en-iso-5360-2008)

<https://standards.iteh.ai/catalog/standards/sist/a5e56b28-ac60-415e-a74d-4c0343d04991/sist-en-iso-5360-2008>

© ISO 2006

All rights reserved. Unless otherwise specified, no part of this publication may be reproduced or utilized in any form or by any means, electronic or mechanical, including photocopying and microfilm, without permission in writing from either ISO at the address below or ISO's member body in the country of the requester.

ISO copyright office
Case postale 56 • CH-1211 Geneva 20
Tel. + 41 22 749 01 11
Fax + 41 22 749 09 47
E-mail copyright@iso.org
Web www.iso.org

Published in Switzerland

Contents

Page

Foreword.....	iv
1 Scope	1
2 Normative references	1
3 Terms and definitions.....	1
4 Bottle	2
5 Bottle collar	4
6 Bottle adaptor.....	8
7 Filler receptacle.....	12
8 Filling rate	14
9 Leakage.....	16
10 Overfilling protection.....	16
11 Colour coding.....	16
12 Information provided by the manufacturer.....	16
Annex A (informative) Types of agent-specific filling systems.....	18
Annex B (normative) Determination of total leakage into atmosphere of anaesthetic agent during filling.....	19
Bibliography	21

Foreword

ISO (the International Organization for Standardization) is a worldwide federation of national standards bodies (ISO member bodies). The work of preparing International Standards is normally carried out through ISO technical committees. Each member body interested in a subject for which a technical committee has been established has the right to be represented on that committee. International organizations, governmental and non-governmental, in liaison with ISO, also take part in the work. ISO collaborates closely with the International Electrotechnical Commission (IEC) on all matters of electrotechnical standardization.

International Standards are drafted in accordance with the rules given in the ISO/IEC Directives, Part 2.

The main task of technical committees is to prepare International Standards. Draft International Standards adopted by the technical committees are circulated to the member bodies for voting. Publication as an International Standard requires approval by at least 75 % of the member bodies casting a vote.

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. ISO shall not be held responsible for identifying any or all such patent rights.

ISO 5360 was prepared by Technical Committee ISO/TC 121, *Anaesthetic and respiratory equipment*, Subcommittee SC 1, *Breathing attachments and anaesthetic machines*.

This second edition cancels and replaces the first edition (ISO 5360:1993) which has been technically revised by virtue of incorporation of Technical Corrigendum 1:1998, inclusion of the changes presented and approved for ISO 5360:1993/Amd 1 (not published) and the need to update cross-references and other factual matters due to the passage of time.

[SIST EN ISO 5360:2008](https://standards.iteh.ai/catalog/standards/sist/a5e56b28-ac60-415e-a74d-4c0343d04991/sist-en-iso-5360-2008)

<https://standards.iteh.ai/catalog/standards/sist/a5e56b28-ac60-415e-a74d-4c0343d04991/sist-en-iso-5360-2008>