

Edition 3.0 2017-06

INTERNATIONAL STANDARD

NORME **INTERNATIONALE**

Medical electrical equipment ANDARD PREVIEW Part 2-28: Particular requirements for the basic safety and essential performance of X-ray tube assemblies for medical diagnosis

IEC 60601-2-28:2017

Appareils électromédicaux en ai/catalog/standards/sist/a511cd40-61b3-4934-a423-Partie 2-28: Exigences particulières pour la sécurité de base et les performances essentielles des gaines équipées pour diagnostic médical





THIS PUBLICATION IS COPYRIGHT PROTECTED Copyright © 2017 IEC, Geneva, Switzerland

All rights reserved. Unless otherwise specified, no part of this publication may be reproduced or utilized in any form or by any means, electronic or mechanical, including photocopying and microfilm, without permission in writing from either IEC or IEC's member National Committee in the country of the requester. If you have any questions about IEC copyright or have an enquiry about obtaining additional rights to this publication, please contact the address below or your local IEC member National Committee for further information.

Droits de reproduction réservés. Sauf indication contraire, aucune partie de cette publication ne peut être reproduite ni utilisée sous quelque forme que ce soit et par aucun procédé, électronique ou mécanique, y compris la photocopie et les microfilms, sans l'accord écrit de l'IEC ou du Comité national de l'IEC du pays du demandeur. Si vous avez des questions sur le copyright de l'IEC ou si vous désirez obtenir des droits supplémentaires sur cette publication, utilisez les coordonnées ci-après ou contactez le Comité national de l'IEC de votre pays de résidence.

IEC Central Office	Tel.: +41 22 919 02 11
3, rue de Varembé	Fax: +41 22 919 03 00
CH-1211 Geneva 20	info@iec.ch
Switzerland	www.iec.ch

About the IEC

The International Electrotechnical Commission (IEC) is the leading global organization that prepares and publishes International Standards for all electrical, electronic and related technologies.

About IEC publications

The technical content of IEC publications is kept under constant review by the IEC. Please make sure that you have the latest edition, a corrigenda or an amendment might have been published.

IEC Catalogue - webstore.iec.ch/catalogue

The stand-alone application for consulting the entire bibliographical information on IEC International Standards, Technical Specifications, Technical Reports and other documents. Available for PC, Mac OS, Android Tablets and iPad.

IEC publications search - www.iec.ch/searchpub

The advanced search enables to find IEC publications by a variety of criteria (reference number, text, technical committee,...). It also gives information on projects, replaced and withdrawn publications.

IEC Just Published - webstore.iec.ch/justpublished

Stay up to date on all new IEC publications. Just Published details all new publications released. Available online and also once a month by email.

Electropedia - www.electropedia.org

The world's leading online dictionary of electronic and electrical terms, containing 20,000 terms and definitions in English and French, with equivalent terms in 16 additional languages. Also known as the International Electrotechnical Vocabulary (IEV) online.

IEC Glossary - std.iec.ch/glossary

65,000 electrotechnical terminology entries in English and French-extracted from the Terms and Definitions clause of IEC publications issued since 2002. Some entries have been collected from earlier publications of IEC TC 37, 77, 86 and CISPR.

IEC Customer Service Centre - webstore.iec.ch/csc

If you wish to give us your feedback on this publication or need further assistance, please contact the Customer Service Centre: csc@iec.ch.

A propos de l'IEC

La Commission Electrotechnique Internationale (IEC) est la première organisation mondiale qui élabore et publie des Normes internationales pour tout ce qui a trait à l'électricité, à l'électronique et aux technologies apparentées.

A propos des publications IEC

Le contenu technique des publications IEC est constamment revu. Veuillez vous assurer que vous possédez l'édition la plus récente, un corrigendum ou amendement peut avoir été publié.

Catalogue IEC - webstore.iec.ch/catalogue

Application autonome pour consulter tous les renseignements bibliographiques sur les Normes internationales, Spécifications techniques, Rapports techniques et autres documents de l'IEC. Disponible pour PC, Mac OS, tablettes Android et iPad.

Recherche de publications IEC - www.iec.ch/searchpub

La recherche avancée permet de trouver des publications IEC en utilisant différents critères (numéro de référence, texte, comité d'études,...). Elle donne aussi des informations sur les projets et les publications remplacées ou retirées.

IEC Just Published - webstore.iec.ch/justpublished

Restez informé sur les nouvelles publications IEC. Just Published détaille les nouvelles publications parues. Disponible en ligne et aussi une fois par mois par email.

Electropedia - www.electropedia.org

Le premier dictionnaire en ligne de termes électroniques et électriques. Il contient 20 000 termes et définitions en anglais et en français, ainsi que les termes équivalents dans 16 langues additionnelles. Egalement appelé Vocabulaire Electrotechnique International (IEV) en ligne.

Glossaire IEC - std.iec.ch/glossary

65 000 entrées terminologiques électrotechniques, en anglais et en français, extraites des articles Termes et Définitions des publications IEC parues depuis 2002. Plus certaines entrées antérieures extraites des publications des CE 37, 77, 86 et CISPR de l'IEC.

Service Clients - webstore.iec.ch/csc

Si vous désirez nous donner des commentaires sur cette publication ou si vous avez des questions contactez-nous: csc@iec.ch.





Edition 3.0 2017-06

INTERNATIONAL STANDARD

NORME INTERNATIONALE

Medical electrical equipment ANDARD PREVIEW Part 2-28: Particular requirements for the basic safety and essential performance of X-ray tube assemblies for medical diagnosis

IEC 60601-2-28:2017

Appareils électromédicauxetrai/catalog/standards/sist/a511cd40-61b3-4934-a423-Partie 2-28: Exigences particulières pour la sécurité de base et les performances essentielles des gaines équipées pour diagnostic médical

INTERNATIONAL ELECTROTECHNICAL COMMISSION

COMMISSION ELECTROTECHNIQUE INTERNATIONALE

ICS 11.040.55

ISBN 978-2-8322-4288-9

Warning! Make sure that you obtained this publication from an authorized distributor. Attention! Veuillez vous assurer que vous avez obtenu cette publication via un distributeur agréé.

 Registered trademark of the International Electrotechnical Commission Marque déposée de la Commission Electrotechnique Internationale

CONTENTS

FOREWO	RD	3
201. 1	Scope, object and related standards	5
201. 2	Normative references	6
201. 3	Terms and definitions	7
201.4	General requirements	7
201. 5	General requirements for testing ME EQUIPMENT	8
201.6	Classification of ME EQUIPMENT and ME SYSTEMS	8
201.7	ME EQUIPMENT identification, marking and documents	9
201.8	Protection against electrical HAZARDS from ME EQUIPMENT	13
201.9	Protection against MECHANICAL HAZARDS of ME EQUIPMENT and ME SYSTEMS	14
201. 10	Protection against unwanted and excessive RADIATION HAZARDS	15
201. 11	Protection against excessive temperatures and other HAZARDS	15
201. 12	Accuracy of controls and instruments and protection against hazardous outputs	16
201. 13	HAZARDOUS SITUATIONS and fault conditions for ME EQUIPMENT	16
201. 14	PROGRAMMABLE ELECTRICAL MEDICAL SYSTEMS (PEMS)	
201. 15	Construction of the EQUIPMENNDARD PREVIEW	17
201.16	ME SYSTEMS	17
201. 17	Electromagnetic compatibility of ME EQUIPMENT and ME SYSTEMS	17
203	RADIATION protection in diagnostic X-RAYEQUIPMENT	18
Annexes	https://standards.iteh.ai/catalog/standards/sist/a511cd40-61b3-4934-a423- 008da5ec6219/iec-60601-2-28-2017	20
	(informative) Test of X-RAY TUBE ASSEMBLIES for expelled parts-related	
	and/or tube implosion-related RISKS	
• •	Bibliography	
Index of d	ndex of defined terms used in this particular standard	

INTERNATIONAL ELECTROTECHNICAL COMMISSION

MEDICAL ELECTRICAL EQUIPMENT –

Part 2-28: Particular requirements for the basic safety and essential performance of X-ray tube assemblies for medical diagnosis

FOREWORD

- 1) The International Electrotechnical Commission (IEC) is a worldwide organization for standardization comprising all national electrotechnical committees (IEC National Committees). The object of IEC is to promote international co-operation on all questions concerning standardization in the electrical and electronic fields. To this end and in addition to other activities, IEC publishes International Standards, Technical Specifications, Technical Reports, Publicly Available Specifications (PAS) and Guides (hereafter referred to as "IEC Publication(s)"). Their preparation is entrusted to technical committees; any IEC National Committee interested in the subject dealt with may participate in this preparatory work. International, governmental and non-governmental organizations liaising with the IEC also participate in this preparation. IEC collaborates closely with the International Organization for Standardization (ISO) in accordance with conditions determined by agreement between the two organizations.
- 2) The formal decisions or agreements of IEC on technical matters express, as nearly as possible, an international consensus of opinion on the relevant subjects since each technical committee has representation from all interested IEC National Committees.
- 3) IEC Publications have the form of recommendations for international use and are accepted by IEC National Committees in that sense. While all reasonable efforts are made to ensure that the technical content of IEC Publications is accurate, IEC cannot be held responsible for the way in which they are used or for any misinterpretation by any end user. (Standards.iten.al)
- 4) In order to promote international uniformity, IEC National Committees undertake to apply IEC Publications transparently to the maximum extent possible in their national and regional publications. Any divergence between any IEC Publication and the corresponding national or regional publication shall be clearly indicated in the latter. https://standards.iteh.ai/catalog/standards/sist/a511cd40-61b3-4934-a423-
- 5) IEC itself does not provide any attestation of conformity. Independent certification bodies provide conformity assessment services and, in some areas, access to IEC marks of conformity. IEC is not responsible for any services carried out by independent certification bodies.
- 6) All users should ensure that they have the latest edition of this publication.
- 7) No liability shall attach to IEC or its directors, employees, servants or agents including individual experts and members of its technical committees and IEC National Committees for any personal injury, property damage or other damage of any nature whatsoever, whether direct or indirect, or for costs (including legal fees) and expenses arising out of the publication, use of, or reliance upon, this IEC Publication or any other IEC Publications.
- 8) Attention is drawn to the Normative references cited in this publication. Use of the referenced publications is indispensable for the correct application of this publication.
- Attention is drawn to the possibility that some of the elements of this IEC Publication may be the subject of patent rights. IEC shall not be held responsible for identifying any or all such patent rights.

International Standard IEC 60601-2-28 has been prepared by subcommittee 62B: Diagnostic imaging equipment, of IEC technical committee 62: Electrical equipment in medical practice.

This third edition cancels and replaces the second edition published in 2010. This edition constitutes a technical revision.

The third edition of this particular standard has been prepared to fit IEC 60601-1:2005 and IEC 60601-1:2005/AMD1:2012 (the amended third edition of IEC 60601-1), which is referred to as the general standard. Apart from the changes related to the amendment of IEC 60601-1, changes related to technical improvements are also included.

The text of this standard is based on the following documents:

FDIS	Report on voting
62B/1040/FDIS	62B/1051/RVD

Full information on the voting for the approval of this particular standard can be found in the report on voting indicated in the above table.

This publication has been drafted in accordance with the ISO/IEC Directives, Part 2.

In this standard, the following print types are used:

- requirements and definitions: roman type;
- test specifications: italic type;
- informative material appearing outside of tables, such as notes, examples and references: in smaller type. Normative text of tables is also in a smaller type;
- TERMS DEFINED IN CLAUSE 3 OF THE GENERAL STANDARD, IN THIS PARTICULAR STANDARD OR AS NOTED: SMALL CAPITALS.

In referring to the structure of this standard, the term

- "clause" means one of the seventeen numbered divisions within the table of contents, inclusive of all subdivisions (e.g. Clause 7 includes subclauses 7.1, 7.2, etc.);
- "subclause" means a numbered subdivision of a clause (e.g. 7.1, 7.2 and 7.2.1 are all subclauses of Clause 7). (standards.iteh.al)

References to clauses within this standard are preceded by the term "Clause" followed by the clause number. References to subclauses within this particular standard are by number only.

008da5ec6219/iec-60601-2-28-2017 In this standard, the conjunctive "or" is used as an "inclusive or" so a statement is true if any combination of the conditions is true.

The verbal forms used in this standard conform to usage described in Clause 7 of the ISO/IEC Directives, Part 2. For the purposes of this standard, the auxiliary verb:

- "shall" means that compliance with a requirement or a test is mandatory for compliance with this standard;
- "should" means that compliance with a requirement or a test is recommended but is not mandatory for compliance with this standard;
- "may" is used to describe a permissible way to achieve compliance with a requirement or test.

A list of all parts of the IEC 60601 series, published under the general title Medical electrical equipment, can be found on the IEC website.

The committee has decided that the contents of this publication will remain unchanged until the stability date indicated on the IEC web site under "http://webstore.iec.ch" in the data related to the specific publication. At this date, the publication will be

- reconfirmed.
- withdrawn,
- replaced by a revised edition, or
- amended.

MEDICAL ELECTRICAL EQUIPMENT -

Part 2-28: Particular requirements for the basic safety and essential performance of X-ray tube assemblies for medical diagnosis

201.1 Scope, object and related standards

Clause 1 of the general standard¹ applies, except as follows:

201.1.1 Scope

Replacement:

This part of IEC 60601 applies to the BASIC SAFETY and ESSENTIAL PERFORMANCE of X-RAY TUBE ASSEMBLIES and to components thereof, intended for medical diagnosis and imaging.

Where the general standard IEC 60601-1 and the collateral standard IEC 60601-1-3 refer to ME EQUIPMENT, this is interpreted as X-RAY TUBE ASSEMBLIES in this particular standard. If a clause or subclause is specifically intended to be applicable to ME EQUIPMENT only, or to ME SYSTEMS only, the title and content of that clause or subclause will say so. If that is not the case, the clause or subclause applies both to ME EQUIPMENT and to ME SYSTEMS, as relevant.

NOTE This document is also applicable to the X-RAY TUBE ASSEMBLY aspects of X-RAY SOURCE ASSEMBLIES and X-RAY TUBE HEADS.

 IEC 60601-2-28:2017

 201.1.2
 Object

 Object
 008da5ec6219/iec-60601-2-28-2017

Replacement:

The object of this particular standard is to establish particular BASIC SAFETY and ESSENTIAL PERFORMANCE requirements for X-RAY TUBE ASSEMBLIES for medical diagnosis.

201.1.3 Collateral standards

Addition:

This particular standard refers to those applicable collateral standards that are listed in Clause 2 of the general standard and Clause 201.2 of this particular standard.

IEC 60601-1-3:2008 and IEC 60601-1-3:2008/AMD1:2013 applies as modified in Clause 203. IEC 60601-1-2, IEC 60601-1-6, IEC 60601-1-8, IEC 60601-1-9, IEC 60601-1-10, IEC 60601-1-11 and IEC 60601-1-12 do not apply. All other published collateral standards in the IEC 60601-1 series apply as published.

NOTE 101 IEC 60601-1-2 does not apply because RISKS for the X-RAY TUBE ASSEMBLY outside the system may only be indicative of RISKS for the system due to the difference in electromagnetic environment.

NOTE 102 IEC 60601-1-6 and IEC 60601-1-8 do not apply because X-RAY TUBE ASSEMBLIES are not operated as a stand-alone device.

NOTE 103 X-RAY TUBE ASSEMBLIES are not in the scope of IEC 60601-1-10, IEC 60601-1-11 and IEC 60601-1-12.

¹ The general standard is IEC 60601-1:2005 and IEC 60601-1:2005/AMD1:2012, *Medical electrical equipment – Part 1: General requirements for basic safety and essential performance.*

201.1.4 Particular standards

Replacement:

In the IEC 60601 series, particular standards may modify, replace or delete requirements contained in the general standard and collateral standards as appropriate for the particular ME EQUIPMENT under consideration, and may add other BASIC SAFETY and ESSENTIAL PERFORMANCE requirements.

A requirement of a particular standard takes priority over the general standard.

For brevity, IEC 60601-1:2005 and IEC 60601-1:2005/AMD1:2012 are referred to in this particular standard as the general standard. Collateral standards are referred to by their document number.

The numbering of clauses and subclauses of this particular standard corresponds to that of the general standard with the prefix "201" (e.g. 201.1 in this document addresses the content of Clause 1 of the general standard) or applicable collateral standard with the prefix "20x" where x is the final digit(s) of the collateral standard document number (e.g. 202.4 in this particular standard addresses the content of Clause 4 of the 60601-1-2 collateral standard, 203.4 in this particular standard addresses the content of Clause 4 of the 60601-1-3 collateral standard, etc.). The changes to the text of the general standard are SPECIFIED by the use of the following words:

"Replacement" means that the clause or subclause of the general standard or applicable collateral standard is replaced completely by the text of this particular standard.

"Addition" means that the text of this particular standard is additional to the requirements of the general standard or applicable collateral(standard)17

https://standards.iteh.ai/catalog/standards/sist/a511cd40-61b3-4934-a423-

"Amendment" means that the clause of subclause of the general standard or applicable collateral standard is amended as indicated by the text of this particular standard.

Subclauses, figures or tables which are additional to those of the general standard are numbered starting from 201.101. However due to the fact that definitions in the general standard are numbered 3.1 through 3.147, additional definitions in this document are numbered beginning from 201.3.201. Additional annexes are lettered AA, BB, etc., and additional items aa), bb), etc.

Subclauses, figures or tables which are additional to those of a collateral standard are numbered starting from 20x, where "x" is the number of the collateral standard, e.g. 202 for IEC 60601-1-2, 203 for IEC 60601-1-3, etc.

The term "this document" is used to make reference to the general standard, any applicable collateral standards and this particular standard taken together.

Where there is no corresponding clause or subclause in this particular standard, the clause or subclause of the general standard or applicable collateral standard, although possibly not relevant, applies without modification; where it is intended that any part of the general standard or applicable collateral standard, although possibly relevant, is not to be applied, a statement to that effect is given in this particular standard.

201.2 Normative references

NOTE Informative references are listed in the Bibliography.

Clause 2 of the general standard applies, except as follows:

IEC 60601-2-28:2017 © IEC 2017 - 7 -

Replacement:

IEC 60601-1-3:2008, Medical electrical equipment – Part 1-3: General requirements for basic safety and essential performance – Collateral Standard: Radiation protection in diagnostic X-ray equipment IEC 60601-1-3:2008/AMD1:2013

Addition:

IEC 60336, Medical electrical equipment – X-ray tube assemblies for medical diagnosis – Characteristics of focal spots

IEC 60522, Determination of the permanent filtration of X-ray tube assemblies

IEC 60613:2010, Electrical and loading characteristics of X-ray tube assemblies for medical diagnosis

IEC TR 60788:2004, Medical electrical equipment – Glossary of defined terms

201.3 Terms and definitions

Clause 3 of the general standard applies, except as follows:

Teh STANDARD PREVIEW For the purposes of this document, the terms and definitions given in the general standard, applicable collateral standards **TEC 60613:2010, aTEC 60522**, IEC 60336, and IEC TR 60788:2004 apply.

IEC 60601-2-28:2017

ISO and IEC maintain/terminological.databases/storasuse14n-standardization at the following addresses: 008da5ec6219/iec-60601-2-28-2017

- IEC Electropedia: available at http://www.electropedia.org/
- ISO Online browsing platform: available at http://www.iso.org/obp

NOTE An index of defined terms is found beginning on page 24.

201.3.71 NORMAL USE

Addition:

Note 1 to entry: Where used in this document, the defined term NORMAL USE is understood to only apply to the X-RAY TUBE ASSEMBLY as it operates in X-RAY EQUIPMENT.

201.4 General requirements

Clause 4 of the general standard applies, except as follows:

201.4.3 ESSENTIAL PERFORMANCE

Addition:

The entity X-RAY TUBE ASSEMBLY itself does not have ESSENTIAL PERFORMANCE. Whether characteristics of an X-RAY TUBE ASSEMBLY shall be considered ESSENTIAL PERFORMANCE, depends on the X-ray system and HIGH-VOLTAGE GENERATOR characteristics combined with the X-RAY TUBE ASSEMBLY.

201.4.4 **EXPECTED SERVICE LIFE**

Addition:

EXPECTED SERVICE LIFE may also be based on metrics related to use.

NOTE 101 Examples of use: number of scans, radiographs, PATIENT exams.

NOTE 102 X-RAY TUBE ASSEMBLIES are consumables, i.e. their use leads ultimately to their replacement. By design, an X-RAY TUBE ASSEMBLY maintains BASIC SAFETY throughout its life and its replacement.

- 8 -

NOTE 103 EXPECTED SERVICE LIFE typifies the estimated replacement times of a population of X-RAY TUBE ASSEMBLIES. EXPECTED SERVICE LIFE is based on a statistical analysis of the survival of e.g. 5 % of the X-RAY TUBE ASSEMBLIES in the population.

201.4.11 **Power input**

Subclause 4.11 of the general standard does not apply.

201.5 General requirements for testing ME EQUIPMENT

Clause 5 of the general standard applies, except as follows:

201.5.5 Supply voltages, type of current, nature of supply, frequency

Addition to paragraph 5.5 f):

(standards.iteh.ai)

A HIGH-VOLTAGE GENERATOR which is not SPECIFIED in the ACCOMPANYING DOCUMENTS can be used if the characteristics which are essential for a given test are equivalent to the SPECIFIED HIGH-VOLTAGE GENERATOR. https://standards.iteh.ai/catalog/standards/sist/a511cd40-61b3-4934-a423-

008da5ec6219/iec-60601-2-28-2017

201.5.7 Humidity preconditioning treatment

Addition:

For those X-RAY TUBE ASSEMBLIES that are to be used only in controlled environments, as to be SPECIFIED in the ACCOMPANYING DOCUMENTS, no humidity preconditioning is required.

201.5.9 Determination of APPLIED PARTS and ACCESSIBLE PARTS

201.5.9.2 ACCESSIBLE PARTS

Subclause 5.9.2 of the general standard does not apply.

NOTE Parts accessibility of the X-RAY TUBE ASSEMBLY will necessarily be evaluated as integrated in specific X-RAY EQUIPMENT.

Classification of ME EQUIPMENT and ME SYSTEMS 201.6

Clause 6 of the general standard applies, except as follows:

201.6.2 Protection against electric shock

Addition:

X-RAY TUBE ASSEMBLIES shall be classified as CLASS I equipment.

201.7 ME EQUIPMENT identification, marking and documents

Clause 7 of the general standard applies, except as follows:

201.7.1 General

201.7.1.1 USABILITY of the identification, marking and documents

Subclause 7.1.1 of the general standard does not apply.

NOTE The user interface is part of the X-RAY EQUIPMENT, but not of the X-RAY TUBE ASSEMBLY.

201.7.2 Marking on the outside of ME EQUIPMENT or ME EQUIPMENT parts

201.7.2.2 Identification

Replacement of the first paragraph by:

The X-RAY TUBE ASSEMBLY shall be marked with:

- the name or trademark and address of the MANUFACTURER;
- a MODEL OR TYPE REFERENCE;
- an individual identification;
 STANDARD PREVIEW
- the date of manufacture.

NOTE 101 See ISO 15223-1 for symbols for MANUFACTURER, serial number, lot or batch, year of manufacture, and use by date.

IEC 60601-2-28:2017

NOTE 102 See also 20175/2102ards.iteh.ai/catalog/standards/sist/a511cd40-61b3-4934-a423-008da5ec6219/iec-60601-2-28-2017

201.7.2.5 ME EQUIPMENT intended to receive power from other equipment

Subclause 7.2.5 of the general standard does not apply.

NOTE For applicable requirements, see 201.7.9.3.101.

201.7.2.11 Mode of operation

Subclause 7.2.11 of the general standard does not apply.

NOTE X-RAY TUBE ASSEMBLIES are not operated as a stand alone device.

201.7.2.14 HIGH VOLTAGE TERMINAL DEVICES

Replacement:

HIGH VOLTAGE cable connections between the X-RAY TUBE ASSEMBLY and the HIGH-VOLTAGE GENERATOR accessible in NORMAL USE shall be marked with symbol IEC 60417-5036 (2002-10) (see Table D.1, symbol 24) unless a tool is required for removal of the cable connection.

201.7.2.15 Cooling conditions

Addition:

Marking of cooling conditions is not required if the cooling unit and the X-RAY TUBE ASSEMBLY have been designed for compatibility.

NOTE A cooling unit is a standalone device or integral part of the X-RAY TUBE ASSEMBLY which provides increased cooling capability of the X-RAY TUBE ASSEMBLY.

Additional subclauses:

201.7.2.101 Marking of X-RAY TUBES

The markings on the X-RAY TUBE shall remain readable when the X-RAY TUBE is dismantled from the X-RAY TUBE HOUSING after a period of NORMAL USE.

The markings shall enable individual products, series or types to be correlated with their ACCOMPANYING DOCUMENTS.

X-RAY TUBES shall be provided with the following markings:

- name or trademark of the MANUFACTURER;
- MODEL OR TYPE REFERENCE;
- individual identification.

The above markings may be given in the form of a combined designation explained in the ACCOMPANYING DOCUMENTS.

201.7.2.102 Marking on the outside of X-RAY TUBE ASSEMBLIES

X-RAY TUBE ASSEMBLIES shall be provided with the following markings:

- (standards.iteh.ai)
- NOMINAL X-RAY TUBE VOLTAGE for which the X-RAY TUBE ASSEMBLY is designed;
- if there is more than one HIGH VOLTAGE)cable2seceptacle, indication of the polarity of the HIGH VOLTAGE cable/seceptacles;catalog/standards/sist/a511cd40-61b3-4934-a423-
- FOCAL SPOT size(s). If the FOCAL⁵SPOT⁹ size(s)¹are² in² the range of NOMINAL FOCAL SPOT VALUES in IEC 60336, then mark the FOCAL SPOT size(s) as NOMINAL FOCAL SPOT VALUE(S) according to IEC 60336.

NOTE See further 201.7.2.2 and 203.7.3.

201.7.3 Marking on the inside of ME EQUIPMENT OR ME EQUIPMENT parts

201.7.3.2 HIGH VOLTAGE parts

Subclause 7.3.2 of the general standard does not apply.

NOTE While the inside of an X-RAY TUBE ASSEMBLY is being worked on, the assembly is normally not energized. Even if the assembly is energized, only trained service personnel are allowed to perform the work, so safe operation is assured.

201.7.9 ACCOMPANYING DOCUMENTS

201.7.9.1 General

Replacement:

ME EQUIPMENT shall be accompanied by documents containing at least the instructions for use and a technical description. The ACCOMPANYING DOCUMENTS shall be regarded as a part of the ME EQUIPMENT.

The ACCOMPANYING DOCUMENTS may be provided with the X-RAY TUBE ASSEMBLY, or they may be integrated into the ACCOMPANYING DOCUMENTS of any ME SYSTEM for which the X-RAY TUBE ASSEMBLY is compatible.

IEC 60601-2-28:2017 © IEC 2017 - 11 -

If an X-RAY TUBE ASSEMBLY is intended to receive its power from other equipment in an ME SYSTEM, or otherwise puts special requirements on the supporting ME SYSTEM, the ACCOMPANYING DOCUMENTS shall sufficiently specify such other equipment to ensure compliance with the requirements of this document.

NOTE 101 The purpose of the ACCOMPANYING DOCUMENTS is to promote the safe use of the ME EQUIPMENT during its EXPECTED SERVICE LIFE.

The ACCOMPANYING DOCUMENTS shall identify the ME EQUIPMENT by including, as applicable, the following:

- name or trade-name of the MANUFACTURER and contact information to which the RESPONSIBLE ORGANIZATION can refer;
- MODEL OR TYPE REFERENCE.

NOTE 102 Contact information can be e.g. a telephone number, email address, address or website, where the MANUFACTURER can be contacted.

ACCOMPANYING DOCUMENTS may be provided electronically, e.g. a file on an electronic medium. If the ACCOMPANYING DOCUMENTS are provided electronically, the RISK MANAGEMENT FILE shall include consideration on which information also needs to be provided as hard copy or as markings on the ME EQUIPMENT.

EXAMPLE Information to cover emergency operation.

NOTE 103 ACCOMPANYING DOCUMENTS provided electronically might not be acceptable in all jurisdictions.

NOTE 104 Instead of the USABILITY ENGINEERING PROCESS (JEC 60601-1-6 is not applicable for an X-RAX TUBE ASSEMBLY), the RISK MANAGEMENT PROCESS (based on ISO 14971) is the carrier of the considerations. ISO 14971:2007, 4.2 (INTENDED USE) and C.2.29 (human factors) cover the USABILITY aspects of X-RAY TUBE ASSEMBLIES sufficiently.

IEC 60601-2-28:2017

The ACCOMPANYING DOCUMENTS shall specify special skills training and knowledge required of the intended OPERATOR or the RESPONSIBLE ORGANIZATION and restrictions on locations or environments in which the ME EQUIPMENT can be used.

The ACCOMPANYING DOCUMENTS shall be written at a level consistent with the education, training and any special needs of the person(s) for whom they are intended.

Compliance is checked by inspection of the ACCOMPANYING DOCUMENTS and, when provided electronically, by inspection of the RISK MANAGEMENT FILE.

201.7.9.2 Instructions for use

201.7.9.2.2 Warning and safety notices

Replacement of the second paragraph of this subclause:

For X-RAY TUBE ASSEMBLIES, the ACCOMPANYING DOCUMENTS shall include a warning statement to the effect: "WARNING: To avoid the risk of electric shock, this equipment must only be connected to a supply with protective earth."

NOTE 101 X-RAY TUBE ASSEMBLIES normally do not connect to a SUPPLY MAINS.

201.7.9.2.3 ME EQUIPMENT SPECIFIED for connection to a separate power supply

Subclause 7.9.2.3 of the general standard does not apply.

201.7.9.2.14 ACCESSORIES, supplementary equipment, used material

Subclause 7.9.2.14 of the general standard applies, except as follows:

- the second paragraph;
- the Note.

201.7.9.2.17 ME EQUIPMENT emitting RADIATION

Subclause 7.9.2.17 of the general standard does not apply.

NOTE RADIATION intensity and distribution are governed at system level. Nature and type of the RADIATION are SPECIFIED in 201.7.9.3.101 a).

Additional subclause:

201.7.9.2.101 Instructions for use of X-RAY TUBE ASSEMBLIES

The instructions for use of an X-RAY TUBE ASSEMBLY shall state the following data as appropriate to the INTENDED USE:

- a) SINGLE LOAD RATING;
- b) SERIAL LOAD RATING;
- c) NOMINAL RADIOGRAPHIC ANODE INPUT POWER according to IEC 60613:2010;
- d) NOMINAL CT ANODE INPUT POWER according to IEC 60613:2010;
- e) NOMINAL CT SCAN POWER INDEX according to IEC 60613:2010.

201.7.9.3 Technical description IEC 60601-2-28:2017

Additional subclause: 008da5ec6219/iec-60601-2-28-2017

201.7.9.3.101 Technical description of X-RAY TUBE ASSEMBLIES

The technical descriptions of X-RAY TUBE ASSEMBLIES shall specify the following data:

- a) the identity of the TARGET material(s) that characterize the RADIATION SPECTRUM;
- b) the REFERENCE AXIS;
- c) the TARGET ANGLE(S);
- d) FOCAL SPOT size(s);

If the FOCAL SPOT size(s) are in the range of NOMINAL FOCAL SPOT VALUES in IEC 60336, then state the FOCAL SPOT size(s) as NOMINAL FOCAL SPOT VALUE(S) according to IEC 60336.

- e) PERMANENT FILTRATION according to IEC 60522 or the thickness(es) of the material(s) concerned, together with its/their chemical symbol(s);
- f) QUALITY EQUIVALENT FILTRATION of parts which are or could become ADDED FILTERS and method of mounting/dismounting such, if applicable;

NOTE 101 The preceding two FILTRATION requirements cover requirements SPECIFIED in 7.3 of IEC 60601-1-3:2008.

- g) NOMINAL X-RAY TUBE VOLTAGE;
- h) data concerning the HIGH VOLTAGE required from the HIGH-VOLTAGE GENERATOR or the type designation of suitable supply equipment;
- i) type-designation or specification of the HIGH VOLTAGE connectors;
- j) requirements for the HIGH-VOLTAGE GENERATOR, for supplying the filament(s), for rotating the ANODE (when appropriate) and for auxiliary equipment (such as for cooling purposes),

appropriate for the safe application of the X-RAY TUBE ASSEMBLY as defined in the RISK MANAGEMENT FILE;

k) CATHODE EMISSION CHARACTERISTIC;

NOTE 102 For the 4 preceding items,(h) to k)), for an X-RAY TUBE ASSEMBLY which comes built in into an X-ray system with HIGH-VOLTAGE GENERATOR, normally no data are required. If the X-RAY TUBE ASSEMBLY is sold to an OEM-system MANUFACTURER, then normally an elaborate interface specification will be included.

- I) ENVELOPE VOLTAGE according to IEC 60613:2010, if applicable;
- m) ENVELOPE CURRENT according to IEC 60613:2010, if applicable;
- n) principal dimensions and interfaces in the form of a drawing; this drawing also shows the REFERENCE AXIS, the position and the accuracy of the position of the FOCAL SPOT(s);
- o) mass with and without additional components;
- p) CONTINUOUS ANODE INPUT POWER according to IEC 60613:2010 at the highest value of NOMINAL X-RAY TUBE VOLTAGE under any operating condition;
- q) classifications according to Clause 6 of the general standard;
- r) polarity of HIGH-VOLTAGE connections;
- s) limits for the conditions for transport and storage;
- t) if applicable, any requirements that shall be fulfilled prior to applying power to the X-RAY TUBE ASSEMBLY, e.g. the time period that the room environmental operating conditions shall be maintained, precautions to be observed before the first LOADING upon completion of the installation of an X-RAY TUBE ASSEMBLY, special procedures for conditioning the X-RAY TUBE;
 TUBE;
- u) NOMINAL CONTINUOUS INPUT POWER according to IEC 60613:2010.
- (standards.iteh.ai)

NOTE 103 As equipment (example: BEAM LIMITING DEVICE) which is – electrically or mechanically – attached to the X-RAY TUBE ASSEMBLY can affect compliance of the X-RAY TUBE ASSEMBLY with this document, the technical description of the X-RAY TUBE ASSEMBLY in this clause lists those specifications and interfaces which might affect compliance of the X-RAY TUBE ASSEMBLY. This is not an exhaustive list of technical descriptions, as such equipment attached to the X-RAY TUBE ASSEMBLY might pose additional requirements on interfacing.

201.8 Protection against electrical HAZARDS from ME EQUIPMENT

Clause 8 of the general standard applies, except as follows:

201.8.2 Requirements related to power sources

201.8.2.1 Connection to a separate power source

Subclause 8.2.1 of the general standard does not apply.

201.8.7 LEAKAGE CURRENTS and PATIENT AUXILIARY CURRENTS

Addition:

NOTE Measurements on the X-RAY TUBE ASSEMBLY outside the system are only indicative of measurements on the system, due to the difference in electrical connections.

201.8.8 Insulation

201.8.8.3 Dielectric strength

Amendment to Table 6 of the general standard:

For PEAK WORKING VOLTAGES $U > 14 \ 140 \ V$, the HIGH VOLTAGE circuits of the X-RAY TUBE ASSEMBLY are tested at 110 % of the NOMINAL X-RAY TUBE VOLTAGE of the X-RAY TUBE ASSEMBLY, where the voltage is raised over a period of 10 s or less, and is then maintained for 3 min.