



SLOVENSKI STANDARD

SIST EN 50379-3:2012

01-junij-2012

Specifikacija za prenosne električne naprave za merjenje parametrov vnetljivosti izpušnega plina grelnih naprav - 3. del: Zahtevane lastnosti za naprave, ki se uporabljajo pri vzdrževanju plinskih grelnih naprav, ki ni zakonsko predpisano

Specification for portable electrical apparatus designed to measure combustion flue gas parameters of heating appliances - Part 3: Performance requirements for apparatus used in non-statutory servicing of gas fired heating appliances

Anforderungen an tragbare elektrische Geräte zur Messung von Verbrennungsparametern von Heizungsanlagen - Teil 3: Anforderungen an das Betriebsverhalten von Geräten für den Einsatz im nicht-geregelten Bereich bei Wartungen von gasbefeuerten Heizungsanlagen

Spécification pour les appareils électriques portatifs conçus pour mesurer les paramètres des gaz de combustion dans les conduits d'évacuation des appareils de chauffage - Partie 3: Prescriptions des caractéristiques des appareils utilisés dans le service après-vente hors champ réglementaire des appareils de chauffage à gaz

Ta slovenski standard je istoveten z: EN 50379-3:2012

ICS:

13.040.40	Emisije nepremičnih virov	Stationary source emissions
91.140.10	Sistemi centralnega ogrevanja	Central heating systems

SIST EN 50379-3:2012

en

EUROPEAN STANDARD
NORME EUROPÉENNE
EUROPÄISCHE NORM

EN 50379-3

April 2012

ICS 13.320

Supersedes EN 50379-3:2004

English version

**Specification for portable electrical apparatus designed to measure
combustion flue gas parameters of heating appliances -
Part 3: Performance requirements for apparatus used in non-statutory
servicing of gas fired heating appliances**

Spécification pour les appareils
électriques portatifs conçus pour mesurer
les paramètres des gaz de combustion
dans les conduits d'évacuation des
appareils de chauffage -
Partie 3: Prescriptions des
caractéristiques des appareils utilisés
dans le service après-vente hors champ
réglementaire des appareils de chauffage
à gaz

Anforderungen an tragbare elektrische
Geräte zur Messung von
Verbrennungsparametern von
Heizungsanlagen -
Teil 3: Anforderungen an das
Betriebsverhalten von Geräten für den
Einsatz im nicht-geregelten Bereich bei
Wartungen von gasbefeuelten
Heizungsanlagen

This European Standard was approved by CENELEC on 2012-03-19. CENELEC members are bound to comply with the CEN/CENELEC Internal Regulations which stipulate the conditions for giving this European Standard the status of a national standard without any alteration.

Up-to-date lists and bibliographical references concerning such national standards may be obtained on application to the CEN-CENELEC Management Centre or to any CENELEC member.

This European Standard exists in three official versions (English, French, German). A version in any other language made by translation under the responsibility of a CENELEC member into its own language and notified to the CEN-CENELEC Management Centre has the same status as the official versions.

CENELEC members are the national electrotechnical committees of Austria, Belgium, Bulgaria, Croatia, Cyprus, the Czech Republic, Denmark, Estonia, Finland, France, Germany, Greece, Hungary, Iceland, Ireland, Italy, Latvia, Lithuania, Luxembourg, Malta, the Netherlands, Norway, Poland, Portugal, Romania, Slovakia, Slovenia, Spain, Sweden, Switzerland, Turkey and the United Kingdom.

CENELEC

European Committee for Electrotechnical Standardization
Comité Européen de Normalisation Electrotechnique
Europäisches Komitee für Elektrotechnische Normung

Management Centre: Avenue Marnix 17, B - 1000 Brussels

Contents

Foreword	3
Introduction.....	4
1 Scope.....	5
2 Normative references	5
3 Terms and definitions	5
4 General requirements	6
5 Test methods and performance requirements	6
5.1 General requirements for tests	6
5.2 Normal conditions for tests	6
5.3 Mechanical tests	6
5.4 Electrical and software tests.....	7
5.5 Tests with test gases.....	7
5.6 Tests with real flue gases.....	8
5.7 Calculated values	9
5.8 Temperature.....	9
5.9 Pressure	10
Bibliography.....	11

Foreword

This document (EN 50379-3:2012) has been prepared by CLC/TC 216 "Gas detectors".

The following dates are fixed:

- latest date by which this document has to be implemented at national level by publication of an identical national standard or by endorsement (dop) 2013-03-19
- latest date by which the national standards conflicting with this document have to be withdrawn (dow) 2015-03-19

This document supersedes EN 50379-3:2004.

EN 50379-3:2012 includes the following significant technical changes with respect to EN 50379-3:2004:

- 5.5.7 considers calibration curves for sensors with nonlinear signal;
- 5.5.8 considers influence of pressure variations;
- 5.5.9 considers the influence of water vapour on the gas signal;
- 5.7.2 for calculated values was amended;
- 5.9.1 was amended to cover measurement at the circular orifice.

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. CENELEC [and/or CEN] shall not be held responsible for identifying any or all such patent rights.

Introduction

This European Standard covers apparatus for measuring gas concentrations and other combustion parameters, as used in the installation and maintenance of heating appliances. It forms a specification for portable electrical apparatus designed to measure combustion flue gas parameters of heating appliances, and includes the following parts under the generic title *Specification for portable electrical apparatus designed to measure combustion flue gas parameters of heating appliances*:

Part 1: General requirements and test methods;

Part 2: Performance requirements for apparatus used in statutory inspections and assessments;

Part 3: Performance requirements for apparatus used in non-statutory servicing of gas fired heating appliances.

EN 50379-1 specifies general requirements for the construction, testing and performance of portable spot reading apparatus, designed to give an assessment of specific combustion flue gas parameters such as concentration of gaseous compounds, temperature and/or pressure, to check the combustion performance of heating appliances for domestic residential and commercial applications using commercially available fuels.

EN 50379-2 is for apparatus intended to be used for statutory measurement. In several European countries, legal requirements exist for the performance of heating appliances. Authorised inspectors use these apparatus to measure the flue gas parameters, in order to test the compliance with national regulations. Due to the legal consequences resulting from the measurement there are strict requirements regarding the measuring uncertainty of these apparatus. Therefore EN 50379-2 includes maximum values for measuring uncertainty of the apparatus. Tests with real flue gases form a key part of the verification of the performance of the apparatus for statutory measurement. The measuring uncertainty has to be justified by internationally accepted methods over the whole measuring range.

EN 50379-3 is for apparatus intended to be used for non-statutory applications. There are reduced performance requirements, because the apparatus are designed to decide whether maintenance for a gas fired appliance is required, and for adjusting the appliance during maintenance. There will be no determination of the measuring uncertainty for the apparatus.

1 Scope

This European Standard covers apparatus designed for checking the performance of heating appliances by measuring flue gas parameters of gas fired heating appliances for domestic residential and commercial applications.

The apparatus may consist of different functional modules which may be tested separately for complying with this standard, and will be combined in different ways according to the different applications. Part 1 of EN 50379 specifies the general requirements and is supplemented by the requirements in EN 50379-2 and/or EN 50379-3.

This European Standard specifies the performance requirements of portable spot reading apparatus designed to detect specific combustion flue gas parameters, such as concentration of gaseous compounds, temperature and/or pressure, to be used to decide if maintenance for the appliance is required and for adjusting the appliance during maintenance.

This European Standard excludes apparatus for

- checking appliances using fuels other than gas,
- continuous emission, safety monitoring and control, and
- use in vessels with an international load line.

NOTE 1 When this apparatus is used in industrial premises, national regulations should be observed.

NOTE 2 Apparatus may contain functional modules which are not covered by this standard e.g. measurement of smoke spot number (see EN 267:2009+A1:2011, Annex A) and/or measurement of indoor ambient air (see EN 50543).

2 Normative references

The following documents, in whole or in part, are normatively referenced in this document and are indispensable for its application. For dated references, only the edition cited applies. For undated references, the latest edition of the referenced document (including any amendments) applies.

EN 50270:2006, *Electromagnetic compatibility – Electrical apparatus for the detection and measurement of combustible gases, toxic gases or oxygen*

EN 50271:2010, *Electrical apparatus for the detection and measurement of combustible gases, toxic gases or oxygen – Requirements and tests for apparatus using software and/or digital technologies*

EN 50379-1:2012, *Specification for portable electrical apparatus designed to measure combustion flue gas parameters of heating appliances – Part 1: General requirements and test methods*

EN 60335-1:2002, *Safety of household and similar electrical appliances – Part 1: General requirements (IEC 60335-1:2001, mod.)*

EN 60529:1991, *Degrees of protection provided by enclosures (IP Code) (IEC 60529:1989)*

3 Terms and definitions

For the purposes of this document, the terms and definitions given in EN 50379-1:2012 apply.

4 General requirements

Unless otherwise stated, the general requirements of EN 50379-1:2012 are applicable and shall be checked by visual inspection.

5 Test methods and performance requirements

5.1 General requirements for tests

5.1.1 General

The requirements specified in EN 50379-1:2012, 5.1.1 are applicable.

5.1.2 Samples and sequence of tests

For the purposes of type testing, for each sample the first test to be carried out shall be the initial performance test (5.5.3). One sample of the apparatus shall be subjected to the relevant tests given in 5.3 and 5.4. The mechanical tests of 5.3 shall be carried out before the tests of 5.4 to 5.9. After finishing the mechanical tests of 5.3 the initial performance test (5.5.3) shall be repeated at room temperature.

NOTE Due to the use of different calibration gases at pre-calibration by the manufacturer and type testing according to this standard the instruments may show a discrepancy. In this case, it is permissible to make minor adjustments to the apparatus during the initial performance test (5.5.3).

The EMC testing according to 5.4.1 may be carried out independently from all other tests of the standard.

The tests within 5.3, 5.4 and 5.5 may be performed in any sequence, but 5.6 to 5.9 shall be performed in the sequence listed in this standard, where relevant.

5.1.3 Preparation of samples

The requirements specified in EN 50379-1:2012, 5.1.3 are applicable.

5.1.4 Test facility

The requirements specified in EN 50379-1:2012, 5.1.4 are applicable.

5.2 Normal conditions for tests

The requirements specified in EN 50379-1:2012, 5.2 are applicable.

5.3 Mechanical tests

5.3.1 General

The requirements specified in EN 50379-1:2012, 5.3.1 are applicable

5.3.2 Degree of protection

The enclosure of the apparatus shall provide at least an IP40 degree of protection when all probes etc. are connected, in accordance with EN 60529:1991, Clauses 12 and 14. If an apparatus is designed for outdoor use it shall provide at least an IP42 degree of protection when connected similarly.

5.3.3 Impact strength

The apparatus shall meet the requirements specified in EN 60335-1:2002, Clause 21, as modified by EN 50379-1:2012, 5.3.3. The function of the apparatus shall not be affected after the test. Visible damage to parts of the housing are acceptable, providing the functionality remains unimpaired.