



SLOVENSKI STANDARD SIST EN ISO 11137-2:2006

01-julij-2006

BUKca Yý U
OSIST prEN ISO 11137-2:2004
SIST EN 552:2000
SIST EN 552:2000/A1:2000
SIST EN 552:2000/A2:2001

ghfj]nUW^U]nXY_cj`nUnXfUj ghj Ybc`bY[c`!`GYj Ub^Y!`&^XY.`8c`c Ub^Y`cXa Yf_U
ghfj]nUW^Y fIGC`%%%`+!&.\$\$*L

Sterilization of health care products - Radiation - Part 2: Establishing the sterilization dose (ISO 11137-2:2006)

standards.iteh.ai

Sterilisation von Produkten für die Gesundheitsfürsorge - Strahlen - Teil 2: Festlegung der Sterilisationsdosis (ISO 11137-2:2006)

[standards/sist/171c206f-ae00-439f-a206-3f7911ca1973/sist-en-iso-11137-2-2006](http://standards.sist/171c206f-ae00-439f-a206-3f7911ca1973/sist-en-iso-11137-2-2006)

Stérilisation des produits de santé - Irradiation - Partie 2: Établissement de la dose de stérilisation (ISO 11137-2:2006)

Ta slovenski standard je istoveten z: EN ISO 11137-2:2006

ICS:

11.080.01 Sterilizacija in dezinfekcija na splošno Sterilization and disinfection in general

SIST EN ISO 11137-2:2006 en

iTeh STANDARD PREVIEW
(standards.iteh.ai)

SIST EN ISO 11137-2:2006

<https://standards.iteh.ai/catalog/standards/sist/171c206f-ae00-439f-a206-3f7911ca1973/sist-en-iso-11137-2-2006>

English Version

**Sterilization of health care products - Radiation - Part 2:
Establishing the sterilization dose (ISO 11137-2:2006)**

Stérilisation des produits de santé - Irradiation - Partie 2:
Établissement de la dose de stérilisation (ISO 11137-
2:2006)

Sterilisation von Produkten für die Gesundheitsfürsorge -
Strahlen - Teil 2: Festlegung der Sterilisationsdosis (ISO
11137-2:2006)

This European Standard was approved by CEN on 13 April 2006.

CEN members are bound to comply with the CEN/CENELEC Internal Regulations which stipulate the conditions for giving this European Standard the status of a national standard without any alteration. Up-to-date lists and bibliographical references concerning such national standards may be obtained on application to the Central Secretariat or to any CEN member.

This European Standard exists in three official versions (English, French, German). A version in any other language made by translation under the responsibility of a CEN member into its own language and notified to the Central Secretariat has the same status as the official versions.

CEN members are the national standards bodies of Austria, Belgium, Cyprus, Czech Republic, Denmark, Estonia, Finland, France, Germany, Greece, Hungary, Iceland, Ireland, Italy, Latvia, Lithuania, Luxembourg, Malta, Netherlands, Norway, Poland, Portugal, Romania, Slovakia, Slovenia, Spain, Sweden, Switzerland and United Kingdom.

[SIST EN ISO 11137-2:2006](https://standards.iteh.ai/catalog/standards/sist/171c206f-ae00-439f-a206-3f7911ca1973/sist-en-iso-11137-2-2006)

<https://standards.iteh.ai/catalog/standards/sist/171c206f-ae00-439f-a206-3f7911ca1973/sist-en-iso-11137-2-2006>



EUROPEAN COMMITTEE FOR STANDARDIZATION
COMITÉ EUROPÉEN DE NORMALISATION
EUROPÄISCHES KOMITEE FÜR NORMUNG

Management Centre: rue de Stassart, 36 B-1050 Brussels

Foreword

This document (EN ISO 11137-2:2006) has been prepared by Technical Committee ISO/TC 198 "Sterilization of health care products" in collaboration with Technical Committee CEN/TC 204 "Sterilization of medical devices", the secretariat of which is held by BSI.

This European Standard shall be given the status of a national standard, either by publication of an identical text or by endorsement, at the latest by October 2006, and conflicting national standards shall be withdrawn at the latest by April 2009.

This document supersedes EN 552:1994.

This document has been prepared under a mandate given to CEN by the European Commission and the European Free Trade Association, and supports essential requirements of EU Directive(s).

For relationship with EU Directive(s), see informative Annex ZA, which is an integral part of this document.

According to the CEN/CENELEC Internal Regulations, the national standards organizations of the following countries are bound to implement this European Standard: Austria, Belgium, Cyprus, Czech Republic, Denmark, Estonia, Finland, France, Germany, Greece, Hungary, Iceland, Ireland, Italy, Latvia, Lithuania, Luxembourg, Malta, Netherlands, Norway, Poland, Portugal, Romania, Slovakia, Slovenia, Spain, Sweden, Switzerland and United Kingdom.

(standards.iteh.ai)

Endorsement notice

[SIST EN ISO 11137-2:2006](https://standards.iteh.ai/catalog/standards/sist/en-iso-11137-2-2006)

The text of ISO 11137-2:2006 has been approved by CEN as EN ISO 11137-2:2006 without any modifications.

<https://standards.iteh.ai/catalog/standards/sist/en-iso-11137-2-2006>
3f7911ca1973/sist-en-iso-11137-2-2006

ANNEX ZA

(informative)

Relationship between this European Standard and the Essential Requirements of EU Directives 90/385/EEC concerning active implantable medical devices, 93/42/EEC concerning medical devices and 98/79/EEC concerning *in vitro* diagnostic medical devices

This European Standard has been prepared under a mandate given to CEN by the European Commission and the European Free Trade Association to provide one means of conforming to Essential Requirements of the New Approach Directive, EU Directives 90/385/EEC concerning active implantable medical devices, 93/42/EEC concerning medical devices and 98/79/EEC concerning *in vitro* diagnostic medical devices.

Once this standard is cited in the Official Journal of the European Communities under that Directive and has been implemented as a national standard in at least one Member State, compliance with the normative clauses of this standard given in Table ZA.1 confers, within the limits of the scope of this standard, a presumption of conformity with the corresponding Essential Requirements of that Directive and associated EFTA regulations.

Table ZA.1 — Correspondence between this European Standard and EU Directives 90/385/EEC concerning active implantable medical devices, 93/42/EEC concerning medical devices and 98/79/EEC concerning *in vitro* diagnostic medical devices

Clause(s)/Sub-clause(s) of this European Standard	Essential Requirements (ERs) of Directive 90/385/EEC	Essential Requirements (ERs) of Directive 93/42/EEC	Essential Requirements (ERs) of Directive 98/79/EEC	Qualifying remarks/Notes
4, 5, 6, 7, 8, 9, 10, 11, 12	7	8.3	B.2.3	In part
4, 5, 6, 7, 8, 9, 10, 11, 12		8.4	B.2.4	In part

WARNING: Other requirements and other EU Directives may be applicable to the product(s) falling within the scope of this standard.

iTeh STANDARD PREVIEW
(standards.iteh.ai)

SIST EN ISO 11137-2:2006

<https://standards.iteh.ai/catalog/standards/sist/171c206f-ae00-439f-a206-3f7911ca1973/sist-en-iso-11137-2-2006>

**Sterilization of health care products —
Radiation —**

Part 2:
Establishing the sterilization dose

Stérilisation des produits de santé — Irradiation —

Partie 2: Établissement de la dose stérilisante

iTeh STANDARD PREVIEW
(standards.iteh.ai)

SIST EN ISO 11137-2:2006

<https://standards.iteh.ai/catalog/standards/sist/171c206f-ae00-439f-a206-3f7911ca1973/sist-en-iso-11137-2-2006>



PDF disclaimer

This PDF file may contain embedded typefaces. In accordance with Adobe's licensing policy, this file may be printed or viewed but shall not be edited unless the typefaces which are embedded are licensed to and installed on the computer performing the editing. In downloading this file, parties accept therein the responsibility of not infringing Adobe's licensing policy. The ISO Central Secretariat accepts no liability in this area.

Adobe is a trademark of Adobe Systems Incorporated.

Details of the software products used to create this PDF file can be found in the General Info relative to the file; the PDF-creation parameters were optimized for printing. Every care has been taken to ensure that the file is suitable for use by ISO member bodies. In the unlikely event that a problem relating to it is found, please inform the Central Secretariat at the address given below.

iTeh STANDARD PREVIEW
(standards.iteh.ai)

[SIST EN ISO 11137-2:2006](https://standards.iteh.ai/catalog/standards/sist/171c206f-ae00-439f-a206-3f7911ca1973/sist-en-iso-11137-2-2006)

<https://standards.iteh.ai/catalog/standards/sist/171c206f-ae00-439f-a206-3f7911ca1973/sist-en-iso-11137-2-2006>

© ISO 2006

All rights reserved. Unless otherwise specified, no part of this publication may be reproduced or utilized in any form or by any means, electronic or mechanical, including photocopying and microfilm, without permission in writing from either ISO at the address below or ISO's member body in the country of the requester.

ISO copyright office
Case postale 56 • CH-1211 Geneva 20
Tel. + 41 22 749 01 11
Fax + 41 22 749 09 47
E-mail copyright@iso.org
Web www.iso.org

Published in Switzerland

Contents

Page

Foreword.....	v
Introduction	vi
1 Scope	1
2 Normative references	1
3 Abbreviations, terms and definitions	1
3.1 Abbreviations	1
3.2 Terms	3
4 Definition and maintenance of product families for dose setting, dose substantiation and sterilization dose auditing	4
4.1 General.....	4
4.2 Defining product families.....	4
4.3 Designation of product to represent a product family for performance of a verification dose experiment or sterilization dose audit	5
4.4 Maintaining product families	6
4.5 Effect of failure of establishment of sterilization dose or of a sterilization dose audit on a product family.....	7
5 Selection and testing of product for establishing and verifying the sterilization dose	7
5.1 Nature of product.....	7
5.2 Sample item portion (SIP)	8
5.3 Manner of sampling	8
5.4 Microbiological testing	9
5.5 Irradiation	9
6 Methods of dose establishment.....	9
7 Method 1: dose setting using bioburden information	10
7.1 Rationale.....	10
7.2 Procedure for Method 1 for product with an average bioburden $\geq 1,0$ for multiple production batches.....	11
7.3 Procedure for Method 1 for product with an average bioburden $\geq 1,0$ for a single production batch.....	16
7.4 Procedure for Method 1 for product with an average bioburden in the range 0,1 to 0,9 for multiple or single production batches.....	18
8 Method 2: Dose setting using fraction positive information from incremental dosing to determine an extrapolation factor.....	18
8.1 Rationale.....	18
8.2 Procedure for Method 2A.....	19
8.3 Procedure for Method 2B.....	22
9 Method VD_{max} — Substantiation of 25 kGy or 15 kGy as the sterilization dose	25
9.1 Rationale.....	25
9.2 Procedure for Method VD _{max} ²⁵ for multiple production batches.....	26
9.3 Procedure for Method VD _{max} ²⁵ for a single production batch	29
9.4 Procedure for Method VD _{max} ¹⁵ for multiple production batches.....	30
9.5 Procedure for Method VD _{max} ¹⁵ for a single production batch	33
10 Auditing sterilization dose.....	34
10.1 Purpose and frequency.....	34

10.2	Procedure for auditing a sterilization dose established using Method 1 or Method 2.....	35
10.3	Procedure for auditing a sterilization dose substantiated using VD_{max}	37
11	Worked examples.....	41
11.1	Worked examples for Method 1.....	41
11.2	Worked examples for Method 2.....	44
11.3	Worked examples for Method VD_{max}	53
11.4	Worked example of a sterilization dose audit for a dose established using Method 1, the findings from which necessitated augmentation of the sterilization dose.....	55
11.5	Worked example of a sterilization dose audit for a dose established using Method 2A, the findings from which necessitated augmentation of the sterilization dose.....	56
11.6	Worked example of a sterilization dose audit for a sterilization dose substantiated using Method VD_{max}^{25}	58
	Bibliography.....	59

iTeh STANDARD PREVIEW
(standards.iteh.ai)

[SIST EN ISO 11137-2:2006](https://standards.iteh.ai/catalog/standards/sist/171c206f-ae00-439f-a206-3f7911ca1973/sist-en-iso-11137-2-2006)
<https://standards.iteh.ai/catalog/standards/sist/171c206f-ae00-439f-a206-3f7911ca1973/sist-en-iso-11137-2-2006>

Foreword

ISO (the International Organization for Standardization) is a worldwide federation of national standards bodies (ISO member bodies). The work of preparing International Standards is normally carried out through ISO technical committees. Each member body interested in a subject for which a technical committee has been established has the right to be represented on that committee. International organizations, governmental and non-governmental, in liaison with ISO, also take part in the work. ISO collaborates closely with the International Electrotechnical Commission (IEC) on all matters of electrotechnical standardization.

International Standards are drafted in accordance with the rules given in the ISO/IEC Directives, Part 2.

The main task of technical committees is to prepare International Standards. Draft International Standards adopted by the technical committees are circulated to the member bodies for voting. Publication as an International Standard requires approval by at least 75 % of the member bodies casting a vote.

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. ISO shall not be held responsible for identifying any or all such patent rights.

ISO 11137-2 was prepared by Technical Committee ISO/TC 198, *Sterilization of health care products*.

This first edition, together with ISO 11137-1 and ISO 11137-3, cancels and replaces ISO 11137:1995.

ISO 11137 consists of the following parts under the general title *Sterilization of health care products — Radiation*:

- *Part 1: Requirements for development, validation and routine control of a sterilization process for medical devices*
- *Part 2: Establishing the sterilization dose*
- *Part 3: Guidance on dosimetric aspects*

Introduction

This part of ISO 11137 describes methods that may be used to establish the sterilization dose in accordance with one of the two approaches specified in 8.2 of ISO 11137-1:2006. The methods used in these approaches are:

- a) dose setting to obtain a product-specific dose;
- b) dose substantiation to verify a preselected dose of 25 kGy or 15 kGy.

The basis of the dose setting methods described in this part of ISO 11137 (Methods 1 and 2) owe much to the ideas first propounded by Tallentire (Tallentire, 1973 [17]; Tallentire, Dwyer and Ley, 1971 [18]; Tallentire and Khan, 1978 [19]). Subsequently, standardized protocols were developed (Davis *et al.*, 1981 [8]; Davis, Strawderman and Whitby, 1984 [9]) which formed the basis of the dose setting methods detailed in the AAMI *Recommended Practice for Sterilization by Gamma Radiation* (AAMI 1984, 1991 [4], [6]).

Methods 1 and 2 and the associated sterilization dose audit procedures use data derived from the inactivation of the microbial population in its natural state on product. The methods are based on a probability model for the inactivation of microbial populations. The probability model, as applied to bioburden made up of a mixture of various microbial species, assumes that each such species has its own unique D_{10} value. In the model, the probability that an item will possess a surviving microorganism after exposure to a given dose of radiation is defined in terms of the initial number of microorganisms on the item prior to irradiation and the D_{10} values of the microorganisms. The methods involve performance of tests of sterility on product items that have received doses of radiation lower than the sterilization dose. The outcome of these tests is used to predict the dose needed to achieve a predetermined sterility assurance level, SAL.

Methods 1 and 2 may also be used to substantiate 25 kGy if, on performing a dose setting exercise, the derived sterilization dose for an SAL of 10^{-6} is ≤ 25 kGy. The basis of the method devised specifically for substantiation of 25 kGy, Method VD_{max} , was put forward by Kowalski and Tallentire (1999) [14]. Subsequent evaluations involving computational techniques demonstrated that the underlying principles were soundly based (Kowalski, Aoshuang and Tallentire, 2000) [13] and field trials confirmed that Method VD_{max} is effective in substantiating 25 kGy for a wide variety of medical devices manufactured and assembled in different ways (Kowalski *et al.*, 2002) [16].

A standardized procedure for the use of VD_{max} for substantiation of a sterilization dose of 25 kGy has been published in the AAMI Technical Information Report *Sterilization of health care products — Radiation sterilization — Substantiation of 25 kGy as a sterilization dose — Method VD_{max}* (AAMI TIR27:2001) [5], a text on which the method described herein is largely based. Method VD_{max} is founded on dose setting Method 1 and, as such, it possesses the high level of conservativeness characteristic of Method 1. In a similar manner to the dose setting methods, it involves performance of tests of sterility on product items that have received a dose of radiation lower than the sterilization dose. The outcomes of these tests are used to substantiate that 25 kGy achieves an SAL of 10^{-6} .

To link the use of VD_{max} for the substantiation of a particular preselected sterilization dose, the numerical value of the latter, expressed in kGy, is included as a superscript to the VD_{max} symbol. Thus, for substantiation of a sterilization dose of 25 kGy the method is designated VD_{max}^{25} .

Method VD_{max}^{15} is based on the same principles as Method VD_{max}^{25} described above. The test procedure is the same as Method VD_{max}^{25} , but VD_{max}^{15} is limited to product with average bioburden ≤ 1.5 . The outcomes of these tests are used to substantiate that 15 kGy achieves a sterility assurance level of 10^{-6} .

This part of ISO 11137 also describes methods that may be used to carry out sterilization dose audits in accordance with ISO 11137-1:2006, Clause 12. Following establishment of the sterilization dose, sterilization dose audits are performed routinely to confirm that the sterilization dose continues to achieve the desired SAL.

Sterilization of health care products — Radiation —

Part 2: Establishing the sterilization dose

1 Scope

This part of ISO 11137 specifies methods of determining the minimum dose needed to achieve a specified requirement for sterility and methods to substantiate the use of 25 kGy or 15 kGy as the sterilization dose to achieve a sterility assurance level, SAL, of 10^{-6} . This part of ISO 11137 also specifies methods of dose auditing in order to demonstrate the continued effectiveness of the sterilization dose.

This part of ISO 11137 defines product families for dose establishment and dose auditing.

2 Normative references

The following referenced documents are indispensable for the application of this document. For dated references, only the edition cited applies. For undated references, the latest edition of the referenced document (including any amendments) applies.

ISO 11137-1:2006, *Sterilization of health care products — Radiation — Part 1: Requirements for the development, validation and routine control of a sterilization process for medical devices*

ISO 11737-1, *Sterilization of medical devices — Microbiological methods — Part 1: Determination of the population of microorganisms on product*

ISO 11737-2, *Sterilization of medical devices — Microbiological methods — Part 2: Test of sterility performed in the validation of a sterilization process*

ISO 13485, *Medical devices — Quality management systems — Requirements for regulatory purposes*

3 Abbreviations, terms and definitions

For purposes of this document, the terms and definitions given in ISO 11137-1 and the following apply.

3.1 Abbreviations

3.1.1

A

dose to adjust the median ffp dose downwards, to the FFP dose

3.1.2

*CD**

number of positive tests of sterility obtained from tests performed individually on 100 product items irradiated in a Method 2 verification dose experiment

3.1.3

*d**
dose derived from an incremental dose experiment performed on product items drawn from a given production batch

3.1.4

*D**
initial estimate of the dose to provide an SAL of 10^{-2} for the test items

NOTE Generally, it is the median of the 3 *d** values derived for a given product.

3.1.5

*D***
final estimate of the dose to provide an SAL of 10^{-2} for the test items, which is used in the calculation of the sterilization dose

3.1.6

*DD**
dose delivered in a Method 2 verification dose experiment

3.1.7

DS
estimate of D_{10} value of microorganisms present on product after exposure to *DD**

3.1.8

D value
D₁₀ value
time or dose required to achieve inactivation of 90 % of a population of the test microorganism under stated conditions

iTeh STANDARD PREVIEW
(standards.iteh.ai)

[ISO/TS 11139:2006]

SIST EN ISO 11137-2:2006
<https://standards.iteh.ai/catalog/standards/sist/171c206f-ae00-439f-a206-375911ee1973/sist-en-iso-11137-2-2006>

NOTE For the purposes of this document, D_{10} applies to the radiation dose only and not to time.

3.1.9

first fraction positive dose

ffp
lowest dose of an incremental dose series, applied to product items drawn from a given production batch, at which at least one of the associated 20 tests of sterility is negative

3.1.10

First Fraction Positive dose

FFP
dose at which 19 positives out of the 20 tests of sterility are expected to occur, calculated by subtracting *A* from the median of 3 ffp doses

3.1.11

First No Positive dose

FNP
estimate of the dose to provide an SAL of 10^{-2} for the test items, which is used in the calculation of *DS*

3.1.12

VD_{max}¹⁵
maximal verification dose for a given bioburden, consistent with the attainment of an SAL of 10^{-6} at a specified sterilization dose of 15 kGy

3.1.13

VD_{max}²⁵
maximal verification dose for a given bioburden, consistent with the attainment of an SAL of 10^{-6} at a specified sterilization dose of 25 kGy

3.2 Terms

3.2.1

batch

defined quantity of product, intended or purported to be uniform in character and quality, which has been produced during a defined cycle of manufacture

[ISO/TS 11139:2006]

3.2.2

bioburden

population of viable microorganisms on or in product and/or sterile barrier system

[ISO/TS 11139:2006]

3.2.3

false positive

test result interpreted as growth arising from the product, or portions thereof, tested when either growth resulted from extraneous microbial contamination or turbidity occurred from interaction between the product, or portions thereof, and the test medium

3.2.4

fraction positive

quotient in which the number of positive tests of sterility is given by the numerator and the number of tests performed is given by the denominator

3.2.5

incremental dose

dose within a series of doses applied to a number of product, or portions thereof, and used in a dose setting method to obtain or confirm the sterilization dose

3.2.6

negative test of sterility

test result for which there is no detectable microbial growth from product, or portion thereof, subjected to a test of sterility

3.2.7

packaging system

combination of the sterile barrier system and protective packaging

[ISO/TS 11139:2006]

3.2.8

positive test of sterility

test result for which there is detectable microbial growth from product, or portion thereof, subjected to a test of sterility

3.2.9

sample item portion

SIP

defined portion of a health care product that is tested

3.2.10

sterile barrier system

minimum package that prevents ingress of microorganisms and allows aseptic presentation of product at the point of use

iTeh STANDARD PREVIEW
(standards.iteh.ai)

<https://standards.iteh.ai/catalog/standards/sist/171c206f-ae00-439f-a206-3f7911ca1973/sist-en-iso-11137-2-2006>