



**SLOVENSKI STANDARD**  
**SIST ISO 14235:1999**  
**01-marec-1999**

Učinek organske snovi na kislinsko kapaciteto talne kislinskega tona

Soil quality -- Determination of organic carbon by sulfochromic oxidation

Qualité du sol -- Dosage du carbone organique par oxydation sulfochromique

ITEH STANDARD PREVIEW  
 (standards.iteh.ai)

**Ta slovenski standard je istoveten z: ISO 14235:1998**

SIST ISO 14235:1999

<https://standards.iteh.ai/catalog/standards/sist/215eea8b-6056-4b77-942d-e4907c003860/sist-iso-14235-1999>

**ICS:**

13.080.10      Sestava talne kislinskega tona      Chemical characteristics of soils

**SIST ISO 14235:1999**      en

**iTeh STANDARD PREVIEW**  
**(standards.iteh.ai)**

SIST ISO 14235:1999

<https://standards.iteh.ai/catalog/standards/sist/215eea8b-6056-4b77-942d-e4907c003860/sist-iso-14235-1999>

INTERNATIONAL  
STANDARD

ISO  
14235

First edition  
1998-08-01

---

---

**Soil quality — Determination of organic  
carbon by sulfochromic oxidation**

*Qualité du sol — Dosage du carbone organique par oxydation  
sulfochromique*

**iTeh STANDARD PREVIEW**  
**(standards.iteh.ai)**

[SIST ISO 14235:1999](https://standards.iteh.ai/catalog/standards/sist/215eea8b-6056-4b77-942d-e4907c003860/sist-iso-14235-1999)

[https://standards.iteh.ai/catalog/standards/sist/215eea8b-6056-4b77-942d-  
e4907c003860/sist-iso-14235-1999](https://standards.iteh.ai/catalog/standards/sist/215eea8b-6056-4b77-942d-e4907c003860/sist-iso-14235-1999)



Reference number  
ISO 14235:1998(E)

## ISO 14235:1998(E)

## Contents

1	Scope .....	1
2	Normative references .....	1
3	Principle .....	1
4	Reagents .....	2
5	Apparatus and glassware .....	2
6	Laboratory sample .....	2
7	Procedure .....	3
8	Calculation of organic carbon content .....	4
9	Repeatability .....	4
10	Test report .....	5

iTeh STANDARD PREVIEW  
 (standards.iteh.ai)  
 SIST ISO 14235:1999  
<https://standards.iteh.ai/catalog/standards/sist/215eea8b-6056-4b77-942d-e4907c003860/sist-iso-14235-1999>

© ISO 1998

All rights reserved. Unless otherwise specified, no part of this publication may be reproduced or utilized in any form or by any means, electronic or mechanical, including photocopying and microfilm, without permission in writing from the publisher.

International Organization for Standardization  
 Case postale 56 • CH-1211 Genève 20 • Switzerland  
 Internet iso@iso.ch

Printed in Switzerland

## Foreword

ISO (the International Organization for Standardization) is a worldwide federation of national standards bodies (ISO member bodies). The work of preparing International Standards is normally carried out through ISO technical committees. Each member body interested in a subject for which a technical committee has been established has the right to be represented on that committee. International organizations, governmental and non-governmental, in liaison with ISO, also take part in the work. ISO collaborates closely with the International Electrotechnical Commission (IEC) on all matters of electrotechnical standardization.

Draft International Standards adopted by the technical committees are circulated to the member bodies for voting. Publication as an International Standard requires approval by at least 75 % of the member bodies casting a vote.

International Standard ISO 14235 was prepared by Technical Committee ISO/TC 190, *Soil quality*, Subcommittee SC 3, *Chemical methods and soil characteristics*.

## iTeh STANDARD PREVIEW (standards.iteh.ai)

[SIST ISO 14235:1999](https://standards.iteh.ai/catalog/standards/sist/215eea8b-6056-4b77-942d-e4907c003860/sist-iso-14235-1999)

<https://standards.iteh.ai/catalog/standards/sist/215eea8b-6056-4b77-942d-e4907c003860/sist-iso-14235-1999>

**iTeh STANDARD PREVIEW**  
**(standards.iteh.ai)**

SIST ISO 14235:1999

<https://standards.iteh.ai/catalog/standards/sist/215eea8b-6056-4b77-942d-e4907c003860/sist-iso-14235-1999>

# Soil quality — Determination of organic carbon in soil by sulfochromic oxidation

## 1 Scope

This International Standard specifies a method for the spectrometric determination of organic carbon content in soil by oxidation in a sulfochromic medium.

This International Standard is applicable to all types of air-dry soil samples.

This International Standard is not applicable to soils containing mineral-reducing compounds, e.g.  $\text{Cl}^-$  or  $\text{Fe}^{2+}$ . By convention the chloride content in the test portion shall not exceed 2 mg.

(standards.iteh.ai)

## 2 Normative references

SIST ISO 14235:1999

The following standards contain provisions which, through reference in this text, constitute provisions of this International Standard. At the time of publication, the editions indicated were valid. All standards are subject to revisions, and parties to agreements based on this International Standard are encouraged to investigate the possibility of applying the most recent additions of the standard indicated below. Members of ICE and ISO maintain registers of currently valid International Standards.

ISO 3696:1987, *Water for analytical use — Specification and test methods*.

ISO 10694:1995, *Soil quality — Determination of organic and total carbon after dry combustion (elemental analysis)*.

ISO 11464:1994, *Soil quality — Pretreatment of samples for physico-chemical analysis*.

ISO 11465:1993, *Soil quality — Determination of dry matter and water content on a mass basis — Gravimetric method*.

## 3 Principle

The organic carbon present in the soil is oxidized in a mixture of potassium dichromate solution (in excess) and sulfuric acid at a temperature of 135 °C. The dichromate ions, which colour the solution orange-red, are reduced to  $\text{Cr}^{3+}$  ions which colour the solution green. The intensity of this green colour is measured spectrophotometrically. As it is assumed that the oxidation of one carbon atom of the organic matter produces four electrons, there is a direct relationship between the  $\text{Cr}^{3+}$  formed and the amount of organic carbon. The method is calibrated using glucose as a source of readily oxidizable carbon.

NOTE An amount of 2 mg of chloride present in the test sample corresponds to an apparent carbon content of about 0,7 mg.

## 4 Reagents

Use only reagents of recognized analytical grade and use distilled or deionized water for all solutions.

**4.1 Water**, having a specific conductivity not higher than 0,2 mS/m at 25 °C (grade 2 water in accordance with ISO 3696).

**4.2 Sulfuric acid**,  $\text{H}_2\text{SO}_4$ , concentrated ( $\rho = 1,84 \text{ g/cm}^3$ ).

**4.3 Potassium dichromate solution**,  $c(\text{K}_2\text{Cr}_2\text{O}_7) = 0,27 \text{ mol/l}$ .

Dissolve 80 g of potassium dichromate,  $\text{K}_2\text{Cr}_2\text{O}_7$ , in ca. 800 ml of water (4.1) in a 1000 ml volumetric flask and make up to the mark with water (4.1).

**WARNING:** The chromate ion present in potassium dichromate is potentially toxic. Laboratory personnel working with this chemical should take appropriate precautions to avoid contact with, or ingestion of, this chemical, following national or international safety regulations where applicable. Such regulations can extend to the disposal of solutions containing chromate ions, as these may also damage the environment. If in doubt, seek professional advice.

**4.4 Glucose, anhydrous**,  $\text{C}_6\text{H}_{12}\text{O}_6$ .

## 5 Apparatus and glassware

**5.1 Standard laboratory glassware.**

**5.2 Analytical balance**, capable of weighing accurately to 0,1 mg.

**5.3 Volumetric glass tubes** of 75,0 ml  $\pm$  0,2 ml, adapted to fit the heating block (5.4).

**5.4 Heating block**, capable of maintaining a uniform temperature of 135 °C  $\pm$  2 °C.

The block shall be drilled with holes such that, if  $h$  is the height, in millimetres, of the sulfochromic solution in the tube (5.3), the depth of the holes shall be at least  $(h + 10)$  mm and the upper part of the tube emerging from the holes shall be at least 150 mm long. The temperature shall be measured with the tubes containing oxidation solution and a thermometer in the holes of the block. To minimize heterogeneity of the temperature within the block, all holes not utilized by tubes containing samples shall be occupied by tubes containing 15 ml of sulfuric acid (4.2).

**5.5 Centrifuge.**

**5.6 Glass fibre filters.**

**5.7 Automatic pipette.**

**5.8 Spectrophotometer**, equipped with a 10 mm cuvette and adjusted to a wavelength of 585 nm.

**5.9 Water bath.**

## 6 Laboratory sample

Use the fraction of particles < 2 mm of air-dry soil samples pretreated in accordance with ISO 11464. Use part of the sample to determine the water content in accordance with ISO 11465. For the determination of carbon, a representative subsample of the laboratory sample shall be milled until it passes a 250  $\mu\text{m}$  aperture sieve in accordance with ISO 11464.



## 7 Procedure

### 7.1 General

The efficiency of the oxidation depends on the mass of the test portion, as well as the mass of carbon within it, even if the potassium dichromate remains in excess. Experience has shown that, under the conditions given in this International Standard, the mass of carbon in the test portion should not exceed 20 mg. Therefore the mass of the test portion used shall be in accordance with table 1.

**Table 1 — Mass of test portion in relation with the estimated carbon content of the soil**

<b>Estimated carbon content</b> g/kg	0 to 40	40 to 80	80 to 160	160 to 400	> 400
<b>Mass of the test portion</b> mg	400 to 500	200 to 250	100 to 125	45 to 50	20 to 25

## iTeh STANDARD PREVIEW (standards.iteh.ai)

### 7.2 Oxidation

Transfer the test portion to a volumetric tube (5.3). Using a pipette (5.7), add first 5,0 ml of the dichromate solution (4.3) and then 7,5 ml of sulfuric acid (4.2). Homogenize carefully.

Place the tubes in the heating block (5.4) which should be preheated to a temperature of 135 °C. Leave the tubes in the heated block for 30 min, then remove them and cool rapidly to room temperature in the water bath (5.9). Slowly add about 50 ml of water (4.1) to each tube and again cool rapidly in the water bath. Transfer the contents of the tube quantitatively into a 100 ml volumetric flask, make up to volume with water (4.1) and mix thoroughly.

Allow the suspension to settle for 1 h. Decant part of the supernatant into a centrifuge tube and centrifuge (5.5) for 10 min at 2000 g. If solid particles still remain in the suspension after centrifugation, filter the supernatant (5.6).

**CAUTION:** The addition of water to sulfuric acid solution is potentially hazardous, as is the use of dichromate solutions. Suitable protection, e.g. gloves, safety mask or goggles, and laboratory coat is essential. If in doubt, obtain professional advice.

### 7.3 Preparation of the standard series

Prepare a standard series according to table 2 by dissolving the mass of glucose indicated in 5,00 ml of dichromate solution (4.3) in a series of 100 ml volumetric tubes and mix to dissolve completely.

Add 7,5 ml of sulfuric acid as in 7.2 and continue the digestion step accordingly. The resulting solution should be completely clear, and should not require centrifugation or filtration. If this is not so, then the digestion has failed and shall be repeated.