

INTERNATIONAL STANDARD

NORME INTERNATIONALE

AMENDMENT 1
AMENDEMENT 1

Ultrasonics – Field characterization – Test methods for the determination of thermal and mechanical indices related to medical diagnostic ultrasonic fields

Ultrasons – Caractérisation du champ – Méthodes d'essai pour la détermination d'indices thermique et mécanique des champs d'ultrasons utilisés pour le diagnostic médical

ITIH STANDARD PREVIEW
(standards.iteh.ai)

IEC 62359-2010/AMD1:2017
<https://standards.iteh.ai/catalog/standards/sist/6c88226f-0136-4667-a1cc-071d05236ca3/iec-62359-2010-amd1-2017>





THIS PUBLICATION IS COPYRIGHT PROTECTED

Copyright © 2017 IEC, Geneva, Switzerland

All rights reserved. Unless otherwise specified, no part of this publication may be reproduced or utilized in any form or by any means, electronic or mechanical, including photocopying and microfilm, without permission in writing from either IEC or IEC's member National Committee in the country of the requester. If you have any questions about IEC copyright or have an enquiry about obtaining additional rights to this publication, please contact the address below or your local IEC member National Committee for further information.

Droits de reproduction réservés. Sauf indication contraire, aucune partie de cette publication ne peut être reproduite ni utilisée sous quelque forme que ce soit et par aucun procédé, électronique ou mécanique, y compris la photocopie et les microfilms, sans l'accord écrit de l'IEC ou du Comité national de l'IEC du pays du demandeur. Si vous avez des questions sur le copyright de l'IEC ou si vous désirez obtenir des droits supplémentaires sur cette publication, utilisez les coordonnées ci-après ou contactez le Comité national de l'IEC de votre pays de résidence.

IEC Central Office
3, rue de Varembe
CH-1211 Geneva 20
Switzerland

Tel.: +41 22 919 02 11
info@iec.ch
www.iec.ch

About the IEC

The International Electrotechnical Commission (IEC) is the leading global organization that prepares and publishes International Standards for all electrical, electronic and related technologies.

About IEC publications

The technical content of IEC publications is kept under constant review by the IEC. Please make sure that you have the latest edition, a corrigenda or an amendment might have been published.

IEC Catalogue - webstore.iec.ch/catalogue

The stand-alone application for consulting the entire bibliographical information on IEC International Standards, Technical Specifications, Technical Reports and other documents. Available for PC, Mac OS, Android Tablets and iPad.

IEC publications search - webstore.iec.ch/advsearchform

The advanced search enables to find IEC publications by a variety of criteria (reference number, text, technical committee,...). It also gives information on projects, replaced and withdrawn publications.

IEC Just Published - webstore.iec.ch/justpublished

Stay up to date on all new IEC publications. Just Published details all new publications released. Available online and also once a month by email.

Electropedia - www.electropedia.org

The world's leading online dictionary of electronic and electrical terms, containing 21 000 terms and definitions in English and French, with equivalent terms in 16 additional languages. Also known as the International Electrotechnical Vocabulary (IEV) online.

IEC Glossary - std.iec.ch/glossary

67 000 electrotechnical terminology entries in English and French extracted from the Terms and Definitions clause of IEC publications issued since 2002. Some entries have been collected from earlier publications of IEC TC 37, 77, 86 and CISPR.

IEC Customer Service Centre - webstore.iec.ch/csc

If you wish to give us your feedback on this publication or need further assistance, please contact the Customer Service Centre: sales@iec.ch.

A propos de l'IEC

La Commission Electrotechnique Internationale (IEC) est la première organisation mondiale qui élabore et publie des Normes internationales pour tout ce qui a trait à l'électricité, à l'électronique et aux technologies apparentées.

A propos des publications IEC

Le contenu technique des publications IEC est constamment revu. Veuillez vous assurer que vous possédez l'édition la plus récente, un corrigendum ou amendement peut avoir été publié.

Catalogue IEC - webstore.iec.ch/catalogue

Application autonome pour consulter tous les renseignements bibliographiques sur les Normes internationales, Spécifications techniques, Rapports techniques et autres documents de l'IEC. Disponible pour PC, Mac OS, tablettes Android et iPad.

Recherche de publications IEC - webstore.iec.ch/advsearchform

La recherche avancée permet de trouver des publications IEC en utilisant différents critères (numéro de référence, texte, comité d'études,...). Elle donne aussi des informations sur les projets et les publications remplacées ou retirées.

IEC Just Published - webstore.iec.ch/justpublished

Restez informé sur les nouvelles publications IEC. Just Published détaille les nouvelles publications parues. Disponible en ligne et aussi une fois par mois par email.

Electropedia - www.electropedia.org

Le premier dictionnaire en ligne de termes électroniques et électriques. Il contient 21 000 termes et définitions en anglais et en français, ainsi que les termes équivalents dans 16 langues additionnelles. Egalement appelé Vocabulaire Electrotechnique International (IEV) en ligne.

Glossaire IEC - std.iec.ch/glossary

67 000 entrées terminologiques électrotechniques, en anglais et en français, extraites des articles Termes et Définitions des publications IEC parues depuis 2002. Plus certaines entrées antérieures extraites des publications des CE 37, 77, 86 et CISPR de l'IEC.

Service Clients - webstore.iec.ch/csc

Si vous désirez nous donner des commentaires sur cette publication ou si vous avez des questions contactez-nous: sales@iec.ch.

INTERNATIONAL STANDARD

NORME INTERNATIONALE

AMENDMENT 1
AMENDEMENT 1

Ultrasonics – Field characterization – Test methods for the determination of thermal and mechanical indices related to medical diagnostic ultrasonic fields

Ultrasons – Caractérisation du champ – Méthodes d'essai pour la détermination d'indices thermique et mécanique des champs d'ultrasons utilisés pour le diagnostic médical

INTERNATIONAL
ELECTROTECHNICAL
COMMISSION

COMMISSION
ELECTROTECHNIQUE
INTERNATIONALE

ICS 17.140.50

ISBN 978-2-8322-5301-4

**Warning! Make sure that you obtained this publication from an authorized distributor.
Attention! Veuillez vous assurer que vous avez obtenu cette publication via un distributeur agréé.**

FOREWORD

This amendment has been prepared by IEC technical committee 87: Ultrasonics.

This bilingual version (2018-01) corresponds to the monolingual English version, published in 2017-09.

The text of this amendment is based on the following documents:

FDIS	Report on voting
87/661/FDIS	87/665/RVD

Full information on the voting for the approval of this amendment can be found in the report on voting indicated in the above table.

The French version of this amendment has not been voted upon.

The committee has decided that the contents of this amendment and the base publication will remain unchanged until the stability date indicated on the IEC website under "<http://webstore.iec.ch>" in the data related to the specific publication. At this date, the publication will be

iTeh STANDARD PREVIEW
(standards.iteh.ai)

- reconfirmed,
- withdrawn,
- replaced by a revised edition, or [IEC 62359:2010/AMD1:2017](https://standards.iteh.ai/catalog/standards/sist/ec88228f-c130-4b67-a1ee-071d05236ca3/iec-62359-2010-amd1-2017)
- amended. <https://standards.iteh.ai/catalog/standards/sist/ec88228f-c130-4b67-a1ee-071d05236ca3/iec-62359-2010-amd1-2017>

Introduction to Amendment

The second edition of IEC 62359 was published in 2010. Since then, IEC 60601-2-37:2007/AMD1:2015 has been published and calls for provision of **attenuated spatial peak temporal average intensity**, $I_{\text{spta},\alpha}$, and **attenuated spatial peak pulse average intensity**, $I_{\text{sppa},\alpha}$, at specific spatial maximum points in the ultrasonic field on the **beam axis**. No IEC standard describes the determination of these quantities at these specific positions. IEC 62359 for determining the thermal indices currently uses similar values at other positions, therefore, the determination of **attenuated spatial peak temporal average intensity**, $I_{\text{spta},\alpha}$, and **attenuated spatial peak pulse average intensity**, $I_{\text{sppa},\alpha}$, has been added as an annex in this amendment.

Additionally, references to newly published collateral standards have been updated.

2 Normative references

Replace the first reference by the following:

IEC 60601-2-37:2007, *Medical electrical equipment – Part 2-37: Particular requirements for the basic safety and essential performance of ultrasonic medical diagnostic and monitoring equipment*
IEC 60601-2-37:2007/AMD1:2015

Replace the second reference by the following:

IEC 61157:2007, *Standard means for the reporting of the acoustic output of medical diagnostic ultrasonic equipment*
IEC 61157:2007/AMD1:2013

Replace, in the third reference, "IEC 61161:2006" by "IEC 61161:2013".

Replace the fifth reference by the following:

IEC 62127-1:2007, *Ultrasonics – Hydrophones – Part 1: Measurement and characterization of medical ultrasonic fields up to 40 MHz*
IEC 62127-1:2007/AMD1:2013

3 Terms and definitions

iTeh STANDARD PREVIEW
(standards.iteh.ai)

Replace the first existing paragraph by the following new paragraph:

[IEC 62359:2010/AMD1:2017](#)

For the purposes of this document, the terms and definitions given in IEC 60601-2-37, IEC 62127-1, IEC 62127-2, IEC 62127-3, IEC 61157 and IEC 61161 apply. Several of these are repeated below for convenience and others are listed because they have been modified for application to this standard.

3.3

acoustic repetition period

Replace the existing definition by the following new definition:

time interval between corresponding points of consecutive cycles, pulses or scans, depending on the current operating mode

Add the following new NOTES 2 and 3:

NOTE 2 For continuous wave modes, the **acoustic repetition period** is the time interval between corresponding points of consecutive cycles

NOTE 3 For **combined operating modes** where transmit pulsing of the constituent modes may be interrupted, the *arp* determination should take into account non-pulsing time to calculate an average period.

Replace "NOTE 2" by NOTE 4".

3.4

acoustic working frequency

Replace, in the existing definition, the words "at the position corresponding to the **spatial-peak temporal-peak acoustic pressure**" by "on the **beam axis**, beyond the **break-point depth**, corresponding to **depth of maximum pulse-intensity integral** z_{pii} "

Replace, in existing NOTE 2, the words "at the position of maximum **pulse-pressure-squared integral**" by "at the **depth for peak pulse-intensity integral**".

3.7 attenuated peak-rarefactional acoustic pressure

Replace, in the existing definition, the words "at a specified distance from" by "on a plane perpendicular to the **beam axis** at a specified distance z from".

Replace, in the definition list after Equation (2), the words " z is the distance from the **external transducer aperture** to the point of interest" by " z is the distance from the **external transducer aperture** along the **beam axis** to the plane containing the point of interest".

3.8 attenuated pulse-intensity integral

Replace, in the definition, the words "at a specified distance from" by "on a plane perpendicular to the **beam axis** at a specified distance z from".

Replace, in the definition list after Equation (3), the words " z is the distance from the **external transducer aperture** to the point of interest" by " z is the distance from the **external transducer aperture** along the **beam axis** to the plane containing the point of interest".

Replace "NOTE" by "NOTE 1".

Add new NOTE 2:

NOTE 2 For measurement purposes of this standard, p_{i_a} is equivalent to $1/(\rho c)$ times the **attenuated pulse-pressure-squared integral** at depth z , with ρc denoting the characteristic acoustic impedance of pure water.

3.9 attenuated spatial-average temporal-average intensity

Replace, in the definition, the words "at a specified distance from" by "on a plane perpendicular to the **beam axis** at a specified distance z from".

Replace, in the definition list after Equation (4), the words " z is the distance from the **external transducer aperture** to the point of interest" by " z is the distance from the **external transducer aperture** along the **beam axis** to the plane containing the point of interest".

3.10 attenuated spatial-peak temporal-average intensity

Replace, in the definition, the words "at a specified distance from" by "on a plane perpendicular to the **beam axis** at a specified distance z from".

Replace, in the definition list after Equation (5), the words " z is the distance from the **external transducer aperture** to the point of interest" by " z is the distance from the **external transducer aperture** along the **beam axis** to the plane containing the point of interest".

3.11 attenuated temporal-average intensity

Replace, in the definition, the words "at a specified distance from" by "on a plane perpendicular to the **beam axis** at a specified distance z from".

Replace, in the definition list after Equation (6), the words " z is the distance from the **external transducer aperture** to the point of interest" by " z is the distance from the **external transducer aperture** along the **beam axis** to the plane containing the point of interest".

3.13 beam-axis

Replace the term "**beam-axis**" by "**beam axis**".

3.17 bone thermal index

Delete, in the definition, "or neonatal cephalic (through the fontanelle)".

3.19 break-point depth

Add, after the existing definition, the words "**acoustic working frequency** and intensity parameters (such as **attenuated spatial-peak temporal-average intensity**)"

Add, after "where D_{eq} is the **equivalent aperture diameter**" the words "for **non-scanning modes**".

Replace the existing NOTE 2 by the following new note:

NOTE 2 For **scanning modes**, use the **non-scanning mode** D_{eq} value calculation [Equation (8)]. Do this using the **output beam area** of one **ultrasonic scan line**; the central $scan$ line, corresponding to the **beam axis** (i.e. the line where pii , MI , and f_{awt} are measured).

3.21 cranial-bone thermal index

Replace the existing definition by the following new definition:

thermal index for applications in which the ultrasound beam passes through bone near the beam entrance into the body such as paediatric and adult cranial or neonatal cephalic applications

3.22 default setting

Replace, in the definition, "**ultrasonic diagnostic equipment**" by "**medical diagnostic ultrasonic equipment**".

3.23 depth for mechanical index

Replace, in the definition, the words "to the plane of maximum **attenuated pulse-intensity integral** (pii_a)" by "to the plane of maximum **attenuated pulse-pressure-squared-integral** ($ppsi_a$)".

Add new NOTE 1:

NOTE 1 Because z_{MI} may occur closer to the transducer than the **break-point depth** z_{bp} , use of $ppsi_a$ rather than pii_a is technically more appropriate. If $z_{ppsi,a}$ is larger than z_{bp} , then $z_{ppsi,a}$ and $z_{pii,a}$ are equal.

Replace "NOTE" by "NOTE 2".

3.24 depth for peak pulse-intensity integral

Replace the existing term, definition and note by the following new term, definition and notes:

3.24 depth for maximum pulse intensity integral

depth z on the **beam axis** and beyond the **break-point depth** z_{bp} from the **external transducer aperture** to the plane of maximum **pulse-intensity integral** (pii) as approximated by the **pulse-pressure-squared integral** ($ppsi$)

NOTE 1 **Depth for maximum pii** is expressed in metres (m).

NOTE 2 **Depth for maximum pii** is termed "depth for peak pulse-intensity integral" in IEC 60601-2-37:2007/AMD1:2015.

NOTE 3 At this depth the **acoustic working frequency** is determined.

3.27

Discrete-perating mode

Replace the term "**Discrete-perating mode**" by "**discrete-operating mode**".

Replace, in the definition, "**ultrasonic diagnostic equipment**" by "**medical diagnostic ultrasonic equipment**".

3.28

equivalent aperture diameter

Replace the existing definition and equation by the following new definition and equation:

diameter of a circle the area A of which is the -12 dB **output beam area** A_{ob} for **non-scanning modes** and the -12 dB **scanned aperture area** A_{sa} for **scanning modes**, given by

$$D_{eq} = \sqrt{\frac{4}{\pi} A} \quad (8)$$

Replace the existing NOTE 1 by the following new note:

NOTE 1 Equation (8) is used in the calculation of the **cranial-bone thermal index**; for **non-scanning modes** with $A = A_{ob}$ and for **scanning modes** with $A = A_{sa}$.

Add new NOTE 2:

iTeh STANDARD PREVIEW
(standards.iteh.ai)

NOTE 2 Equation (8) with $A = A_{ob}$ is also used in calculating the **break-point depth**.

Replace "NOTE 2" by "NOTE 3". [IEC 62359:2010/AMD1:2017](https://standards.iteh.ai/catalog/standards/sist/ec88228f-c130-4b67-a1ee-071d05236ca3/iec-62359-2010-amd1-2017)
<https://standards.iteh.ai/catalog/standards/sist/ec88228f-c130-4b67-a1ee-071d05236ca3/iec-62359-2010-amd1-2017>

3.33

medical diagnostic ultrasonic equipment

Replace "NOTE" by "NOTE 1".

Add new NOTE 2:

NOTE 2 IEC 60601-2-37:2007 uses the term "ultrasonic diagnostic equipment" instead of **medical diagnostic ultrasonic equipment**.

3.34

non-scanning mode

Replace, in the definition, the words "**ultrasonic diagnostic equipment**" by "**medical diagnostic ultrasonic equipment**".

3.37

output power

Replace, in the source, "IEC 61161:2006" by "IEC 61161:2013".

3.40

prudent-use statements

Replace the existing definition by the following new definition:

affirmations of the principle that only necessary clinical information should be acquired and that high exposure levels and long exposure times should be avoided

3.44 pulse repetition period

Add new NOTE 1:

NOTE 1 In general, for **non-scanning modes** the **pulse repetition period** needs to be adjusted to represent a 'per-second' average taking into account interruptions-in, or non-constant, pulsing; e.g. such as may occur in **combined operating modes**.

Replace "NOTE" by "NOTE 2".

3.49 scanning mode

Replace, in the definition, the words "an **ultrasonic diagnostic equipment**" by "a **medical diagnostic ultrasonic equipment**".

3.54 spatial-peak temporal-average intensity

Replace the existing NOTE 1 by the following new note:

NOTE 1 For systems in **combined-operating modes**, the averaging time period needs to be sufficient to include periods in which scanning or pulsing is interrupted.

3.56 thermal index

Replace, in the definition, the two occurrences of the words "attenuated acoustic power" by "attenuated output power"

(standards.iteh.ai)

Add the following new definitions:

[IEC 62359:2010/AMD1:2017](https://standards.iteh.ai/catalog/standards/sist/ec88228f-c130-4b67-a1ee-0711405236ca3/iec-62359-2010-amd1-2017)

3.61 instantaneous acoustic pressure

$p(t)$

pressure minus the ambient pressure at a particular instant in time and at a particular point in an acoustic field

NOTE **Instantaneous acoustic pressure** is expressed in pascals (Pa).

[SOURCE: IEC 62127-1:2007, 3.33, modified – The reference to IEV 801-01-19 has been removed in the definition]

3.62 attenuated instantaneous acoustic pressure

$p_{\alpha}(z,t)$

value of the **instantaneous acoustic pressure** at time t after attenuation on a plane perpendicular to the **beam axis** at a specified distance z from the source, and given by

$$p_{\alpha}(z,t) = p(z,t)10^{(-\alpha z f_{awf}/20 \text{ dB})} \quad (26)$$

where

α is the **acoustic attenuation coefficient**;

z is the distance from the source to the point (plane) of interest;

f_{awf} is the **acoustic working frequency**;

$p(z,t)$ is the **instantaneous acoustic pressure**

NOTE **Attenuated instantaneous acoustic pressure** is expressed in pascals (Pa).

3.63
attenuated pulse-pressure-squared integral

$ppsi_{\alpha}(z)$

time integral of the square of the **attenuated instantaneous acoustic pressure**, integrated over the acoustic pulse waveform, on a plane perpendicular to the **beam axis** at a specified distance z in an acoustic field

$$ppsi_{\alpha}(z) = \int p_{\alpha}^2(z, t) dt = \int [p(z, t) 10^{(-\alpha z f_{awf}/20 \text{ dB})}]^2 dt = ppsi(z) \times 10^{(-\alpha z f_{awf}/10 \text{ dB})} \quad (27)$$

where

- $p(z, t)$ is the **instantaneous acoustic pressure** at depth z .
- α is the **acoustic attenuation coefficient**;
- z is the distance from the source to the point (plane) of interest;
- f_{awf} is the **acoustic working frequency**;
- $ppsi$ is the **pulse-pressure-squared integral**

NOTE 1 **Attenuated pulse-pressure-squared integral** is expressed in pascal squared seconds (Pa²s).

NOTE 2 See definition 3.43 for the non-attenuated version; with the addition here of the perpendicular plane at depth z .

3.64
attenuated scan intensity integral

$sii_{\alpha}(z)$

sum of the **attenuated pulse intensity integrals** in one scan (one frame of **ultrasonic scan lines**) at depth z

NOTE 1 **Attenuated scan intensity integral** is expressed in joules per square metre (Jm⁻²).
<https://standards.iteh.ai/catalog/standards/sist/ec88228f-c130-4b67-a1ee-07195336a36a/iec-62359-3910-amd1-2017>

NOTE 2 For measurement purposes of this standard, $sii_{\alpha}(z)$ is equivalent to $1/(\rho c)$ times the **sum of attenuated pulse-pressure-squared integrals** at depth z , for $z \geq z_{bp}$, with ρc denoting the characteristic acoustic impedance of pure water.

NOTE 3 See definition 3.79 for the non-attenuated version.

3.65
attenuated spatial-peak pulse-average intensity

$I_{sppa, \alpha}(z)$

maximum value of the **spatial-peak pulse-average intensity** after attenuation, on a plane perpendicular to the **beam axis** at a specified distance z from the source, and given by

$$I_{sppa, \alpha}(z) = \frac{1}{t_d(z)} pii_{\alpha}(z) \quad (28)$$

where

- $t_d(z)$ is the **pulse duration** at the same depth z ;
- $pii_{\alpha}(z)$ is the **attenuated pulse-intensity integral** at depth z .

NOTE 1 **Attenuated spatial-peak pulse-average intensity** is expressed in watts per square metre (Wm⁻²).

NOTE 2 For measurement purposes of this standard, $pii_{\alpha}(z)$ is equivalent to $1/(\rho c)$ times the **attenuated pulse-pressure-squared integrals** at depth z , $ppsi_{\alpha}(z)$, for $z \geq z_{bp}$, with ρc denoting the characteristic acoustic impedance of pure water.

NOTE 3 See definition 3.81 for the non-attenuated version.

3.66

attenuated sum of pulse-pressure-squared integrals

$s_{\text{appsi}}(z)$

attenuated value of the sum of **pulse-pressure-squared integrals** in one scan (one frame of **ultrasonic scan lines**) at depth z

NOTE 1 **Attenuated sum of pulse-pressure-squared integrals** is expressed in pascal squared seconds (Pa^2s).

NOTE 2 The **attenuated sum of pulse-pressure-squared integrals** at depth z will be equal to the **sum of attenuated pulse-pressure-squared integrals** if each **ultrasonic scan line** in the frame which is included in the sum has the same **acoustic working frequency**.

NOTE 3 See F.3.1.4.2 for additional explanation.

NOTE 4 See definition 3.83 for the non-attenuated version.

3.67

depth for maximum I_{sppa}

$z_{\text{sppa,max}}$

depth z on the **beam axis** and beyond the **break-point depth** z_{bp} of maximum **spatial-peak pulse-average intensity**

NOTE 1 **Depth for maximum I_{sppa}** is expressed in metres (m).

NOTE 2 This depth is equivalent to the **depth for maximum pii** .

3.68

depth for maximum $I_{\text{sppa},\alpha}$

$z_{\text{sppa},\alpha,\text{max}}$

depth z on the **beam axis** and beyond the **break-point depth** z_{bp} of maximum **attenuated spatial-peak pulse-average intensity**

NOTE 1 **Depth for maximum $I_{\text{sppa},\alpha}$** is expressed in metres (m).

NOTE 2 This depth is equivalent to the **depth for maximum pii_{α}** .

3.69

depth for maximum I_{spta}

$z_{\text{spta,max}}$

depth z on the **beam axis** and beyond the **break-point depth** z_{bp} of maximum **spatial-peak temporal-average intensity**

NOTE 1 **Depth for maximum I_{spta}** is expressed in metres (m).

NOTE 2 For **non-scanning modes**, this depth is equivalent to the **depth for maximum pii** . For **scanning modes**, it is equivalent to the **depth for maximum sii** .

3.70

depth for maximum $I_{\text{spta},\alpha}$

$z_{\text{spta},\alpha,\text{max}}$

depth z on the **beam axis** and beyond the **break-point depth** z_{bp} of maximum **attenuated spatial-peak temporal-average intensity**

NOTE 1 **Depth for maximum $I_{\text{spta},\alpha}$** is expressed in metres (m).

NOTE 2 For **non-scanning modes**, this depth is equivalent to the **depth for maximum pii_{α}** . For **scanning modes**, it is equivalent to the **depth for maximum sii_{α}** .

3.71

depth for maximum pii_{α}

$z_{\text{pii},\alpha}$

depth z on the **beam axis** and beyond the **break-point depth** z_{bp} of maximum **attenuated pulse-intensity integral**

NOTE 1 **Depth for maximum p_{ii_a}** is expressed in metres (m).

NOTE 2 This depth is equivalent to the **depth for maximum p_{psi_a}** , $z_{p_{psi,a}}$ when $z_{p_{psi,a}}$ occurs beyond the **break-point depth** (see 3.73).

NOTE 3 **Depth for maximum p_{ii_a}** is termed "depth for peak attenuated pulse-intensity integral" in IEC 60601-2-37:2007+IEC 60601-2-37:2007/AMD1:2015.

3.72

depth for maximum p_{psi}

$z_{p_{psi}}$
depth z on the **beam axis** of maximum **pulse-pressure-squared integral**

NOTE 1 **Depth for maximum p_{psi}** is expressed in metres (m).

NOTE 2 When it occurs beyond the **break-point depth**, this depth is equivalent to the **depth for maximum p_{ii}** , $z_{p_{ii}}$ (see 3.24).

3.73

depth for maximum p_{psi_a}

$z_{p_{psi,a}}$
depth z on the **beam axis** of maximum **attenuated pulse-pressure-squared integral**

NOTE 1 **Depth for maximum p_{psi_a}** is expressed in metres (m).

NOTE 2 When it occurs beyond the **break-point depth**, this depth is equivalent to the **depth for maximum p_{ii_a}** , $z_{p_{ii,a}}$ (i.e. depth for maximum **attenuated pulse-intensity integral**).

NOTE 3 This depth is the **depth for mechanical index**, z_{MI} (see 3.23).

3.74

depth for maximum s_{ii}

$z_{s_{ii}}$
depth z on the **beam axis** and beyond the **break-point depth** z_{bp} of maximum **scan-intensity integral**

NOTE 1 **Depth for maximum s_{ii}** is expressed in metres (m).

NOTE 2 This depth is equivalent to the **depth for maximum s_{psi}** , $z_{s_{psi}}$ when $z_{s_{psi}}$ occurs beyond the **break-point depth** (see 3.76).

NOTE 3 **Depth for maximum s_{ii}** is termed "depth for peak sum of pulse-intensity integrals" in IEC 60601-2-37:2007+IEC 60601-2-37:2007/AMD1:2015.

3.75

depth for maximum s_{ii_a}

$z_{s_{ii,a}}$
depth z on the **beam axis** and beyond the **break-point depth** z_{bp} of maximum **attenuated scan-intensity integral**

NOTE 1 **Depth for maximum s_{ii_a}** is expressed in metres (m).

NOTE 2 This depth is equivalent to the **depth for maximum s_{psi_a}** , $z_{s_{psi,a}}$ when $z_{s_{psi,a}}$ occurs beyond the **break-point depth** (see 3.77).

NOTE 3 **Depth for maximum s_{ii_a}** is termed "depth for peak sum of attenuated pulse-intensity integrals" in IEC 60601-2-37:2007+IEC 60601-2-37:2007/AMD1:2015.

3.76

depth for maximum s_{psi}

$z_{s_{psi}}$
depth z on the **beam axis** of maximum **sum of pulse-pressure-squared integrals**

NOTE 1 **Depth for maximum s_{psi}** is expressed in metres (m).

NOTE 2 When it occurs beyond the **break-point depth**, this depth is equivalent to the **depth for maximum sii** , z_{sii} (see 3.74).

NOTE 3 **Depth for maximum $sppsi$** is termed "depth for peak sum of pulse-intensity integrals" in IEC 60601-2-37:2007+IEC 60601-2-37:2007/AMD1:2015.

3.77

depth for maximum $sppsi_\alpha$

$z_{sppsi,\alpha}$

depth z on the **beam axis** of maximum **sum of attenuated pulse-pressure-squared integrals**

NOTE 1 **Depth for maximum $sppsi_\alpha$** is expressed in metres (m).

NOTE 2 When it occurs beyond the **break-point depth**, this depth is equivalent to the **depth for maximum sii_α** , $z_{sii,\alpha}$ (see 3.75).

NOTE 3 **Depth for maximum $sppsi_\alpha$** is termed "depth for peak sum of attenuated pulse-intensity integrals" in IEC 60601-2-37:2007+IEC 60601-2-37:2007/AMD1:2015.

3.78

pulse-average intensity

I_{pa}

quotient of the **pulse-intensity integral** to the **pulse duration** at a particular point in an acoustic field

NOTE 1 This definition applies to pulses and bursts.

NOTE 2 **Pulse-average intensity** is expressed in watts per square metre (Wm^{-2}).

[IEC 62127-1:2007+IEC 62127-1:2007/AMD1:2013, 3.47]

3.79

scan intensity integral

sii

sum of the **pulse intensity integrals** in one scan (one frame of **ultrasonic scan lines**) at depth z in the acoustic field

NOTE 1 **Scan intensity integral** is expressed in joules per square metre (Jm^{-2}).

NOTE 2 For measurement purposes of this standard, sii is equivalent to $1/(\rho c)$ times the **sum of pulse-pressure-squared integral** at depth z where ρc is the characteristic acoustic impedance of pure water.

3.80

scan repetition rate

srr

inverse of the **scan repetition period** (see 3.50)

NOTE **Scan repetition rate** is expressed in hertz (Hz).

3.81

spatial-peak pulse-average intensity

$I_{sppa}(z)$

maximum value of the **pulse-average intensity** on a plane perpendicular to the **beam axis** at a specified distance z from the source, and given by

$$I_{sppa}(z) = \frac{1}{t_d(z)} pii(z) \quad (29)$$

where

$t_d(z)$ is the **pulse duration** at the same depth z ;

$pii(z)$ is the **pulse-intensity integral** at depth z .

NOTE 1 **Spatial-peak pulse-average intensity** is expressed in watts per square metre (Wm^{-2}).

NOTE 2 For measurement purposes of this standard, $pii(z)$ is equivalent to $1/(\rho c)$ times the **pulse-pressure-squared integral** at depth z , $ppsi(z)$, for $z \geq z_{bp}$, with ρc denoting the characteristic acoustic impedance of pure water.

NOTE 3 See IEC 62127-1:2007, 3.60, which has been modified to specify the perpendicular plane at depth z . Equation (29) and Note 2 have been added, in accordance with IEC 62127-1:2007, Equation (15).

3.82
sum of attenuated pulse-pressure-squared integrals

$sppsi_a(z)$

sum of the **attenuated pulse-pressure-squared integrals** in one scan (one frame of **ultrasonic scan lines**) at depth z

NOTE 1 **Sum of attenuated pulse-pressure-squared integrals** is expressed in pascal squared seconds (Pa^2s).

NOTE 2 Closely related to the **attenuated scan intensity integral**, see 3.64.

NOTE 3 The **sum of attenuated pulse-pressure-squared integrals** at depth z will be equal to the **attenuated sum of pulse-pressure-squared integrals** at depth z if each **ultrasonic scan line** in the frame which is included in the sum has the same **acoustic working frequency**.

3.83
sum of pulse-pressure-squared integrals

$sppsi(z)$

sum of **pulse-pressure-squared integrals** in one scan (one frame of **ultrasonic scan lines**) at depth z

iTeh STANDARD PREVIEW

NOTE 1 **Sum of pulse-pressure-squared integrals** is expressed in pascal squared seconds (Pa^2s).

NOTE 2 The $sppsi$ at depth z may also be referred to as the scan pulse pressure squared integral and is proportional to the **scan intensity integral** for $z \geq z_{bp}$.

(standards.iteh.ai)

IEC 62359:2010/AMD1:2017

<https://standards.iteh.ai/catalog/standards/sist/ec88228f-c130-4b67-a1ee-071d05236ca3/iec-62359-2010-amd1-2017>

4 List of symbols

Replace the symbols " $TIB_{as,sc}$ " and " $TIB_{as,ns}$ " by " TIC_{sc} " and " TIC_{ns} ".

Replace "depth for peak pulse-intensity integral" by "depth for maximum pii ".

Add the following new symbols:

I_{sppa}	spatial-peak pulse-average intensity
$I_{\text{sppa},\alpha}$	attenuated spatial-peak pulse-average intensity
I_{ta}	temporal-average intensity
$ppsi_{\alpha}$	attenuated pulse-pressure-squared integral
sii	scan intensity integral
sii_{α}	attenuated scan intensity integral
$sppsi$	sum of pulse-pressure-squared integrals
$s_{\alpha}ppsi$	attenuated sum of pulse-pressure-squared integrals
$sppsi_{\alpha}$	sum of attenuated pulse-pressure-squared integrals
srr	scan repetition rate
$z_{pii,\alpha}$	depth for maximum pii_{α}
z_{ppsi}	depth for maximum $ppsi$
$z_{ppsi,\alpha}$	depth for maximum $ppsi_{\alpha}$
z_{sii}	depth for maximum sii
$z_{sii,\alpha}$	depth for maximum sii_{α}
$z_{\text{sppa,max}}$	depth for maximum I_{sppa}
$z_{\text{sppa},\alpha,\text{max}}$	depth for maximum $I_{\text{sppa},\alpha}$
$z_{\text{spta,max}}$	depth for maximum I_{spta}
$z_{\text{spta},\alpha,\text{max}}$	depth for maximum $I_{\text{spta},\alpha}$
z_{sppsi}	depth for maximum $sppsi$
$z_{\text{sppsi},\alpha}$	depth for maximum $sppsi_{\alpha}$

TECHNICAL STANDARD PREVIEW
 (standards.iteh.ai)

5.1 General

Replace, in the first and second paragraphs, the words "ultrasonic diagnostic equipment" by "medical diagnostic ultrasonic equipment".

5.4.2.1 Determination of bone thermal index at surface for non-scanning modes, $TIC_{\text{ns}} (= TIB_{\text{as,ns}})$

Delete, in the subclause's title, "(= $TIB_{\text{as,ns}}$)".

Replace the existing Equation (17) by the following new equation:

$$TIC_{\text{ns}} = \frac{P / D_{\text{eq}}}{C_{\text{TIC}}} \quad (17)$$

Replace, in the definition list below Equation (17), the words "is the equivalent aperture diameter" by "is the equivalent aperture diameter; calculation for non-scanning modes as described in 3.28 using the output beam area A_{ob} ".

Replace the existing note by the following new note:

NOTE TIC_{ns} applies to the bone-at-surface case for non-scanning modes.