

INTERNATIONAL STANDARD

NORME INTERNATIONALE

Medical electrical equipment –
Part 2-60: Particular requirements for the basic safety and essential performance
of dental equipment [\(standards.iteh.ai\)](https://standards.iteh.ai/)

Appareils électromédicaux – [IEC 80601-2-60:2019](https://standards.iteh.ai/catalog/standards/sist/80569-445-4-4d-825b-a15ec9a8862/iec-80601-2-60-2019)
Partie 2-60: Exigences particulières pour la sécurité de base et les performances
essentiels des équipements dentaires



THIS PUBLICATION IS COPYRIGHT PROTECTED
Copyright © 2019 IEC, Geneva, Switzerland

All rights reserved. Unless otherwise specified, no part of this publication may be reproduced or utilized in any form or by any means, electronic or mechanical, including photocopying and microfilm, without permission in writing from either IEC or IEC's member National Committee in the country of the requester. If you have any questions about IEC copyright or have an enquiry about obtaining additional rights to this publication, please contact the address below or your local IEC member National Committee for further information.

Droits de reproduction réservés. Sauf indication contraire, aucune partie de cette publication ne peut être reproduite ni utilisée sous quelque forme que ce soit et par aucun procédé, électronique ou mécanique, y compris la photocopie et les microfilms, sans l'accord écrit de l'IEC ou du Comité national de l'IEC du pays du demandeur. Si vous avez des questions sur le copyright de l'IEC ou si vous désirez obtenir des droits supplémentaires sur cette publication, utilisez les coordonnées ci-après ou contactez le Comité national de l'IEC de votre pays de résidence.

IEC Central Office
3, rue de Varembe
CH-1211 Geneva 20
Switzerland

Tel.: +41 22 919 02 11
info@iec.ch
www.iec.ch

About the IEC

The International Electrotechnical Commission (IEC) is the leading global organization that prepares and publishes International Standards for all electrical, electronic and related technologies.

About IEC publications

The technical content of IEC publications is kept under constant review by the IEC. Please make sure that you have the latest edition, a corrigendum or an amendment might have been published.

IEC publications search - webstore.iec.ch/advsearchform

The advanced search enables to find IEC publications by a variety of criteria (reference number, text, technical committee,...). It also gives information on projects, replaced and withdrawn publications.

IEC Just Published - webstore.iec.ch/justpublished

Stay up to date on all new IEC publications. Just Published details all new publications released. Available online and once a month by email.

IEC Customer Service Centre - webstore.iec.ch/csc

If you wish to give us your feedback on this publication or need further assistance, please contact the Customer Service Centre: sales@iec.ch.

Electropedia - www.electropedia.org

The world's leading online dictionary on electrotechnology, containing more than 22 000 terminological entries in English and French, with equivalent terms in 16 additional languages. Also known as the International Electrotechnical Vocabulary (IEV) online.

IEC Glossary - std.iec.ch/glossary

67 000 electrotechnical terminology entries in English and French extracted from the Terms and Definitions clause of IEC publications issued since 2002. Some entries have been collected from earlier publications of IEC TC 37, 77, 86 and CISPR-60-2019

A propos de l'IEC

La Commission Electrotechnique Internationale (IEC) est la première organisation mondiale qui élabore et publie des Normes internationales pour tout ce qui a trait à l'électricité, à l'électronique et aux technologies apparentées.

A propos des publications IEC

Le contenu technique des publications IEC est constamment revu. Veuillez vous assurer que vous possédez l'édition la plus récente, un corrigendum ou amendement peut avoir été publié.

Recherche de publications IEC - webstore.iec.ch/advsearchform

La recherche avancée permet de trouver des publications IEC en utilisant différents critères (numéro de référence, texte, comité d'études,...). Elle donne aussi des informations sur les projets et les publications remplacées ou retirées.

IEC Just Published - webstore.iec.ch/justpublished

Restez informé sur les nouvelles publications IEC. Just Published détaille les nouvelles publications parues. Disponible en ligne et une fois par mois par email.

Service Clients - webstore.iec.ch/csc

Si vous désirez nous donner des commentaires sur cette publication ou si vous avez des questions contactez-nous: sales@iec.ch.

Electropedia - www.electropedia.org

Le premier dictionnaire d'électrotechnologie en ligne au monde, avec plus de 22 000 articles terminologiques en anglais et en français, ainsi que les termes équivalents dans 16 langues additionnelles. Egalement appelé Vocabulaire Electrotechnique International (IEV) en ligne.

Glossaire IEC - std.iec.ch/glossary

67 000 entrées terminologiques électrotechniques, en anglais et en français, extraites des articles Termes et Définitions des publications IEC parues depuis 2002. Plus certaines entrées antérieures extraites des publications des CE 37, 77, 86 et CISPR de l'IEC.



IEC 80601-2-60

Edition 2.0 2019-06

INTERNATIONAL STANDARD

NORME INTERNATIONALE

Medical electrical equipment –
Part 2-60: Particular requirements for the basic safety and essential performance
of dental equipment

Appareils électromédicaux –
Partie 2-60: Exigences particulières pour la sécurité de base et les performances
essentiels des équipements dentaires

INTERNATIONAL
ELECTROTECHNICAL
COMMISSION

COMMISSION
ELECTROTECHNIQUE
INTERNATIONALE

ICS 11.040.01

ISBN 978-2-8322-7049-3

Warning! Make sure that you obtained this publication from an authorized distributor.
Attention! Veuillez vous assurer que vous avez obtenu cette publication via un distributeur agréé.

CONTENTS

FOREWORD.....	3
201.1 Scope, object and related standards	6
201.2 Normative references.....	8
201.3 Terms and definitions.....	8
201.4 General requirements	10
201.5 General requirements for testing of ME EQUIPMENT	10
201.6 Classification of ME EQUIPMENT and ME SYSTEMS.....	10
201.7 ME EQUIPMENT identification, marking and documents	10
201.8 Protection against electrical HAZARDS from ME EQUIPMENT	11
201.9 Protection against MECHANICAL HAZARDS of ME EQUIPMENT and ME SYSTEMS	16
201.10 Protection against unwanted and excessive radiation HAZARDS	19
201.11 Protection against excessive temperatures and other HAZARDS	19
201.12 Accuracy of controls and instruments and protection against hazardous outputs	23
201.13 HAZARDOUS SITUATIONS and fault conditions for ME EQUIPMENT	23
201.14 PROGRAMMABLE ELECTRICAL MEDICAL SYSTEMS (PEMS).....	24
201.15 Construction of ME EQUIPMENT.....	24
201.16 ME SYSTEMS	25
201.17 Electromagnetic compatibility of ME EQUIPMENT and ME SYSTEMS	25
201.101 Cordless HAND-HELD and foot-operated control devices	25
Annexes	26
Annex AA (informative) Particular guidance and rationale.....	27
Bibliography.....	39
Index of defined terms used in this document	40
Figure AA.1 – Example of APPLIED PARTS for DENTAL EQUIPMENT	28
Figure AA.2 – Calculation of LEAKAGE CURRENT	29
Figure AA.3 – Insulation problem of commutator DENTAL ELECTRICAL MOTOR.....	31
Figure AA.4 – Loading fan construction.....	37
Figure AA.5 – Load diagram with loading fan	37
Table 201.101 – Test voltages for solid insulation for SECONDARY CIRCUITS according to 201.8.9.1.12	12
Table 201.102 – Determination of TENSILE SAFETY FACTOR.....	18
Table 201.103 – Mass distribution.....	19
Table 201.104 – Allowable maximum temperatures for the OPERATOR SIDE of DENTAL HANDPIECES	20
Table AA.1 – RATED impulse voltage for equipment energized directly from the low-voltage mains	32

INTERNATIONAL ELECTROTECHNICAL COMMISSION

MEDICAL ELECTRICAL EQUIPMENT –**Part 2-60: Particular requirements for the basic safety
and essential performance of dental equipment**

FOREWORD

- 1) The International Electrotechnical Commission (IEC) is a worldwide organization for standardization comprising all national electrotechnical committees (IEC National Committees). The object of IEC is to promote international co-operation on all questions concerning standardization in the electrical and electronic fields. To this end and in addition to other activities, IEC publishes International Standards, Technical Specifications, Technical Reports, Publicly Available Specifications (PAS) and Guides (hereafter referred to as “IEC Publication(s)”). Their preparation is entrusted to technical committees; any IEC National Committee interested in the subject dealt with may participate in this preparatory work. International, governmental and non-governmental organizations liaising with the IEC also participate in this preparation. IEC collaborates closely with the International Organization for Standardization (ISO) in accordance with conditions determined by agreement between the two organizations.
- 2) The formal decisions or agreements of IEC on technical matters express, as nearly as possible, an international consensus of opinion on the relevant subjects since each technical committee has representation from all interested IEC National Committees.
- 3) IEC Publications have the form of recommendations for international use and are accepted by IEC National Committees in that sense. While all reasonable efforts are made to ensure that the technical content of IEC Publications is accurate, IEC cannot be held responsible for the way in which they are used or for any misinterpretation by any end user.
- 4) In order to promote international uniformity, IEC National Committees undertake to apply IEC Publications transparently to the maximum extent possible in their national and regional publications. Any divergence between any IEC Publication and the corresponding national or regional publication shall be clearly indicated in the latter.
- 5) IEC itself does not provide any attestation of conformity. Independent certification bodies provide conformity assessment services and, in some areas, access to IEC marks of conformity. IEC is not responsible for any services carried out by independent certification bodies.
- 6) All users should ensure that they have the latest edition of this publication.
- 7) No liability shall attach to IEC or its directors, employees, servants or agents including individual experts and members of its technical committees and IEC National Committees for any personal injury, property damage or other damage of any nature whatsoever, whether direct or indirect, or for costs (including legal fees) and expenses arising out of the publication, use of, or reliance upon, this IEC Publication or any other IEC Publications.
- 8) Attention is drawn to the Normative references cited in this publication. Use of the referenced publications is indispensable for the correct application of this publication.
- 9) Attention is drawn to the possibility that some of the elements of this IEC Publication may be the subject of patent rights. IEC shall not be held responsible for identifying any or all such patent rights.

International standard IEC 80601-2-60 has been prepared by a Joint Working Group of subcommittee 62D: Electromedical equipment, of IEC technical committee 62: Electrical equipment in medical practice, and subcommittee 6: Dental equipment, of ISO technical committee 106: Dentistry.

This second edition cancels and replaces the first edition published in 2012. This edition constitutes a technical revision.

This edition includes the following significant technical changes with respect to the previous edition:

- a) alignment with IEC 60601-1:2005 and IEC 60601-1:2005/AMD1:2012.

The text of this International Standard is based on the following documents:

FDIS	Report on voting
62D/1683/FDIS	62D/1691/RVD

Full information on the voting for the approval of this International Standard can be found in the report on voting indicated in the above table.

This document has been drafted in accordance with the ISO/IEC Directives, Part 2.

This publication is published as a double logo standard.

In this document, the following print types are used:

- requirements and definitions: roman type;
- *test specifications: italic type;*
- informative material appearing outside of tables, such as notes, examples and references: in smaller type. Normative text of tables is also in a smaller type;
- TERMS DEFINED IN CLAUSE 3 OF THE GENERAL STANDARD, IN THIS PARTICULAR STANDARD OR AS NOTED: SMALL CAPITALS.

In referring to the structure of this document, the term

- “clause” means one of the seventeen numbered divisions within the table of contents, inclusive of all subdivisions (e.g. Clause 7 includes subclauses 7.1, 7.2, etc.);
- “subclause” means a numbered subdivision of a clause (e.g. 7.1, 7.2 and 7.2.1 are all subclauses of Clause 7).

References to clauses within this document are preceded by the term “Clause” followed by the clause number. References to subclauses within this particular standard are by number only.

In this document, the conjunctive “or” is used as an “inclusive or” so a statement is true if any combination of the conditions is true.

The verbal forms used in this document conform to usage described in Clause 7 of the ISO/IEC Directives, Part 2. For the purposes of this document, the auxiliary verb:

- “shall” means that compliance with a requirement or a test is mandatory for compliance with this document;
- “should” means that compliance with a requirement or a test is recommended but is not mandatory for compliance with this document;
- “may” is used to describe a permissible way to achieve compliance with a requirement or test.

An asterisk (*) as the first character of a title or at the beginning of a paragraph or table title indicates that there is guidance or rationale related to that item in Annex AA.

A list of all parts of the IEC 60601 series, published under the general title: *Medical electrical equipment*, can be found on the IEC website.

The committee has decided that the contents of this document will remain unchanged until the stability date indicated on the IEC website under "<http://webstore.iec.ch>" in the data related to the specific document. At this date, the document will be

- reconfirmed,
- withdrawn,
- replaced by a revised edition, or
- amended.

iTeh STANDARD PREVIEW (standards.iteh.ai)

[IEC 80601-2-60:2019](#)

<https://standards.iteh.ai/catalog/standards/sist/a80f56fb-44f5-4a4d-825b-a15ecf9a8862/iec-80601-2-60-2019>

MEDICAL ELECTRICAL EQUIPMENT –

Part 2-60: Particular requirements for the basic safety and essential performance of dental equipment

201.1 Scope, object and related standards

Clause 1 of the general standard¹ applies, except as follows:

201.1.1 Scope

Replacement:

This part of IEC 80601 applies to the BASIC SAFETY and ESSENTIAL PERFORMANCE OF DENTAL UNITS, DENTAL PATIENT CHAIRS, DENTAL HANDPIECES AND DENTAL OPERATING LIGHTS, hereafter referred to as DENTAL EQUIPMENT.

If a clause or subclause is specifically intended to be applicable to ME EQUIPMENT only, or to ME SYSTEMS only, the title and content of that clause or subclause will say so. If that is not the case, the clause or subclause applies both to ME EQUIPMENT and to ME SYSTEMS, as relevant.

HAZARDS inherent in the intended physiological function of ME EQUIPMENT or ME SYSTEMS within the scope of this document are not covered by specific requirements in this document except in 7.2.13 and 8.4.1 of the general standard.

NOTE See also 4.2 of the general standard. <https://standards.iteh.ai/catalog/standards/sist/a80f56fb-44f5-4a4d-825b-a15ecf9a8862/iec-80601-2-60-2019>

201.1.2 Object

Replacement:

The object of this particular standard is to establish BASIC SAFETY and ESSENTIAL PERFORMANCE requirements for DENTAL EQUIPMENT (as defined in 201.3.202.)

201.1.3 Collateral standards

Addition:

This particular standard refers to those applicable collateral standards that are listed in Clause 2 of the general standard and Clause 201.2 of this particular standard.

IEC 60601-1-3 and IEC 60601-1-10 do not apply. All other published collateral standards in the IEC 60601-1 series apply as published.

¹ The general standard is IEC 60601-1:2005 and IEC 60601-1:2005/AMD1:2012, *Medical electrical equipment – Part 1: General requirements for basic safety and essential performance*.

201.1.4 Particular standards

Replacement:

In the IEC 60601 series, particular standards may modify, replace or delete requirements contained in the general standard and collateral standards as appropriate for the particular ME EQUIPMENT under consideration, and may add other BASIC SAFETY and ESSENTIAL PERFORMANCE requirements.

A requirement of a particular standard takes priority over the general standard.

For brevity, IEC 60601-1:2005 and IEC 60601-1:2005/AMD1:2012 are referred to in this particular standard as the general standard. Collateral standards are referred to by their document number.

The numbering of clauses and subclauses of this particular standard corresponds to that of the general standard with the prefix “201” (e.g. 201.1 in this document addresses the content of Clause 1 of the general standard) or applicable collateral standard with the prefix “20x” where x is the final digit(s) of the collateral standard document number (e.g. 202.4 in this particular standard addresses the content of Clause 4 of the IEC 60601-1-2 collateral standard, 203.4 in this particular standard addresses the content of Clause 4 of the IEC 60601-1-3 collateral standard, etc.). The changes to the text of the general standard are specified by the use of the following words:

"Replacement" means that the clause or subclause of the general standard or applicable collateral standard is replaced completely by the text of this particular standard.

"Addition" means that the text of this particular standard is additional to the requirements of the general standard or applicable collateral standard.

"Amendment" means that the clause or subclause of the general standard or applicable collateral standard is amended as indicated by the text of this particular standard.

Subclauses, figures or tables which are additional to those of the general standard are numbered starting from 201.101. However, due to the fact that definitions in the general standard are numbered 3.1 through 3.147, additional definitions in this document are numbered beginning from 201.3.201. Additional annexes are lettered AA, BB, etc., and additional items aa), bb), etc.

Subclauses, figures or tables which are additional to those of a collateral standard are numbered starting from 20x, where “x” is the number of the collateral standard, e.g. 202 for IEC 60601-1-2, 203 for IEC 60601-1-3, etc.

The term "this document" is used to make reference to the general standard, any applicable collateral standards and this particular standard taken together.

Where there is no corresponding clause or subclause in this particular standard, the clause or subclause of the general standard or applicable collateral standard, although possibly not relevant, applies without modification; where it is intended that any part of the general standard or applicable collateral standard, although possibly relevant, is not to be applied, a statement to that effect is given in this particular standard.

201.2 Normative references

NOTE Informative references are listed in the Bibliography.

Clause 2 of the general standard applies, except as follows:

Replacement:

IEC 60825-1:2014, *Safety of laser products – Part 1: Equipment classification and requirements*

Addition:

IEC 60601-1:2005, *Medical electrical equipment – Part 1: General requirements for basic safety and essential performance*
IEC 60601-1:2005/AMD1:2012

IEC 60601-2-2:2017, *Medical electrical equipment – Part 2-2: Particular requirements for the basic safety and essential performance of high frequency surgical equipment and high frequency surgical accessories*

IEC 60601-2-22:2007, *Medical electrical equipment – Part 2-22: Particular requirements for basic safety and essential performance of surgical, cosmetic, therapeutic and diagnostic laser equipment*
IEC 60601-2-22:2007/AMD1:2012

IEC 60601-2-57:2011, *Medical electrical equipment – Part 2-57: Particular requirements for the basic safety and essential performance of non-laser light source equipment intended for therapeutic, diagnostic, monitoring and cosmetic/aesthetic use*

IEC 60664-1:2007, *Insulation coordination for equipment within low-voltage systems – Part 1: Principles, requirements and tests*

IEC 60664-4:2005, *Insulation coordination for equipment within low-voltage systems – Part 4: Consideration of high-frequency voltage stress*

IEC 61180:2016, *High-voltage test techniques for low-voltage equipment – Definitions, test and procedure requirements, test equipment*

IEC 61810-1:2015, *Electromechanical elementary relays – Part 1: General and safety requirements*

ISO 1942:2009, *Dentistry – Vocabulary*

ISO 14457: 2017, *Dentistry – Handpieces and motors*

201.3 Terms and definitions

For the purposes of this document, the terms and definitions given in IEC 60601-1:2005 and IEC 60601-1:2005/AMD1:2012, IEC 60601-2-2:2017 and ISO 1942 and the following apply.

ISO and IEC maintain terminological databases for use in standardization at the following addresses:

- IEC Electropedia: available at <http://www.electropedia.org/>
- ISO Online browsing platform: available at <http://www.iso.org/obp>

NOTE An index of defined terms is found beginning on page 40.

Addition:

201.3.201

DENTAL ELECTRICAL MOTOR

HAND-HELD part of the DENTAL HANDPIECE electrically powered by the DENTAL UNIT

201.3.202

DENTAL EQUIPMENT

ME EQUIPMENT with any combination of DENTAL HANDPIECES, DENTAL UNITS, DENTAL PATIENT CHAIRS and DENTAL OPERATING LIGHTS

201.3.203

DENTAL HANDPIECE

HAND-HELD instrument used in dentistry for use in PATIENT treatment and connected to the DENTAL UNIT

201.3.204

DENTAL OPERATING LIGHT

device designed for use by an OPERATOR for illuminating the oral cavity, consisting of a luminaire and one or more lamps

201.3.205

DENTAL PATIENT CHAIR

device designed to support and position the PATIENT for treatment and therefore provided with a range of movements

201.3.206

DENTAL UNIT

assembly of devices designed to provide utilities and amenities for dental treatment, such as compressed air, water or other liquids, suction, electricity, hand- or foot-activated controllers, work surface(s), tray support(s), cuspidor or gasses

Note 1 to entry: This device is usually fitted with conveniently oriented instrument holders and controls, and consists of interconnected sub-units of DENTAL EQUIPMENT and instruments providing a functional unit for dental use.

201.3.207

OPERATOR SIDE OF DENTAL HANDPIECE

part of DENTAL HANDPIECE which is designed to be handheld by the OPERATOR in NORMAL USE

201.3.208

PATIENT SIDE OF DENTAL HANDPIECE

part of DENTAL HANDPIECE which is designed to be introduced into the oral cavity where all parts of the DENTAL HANDPIECE within 80 mm of the tip shall be considered as an APPLIED PART

201.3.209

CORD-CONNECTED MOBILE PARTS OF DENTAL EQUIPMENT

DENTAL EQUIPMENT which is permanently connected to FIXED parts of DENTAL EQUIPMENT and is equipped with castor or wheels for positioning by the OPERATOR and which is intended for the use on an even, obstacle free, floor

201.4 General requirements

Clause 4 of the general standard applies, except as follows:

201.4.3 ESSENTIAL PERFORMANCE

Replacement:

DENTAL EQUIPMENT does not have ESSENTIAL PERFORMANCE unless specified by the manufacturer.

201.5 General requirements for testing of ME EQUIPMENT

Clause 5 of the general standard applies, except as follows:

201.5.3 Ambient temperature, humidity, atmospheric pressure

Amendment of item a):

After the ME EQUIPMENT to be tested has been set up for NORMAL USE (according to 5.7 of general standard), tests are performed within the range of environmental conditions indicated in the technical description (see 7.9.3.1 of general standard) but at least at one temperature within an ambient temperature range +10 °C to 35 °C.

201.6 Classification of ME EQUIPMENT and ME SYSTEMS

Clause 6 of the general standard applies, except as follows:

201.6.2 Protection against electric shock

Replacement:

ME EQUIPMENT energized from an external electrical power source shall be classified as CLASS I ME EQUIPMENT or CLASS II ME EQUIPMENT (see 7.2.6 of IEC 60601-1:2005 and IEC 60601-1:2005/AMD1:2012). Other ME EQUIPMENT shall be classified as INTERNALLY POWERED ME EQUIPMENT.

INTERNALLY POWERED ME EQUIPMENT having a means of connection to a SUPPLY MAINS shall comply with the requirements for CLASS I ME EQUIPMENT or CLASS II ME EQUIPMENT while so connected, and with the requirements for INTERNALLY POWERED ME EQUIPMENT while not so connected.

APPLIED PARTS of DENTAL EQUIPMENT which are connected through water lines shall be considered as TYPE B APPLIED PARTS.

201.7 ME EQUIPMENT identification, marking and documents

Clause 7 of the general standard applies, except as follows:

201.7.2.10 APPLIED PARTS

Addition:

If a DENTAL EQUIPMENT has only one APPLIED PART or APPLIED PARTS of different degree of protection against electric shock, it is sufficient to place the appropriate symbol of the lowest degree of protection against electrical shock once on the ENCLOSURE of the DENTAL EQUIPMENT. APPLIED PARTS with higher degree of protection shall be marked individually.

201.7.2.11 * Mode of operation

Addition:

For DENTAL HANDPIECES, no marking is necessary.

For DENTAL ELECTRICAL MOTORS, no marking is necessary.

201.7.9 ACCOMPANYING DOCUMENTS

201.7.9.2 Instructions for use

Additional subclause:

201.7.9.2.101 * Additional instructions for use

For DENTAL HANDPIECES intended for non-CONTINUOUS OPERATION, the duty cycle shall be provided.

For DENTAL ELECTRICAL MOTORS no explanation of the duty cycle is necessary.

<https://standards.iteh.ai/catalog/standards/sist/a80f56fb-44f5-4a4d-825b-a15ecf9a8862/iec-80601-2-60-2019>

201.8 Protection against electrical HAZARDS from ME EQUIPMENT

Clause 8 of the general standard applies, except as follows:

Additional subclause:

201.8.4.101 * NEUTRAL ELECTRODE monitoring circuit

The requirements specified in 201.8.4.101 of IEC 60601-2-2:2017 do not apply for HIGH FREQUENCY SURGICAL EQUIPMENT having a RATED OUTPUT POWER not exceeding 50W.

201.8.5.2 Separation of PATIENT CONNECTIONS

Additional subclause:

201.8.5.2.101 * APPLIED PARTS that form one single APPLIED PART

In a DENTAL EQUIPMENT the DENTAL HANDPIECES may be considered as multiple functions of one APPLIED PART or PATIENT CONNECTIONS of one APPLIED PART.

The combination of several APPLIED PARTS to form one single APPLIED PART is only permitted if no HAZARDOUS SITUATION occurs when applying several APPLIED PARTS concurrently.

201.8.7 LEAKAGE CURRENTS and PATIENT AUXILIARY CURRENTS

201.8.7.1 * General requirements

Addition:

bb) Water columns inside of an APPLIED PARTS or its multiple connections can be considered as electric impedances according to their geometric dimensions and the resistance of the water. If this method is used the MANUFACTURER shall state the maximum fluid conductivity in the instructions for use and technical description (e.g. installation instructions).

Additional subclause:

201.8.7.3.101 * Thermal effects of HF LEAKAGE CURRENTS

Requirements and tests 1) and 2) specified in 201.8.7.3.101 of IEC 60601-2-2:2017 do not apply for HF SURGICAL EQUIPMENT having a RATED OUTPUT POWER not exceeding 50 W and intended for use without a NEUTRAL ELECTRODE.

201.8.8.3 * Dielectric strength

Addition:

For circuits according to 201.8.9.1.12 SECONDARY CIRCUITS, the test voltages indicated in Table 201.101 shall be used for testing solid insulation:

Table 201.101 – Test voltages for solid insulation for SECONDARY CIRCUITS according to 201.8.9.1.12

PEAK WORKING VOLTAGE (U) in V	One MOPP U in V RMS	Two MOPP U in V RMS
≤ 71	500	500
≤ 50 (RMS)		

For higher PEAK WORKING VOLTAGES, IEC 60601-1:2005 and IEC 60601-1:2005/AMD1:2012 shall be applied.

201.8.9 * CREEPAGE DISTANCES and AIR CLEARANCES

Amendment:

For CREEPAGE DISTANCES and AIR CLEARANCES, 8.9 of the general standard shall be used without modification or with the following changes.

Alternatively CREEPAGE DISTANCES and AIR CLEARANCES of this particular standard may be applied. In this case, 8.9 of the general standard applies except as follows:

201.8.9.1 Values

201.8.9.1.2 CREEPAGE DISTANCES and AIR CLEARANCES complying with IEC 60950-1

The subclause of the general standard does not apply.

201.8.9.1.3 CREEPAGE DISTANCES across glass, mica, ceramic and similar materials

Amendment:

For SECONDARY CIRCUITS, the subclause of the general standard does not apply.

201.8.9.1.4 Minimum CREEPAGE DISTANCE

Amendment:

For SECONDARY CIRCUITS, the subclause of the general standard does not apply.

201.8.9.1.6 Interpolation

Amendment:

For SECONDARY CIRCUITS, the subclause of the general standard does not apply.

201.8.9.1.7 Material groups classification

Amendment:

For SECONDARY CIRCUITS, the subclause of the general standard does not apply.

201.8.9.1.8 Pollution degree classification

Amendment:

For SECONDARY CIRCUITS, 4.6.2 of IEC 60664-1:2007 applies.
IEC 80601-2-60:2019
https://standards.iteh.ai/catalog/standards/sist/a80f561b-44f5-4a4d-825b-a15ecf9a8862/iec-80601-2-60-2019

201.8.9.1.9 Overvoltage category classification

The subclause of the general standard does not apply.

201.8.9.1.11 SUPPLY MAINS overvoltage

Replacement:

This particular standard relates to overvoltage category II according to IEC 60664-1:2007.

201.8.9.1.12 SECONDARY CIRCUITS

Replacement:

For separation between SUPPLY MAINS and SECONDARY CIRCUITS two MOPP according to Table 12 of IEC 60601-1:2005 and IEC 60601-1:2005/AMD1:2012 applies.

For the separation between SUPPLY MAINS and APPLIED PARTS two MOPP according to Table 12 of IEC 60601-1:2005 and IEC 60601-1:2005/AMD1:2012 shall be applied.

Within SECONDARY CIRCUITS, for the separation within and between APPLIED PARTS of DENTAL EQUIPMENT for AIR CLEARANCES and CREEPAGE DISTANCES, the following apply:

- a) IEC 60664-1:2007 for RATED frequencies up to 30 kHz with the following tables and conditions:

The DENTAL UNIT, DENTAL PATIENT CHAIRS and DENTAL OPERATING LIGHTS shall be capable of withstanding 4 kV withstand impulse voltage in SUPPLY MAINS circuits.