

SLOVENSKI STANDARD SIST EN 55016-1-4:2011/A1:2013

01-januar-2013

Specifikacija merilnih naprav in metod za merjenje radiofrekvenčnih motenj in odpornosti - 1-4. del: Merilne naprave za merjenje radiofrekvenčnih motenj in odpornosti - Antene in preskuševališča za meritve sevanih motenj - Dopolnilo A1

Specification for radio disturbance and immunity measuring apparatus and methods -Part 1-4: Radio disturbance and immunity measuring apparatus - Antennas and test sites for radiated disturbance measurements

Anforderungen an Geräte und Einrichtungen sowie Festlegung der Verfahren zur

Anforderungen an Geräte und Einrichtungen sowie Festlegung der Verfahren zur Messung der hochfrequenten Störaussendung (Funkstörungen) und Störfestigkeit - Teil 1-4: Geräte und Einrichtungen zur Messung der hochfrequenten Störaussendung (Funkstörungen) und Störfestigkeit - Zusatz-/Hilfseinrichtungen - Gestrahlte Störaussendung Köraussendung

Spécifications des méthodes et des appareils de mesure des perturbations radioélectriques et de l'immunité aux perturbations radioélectriques - Partie 1-4: Appareils de mesure des perturbations radioélectriques et de l'immunité aux perturbations radioélectriques - Antennes et emplacements d'essai pour les mesures des perturbations rayonnées

Ta slovenski standard je istoveten z: EN 55016-1-4:2010/A1:2012

<u>ICS:</u>

17.240Merjenje sevanja33.100.20Imunost

Radiation measurements Immunity

SIST EN 55016-1-4:2011/A1:2013

en,fr,de

iTeh STANDARD PREVIEW (standards.iteh.ai)

SIST EN 55016-1-4:2011/A1:2013

EUROPEAN STANDARD NORME EUROPÉENNE EUROPÄISCHE NORM

EN 55016-1-4/A1

November 2012

ICS 33.100.10; 33.100.20

English version

Specification for radio disturbance and immunity measuring apparatus and methods -Part 1-4: Radio disturbance and immunity measuring apparatus -Antennas and test sites for radiated disturbance measurements (CISPR 16-1-4:2010/A1:2012)

Spécifications des méthodes et des Anforderungen an Geräte und appareils de mesure des perturbations Einrichtungen sowie Festlegung radioélectriques et de l'immunité der Verfahren zur Messung der aux perturbations radioélectriques hochfrequenten Störaussendung Partie 1-4: Appareils de mesure des (Funkstörungen) und Störfestigkeit perturbations radioélectriques Teil 1-4: Geräte und Einrichtungen et de l'immunité aux perturbations zur Messung der hochfrequenten radioélectriques - Antennes et emplacements d'essai pour radioélectriques -Störaussendung (Funkstörungen) und Störfestigkeit les mesures des perturbations rayonnées des ite Antennen und Messplätze für Messungen (CISPR 16-1-4:2010/A1:2012) der gesterahlten Störaussendung (CISPR 16-1-4:2010/A1:2012)

<u>SIST EN 55016-1-4:2011/A12013</u> **FK 10-1-4:20** https://standards.iteh.ai/catalog/standards/sist/af1cba3d-4978-417a-af0efc43db8b215a/sist-en-55016-1-4-2011-a1-2013

This amendment A1 modifies the European Standard EN 55016-1-4:2010; it was approved by CENELEC on 2012-08-15. CENELEC members are bound to comply with the CEN/CENELEC Internal Regulations which stipulate the conditions for giving this amendment the status of a national standard without any alteration.

Up-to-date lists and bibliographical references concerning such national standards may be obtained on application to the CEN-CENELEC Management Centre or to any CENELEC member.

This amendment exists in three official versions (English, French, German). A version in any other language made by translation under the responsibility of a CENELEC member into its own language and notified to the CEN-CENELEC Management Centre has the same status as the official versions.

CENELEC members are the national electrotechnical committees of Austria, Belgium, Bulgaria, Croatia, Cyprus, the Czech Republic, Denmark, Estonia, Finland, Former Yugoslav Republic of Macedonia, France, Germany, Greece, Hungary, Iceland, Ireland, Italy, Latvia, Lithuania, Luxembourg, Malta, the Netherlands, Norway, Poland, Portugal, Romania, Slovakia, Slovenia, Spain, Sweden, Switzerland, Turkey and the United Kingdom.

CENELEC

European Committee for Electrotechnical Standardization Comité Européen de Normalisation Electrotechnique Europäisches Komitee für Elektrotechnische Normung

Management Centre: Avenue Marnix 17, B - 1000 Brussels

© 2012 CENELEC - All rights of exploitation in any form and by any means reserved worldwide for CENELEC members.

Foreword

The text of document CISPR/A/995/FDIS, future amendment 1 to edition 3 of CISPR 16-1-4, prepared by SC A "Radio-interference measurements and statistical methods" of IEC/TC CISPR "International special committee on radio interference" was submitted to the IEC-CENELEC parallel vote and approved by CENELEC as EN 55016-1-4:2010/A1:2012.

The following dates are fixed:

•	latest date by which the document has to be implemented at national level by publication of an identical national standard or by endorsement	(dop)	2013-05-16
•	latest date by which the national standards conflicting with the document have to be withdrawn	(dow)	2015-08-15

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. CENELEC [and/or CEN] shall not be held responsible for identifying any or all such patent rights.

Endorsement notice

The text of the International Standard CISPR 16-1-4:2010/A1:2012 was approved by CENELEC as a European Standard without any modification. DARD PREVE

(standards.iteh.ai)

Annex ZA (normative)

Normative references to international publications with their corresponding European publications

In Annex ZA of EN 55016-1-4:2010, **Delete** the existing date and **delete** the two existing amendments from CISPR/TR 16-3:

Publication	Year	Title	<u>EN/HD</u>	Year
CISPR/TR 16-3	-	Specification for radio disturbance and immunity measuring apparatus and methods - Part 3: CISPR technical reports	-	-

iTeh STANDARD PREVIEW (standards.iteh.ai)

iTeh STANDARD PREVIEW (standards.iteh.ai)



CISPR 16-1-4

Edition 3.0 2012-07

INTERNATIONAL STANDARD

NORME INTERNATIONALE

INTERNATIONAL SPECIAL COMMITTEE ON RADIO INTERFERENCE COMITÉ INTERNATIONAL SPÉCIAL DES PERTURBATIONS RADIOÉLECTRIQUES

BASIC EMC PUBLICATION

PUBLICATION FONDAMENTALE EN CEM

AMENDMENT 1 11 EN STANDARD PREV AMENDEMENT 1 (standards.iteh.ai)

Specification for radio disturbance and immunity measuring apparatus and methods – Part 1-4: Radio disturbance and immunity measuring apparatus – Antennas and test https://standards.iteh.ai/catalog/standards/sist/afl/cba3d-4978-417a-af0esites for radiated disturbance measurements intervention of the standards/sist/afl/cba3d-4978-417a-af0e-

Spécifications des méthodes et des appareils de mesure des perturbations radioélectriques et de l'immunité aux perturbations radioélectriques – Partie 1-4: Appareils de mesure des perturbations radioélectriques et de l'immunité aux perturbations radioélectriques – Antennes et emplacements d'essai pour les mesures des perturbations rayonnées

INTERNATIONAL ELECTROTECHNICAL COMMISSION

COMMISSION ELECTROTECHNIQUE INTERNATIONALE

PRICE CODE CODE PRIX



ICS 33.100.10; 33.100.20

ISBN 978-2-83220-150-3

Warning! Make sure that you obtained this publication from an authorized distributor. Attention! Veuillez vous assurer que vous avez obtenu cette publication via un distributeur agréé.

 Registered trademark of the International Electrotechnical Commission Marque déposée de la Commission Electrotechnique Internationale

- 2 - CISPR 16-1-4 Amend.1 © IEC:2012

FOREWORD

This amendment has been prepared by subcommittee A: Radio-interference measurements and statistical methods, of IEC technical committee CISPR: International special committee on radio interference.

The text of this amendment is based on the following documents:

FDIS	Report on voting
CISPR/A/995/FDIS	CISPR/A/1005/RVD

Full information on the voting for the approval of this amendment can be found in the report on voting indicated in the above table.

The committee has decided that the contents of this publication will remain unchanged until the stability date indicated on the IEC web site under "http://webstore.iec.ch" in the data related to the specific publication. At this date, the publication will be

- reconfirmed,
- withdrawn,
- replaced by a revised edition, or
- amended.

iTeh STANDARD PREVIEW (standards.iteh.ai)

CISPR 16-1-4 Amend.1 © IEC:2012

- 3 -

INTRODUCTION (to Amendment 1)

This amendment introduces the Reference Site Method (RSM). In addition to introducing new content, Clause 5 is significantly restructured. To aid the reader in navigating this amendment, the following table provides a comparison of subclauses in the existing Edition 3.0 with those in this amendment. This introduction will be removed before the subsequent edition is published.

Comparison of Clause 5 between original Edition 3.0 and Amendment 1

	Original Edition 3.0	Amendment 1		
5	Test sites for the measurement of radio disturbance field strength for the frequency range of 30 MHz to 1 000 MHz	5	Test sites for the measurement of radio disturbance field strength for the frequency range of 30 MHz to 1000 MHz	
5.1	General	5.1	General	
5.2	OATS	5.2	OATS	
5.2.1	General	5.2.1	General	
5.2.2	Weather protection enclosure	5.2.2	Weather protection enclosure	
5.2.3	Obstruction-free area		Obstruction-free area	
5.2.4	Ambient radio frequency environment of a test site	5.2.4	Ambient radio frequency environment of a test site	
5.2.5	Ground plane ITeh STANDA	5.2 .5	Ground plane	
5.2.6	OATS validation procedure (standard	s itel	nai)	
5.3	Test site suitability for other ground-plane test sites	5.3	Suitability of other test sites	
5.3.1	General https://standards.iteh.ai/catalog/standa	-4.2011/A 5.3.1 rds/sist/afl	Other ground-plane test sites	
5.3.2	Normalized site attenuation for alternative test5 sites)1(5.3.2-2()∏est sites∋without ground plane (FAR)	
5.3.3	Site attenuation			
5.3.4	Conducting ground plane			
5.4	Test site suitability without ground plane	5.4	Test site validation	
5.4.1	Measurement considerations for free space test sites, as realized by fully-absorber-lined shielded enclosures	5.4.1	General	
5.4.2	Site performance	5.4.2	Overview of test site validations	
5.4.3	Site validation criteria	5.4.3	Principles and values of the NSA method for OATS and SAC	
		5.4.4	Reference site method for OATS and SAC	
		5.4.5	Validation of an OATS by the NSA method	
		5.4.6	Validation of a weather-protection-enclosed OATS or a SAC	
		5.4.7	Site validation for FARs	
5.5	Evaluation of set-up table and antenna tower	5.5	Evaluation of set-up table and antenna tower	
5.5.1	General	5.5.1	General	
5.5.2	Evaluation procedure for set-up table influences	5.5.2	Evaluation procedure for set-up table influences	

2 Normative references

Remove the existing date and remove the two existing amendment references from the existing reference to CISPR/TR 16-3.

3.1.12

quasi-free space test-site

Replace the existing term by the new term "quasi free-space test site".

3.1.15 semi-anechoic chamber

SAC

Replace the existing definition of this term by the following new definition:

shielded enclosure in which all surfaces except the metal floor are covered with material that absorbs electromagnetic energy (i.e. RF absorber) in the frequency range of interest

3.1.19

test volume

Replace, in the existing definition of this term, "the FAR" by "a FAR".

Replace, in the existing Note of this term, "quasi-free space condition" by "quasi free-space condition", and "the FAR" by "a FAR".

Add, after the existing definition 3.1.21, the following new terms and definitions 3.1.22, 3.1.23, 3.1.24, 3.1.25, 3.1.26 and 3.1.27: (standards.iteh.ai)

3.1.22

antenna factorSIST EN 55016-1-4:2011/A1:2013AFhttps://standards.iteh.ai/catalog/standards/sist/af1cba3d-4978-417a-af0e-
fc43db8b215a/sist-en-55016-1-4-2011-a1-2013

ratio of the electric field strength of an incident plane wave to the voltage induced across a specified load (typically 50 Ω) connected to the antenna

NOTE 1 F_a is affected by the load impedance connected to the antenna radiating elements, and is frequency dependent. For a biconical antenna this impedance could be up to 200 Ω . For antennas with no balun the impedance is equal to the load impedance, typically 50 Ω .

NOTE 2 Usually, the AF is defined for the plane wave incident from the direction corresponding with the maximum gain of the antenna and at a specified point of the antenna.

NOTE 3 The AF has the physical dimension of inverse metres (m^{-1}) and measured data are normally expressed in dB(m^{-1}). In radiated emission measurements, if F_a is known, the strength of an incident field, E, can be estimated from a reading, V, of a measuring receiver connected to the antenna as follows:

$E = V + F_a$

where *E* is in dB(μ V/m), *V* is in dB(μ V) and *F*_a is in dB(m⁻¹).

3.1.23 antenna factor, free-space

 $F_{a \text{ fs}}$ AF of an antenna located in a free-space environment

NOTE $F_{a fs}$ is a measurand for uncertainty calculation for antenna calibration. For NSA measurements $F_{a fs}$ is an input quantity for uncertainty calculation.

CISPR 16-1-4 Amend.1 © IEC:2012 - 5 -

3.1.24

antenna pair reference site attenuation

^AAPR

set of site attenuation measurement results for both vertical and horizontal polarizations using a pair of antennas separated by a defined distance at an ideal open-area test site, with one antenna at a specified fixed height above the ground plane, and the other antenna scanned over a specified height range in which the minimum insertion loss is recorded

NOTE 1 A_{APR} is a measurand for uncertainty calculation.

NOTE 2 A_{APR} measurements are used for comparison to corresponding site attenuation measurements of a COMTS to evaluate the performance of the COMTS.

3.1.25

antenna reference point

midpoint of an antenna from which the distance to the EUT or second antenna is measured

NOTE The antenna reference point is either defined by the manufacturer using a marker on LPDA antennas or by the calibration laboratory.

3.1.26

ideal open-area test site

open-area test site having a perfectly flat, perfectly conducting ground plane of infinite area, and with no reflecting objects except the ground plane

NOTE An ideal OATS is a theoretical construct that is used in the definition of the measurand A_{APR} and in the calculation of the theoretical normalized site attenuation A_{N} for ground plane sites.

(standards.iteh.ai)

3.1.27 reference test site REFTS

open-area test site with metallic ground blane and tightly specified site attenuation performance in horizontal and vertical electric field polarizations 8-41/a-af0etc43db8b215a/sist-en-55016-1-4-2011-a1-2013

3.2 Abbreviations

Add, before the existing abbreviation EUT of this subclause, the following new sentence:

The following are abbreviations used in this standard that are not already given in 3.1.

Add the following new abbreviation to the existing list:

RSM Reference site method

Delete, in the existing list, the entire abbreviation SAC.

4.5.3 Antenna characteristics

Replace, in item c) 2) of the list, in the paragraph after the variable list of Equation (4), the existing text "corrections or as" by "a correction with associated".

Delete the existing NOTE 3 in the existing item c) 3) of the list.

Replace, in the list of quantities below Equation (4), existing quantity " h_1 " by " h_2 " and replace existing quantity " h_2 " by " h_1 ".

- 6 -

CISPR 16-1-4 Amend.1 © IEC:2012

Replace the existing Figure 1 by the following new figure:



Figure 1 – Schematic of radiation from EUT reaching an LPDA antenna directly and via ground reflections on a 3 m site, showing the half beamwidth, what the reflected ray

4.5.4.2 Balun DM/CM conversion theckrds.iteh.ai)

Replace, in the existing Note 2_{315} from 3_{50} by $_{1-4}$ from 3_{50} by $_{1-4}$ from 3_{50} by $_{1-4}$ from $3_{1/4}$ and g_{13} ±4 dB criterion" by "± 4 dB site validation criterion" https://standards.iteh.ai/catalog/standards/sist/aflcba3d-4978-417a-af0e-

fc43db8b215a/sist-en-55016-1-4-2011-a1-2013

Replace, in the existing Note 3, "fully-anechoic room" by "FAR", and "NSA (normalized site attenuation)" by "site validation".

4.5.5 Cross-polar response of antenna

Replace, in the third paragraph of this subclause, "quasi-free space conditions" by "quasi free-space conditions", and "high-quality anechoic chamber" by "high-quality fully anechoic room".

5.1 General

Replace the existing text of this subclause by the following new text:

An environment is required that assures valid, repeatable measurement results of disturbance field strength from an EUT. For an EUT that can only be tested at its place of use, other provisions shall be utilized (i.e. see details on in-situ measurements in CISPR 16-2-3).

5.2.1 General

Replace the existing text of this subclause by the following new text:

An OATS is an area characterized by cleared level terrain and the presence of a ground plane. To meet the validation requirements of this standard, a metallic ground plane is recommended. Such a test site shall be free of buildings, electric lines, fences, trees, etc. and free from underground cables, pipelines, etc., except as required to supply and operate the EUT. Refer to Annex D for specific construction recommendations of an OATS for disturbance field-strength measurements in the range of 30 MHz to 1 000 MHz. The site validation

CISPR 16-1-4 Amend.1 © IEC:2012 - 7 -

procedures for an OATS are given in 5.4.4 and 5.4.5. Annex F explains the basis for the acceptability criterion.

5.2.3 Obstruction-free area

Replace the third paragraph of this subclause by the following new paragraph:

For this ellipse, the length of the path taken by the undesired indirect ray reflected from any object on the perimeter is twice the length of the path taken by the direct ray between the foci. If a large EUT is installed on the turntable, the obstruction-free area shall be expanded so that the obstruction clearance distances exist from the perimeter of the EUT.

Replace, in the last sentence of the fourth paragraph of this subclause, the existing words "separation distance" by "measurement distance".

Replace the existing Figure 2 by the following new figure:



IEC 1073/12

Figure 2 – Obstruction-free area of a test site with a turntable (see 5.2.3)

- 8 -

CISPR 16-1-4 Amend.1 © IEC:2012



Replace the existing Figure 3 by the following new figure:

5.2.4 Ambient radio frequency environment of a test site 243db8b215a/sist-en-55016-1-4-2011-a1-2013

Replace the first paragraph of this subclause before the list by the following new paragraph:

The ambient radio frequency levels at an OATS shall be sufficiently low compared to the levels of measurements to be performed. The quality of the site in this regard may be evaluated under four categories, listed below in order of merit:

Replace the existing note of this subclause by the following new note:

NOTE A measured ambient level of 20 dB or more below the emission limit is considered optimum.

5.2.5 Ground plane

Replace the existing text in this subclause by the following new text:

The OATS ground plane can be at earth level or elevated on a suitably sized platform or horizontal rooftop site. A metal ground plane is preferred, but for certain equipment and applications, product publications may recommend other site types. Adequacy of the metal ground plane will be dependent on whether the test site meets the site validation requirements of 5.4. If metallic material is not used, caution is required to select a site that does not change its reflective characteristics with time, weather conditions, or effects due to buried metallic material such as pipes, conduits, and non-homogeneous soil. Such sites generally give different SA characteristics compared to those with metallic surfaces.

5.2.6 OATS validation procedure

Delete this existing subclause, as well as the existing subclauses 5.2.6.1 to 5.2.6.4.

CISPR 16-1-4 Amend.1 © IEC:2012 - 9 -

5.3 Test site suitability for other ground-plane test sites

Replace the existing title of this subclause by the following new title:

5.3 Suitability of other test sites

5.3.1 General

Replace the existing title of this subclause by the following new title:

5.3.1 Other ground-plane test sites

Replace the existing text of this subclause by the following new text and new note:

There are many different test sites and facilities that have been constructed and used to make radiated emission measurements. Most are protected from the weather and the adverse effects of the radio frequency ambient. In a SAC, all walls and the ceiling are equipped with appropriate absorbing material. The floor consists of a metallic ground plane to emulate an OATS. A SAC isolates the receiving antenna from the RF ambient environment, and permits EUT testing independent of weather conditions.

Whenever construction material encloses a ground-plane test site, it is possible that the results of a validation measurement at a single location, as specified in 5.4.5, are not adequate to show acceptability of such an alternative site.

iTeh STANDARD PREVIEW

To evaluate suitability of an alternative ground-plane test site, the procedure of 5.4.6 shall be used, which is based on making multiple validation measurements throughout a volume occupied by the EUT. These validation measurement results shall all be within a tolerance of \pm 4 dB for a site to be judged suitable as an equivalent to an OATS.

NOTE SACs typically meet the site quality categories listed in 5.2.4 ic43db80215a/sist-en-55016-1-4-2011-a1-2013

5.3.2 Normalized site attenuation for alternative test sites

Replace the existing title of this subclause by the following new title:

5.3.2 Test sites without ground plane (FAR)

Replace the existing text of this subclause by the following new text and new note:

A fully-absorber-lined shielded enclosure, also known as a fully-anechoic room (FAR), can be used for radiated emission measurements. When a FAR site is used, appropriate radiated emission limits shall be defined in relevant standards (generic, product or product family standards). Compliance of an EUT with the requirements for the protection of radio-services (limits) shall be evaluated at FAR sites using similar methods as for tests done at an OATS.

A FAR is intended to simulate a free-space environment such that only the direct ray from the transmitting antenna or EUT reaches the receiving antenna. All indirect and reflected waves shall be minimized by appropriate placement of absorbing material on all walls, ceiling and floor of a FAR. Like a SAC, a FAR isolates the receiving antenna from the RF ambient environment, and permits EUT testing independent of weather conditions.

NOTE FARs typically meet the site quality categories listed in 5.2.4.

5.3.3 Site attenuation

Delete this existing subclause and its title.