

SLOVENSKI STANDARD SIST EN 14881:2006

01-januar-2006

Dcj fý]bg_c`U_hjj bY`gbcj]`!`8 c`c Yj Ub^Y`j gYVbcghj`B!fl !X]a Yhj`Ua]bcdfcd]`Ł! U TUA XUI U TUA Xcdfcd VYIUbIN ! A YhcXUd bg Y fca Uhc fUZ Y"

Surface active agents - Determination of N-(3-dimethylaminopropyl)-alkylamide content in alkylamidopropylbetaines - Gas chromatographic method

Grenzflächenaktive Stoffe - Bestimmung des Gehaltes an N-(3-dimethyl-aminopropyl)alkylamid in Alkylamidopropylbetainen - Gaschromatographisches Verfahren

Agents de surface - Détermination de la teneur en N-(3-diméthylaminopropyl)-alkylamide dans les alkylamidopropylbétaines - Méthode par chromatographie en phase gazeuse

Ta slovenski standard je istoveten z: EN 14881:2005

ICS:

71.100.40 Površinsko aktivna sredstva Surface active agents

SIST EN 14881:2006

en



iTeh STANDARD PREVIEW (standards.iteh.ai)

<u>SIST EN 14881:2006</u> https://standards.iteh.ai/catalog/standards/sist/1eea7d9c-b89f-46ff-b75ed4a68a19a832/sist-en-14881-2006

SIST EN 14881:2006

EUROPEAN STANDARD NORME EUROPÉENNE EUROPÄISCHE NORM

EN 14881

August 2005

ICS 71.100.40

English Version

Surface active agents - Determination of N-(3dimethylaminopropyl)-alkylamide content in alkylamidopropylbetaines - Gas chromatographic method

Agents de surface - Détermination de la teneur en N-(3diméthylaminopropyl)-alkylamide dans les alkylamidopropylbétaïnes - Méthode par chromatographie en phase gazeuse Grenzflächenaktive Stoffe - Bestimmung des Gehaltes an N-(3-dimethyl-aminopropyl)-alkylamid in Alkylamidopropylbetainen - Gaschromatographisches Verfahren

This European Standard was approved by CEN on 8 July 2005.

CEN members are bound to comply with the CEN/CENELEC Internal Regulations which stipulate the conditions for giving this European Standard the status of a national standard without any alteration. Up-to-date lists and bibliographical references concerning such national standards may be obtained on application to the Central Secretariat or to any CEN member.

This European Standard exists in three official versions (English, French, German). A version in any other language made by translation under the responsibility of a CEN member into its own language and notified to the Central Secretariat has the same status as the official versions.

CEN members are the national standards bodies of <u>Austria, Belgium, Cypr</u>us, Czech Republic, Denmark, Estonia, Finland, France, Germany, Greece, Hungary, Iceland, Ireland, Italy, Latvia, Lithuania, Luxembourg, Malta, Netherlands, Norway, Poland, Portugal, Slovakia, Slovenia, Spain, Sweden, Switzerland and United Kingdom, 4468a19a832/sist-en-14881-2006



EUROPEAN COMMITTEE FOR STANDARDIZATION COMITÉ EUROPÉEN DE NORMALISATION EUROPÄISCHES KOMITEE FÜR NORMUNG

Management Centre: rue de Stassart, 36 B-1050 Brussels

© 2005 CEN All rights of exploitation in any form and by any means reserved worldwide for CEN national Members. Ref. No. EN 14881:2005: E

SIST EN 14881:2006

EN 14881:2005 (E)

Contents

		Page
Foreword3		
1	Scope	4
2	Normative references	4
3	Principle	4
4	Reagents	4
5	Apparatus	5
6	Sampling and preparation of the sample	5
7	Procedure	5
8	Calculation and expression of results	6
9	Precision	6
10	Test report	6
Annex	A (informative) N-(3-(dimethylaminopropyl)-undecylamide.	7
Annex	B (informative) Reference gas chromatogram: ds. iteh. ai)	8
Annex	C (informative) Validation method	10
Annex	D (informative) Inter-laboratory test results	11
Bibliog	https://standards.iteh.ai/catalog/standards/sist/1eea7d9c-b89f-46ff-b75e- graphyd4a68a19a832/sist-en-14881-2006	12

Foreword

This European Standard (EN 14881:2005) has been prepared by Technical Committee CEN/TC 276 "Surface active agents", the secretariat of which is held by AFNOR.

This European Standard shall be given the status of a national standard, either by publication of an identical text or by endorsement, at the latest by February 2006, and conflicting national standards shall be withdrawn at the latest by February 2006.

According to the CEN/CENELEC Internal Regulations, the national standards organizations of the following countries are bound to implement this European Standard: Austria, Belgium, Cyprus, Czech Republic, Denmark, Estonia, Finland, France, Germany, Greece, Hungary, Iceland, Ireland, Italy, Latvia, Lithuania, Luxembourg, Malta, Netherlands, Norway, Poland, Portugal, Slovakia, Slovenia, Spain, Sweden, Switzerland and United Kingdom.

iTeh STANDARD PREVIEW (standards.iteh.ai)

<u>SIST EN 14881:2006</u> https://standards.iteh.ai/catalog/standards/sist/1eea7d9c-b89f-46ff-b75ed4a68a19a832/sist-en-14881-2006

1 Scope

This European Standard specifies a method for the determination of the content of free N-(3-dimethylaminopropyl)alkylamides (amidoamine) in alkylamidopropylbetaines, expressed in grams per 100 g of product.

This method is applicable in the range between 0,02 g and 1,0 g of amidoamine per 100 g of product.

2 Normative references

The following referenced documents are indispensable for the application of this European Standard. For dated references, only the edition cited applies. For undated references, the latest edition of the referenced document (including any amendments) applies.

EN ISO 3696, Water for analytical laboratory use – Specification and test methods (ISO 3696:1987)

ISO 607, Surface active agents and detergents – Methods of sample division

3 Principle

The method is based on the amidoamine extraction with diethyl ether at alkaline pH and subsequent analysis of the organic extract by GLC-FID. The chromatogram resolves the different amidoamines according to their alkyl chain length. The result is calculated from the sum of all the chain homologues.

(standards.iteh.ai)

4 Reagents

SIST EN 14881:2006

WARNING: Your attention is/drawn to the regulations covering the handling of hazardous substances. Technical organisational and personal protection measures should be observed.

During the analysis, unless otherwise specified, use only reagents of recognized analytical grade that have been checked in advance as not interfering with the analytical results.

4.1 Water, complying with grade 3 as defined in EN ISO 3696.

NOTE If the water is purified via ion-exchange resins, ensure that no cationic or anionic species from the resins cause interference.

- 4.2 Ethanol
- 4.3 Diethyl ether
- **4.4** N-(3-dimethylaminopropyl)-undecylamide (amidoamine-C11), purity \ge 98 % m/m.

NOTE The N-(3-dimethylaminopropyl) alkylamide standard preparation and purity determination are given in Annex A.

4.5 Potassium hydroxide, ethanolic solution *c* (KOH)= 1 mol/l.

4.6 Internal standard solution

Weigh in a 50 ml volumetric flask, to the nearest 0,1 mg, approximately 0,2 g of the amidoamine-C11 (4.4). Dissolve and make up to the mark with ethanol (4.2). Keep tightly closed.

- **4.7** Sodium chloride, aqueous solution, 30 % m/m.
- **4.8 Phenolphthalein**, solution of 1 % m/V in ethanol.

5 Apparatus

Ordinary laboratory apparatus and the following.

5.1 Gas chromatograph, with capillary split injector, flame ionisation detector and integrator or computer.

5.2 Fused silica capillary column, coating with bonded deactivated poly (5 % diphenyl/95 % dimethyl) siloxane, length 30 m, internal diameter 0,25 mm, film thickness 0,50 μm (PTA-5¹).

5.3 Syringe, capacity 1 µl.

Sampling and preparation of the sample 6

The laboratory sample shall be prepared and stored in accordance with ISO 607.

Procedure 7

7.1 Calibration

Since it is assumed that the response factor of the different homologues is the same no extra calibration is needed (see Annex C).

7.2 Determination

iTeh STANDARD PREVIEW

(standards.iteh.ai) Weigh to the nearest 0,1 mg, approximately 3,0 g of sample in a 15 ml vial with snap cap.

Add 2 ml of NaCl solution (4.7), and homogenise

g/standards/sist/1eea7d9c-b89f-46ff-b75e-

Add 1 ml of internal standard solution (4.6), measured with a one-mark pipette.

Add two drops of the phenolphthalein solution (4.8) and enough potassium hydroxide solution (4.5) to red change and add few extra drops of potassium hydroxide solution.

Add 4 ml of diethyl ether (4.3) measured with a measuring cylinder.

Stopper and shake the tube vigorously for about 1 min.

Let stand until the phases separate (usually 2 min will be enough).

Analyse the organic phase by injecting 1 μ l of the organic layer directly into the gas chromatograph (5.1).

Set up the gas chromatograph to give results similar to the reference chromatogram in Figure B.1.

¹⁾ PTA-5 is the trade name of product supplied by Supelco (U.S.A). This information is given for the convenience of users of this European Standard and does not constitute an endorsement by CEN of the product named.

8 Calculation and expression of results

The content of total free amidoamine, *c*, in the sample, expressed in grams per 100 g of the product, is calculated by the equation:

$$c = \frac{m_{\rm is} \times A \times w}{m \times A_{\rm is}} \tag{1}$$

where

- $m_{\rm IS}$ is the mass of the internal standard, in grams, present in 1 millilitre;
- *m* is the mass of the test sample, in grams;
- $A_{\rm IS}$ is the peak area of the internal standard in the chromatogram of the sample solution;

A is the sum of the areas of all the peaks corresponding to alkylamidoamine homologues (except that of the internal standard) in the chromatogram of the sample solution;

w is the purity of internal standard in grams per 100 g (see Annex A).

The relative response factor of all the N-(3-dimethylaminopropyl)- alkylamide homologues with respect to the internal standard N-(3-dimethylaminopropyl)- undecylamide is assumed to be 1 (see Annex C).

9 Precision

SIST EN 14881:2006 https://standards.iteh.ai/catalog/standards/sist/1eea7d9c-b89f-46ff-b75ed4a68a19a832/sist-en-14881-2006

9.1 Repeatability limit

The absolute difference between two independent single test results, obtained using the same method on identical test material in the same laboratory by the same operator using the same equipment within a short interval of time, will not exceed the repeatability limit, *r*, in more than 5 % of the cases.

Precision data are given in Annex D.

9.2 Reproducibility limit

The absolute difference between two independent single test results, obtained using the same method on identical test material in different laboratories by different operators using different equipment, will not exceed the reproducibility limit, R, in more than 5 % of the cases.

Precision data are given in Annex D.

10 Test report

The test report shall include the following information:

- a) all information necessary for the complete identification of the sample;
- b) method used (a reference to this European Standard, i.e. EN 14881);
- c) test results;
- d) details of any operation not specified in this European Standard or in the standards to which reference is made, and any operations regarded as optional, as well as any incidents like to have affected the results.

Annex A

(informative)

N-(3-dimethylaminopropyl)-undecylamide

A.1 Preparation of N-(3-dimethylaminopropyl)-undecylamide

250 g (1,34 mol) undecanoic acid of about 98 % purity is weighed together with 1,5 g of 50 % hypophosphorous acid into a 500 ml four-neck round-bottom flask equipped with an immersion thermometer, an addition funnel, a mechanical stirrer, condenser and nitrogen purge systems.

206 g (2,01 mole) dimethylaminopropylamine (DMAPA) is added drop wise during 30 min at 140 °C under N₂ atmosphere. Temperature is then raised up to 200 °C within 3 h while the water is removed. The reaction is completed at 200 °C for an additional 2 h, when the acid value becomes lower than 1 mg KOH/g.

The reactor content is cooled down to 180 °C and the condenser system removed.

Keeping the system at 180 °C under a reduced pressure of 2 mm/Hg for an additional 2 h distils off the residual DMAPA.

(standards.iteh.ai)

A.2 Determination of purity of N-(3-dimethylaminopropyl)-undecylamide standard

SIST EN 14881:2006

The N-(3-(dimethylaminopropyl)-undecylamide standard purity is determined as follows:

$$w = \left(100 - \left(\frac{186 \times AV}{56, 1 \times 10}\right)\right) \times \frac{r}{100}$$

d4a68a19a832/sist-en-14881-2006

(A.1)

where

- *w* is the concentration of the N-(3-dimethylaminopropyl)-undecylamide (internal standard) in grams per 100 grams;
- AV is the acid value of the N-(3-dimethylaminopropyl)-undecylamide, in milligrams KOH per gram;
- *r* is the relative percentage of alkyl chain C 11 obtained by gas chromatograph.
- 186 is the molecular weight of the undecanoic acid.

NOTE 1 The acid value (AV) corresponds to the determination of free undecanoic acid by potentiometric titration with a potassium hydroxide standard volumetric solution c (KOH) = 0,1 mol/l in ethanol (see EN ISO 660).

NOTE 2 The alkyl chain purity (*r*) is determined by analysing an ethanolic solution of 5 % (m/V) by gas chromatography following the same conditions specified in 7.2 and calculated by areas normalisation in percentage.