



SLOVENSKI STANDARD

SIST EN 61788-13:2012

01-december-2012

Nadomešča:

SIST EN 61788-13:2003

Supraprevodnost - 13. del: Meritve izgub pri izmeničnem toku - Metode z magnetometrom za histerezne izgube v supraprevodnih večvlakenskih kompozitih

Superconductivity - Part 13: AC loss measurements - Magnetometer methods for hysteresis loss in superconducting multifilamentary composites

/

iTeh STANDARD PREVIEW
(standards.iteh.ai)

Supraconductivité - Partie 13: Mesure des pertes en courant alternatif - Méthodes de mesure par magnétomètre des pertes par hystérésis dans les composites multifilamentaires supraconducteurs

<http://standards.iteh.ai/catalog/standards/sist/8ce167ff-fe04-4e06-b01c-f82fb024ccaf/sist-en-61788-13-2012>

Ta slovenski standard je istoveten z: EN 61788-13:2012

ICS:

17.220.20	Merjenje električnih in magnetnih veličin	Measurement of electrical and magnetic quantities
29.050	Superprevodnost in prevodni materiali	Superconductivity and conducting materials

SIST EN 61788-13:2012

en

iTeh STANDARD PREVIEW
(standards.iteh.ai)

SIST EN 61788-13:2012

<https://standards.iteh.ai/catalog/standards/sist/8ce167ff-fe04-4e06-b01c-f82fb024ccaf/sist-en-61788-13-2012>

EUROPEAN STANDARD
NORME EUROPÉENNE
EUROPÄISCHE NORM

EN 61788-13

October 2012

ICS 17.220; 29.050

Supersedes EN 61788-13:2003

English version

**Superconductivity -
Part 13: AC loss measurements -
Magnetometer methods for hysteresis loss in superconducting
multifilamentary composites
(IEC 61788-13:2012)**

Supraconductivité -
Partie 13: Mesure des pertes
en courant alternatif -
Méthodes de mesure par magnétomètre
des pertes par hystérésis dans les
composites multifilamentaires
supraconducteurs
(CEI 61788-13:2012)

Supraleitfähigkeit -
Teil 13: Messung der
Wechselstromverluste -
Magnetometerverfahren zur Messung
der Hystereseverluste von supraleitenden
Multifilament-Verbundleitern
(IEC 61788-13:2012)

ITEH STANDARD PREVIEW
(standards.iteh.ai)

[SIST EN 61788-13:2012](https://standards.iteh.ai/catalog/standards/sist/8ce167ff-fe04-4e06-b01c-17220-29050)

<https://standards.iteh.ai/catalog/standards/sist/8ce167ff-fe04-4e06-b01c-17220-29050>

This European Standard was approved by CENELEC on 2012-08-29. CENELEC members are bound to comply with the CEN/CENELEC Internal Regulations which stipulate the conditions for giving this European Standard the status of a national standard without any alteration.

Up-to-date lists and bibliographical references concerning such national standards may be obtained on application to the CEN-CENELEC Management Centre or to any CENELEC member.

This European Standard exists in three official versions (English, French, German). A version in any other language made by translation under the responsibility of a CENELEC member into its own language and notified to the CEN-CENELEC Management Centre has the same status as the official versions.

CENELEC members are the national electrotechnical committees of Austria, Belgium, Bulgaria, Croatia, Cyprus, the Czech Republic, Denmark, Estonia, Finland, Former Yugoslav Republic of Macedonia, France, Germany, Greece, Hungary, Iceland, Ireland, Italy, Latvia, Lithuania, Luxembourg, Malta, the Netherlands, Norway, Poland, Portugal, Romania, Slovakia, Slovenia, Spain, Sweden, Switzerland, Turkey and the United Kingdom.

CENELEC

European Committee for Electrotechnical Standardization
Comité Européen de Normalisation Electrotechnique
Europäisches Komitee für Elektrotechnische Normung

Management Centre: Avenue Marnix 17, B - 1000 Brussels

Foreword

The text of document 90/302/FDIS, future edition 2 of IEC 61788-13, prepared by IEC/TC 90 "Superconductivity" was submitted to the IEC-CENELEC parallel vote and approved by CENELEC as EN 61788-13:2012.

The following dates are fixed:

- latest date by which the document has to be implemented at national level by publication of an identical national standard or by endorsement (dop) 2013-05-29
- latest date by which the national standards conflicting with the document have to be withdrawn (dow) 2015-08-29

This document supersedes EN 61788-13:2003.

EN 61788-13:2012 includes the following significant technical changes with respect to EN 61788-13:2003:

- to extend to the measurement of superconductors in general, in various sample sizes and shapes, and at temperatures other than 4,2 K;
- to use the word "uncertainty" for all quantitative (associated with a number) statistical expressions and eliminate the quantitative use of "precision" and "accuracy" in accordance with the decision at the June 2006 IEC/TC 90 meeting in Kyoto.

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. CENELEC [and/or CEN] shall not be held responsible for identifying any or all such patent rights.

[SIST EN 61788-13:2012](https://standards.iteh.ai/catalog/standards/sist/8ce167ff-fe04-4e06-b01c-f82fb024ccaf/sist-en-61788-13-2012)

[https://standards.iteh.ai/catalog/standards/sist/8ce167ff-fe04-4e06-b01c-](https://standards.iteh.ai/catalog/standards/sist/8ce167ff-fe04-4e06-b01c-f82fb024ccaf/sist-en-61788-13-2012)

[f82fb024ccaf/sist-en-61788-13-2012](https://standards.iteh.ai/catalog/standards/sist/8ce167ff-fe04-4e06-b01c-f82fb024ccaf/sist-en-61788-13-2012)

Endorsement notice

The text of the International Standard IEC 61788-13:2012 was approved by CENELEC as a European Standard without any modification.

Annex ZA (normative)

Normative references to international publications with their corresponding European publications

The following documents, in whole or in part, are normatively referenced in this document and are indispensable for its application. For dated references, only the edition cited applies. For undated references, the latest edition of the referenced document (including any amendments) applies.

NOTE When an international publication has been modified by common modifications, indicated by (mod), the relevant EN/HD applies.

<u>Publication</u>	<u>Year</u>	<u>Title</u>	<u>EN/HD</u>	<u>Year</u>
IEC 60050	Series	International Electrotechnical Vocabulary (IEV)	-	-
IEC 61788-5	-	Superconductivity - Part 5: Matrix to superconductor volume ratio measurement - Copper to superconductor volume ratio of Cu/Nb-Ti composite superconductors	EN 61788-5	-

iTeh STANDARD PREVIEW (standards.iteh.ai)

[SIST EN 61788-13:2012](#)

<https://standards.iteh.ai/catalog/standards/sist/8ce167ff-fe04-4e06-b01c-f82fb024ccaf/sist-en-61788-13-2012>

iTeh STANDARD PREVIEW
(standards.iteh.ai)

SIST EN 61788-13:2012

<https://standards.iteh.ai/catalog/standards/sist/8ce167ff-fe04-4e06-b01c-f82fb024ccaf/sist-en-61788-13-2012>



IEC 61788-13

Edition 2.0 2012-07

INTERNATIONAL STANDARD

NORME INTERNATIONALE

Superconductivity – Part 13: AC loss measurements – Magnetometer methods for hysteresis loss in superconducting multifilamentary composites

Supraconductivité – Partie 13: Mesure des pertes en courant alternatif – Méthodes de mesure par magnétomètre des pertes par hystérésis dans les composites multifilamentaires supraconducteurs

INTERNATIONAL
ELECTROTECHNICAL
COMMISSION

COMMISSION
ELECTROTECHNIQUE
INTERNATIONALE

PRICE CODE
CODE PRIX

T

ICS 17.220, 29.050

ISBN 978-2-83220-292-0

**Warning! Make sure that you obtained this publication from an authorized distributor.
Attention! Veuillez vous assurer que vous avez obtenu cette publication via un distributeur agréé.**

CONTENTS

FOREWORD.....	3
INTRODUCTION.....	5
1 Scope.....	6
2 Normative references.....	6
3 Terms and definitions.....	6
4 General specifications.....	8
4.1 Target uncertainty.....	8
4.2 Uncertainty and uniformity of the applied field.....	8
4.3 VSM calibration.....	8
4.4 Temperature.....	9
4.5 Specimen length.....	9
4.6 Specimen orientation and demagnetization effects.....	9
4.7 Normalization volume.....	9
4.8 Mode of field cycling or sweeping.....	9
5 The VSM method of measurement.....	10
5.1 General.....	10
5.2 VSM measurement principle.....	10
5.3 VSM specimen preparation.....	10
5.4 VSM measurement conditions and calibration.....	12
5.4.1 Field amplitude.....	12
5.4.2 Direction of applied field.....	12
5.4.3 Rate of change of the applied field (sweep rate).....	12
5.4.4 Waveform of the field change.....	12
5.4.5 Specimen size and shape correction.....	12
5.4.6 Allowance for addendum (background subtraction).....	13
5.4.7 Data point density.....	13
6 Test report.....	13
6.1 General.....	13
6.2 Initiation of the test.....	13
6.3 Technical details.....	13
Annex A (informative) The SQUID method of measurement.....	15
Annex B (normative) Extension of the standard to the measurement of superconductors in general.....	16
Annex C (informative) Uncertainty considerations.....	18
Bibliography.....	23
Figure 1 – A typical experimental setup of VSM measurement.....	11
Figure 2 – Three alternative specimen configurations for the VSM measurement.....	11
Table C.1 – Output signals from two nominally identical extensometers.....	19
Table C.2 – Mean values of two output signals.....	19
Table C.3 – Experimental standard deviations of two output signals.....	19
Table C.4 – Standard uncertainties of two output signals.....	20
Table C.5 – Coefficient of variations of two output signals.....	20

INTERNATIONAL ELECTROTECHNICAL COMMISSION

SUPERCONDUCTIVITY –

**Part 13: AC loss measurements –
Magnetometer methods for hysteresis loss
in superconducting multifilamentary composites**

FOREWORD

- 1) The International Electrotechnical Commission (IEC) is a worldwide organization for standardization comprising all national electrotechnical committees (IEC National Committees). The object of IEC is to promote international co-operation on all questions concerning standardization in the electrical and electronic fields. To this end and in addition to other activities, IEC publishes International Standards, Technical Specifications, Technical Reports, Publicly Available Specifications (PAS) and Guides (hereafter referred to as "IEC Publication(s)"). Their preparation is entrusted to technical committees; any IEC National Committee interested in the subject dealt with may participate in this preparatory work. International, governmental and non-governmental organizations liaising with the IEC also participate in this preparation. IEC collaborates closely with the International Organization for Standardization (ISO) in accordance with conditions determined by agreement between the two organizations.
- 2) The formal decisions or agreements of IEC on technical matters express, as nearly as possible, an international consensus of opinion on the relevant subjects since each technical committee has representation from all interested IEC National Committees.
- 3) IEC Publications have the form of recommendations for international use and are accepted by IEC National Committees in that sense. While all reasonable efforts are made to ensure that the technical content of IEC Publications is accurate, IEC cannot be held responsible for the way in which they are used or for any misinterpretation by any end user.
- 4) In order to promote international uniformity, IEC National Committees undertake to apply IEC Publications transparently to the maximum extent possible in their national and regional publications. Any divergence between any IEC Publication and the corresponding national or regional publication shall be clearly indicated in the latter.
- 5) IEC itself does not provide any attestation of conformity. Independent certification bodies provide conformity assessment services and, in some areas, access to IEC marks of conformity. IEC is not responsible for any services carried out by independent certification bodies.
- 6) All users should ensure that they have the latest edition of this publication.
- 7) No liability shall attach to IEC or its directors, employees, servants or agents including individual experts and members of its technical committees and IEC National Committees for any personal injury, property damage or other damage of any nature whatsoever, whether direct or indirect, or for costs (including legal fees) and expenses arising out of the publication, use of, or reliance upon, this IEC Publication or any other IEC Publications.
- 8) Attention is drawn to the Normative references cited in this publication. Use of the referenced publications is indispensable for the correct application of this publication.
- 9) Attention is drawn to the possibility that some of the elements of this IEC Publication may be the subject of patent rights. IEC shall not be held responsible for identifying any or all such patent rights.

International Standard IEC 61788-13 has been prepared by IEC technical committee 90: Superconductivity.

This second edition cancels and replaces the first edition published in 2003. It constitutes a technical revision.

Modifications made to the second edition are

- to extend to the measurement of superconductors in general, in various sample sizes and shapes, and at temperatures other than 4,2 K,
- to use the word "uncertainty" for all quantitative (associated with a number) statistical expressions and eliminate the quantitative use of "precision" and "accuracy" in accordance with the decision at the June 2006 IEC/TC90 meeting in Kyoto.

The text of this standard is based on the following documents:

FDIS	Report on voting
90/302/FDIS	90/306/RVD

Full information on the voting for the approval of this standard can be found in the report on voting indicated in the above table.

This publication has been drafted in accordance with the ISO/IEC Directives, Part 2.

A list of all parts of the IEC 61788 series, under the general title: *Superconductivity*, can be found on the IEC website.

The committee has decided that the contents of this publication will remain unchanged until the stability date indicated on the IEC web site under "<http://webstore.iec.ch>" in the data related to the specific publication. At this date, the publication will be

- reconfirmed,
- withdrawn,
- replaced by a revised edition, or
- amended.

iTeh STANDARD PREVIEW (standards.iteh.ai)

[SIST EN 61788-13:2012](#)

<https://standards.iteh.ai/catalog/standards/sist/8ce167ff-fe04-4e06-b01c-f82fb024ccaf/sist-en-61788-13-2012>

INTRODUCTION

IEC Technical Committee 90 proposes magnetometer and pickup coil methods for measuring the AC losses of Cu/Nb-Ti composite superconducting wires in transverse time-varying magnetic fields. These represent initial steps in standardization of methods for measuring the various contributions to AC loss in transverse fields, the most frequently encountered configuration.

It was decided to split the initial proposal mentioned above into two documents covering two standard methods. One of them describes the magnetometer method for hysteresis loss and low frequency (or sweep rate) total AC loss measurement in a slowly varying magnetic field, and the other describes the pickup coil method for total AC loss measurement in higher frequency (or sweep rate) magnetic fields. The frequency range is 0 Hz – 0,06 Hz for the magnetometer method and 0,005 Hz – 60 Hz for the pickup-coil method. The overlap between 0,005 Hz and 0,06 Hz is a complementary frequency range for the two methods.

This standard deals with the magnetometer method.

iTeh STANDARD PREVIEW (standards.iteh.ai)

[SIST EN 61788-13:2012](https://standards.iteh.ai/catalog/standards/sist/8ce167ff-fe04-4e06-b01c-f82fb024ccaf/sist-en-61788-13-2012)

<https://standards.iteh.ai/catalog/standards/sist/8ce167ff-fe04-4e06-b01c-f82fb024ccaf/sist-en-61788-13-2012>