
Medical electrical equipment —
Part 2-85:
Particular requirements for the basic
safety and essential performance of
cerebral tissue oximeter equipment

iTeh STANDARD PREVIEW
(standards.iteh.ai)

Appareils électromédicaux —

Partie 2-85: Exigences particulières pour la sécurité de base et les performances essentielles des oxymètres pour tissu cérébral

[ISO 80601-2-85:2021](https://standards.iteh.ai/catalog/standards/sist/6e8d079e-a50e-4c3d-b425-bad94244f514/iso-80601-2-85-2021)

<https://standards.iteh.ai/catalog/standards/sist/6e8d079e-a50e-4c3d-b425-bad94244f514/iso-80601-2-85-2021>



iTeh STANDARD PREVIEW
(standards.iteh.ai)

[ISO 80601-2-85:2021](https://standards.iteh.ai/catalog/standards/sist/6e8d079e-a50e-4c3d-b425-bad94244f514/iso-80601-2-85-2021)
<https://standards.iteh.ai/catalog/standards/sist/6e8d079e-a50e-4c3d-b425-bad94244f514/iso-80601-2-85-2021>



COPYRIGHT PROTECTED DOCUMENT

© ISO 2021

All rights reserved. Unless otherwise specified, or required in the context of its implementation, no part of this publication may be reproduced or utilized otherwise in any form or by any means, electronic or mechanical, including photocopying, or posting on the internet or an intranet, without prior written permission. Permission can be requested from either ISO at the address below or ISO's member body in the country of the requester.

ISO copyright office
CP 401 • Ch. de Blandonnet 8
CH-1214 Vernier, Geneva
Phone: +41 22 749 01 11
Email: copyright@iso.org
Website: www.iso.org

Published in Switzerland

Contents

Page

Foreword.....	vi
Introduction.....	vii
201.1 Scope, object and related standards.....	1
201.1.1 * Scope.....	1
201.1.2 Object.....	2
201.1.3 Collateral standards.....	2
201.1.4 Particular standards.....	3
201.2 Normative references.....	4
201.3 Terms and definitions.....	5
201.4 General requirements.....	10
201.4.3.101 * Additional requirements for <i>essential performance</i>	11
201.4.102 Additional requirements for acceptance criteria.....	11
201.4.103 Additional requirements for <i>cerebral tissue oximeter equipment, parts and accessories</i>	11
201.5 General requirements for testing of <i>ME equipment</i>	12
201.6 Classification of <i>ME equipment</i> and <i>ME systems</i>	12
201.7 <i>ME equipment</i> identification, marking and documents.....	12
201.7.1.101 <i>Information to be supplied by the manufacturer</i>	12
201.7.2.3 Consult <i>accompanying documents</i>	12
201.7.2.9.101 IP classification.....	12
201.7.2.101 Additional requirements for marking on the outside of <i>ME equipment</i> parts.....	13
201.7.4.3 Units of measurement.....	13
201.7.9.2 Instructions for use.....	13
201.7.9.2.1.101 Additional general requirements.....	13
201.7.9.2.2.101 Additional requirements for warnings and safety notices.....	15
201.7.9.2.9.101 Additional requirements for operating instructions.....	15
201.7.9.2.14.101 Additional requirements for <i>accessories, supplementary equipment, used material</i>	15
201.7.9.3.1.101 * Additional general requirements.....	16
201.8 Protection against electrical <i>hazards</i> from <i>ME equipment</i>	16
201.8.3.101 Additional requirements for classification of <i>applied parts</i>	16
201.8.5.5.1.101 Defibrillation protection.....	16
201.8.7.4.7.101 Additional requirements for measurement of the <i>patient leakage current</i>	16
201.9 Protection against mechanical <i>hazards</i> of <i>ME equipment</i> and <i>ME systems</i>	17
201.10 Protection against unwanted and excessive radiation <i>hazards</i>	17
201.10.4 Lasers.....	17
201.11 Protection against excessive temperatures and other <i>hazards</i>	17
201.11.1.2.2 <i>Applied parts</i> not intended to supply heat to a <i>patient</i>	17
201.11.6.5.101 * Additional requirements for ingress of water or particulate matter into the <i>ME equipment</i> or <i>ME system</i>	18
201.11.6.7 <i>Sterilization</i> of <i>ME equipment</i> or <i>ME system</i>	18
201.11.8.101 Additional requirements for interruption of the power supply/ <i>supply mains</i> to <i>ME equipment</i>	18

201.11.8.101.1	<i>Technical alarm condition for power supply failure</i>	18
201.11.8.101.2	Settings and data storage following short interruptions or automatic switchover	19
201.11.8.101.3	Operation following long interruptions	19
201.12	Accuracy of controls and instruments and protection against hazardous outputs.....	19
201.12.1.101	* <i>StO₂ accuracy of cerebral tissue oximeter equipment</i>	19
201.12.1.101.1	* Specification	19
201.12.1.101.2	* Data collection for determination of <i>StO₂ accuracy</i>	21
201.12.1.101.3	* Data analysis for determination of <i>StO₂ accuracy</i>	22
201.12.1.101.4	Characteristics of the study used for determination of <i>StO₂ accuracy</i>	23
201.12.4	Protection against hazardous output	23
201.12.4.101	* <i>Data update period</i>	23
201.12.4.102	* <i>Signal inadequacy</i>	23
201.13	<i>Hazardous situations and fault conditions for ME equipment</i>	24
201.13.101	Detection of <i>probe faults and probe cable extender faults</i>	24
201.14	<i>Programmable electrical medical systems (PEMS)</i>	24
201.15	Construction of <i>ME equipment</i>	24
201.15.3.5.101	* Additional requirements for rough handling.....	25
201.15.3.5.101.1	* Shock and vibration (robustness).....	25
201.15.3.5.101.2	* Shock and vibration for a <i>transit-operable cerebral tissue oximeter</i> during operation	26
201.15.101	Mode of operation	27
201.16	<i>ME systems</i>	27
201.17	Electromagnetic compatibility of <i>ME equipment and ME systems</i>	27
201.101	* <i>Cerebral tissue oximeter probes and probe cable extenders</i>	27
201.101.1	General	27
201.101.2	Labelling	28
201.102	<i>Functional connection</i>	28
201.102.1	General	28
201.102.2	* Connection to an electronic health record or <i>integrated clinical environment</i>	28
201.102.3	Connection to a <i>distributed alarm system</i>	28
202	Electromagnetic disturbances — Requirements and tests.....	29
202.4.3.1	Configurations.....	29
202.5.2.2.1	Requirements applicable to all <i>ME equipment and ME systems</i>	29
202.8.1.101	Additional general requirements	29
202.8.2	<i>Patient physiological simulation</i>	29
206	Usability.....	30
208	General requirements, tests and guidance for alarm systems in medical electrical equipment and medical electrical systems.....	30
208.6.1.2.101	* Additional requirements for <i>alarm condition priority</i>	30
208.6.5.4.101	* Additional requirements for <i>default alarm preset</i>	31
208.6.8.5.101	Additional requirements for <i>alarm signal inactivation states, indication and</i> access.....	31
211	Requirements for medical electrical equipment and medical electrical systems used in the home healthcare environment	31
212	Requirements for medical electrical equipment and medical electrical systems used in the emergency medical services environment.....	31

Annex C (informative) Guide to marking and labelling requirements for <i>ME</i> equipment and <i>ME</i> systems	32
Annex D (informative) Symbols on marking	36
Annex AA (informative) Particular guidance and rationale	37
Annex BB (informative) Skin temperature at the <i>cerebral tissue oximeter probe</i>	48
Annex CC (informative) Determination of <i>accuracy</i>	50
Annex DD (informative) Characteristics of a <i>tissue haemoglobin phantom</i> for the verification of the <i>accuracy</i> of <i>cerebral tissue oximeter equipment</i>	56
Annex EE (informative) Guideline for evaluating and documenting <i>StO₂ accuracy</i> in human subjects	66
Annex FF (informative) <i>Functional testers for cerebral tissue oximeter equipment</i>	72
Annex GG (informative) Concepts of <i>ME equipment</i> response time	75
Annex HH (normative) Data interface requirements	80
Annex II (informative) Comparison of methods of performance evaluation	84
Annex JJ (informative) Reference to the <i>IMDRF essential principles</i> and labelling guidances	89
Annex KK (informative) Reference to the <i>essential principles</i>	92
Annex LL (informative) Reference to the general safety and performance requirements	95
Annex MM (informative) Terminology — alphabetized index of defined terms	98
Bibliography	102

ISO 80601-2-85:2021

<https://standards.iteh.ai/catalog/standards/sist/6e8d079e-a50e-4c3d-b425-bad94244f514/iso-80601-2-85-2021>

Foreword

ISO (the International Organization for Standardization) is a worldwide federation of national standards bodies (ISO member bodies). The work of preparing International Standards is normally carried out through ISO technical committees. Each member body interested in a subject for which a technical committee has been established has the right to be represented on that committee. International organizations, governmental and non-governmental, in liaison with ISO, also take part in the work. ISO collaborates closely with the International Electrotechnical Commission (IEC) on all matters of electrotechnical standardization.

The procedures used to develop this document and those intended for its further maintenance are described in the ISO/IEC Directives, Part 1. In particular, the different approval criteria needed for the different types of ISO documents should be noted. This document was drafted in accordance with the editorial rules of the ISO/IEC Directives, Part 2 (see www.iso.org/directives).

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. ISO shall not be held responsible for identifying any or all such patent rights. Details of any patent rights identified during the development of the document will be in the Introduction and/or on the ISO list of patent declarations received (see www.iso.org/patents).

Any trade name used in this document is information given for the convenience of users and does not constitute an endorsement.

iTeh STANDARD PREVIEW

For an explanation of the voluntary nature of standards, the meaning of ISO specific terms and expressions related to conformity assessment, as well as information about ISO's adherence to the World Trade Organization (WTO) principles in the Technical Barriers to Trade (TBT), see www.iso.org/iso/foreword.html.

ISO 80601-2-85:2021
https://standards.iteh.ai/catalog/standards/sist/6e8d079e-a50e-4c3d-b425-bad94244f514/iso-80601-2-85-2021

This document was prepared jointly by Technical Committee ISO/TC 121, *Anaesthetic and respiratory equipment*, Subcommittee SC 3, *Respiratory devices and related equipment used for patient care*, and Technical Committee IEC/TC 62, *Electrical equipment in medical practice*, Subcommittee 62D, *Electromedical equipment*, in collaboration with the European Committee for Standardization (CEN) Technical Committee CEN/TC 215, *Respiratory and anaesthetic equipment*, in accordance with the Agreement on technical cooperation between ISO and CEN (Vienna Agreement).

A list of all parts in the ISO and IEC 80601 series can be found on the ISO and IEC websites.

Any feedback or questions on this document should be directed to the user's national standards body. A complete listing of these bodies can be found at www.iso.org/members.html.

Introduction

The estimation of blood oxygen saturation in the brain tissue by *cerebral tissue oximetry equipment* is increasingly used in many areas of medicine. This document covers *basic safety* and *essential performance* requirements achievable within the limits of existing technology.

Annex AA contains a rationale for some of the requirements. It is included to provide additional insight into the reasoning of the committees that led to a requirement and into the *hazards* that the requirement addresses.

Annex BB is a literature review and provides recommendations relevant to determining the maximum safe temperature of the interface between a *cerebral tissue oximeter probe* and a *patient's* tissue.

Annex CC discusses both the formulae used to evaluate the *StO₂ accuracy* of *cerebral tissue oximeter equipment* measurements, and the names that are assigned to those formulae.

Annex DD presents guidance on using in-vitro methods (phantoms) for *verification* of *StO₂ accuracy* of *cerebral tissue oximeter equipment*.

Annex EE presents a guideline for an in-vivo (human subjects) *controlled desaturation study* for the *verification* of *StO₂ accuracy* of *cerebral tissue oximeter equipment*.

Annex FF is a description of *functional testers* for use with *cerebral tissue oximeter equipment*.

Annex GG describes concepts of *cerebral tissue oximeter equipment* response time.

Annex HH describes data interface requirements.

Annex II is a comparison between human desaturations (in-vivo) and *tissue haemoglobin phantom* desaturations (in-vitro) for assessing *StO₂ accuracy*.

In this document, the following print types are used:

- requirements and definitions: roman type;
- *Instructions, test specifications and terms defined in Clause 3 of the general standard, in this document or as noted: italic type;*
- informative material appearing outside of tables, such as notes, examples and references: in smaller type; normative text of tables is also in a smaller type.

In referring to the structure of this document, the term

- “clause” means one of the numbered divisions within the table of contents, inclusive of all subdivisions (e.g. Clause 201.7 includes subclauses 201.7.1, 201.7.2) and
- “subclause” means a numbered subdivision of a clause (e.g. 201.7.1, 7.2 and 201.7.2.1 are all subclauses of Clause 201.7).

References to clauses within this document are preceded by the term “Clause” followed by the clause number. References to subclauses within this document are by number only.

ISO 80601-2-85:2021(E)

In this document, the conjunctive “or” is used as an “inclusive or” so a statement is true if any combination of the conditions is true.

For the purposes of this document, the auxiliary verb:

- “shall” means that conformance with a requirement or a test is mandatory for conformance with this document;
- “should” means that conformance with a requirement or a test is recommended but is not mandatory for conformance with this document; and
- “may” is used to describe permission (e.g. a permissible way to achieve conformance with a requirement or test);
- “can” is used to describe a possibility or capability; and
- “must” is used to express an external constraint.

Annex C contains a guide to the marking and labelling requirements in this document.

Annex D contains a summary of the symbols referenced in this document.

An asterisk (*) as the first character of a title or at the beginning of a paragraph or table title indicates that there is guidance or rationale related to that item in AA.

ITIH STANDARD PREVIEW
(standards.iteh.ai)
ISO 80601-2-85:2021
<https://standards.iteh.ai/catalog/standards/sist/6e8d079e-a50e-4c3d-b425-bad94244f514/iso-80601-2-85-2021>

Medical electrical equipment —

Part 2-85:

Particular requirements for the basic safety and essential performance of cerebral tissue oximeter equipment

201.1 Scope, object and related standards

Clause 1 of the general standard applies, except as follows.

NOTE The general standard is IEC 60601-1:2005+AMD1:2012+AMD2:2020.

201.1.1 * Scope

Replacement:

This document applies to *basic safety and essential performance of cerebral tissue oximeter equipment*, that employs light at multiple wavelengths to derive a quantitative measure of oxygen saturation of haemoglobin within the volume of tissue sampled under the *probe* attached to the head. The *cerebral tissue oximeter equipment* can be based on continuous light, frequency domain or time domain technologies. This document applies to *ME equipment* used in a hospital environment as well as when used outside the hospital environment, such as in ambulances and air transport. Additional standards may apply to *ME equipment* for those environments of use.

NOTE 1 *Cerebral tissue oximeters* are sometimes referred to as near infrared spectroscopy equipment in medical literature.

Not included within the scope of this document are:

- invasive tissue or vascular oximeters;
- oximeters that require a blood sample from the *patient*;
- equipment measuring dissolved oxygen;
- *ME equipment*, or part thereof, that measures path-length-dependent haemoglobin change. The requirements for functional near-infrared spectroscopy equipment are found in ISO 80601-2-71^[4];
- *ME equipment*, or part thereof, that measures arterial saturation based on pulsatile changes in tissue optical properties (SpO_2). The requirements for pulse oximeter equipment are found in ISO 80601-2-61^[3];
- *ME equipment*, or any part thereof, that claims to monitor tissue in parts of the body other than the head.

This document also applies to *cerebral tissue oximeter equipment*, including *cerebral tissue oximeter monitors*, *cerebral tissue oximeter probes* and *probe cable extenders*, that have been remanufactured.

ISO 80601-2-85:2021(E)

If a clause or subclause is specifically intended to be applicable to *ME equipment* only, or to *ME systems* only, the title and content of that clause or subclause will say so. If that is not the case, the clause or subclause applies both to *ME equipment* and to *ME systems*, as relevant.

Hazards inherent in the intended physiological function of *ME equipment* or *ME systems* within the scope of this document are not covered by specific requirements in this document except in 201.11 and in 201.7.2.13 and 201.8.4.1 of the general standard.

NOTE 2 See also 4.2 of the general standard.

This document can also be applied to *ME equipment* and their *accessories* used for compensation or alleviation of disease, injury or disability.

This document is not applicable to remote or slave (secondary) equipment that displays StO_2 values that are located outside of the *patient environment*.

NOTE 3 *ME equipment* that provides selection between diagnostic and monitoring functions is expected to meet the requirements of the appropriate document when configured for that function.

201.1.2 Object

Replacement:

The object of this document is to establish particular *basic safety* and *essential performance* requirements for *cerebral tissue oximeter equipment* [as defined in 201.3.202] and its *accessories*.

NOTE 1 *Accessories* are included because the combination of the *cerebral tissue oximeter monitor* and the *accessories* needs to be adequately safe. *Accessories* can have a significant impact on the *basic safety* or *essential performance* of *cerebral tissue oximeter equipment*.

NOTE 2 This document has been prepared to address the relevant International Medical Device Regulators Forum (IMDRF) *essential principles* and labelling guidances as indicated in Annex JJ.

NOTE 3 This document has been prepared to address the relevant *essential principles of safety and performance* of ISO 16142-1:2016 as indicated in Annex KK.

NOTE 4 This document has been prepared to address the relevant general safety and performance requirements of European regulation (EU) 2017/745^[20] as indicated in Annex LL.

201.1.3 Collateral standards

Addition:

This document refers to those applicable collateral standards that are listed in Clause 2 of the general standard and Clause 201.2 of this document.

IEC 60601-1-2:2014+AMD1:2020, IEC 60601-1-6:2010+AMD1:2013+AMD2:2020, IEC 60601-1-8:2006+AMD1:2012+AMD2:2020, IEC 60601-1-11:2015+AMD1:2020 and IEC 60601-1-12:2014+AMD1:2020 apply as modified in Clauses 202, 206, 208, 211 and 212 respectively. IEC 60601-1-3 does not apply. All other published collateral standards in the IEC 60601-1 series apply as published.

201.1.4 Particular standards

Replacement:

In the IEC 60601 series, particular standards define *basic safety* and *essential performance* requirements, and may modify, replace or delete requirements contained in the general standard, including the collateral standards, as appropriate for the particular *ME equipment* under consideration.

A requirement of a particular standard takes priority over the general standard or the collateral standards.

For brevity, IEC 60601-1:2005+AMD1:2012+AMD2:2020 is referred to in this document as the general standard. Collateral standards are referred to by their document number.

The numbering of clauses and subclauses of this document corresponds to those of the general standard with the prefix "201" (e.g. 201.1 in this document addresses the content of Clause 1 of the general standard) or applicable collateral standard with the prefix "2xx" where xx is the final digits of the collateral standard document number (e.g. 202.4 in this document addresses the content of Clause 4 of the IEC 60601-1-2 collateral standard, 208.4 in this document addresses the content of Clause 4 of the IEC 60601-1-8 collateral standard, etc.). The changes to the text of the general standard are specified by the use of the following words:

"Replacement" means that the clause or subclause of the general standard or applicable collateral standard is replaced completely by the text of this document.

"Addition" means that the text of this document is additional to the requirements of the general standard or applicable collateral standard.

"Amendment" means that the clause or subclause of the general standard or applicable collateral standard is amended as indicated by the text of this document.

Clauses, subclauses or figures that are additional to those of the general standard are numbered starting from 201.101. However, due to the fact that definitions in the general standard are numbered 3.1 through 3.147, additional definitions in this document are numbered beginning from 201.3.201. Additional annexes are lettered AA, BB, etc., and additional items aa), bb), etc.

Subclauses or figures that are additional to those of a collateral standard are numbered starting from 2xx, where "x" is the number of the collateral standard, e.g. 202 for IEC 60601-1-2, 203 for IEC 60601-1-3, etc.

The term "this document" is used to make reference to the general standard, any applicable collateral standards and this particular document taken together.

Where there is no corresponding clause or subclause in this particular document, the section, clause or subclause of the general standard or applicable collateral standard, although possibly not relevant, applies without modification; where it is intended that any part of the general standard or applicable collateral standard, although possibly relevant, is not to be applied, a statement to that effect is given in this particular document.

201.2 Normative references

The following documents are referred to in the text in such a way that some or all of their content constitutes requirements of this document. For dated references, only the edition cited applies. For undated references, the latest edition of the referenced document (including any amendments) applies.

Clause 2 of the general standard applies, except as follows:

Replacement:

ISO 15223-1:—¹, *Medical devices — Symbols to be used with medical device labels, labelling and information to be supplied — Part 1: General requirements*

Addition:

ISO 14155:2020, *Clinical investigation of medical devices for human subjects — Good clinical practice*

ISO 16142-1:2016, *Medical devices — Recognized essential principles of safety and performance of medical devices — Part 1: General essential principles and additional specific essential principles for all non-IVD medical devices and guidance on the selection of standards*

ISO 17664:2017, *Processing of health care products — Information to be provided by the medical device manufacturer for the processing of medical devices*

ISO 20417:2020, *Medical devices — Information to be supplied by the manufacturer*

IEC 60068-2-31:2008, *Environmental testing — Part 2-31: Tests — Test Ec: Rough handling shocks, primarily for equipment-type specimens*

IEC 60068-2-64:2008+AMD1:2019, *Environmental testing — Part 2-64: Tests — Test Fh: Vibration, broadband random and guidance*

IEC 60601-1:2005+AMD1:2012+AMD2:2020, *Medical electrical equipment — Part 1: General requirements for basic safety and essential performance*

IEC 60601-1-11:2015+AMD1:2020, *Medical electrical equipment — Part 1-11: General requirements for basic safety and essential performance — Collateral Standard: Requirements for medical electrical equipment and medical electrical systems used in the home healthcare environment*

IEC 60601-1-12:2014+AMD1:2020, *Medical electrical equipment — Part 1-12: General requirements for basic safety and essential performance — Collateral Standard: Requirements for medical electrical equipment and medical electrical systems intended for use in the emergency medical services environment*

ISO 80601-2-61:2017, *Medical electrical equipment — Part 2-61: Particular requirements for basic safety and essential performance of pulse oximeter equipment*

IEC 62471:2006, *Photobiological safety of lamps and lamp systems*

¹ Under preparation. Stage at the time of publication: ISO/DIS 15223-1:2020.

AAMI 2700-1:2019², *Medical devices and medical systems — Essential safety requirements for equipment comprising the patient-centric integrated clinical environment (ICE) — Part 1: General requirements and conceptual model*

201.3 Terms and definitions

For the purposes of this document, the terms and definitions given in ISO 16142-1:2016, ISO 17664:2017, ISO 20417:2020, IEC 60601-1:2005+AMD1:2012+AMD2:2020, IEC 60601-1-2:2014+AMD1:2020, IEC 60601-1-6:2010+AMD1:2013+AMD2:2020, IEC 60601-1-8:2006+AMD1:2012+AMD2:2020, IEC 60601-1-11:2015+AMD1:2020, IEC 60601-1-12:2014+AMD2:2020, ISO 80601-2-61:2017, AAMI 2700-1:2019 and the following apply.

ISO and IEC maintain terminological databases for use in standardization at the following addresses:

- ISO Online browsing platform: available at <https://www.iso.org/obp>
- IEC Electropedia: available at <http://www.electropedia.org/>

NOTE An alphabetized index of defined terms is found in Annex MM.

201.3.201

accuracy

A_{rms}

closeness of agreement between a test result and the true value

Note 1 to entry: 201.12.1.101.2 contains methods for estimating the StO_2 accuracy of cerebral tissue oximeter equipment.

Note 2 to entry: Additional information is found in Annexes CC, DD, EE and II.

Note 3 to entry: In this document, accuracy (A_{rms}) is stated in terms of the root mean square difference. See 201.12.1.101.3.

[SOURCE: ISO 3534-2:2006^[6] 3.3.1, modified — Notes to entry replaced.]

201.3.202

cerebral tissue oximeter

cerebral tissue oximeter equipment

ME equipment for the non-invasive estimation of *functional oxygen saturation* of haemoglobin in cerebral tissue below the *probe* (StO_2 or rSO_2), based on light interacting with tissue

Note 1 to entry: *Cerebral tissue oximeter equipment* comprises a *cerebral tissue oximeter monitor*, a *probe cable extender*, if provided, and a *cerebral tissue oximeter probe*, which can be combined in a single assembly.

Note 2 to entry: Light is more technically referred to as electromagnetic radiation (optical radiation). This document uses the common term.

Note 3 to entry: Measurements are based upon light interacting with all tissue under the *probe* to determine StO_2 .

201.3.203

cerebral tissue oximeter monitor

monitor

part of the *cerebral tissue oximeter equipment* that encompasses the measurement electronics, display and *operator interface*, excluding the *cerebral tissue oximeter probe* and *probe cable extender*

² Formerly ASTM F2761-09.

ISO 80601-2-85:2021(E)

Note 1 to entry: A *cerebral tissue oximeter monitor* can consist of multiple pieces of hardware in separate locations, e.g. a telemetry system in which the *applied part* and primary display are physically separated.

201.3.204 cerebral tissue oximeter probe probe

part of the *cerebral tissue oximeter equipment* that includes the *applied part*

Note 1 to entry: The terms sensor and transducer have also been used for *cerebral tissue oximeter probe*.

Note 2 to entry: A reflectance *probe* design is the typical configuration.

201.3.205 controlled desaturation study

hypoxaemia induced in a group of human subjects performed under laboratory conditions

Note 1 to entry: This can also be referred to as a controlled hypoxaemia study. Additional information is found in Annex EE.

[SOURCE: ISO 80601-2-61:2017^[3], 201.3.202]

201.3.206 CO-oximeter

multiwavelength, optical analyser that measures *total haemoglobin concentration* and the concentrations of various haemoglobin derivatives via direct measurement of a blood specimen

Note 1 to entry: The relevant CO-oximetry values for this document are *functional oxygen saturation* of arterial and jugular venous blood, SaO_2 and $SjvO_2$. (standards.iteh.ai)

Note 2 to entry: This excludes indirect measurements that are performed by pulse oximetry and other optical measurements methods on human tissue. ISO 80601-2-85:2021

[SOURCE: CLSI C46-A2: 2009, 4.1.9^[7]]
<https://standards.iteh.ai/catalog/standards/sist/6e8d079e-a50e-4c3d-b425-bad94244f514/iso-80601-2-85-2021>

201.3.207 data update period

interval in which the *cerebral tissue oximeter equipment* algorithm provides new valid data to the display or the *functional connection*

Note 1 to entry: This definition does not refer to the regular refresh period of the display, which is typically on the order of 1 s, but rather to the (typically longer) interval defined above.

[SOURCE: ISO 80601-2-61:2017^[3], 201.3.204, modified.]

201.3.208 declared range

range of the *reference haemoglobin oxygen saturation* (S_R) over which there is specified *accuracy* performance

[SOURCE: ISO 80601-2-61:2017^[3], 201.3.202]

201.3.209 displayed range

range of StO_2 values that can be displayed by the *cerebral tissue oximeter equipment*

Note 1 to entry: The *displayed range* can extend beyond the *declared range*.

[SOURCE: ISO 80601-2-61:2017^[3], 201.3.206, modified.]

201.3.210**functional oxygen saturation**

saturation given by the oxyhaemoglobin concentration (cO_2Hb) divided by the sum of the oxyhaemoglobin concentration and the deoxyhaemoglobin concentration ($cHHb$)

$$\frac{cO_2Hb}{cO_2Hb + cHHb}$$

Note 1 to entry: The CLSI^[7] term for this ratio is haemoglobin oxygen saturation, and its notation is SO_2 .

[SOURCE: CLSI C46-A2: 2009, 4.1.10.2^[7]]

201.3.211**functional tester**

test equipment that presents the *cerebral tissue oximeter equipment* with a tissue analogue representing a predictable value of tissue haemoglobin saturation so that the *operator* can observe the resulting displayed value of StO_2 and compare it to the expected value.

Note 1 to entry: Additional information is found in Annex FF.

Note 2 to entry: Not all *functional testers* and *cerebral tissue oximeter equipment* are compatible.

201.3.212**local bias**

b

difference between the expectation of the results (StO_2) and a reference value

Note 1 to entry: For human desaturation studies, the reference is $SavO_2$.

Note 2 to entry: For phantom desaturation studies, the reference values are obtained from a reference measurement.

Note 3 to entry: For *cerebral tissue oximeter equipment*, this is, at a given value of the reference oxygen saturation, the difference between the y -value of the regression line at that coordinate and the y -value of the line of identity, in a plot of StO_2 versus S_R , or given by:

$$b_i = StO_{2\text{fit},i} - S_{R,i}$$

where $StO_{2\text{fit},i}$ is the value of the curve fitted to the test data at the i^{th} reference oxygen saturation value, $S_{R,i}$.

Note 4 to entry: Additional information is found with the term *mean bias* and in the discussion in Annex CC.

[SOURCE: ISO 80601-2-61: 2017^[3], 201.3.211, modified.]

201.3.213**mean bias**

B

mean difference between the test and reference values, preserving sign

$$B = \frac{\sum_{i=1}^n (StO_{2i} - S_{Ri})}{n}$$

where

n is the number of data pairs in the sample;

StO_{2i} is the i^{th} StO_2 datum; and