

# INTERNATIONAL STANDARD

# IEC 60601-2-12

[ISO 10651-1]

Second edition  
2001-10

---

---

**Medical electrical equipment –**

**Part 2-12:**

**Particular requirements for the safety of lung  
ventilators – Critical care ventilators**

**iTeh STANDARD PREVIEW**

*Appareils électromédicaux –*

*Partie 2-12: IEC 60601-2-12:2001*

*Règles particulières de sécurité pour ventilateurs*

*pulmonaires – Ventilateurs pour utilisation en soins intensifs*



Reference number  
IEC 60601-2-12:2001(E)

## Publication numbering

As from 1 January 1997 all IEC publications are issued with a designation in the 60000 series. For example, IEC 34-1 is now referred to as IEC 60034-1.

## Consolidated editions

The IEC is now publishing consolidated versions of its publications. For example, edition numbers 1.0, 1.1 and 1.2 refer, respectively, to the base publication, the base publication incorporating amendment 1 and the base publication incorporating amendments 1 and 2.

## Further information on IEC publications

The technical content of IEC publications is kept under constant review by the IEC, thus ensuring that the content reflects current technology. Information relating to this publication, including its validity, is available in the IEC Catalogue of publications (see below) in addition to new editions, amendments and corrigenda. Information on the subjects under consideration and work in progress undertaken by the technical committee which has prepared this publication, as well as the list of publications issued, is also available from the following:

- **IEC Web Site** ([www.iec.ch](http://www.iec.ch))

- **Catalogue of IEC publications**

The on-line catalogue on the IEC web site ([www.iec.ch/catalog-e.htm](http://www.iec.ch/catalog-e.htm)) enables you to search by a variety of criteria including text searches, technical committees and date of publication. On-line information is also available on recently issued publications, withdrawn and replaced publications, as well as corrigenda.

- **IEC Just Published** [IEC 60601-2-12:2001](http://www.iec.ch/JP.htm)

This summary of recently issued publications ([www.iec.ch/JP.htm](http://www.iec.ch/JP.htm)) is also available by email. Please contact the Customer Service Centre (see below) for further information.

- **Customer Service Centre**

If you have any questions regarding this publication or need further assistance, please contact the Customer Service Centre:

Email: [custserv@iec.ch](mailto:custserv@iec.ch)  
Tel: +41 22 919 02 11  
Fax: +41 22 919 03 00

# INTERNATIONAL STANDARD

# IEC 60601-2-12

[ISO 10651-1]

Second edition  
2001-10

---

---

## Medical electrical equipment –

### Part 2-12:

### Particular requirements for the safety of lung ventilators – Critical care ventilators

iTeh STANDARD PREVIEW

Appareils électromédicaux –

Partie 2-12: IEC 60601-2-12:2001

<https://www.it-ebooks.info> Règles particulières de sécurité pour ventilateurs

pulmonaires – Ventilateurs pour utilisation en soins intensifs

© IEC 2001 — Copyright - all rights reserved

No part of this publication may be reproduced or utilized in any form or by any means, electronic or mechanical, including photocopying and microfilm, without permission in writing from the publisher.

International Electrotechnical Commission  
Telefax: +41 22 919 0300

3, rue de Varembe Geneva, Switzerland  
e-mail: [inmail@iec.ch](mailto:inmail@iec.ch)

IEC web site <http://www.iec.ch>



PRICE CODE

S

For price, see current catalogue

## CONTENTS

FOREWORD.....	4
INTRODUCTION.....	6
<b>SECTION ONE – GENERAL</b>	
1 Scope and object.....	7
2 Terminology and definitions .....	8
3 General requirements .....	12
4 General requirements for tests.....	12
5 Classification .....	12
6 Identification, marking and documents .....	12
7 Power input .....	16
<b>SECTION TWO – ENVIRONMENTAL CONDITIONS</b>	
8 Basic safety categories.....	16
9 Removable protective means .....	16
10 Environmental conditions .....	16
<b>SECTION THREE – PROTECTION AGAINST ELECTRIC SHOCK HAZARDS</b>	
13 General .....	17
14 Requirements related to classification.....	17
15 Limitation of voltage and/or energy.....	17
16 ENCLOSURES and PROTECTIVE COVERS.....	17
17 Separation.....	18
18 Protective earthing, functional earthing and potential equalization.....	18
19 Continuous LEAKAGE CURRENTS and PATIENT AUXILIARY CURRENTS.....	18
20 Dielectric strength.....	18
<b>SECTION FOUR – PROTECTION AGAINST MECHANICAL HAZARDS</b>	
21 Mechanical strength.....	18
22 Moving parts.....	18
23 Surfaces, corners and edges .....	18
24 Stability in NORMAL USE.....	18
25 Expelled parts.....	18
26 Vibration and noise .....	18
27 Pneumatic and hydraulic power.....	19
28 Suspended masses .....	19
<b>SECTION FIVE – PROTECTION AGAINST HAZARDS FROM UNWANTED OR EXCESSIVE RADIATION</b>	
29 X-radiation.....	19
30 Alpha, beta, gamma, neutron radiation and other particle radiation .....	19
31 Microwave radiation .....	19
32 Light radiation (including lasers).....	19
33 Infra-red radiation .....	19
34 Ultra-violet radiation.....	19

35	Acoustical energy (including ultrasonics) .....	19
36	Electromagnetic compatibility .....	19

#### SECTION SIX – PROTECTION AGAINST HAZARDS OF IGNITION OF FLAMMABLE ANAESTHETIC MIXTURES

37	Locations and basic requirements .....	20
38	Marking, ACCOMPANYING DOCUMENTS .....	20
39	Common requirements for CATEGORY AP and CATEGORY APG EQUIPMENT .....	20
40	Requirements and tests for CATEGORY AP EQUIPMENT, parts and components thereof.....	20
41	Requirements and tests for CATEGORY APG EQUIPMENT, parts and components thereof ...	20

#### SECTION SEVEN – PROTECTION AGAINST EXCESSIVE TEMPERATURES AND OTHER SAFETY HAZARDS

42	Excessive temperatures .....	20
43	* Fire prevention .....	20
44	Overflow, spillage, leakage, humidity, ingress of liquids, cleaning, sterilization, disinfection and compatibility .....	21
45	Pressure vessels and parts subject to pressure .....	21
46	Human errors .....	22
47	Electrostatic charges .....	22
48	Biocompatibility .....	22
49	Interruption of the power supply .....	22

#### SECTION EIGHT – ACCURACY OF OPERATING DATA AND PROTECTION AGAINST HAZARDOUS OUTPUT

50	Accuracy of operating data .....	23
51	Protection against hazardous output .....	24

#### SECTION NINE – ABNORMAL OPERATION AND FAULT CONDITIONS; ENVIRONMENTAL TESTS

52	Abnormal operation and fault conditions .....	26
53	Environmental tests .....	26

#### SECTION TEN – CONSTRUCTIONAL REQUIREMENTS

54	General .....	27
55	ENCLOSURES and covers .....	27
56	Components and general assembly .....	27
57	MAINS PARTS, components and layout .....	29
58	Protective earthing – Terminals and connections .....	30
59	Construction and layout .....	30

Appendix L	References – Publications mentioned in this standard .....	31
Annex AA (informative)	Rationale .....	34
Annex BB (normative)	Legibility and visibility of visual indications .....	39
Annex CC (informative)	Intelligent alarm systems .....	40
Bibliography	.....	41
Terminology	– Index of defined terms .....	42

## INTERNATIONAL ELECTROTECHNICAL COMMISSION

## MEDICAL ELECTRICAL EQUIPMENT –

**Part 2-12: Particular requirements for the safety of lung ventilators –  
Critical care ventilators**

## FOREWORD

- 1) The IEC (International Electrotechnical Commission) is a worldwide organization for standardization, comprising all national electrotechnical committees (IEC National Committees). The object of the IEC is to promote international cooperation on all questions concerning standardization in the electrical and electronic fields. To this end and in addition to other activities, the IEC publishes International Standards. Their preparation is entrusted to technical committees; any IEC National Committee interested in the subject dealt with may participate in this preparatory work. International, governmental and non-governmental organizations liaising with the IEC also participate in this preparation. The IEC collaborates closely with the International Organization for Standardization (ISO) in accordance with conditions determined by agreement between the two organizations.
- 2) The formal decisions or agreements of the IEC on technical matters, express as nearly as possible, an international consensus of opinion on the relevant subjects since each technical committee has representation from all interested National Committees.
- 3) The documents produced have the form of recommendations for international use and are published in the form of standards, technical specifications, technical reports or guides, and they are accepted by the National Committees in that sense.
- 4) In order to promote international unification, IEC National Committees undertake to apply IEC International Standards transparently to the maximum extent possible in their national and regional standards. Any divergence between the IEC Standard and the corresponding national standard shall be clearly indicated in the latter.
- 5) The IEC provides no marking procedure to indicate its approval and cannot be rendered responsible for any equipment declared to be in conformity with one of its standards.
- 6) Attention is drawn to the possibility that some of the elements of this International Standard may be the subject of patent rights. IEC shall not be held responsible for identifying any such patent rights.

International Standard IEC 60601-2-12 has been prepared by subcommittee 62D: Electro-medical equipment, of IEC technical committee 62: Electrical equipment in medical practice.

ISO TC 121/SC 3, Lung ventilators and related equipment, also participated in the preparation of this standard.

This second edition replaces the first edition of IEC 60601-2-12:1988, *Medical electrical equipment – Part 2: Particular requirements for the safety of lung ventilators for medical use*, and ISO 10651-1:1993, *Lung ventilators for medical use – Part 1: Requirements*.

The text of this Particular Standard is based on the following documents:

FDIS	Report on voting
62D/414/FDIS	62D/440/RVD

Full information on the voting for the approval of this standard can be found in the report on voting indicated in the above table.

Annex BB forms an integral part of this standard.

Annexes AA and CC are for information only.

In this Particular Standard, the following print types are used:

- requirements, compliance with which can be tested, and definitions: roman type;
- notes, explanations, advice, introductions, general statements and references: smaller roman type;
- *test specifications: italic type;*
- TERMS DEFINED IN CLAUSE 2 OF THE GENERAL STANDARD IEC 60601-1 OR IN THIS PARTICULAR STANDARD: SMALL CAPITALS.

The committee has decided that the contents of this publication will remain unchanged until 2004. At this date, the publication will be:

- reconfirmed;
- withdrawn;
- replaced by a revised edition, or
- amended.

NOTE IEC 60601-1-8, *Medical electrical equipment – Part 1-8: General requirements for safety – Collateral Standard: General requirements and guidelines for the application of alarms in medical electrical equipment* is currently under development. This Standard will require maintenance to conform to that Collateral Standard.

## iTeh STANDARD PREVIEW (standards.iteh.ai)

[IEC 60601-2-12:2001](https://standards.iteh.ai/catalog/standards/sist/2fd258b7-fc51-45ce-ba60-fb2ffcdcf39/iec-60601-2-12-2001)

<https://standards.iteh.ai/catalog/standards/sist/2fd258b7-fc51-45ce-ba60-fb2ffcdcf39/iec-60601-2-12-2001>

## INTRODUCTION

Critical care VENTILATORS are an essential medical device in every intensive care unit (ICU). Approximately half of all PATIENTS in ICUs receive partial to full ventilatory support with this EQUIPMENT. Given the vulnerable status of these PATIENTS, EQUIPMENT safety is of fundamental importance. Accordingly, this Particular Standard, by building on other standards and specifically on IEC 60601-1: *Medical electrical equipment – Part 1: General requirements for safety*, herein referred to as the “General Standard”, sets the minimum requirements that should be met by every critical care VENTILATOR that is designed after the publication of this Particular Standard.

A rationale for the most important requirements is given in Annex AA.

Continuous positive airway pressure (CPAP) devices, sleep apnea therapy devices, support-care VENTILATORS, anaesthesia, emergency and transport VENTILATORS, jet and high frequency VENTILATOR and oscillators are not covered by this Particular Standard, nor are devices that may be used within hospitals, intended solely to augment the ventilation of spontaneously breathing PATIENTS.

## iTeh STANDARD PREVIEW (standards.iteh.ai)

[IEC 60601-2-12:2001](https://standards.iteh.ai/catalog/standards/sist/2fd258b7-fc51-45ce-ba60-fb2ffcdcf39/iec-60601-2-12-2001)

<https://standards.iteh.ai/catalog/standards/sist/2fd258b7-fc51-45ce-ba60-fb2ffcdcf39/iec-60601-2-12-2001>



## MEDICAL ELECTRICAL EQUIPMENT –

### Part 2-12: Particular requirements for the safety of lung ventilators – Critical care ventilators

#### SECTION ONE – GENERAL

The clauses and subclauses of this section of the General Standard apply except as follows:

#### 1 Scope and object

This clause of the General Standard applies, except as follows:

##### 1.1 Scope

*Addition:*

This Particular Standard specifies the safety requirements for VENTILATORS, as defined in 2.1.125, intended for use in critical care settings.

Continuous positive airway pressure (CPAP) devices, sleep apnea therapy devices, support-care VENTILATORS, emergency and transport VENTILATORS, jet and high frequency VENTILATORS and oscillators are outside the scope of this Particular Standard, nor are devices that may be used within hospitals, intended solely to augment the ventilation of spontaneously breathing PATIENTS. Standards for other types of VENTILATORS, e.g. high frequency jet and oscillation ventilators, are under consideration.

Requirements for VENTILATORS intended for anaesthetic applications are given in IEC 60601-2-13.

##### 1.2 Object

*Addition:*

The object of this standard is to specify particular safety requirements for VENTILATORS intended for use in critical care settings.

##### 1.3 Particular standards

*Addition:*

This Particular Standard refers to IEC 60601-1:1988, *Medical electrical equipment – Part 1: General requirements for safety*, as amended by its amendment 1 (1991) and its amendment 2 (1995), herein referred to as the “General Standard”.

The General Standard takes into account a set of Collateral Standards:

IEC 60601-1-1:2000, *Medical electrical equipment – Part 1-1: General requirements for safety, Collateral standard: Safety requirements for medical electrical systems*

IEC 60601-1-2:2001, *Medical electrical equipment – Part 1-2: General requirements for safety – Collateral standard: Electromagnetic compatibility – Requirements and tests*

IEC 60601-1-4:1996, *Medical electrical equipment – Part 1: General requirements for safety, 4. Collateral standard: Programmable electrical medical systems*  
Amendment 1<sup>1</sup>

The term “this Standard” covers this Particular Standard, used together with the General Standard and the Collateral Standards.

The numbering of sections, clauses, and subclauses of this Particular Standard corresponds with that of the General Standard. The changes to the text of the General Standard are specified by the use of the following words:

- “Replacement” means that the clause or subclause of the General Standard is replaced completely by the text of this Particular Standard.
- “Addition” means that the text of this Particular Standard is additional to the requirements of the General Standard.
- “Amendment” means that the clause or subclause of the General Standard is amended as indicated by the text of this Particular Standard.

Subclauses or figures that are additional to those of the General Standard are numbered starting from 101, additional annexes are lettered AA, BB, etc., and additional items aa), bb), etc.

Clauses and subclauses to which there is a rationale are marked with an asterisk \*. These rationales can be found in an informative Annex AA.

Annexes AA and CC are not normative parts of this Particular Standard and only provide additional information; they can never be the subjects of testing.

Where there is no corresponding section, clause or subclause in this Particular Standard, the section, clause or subclause of the General Standard or specified Collateral Standard applies without modification.

Where it is intended that any part of the General Standard or Collateral Standard, although possibly relevant, is not to be applied, a statement to that effect is given in this Particular Standard.

A requirement of this Particular Standard replacing or modifying requirements of the General Standard or a Collateral Standard takes precedence over the corresponding general requirement(s).

## 2 Terminology and definitions

This clause of the General Standard applies, except as follows:

### \* 2.1.5 applied part

*Addition:*

or any part of the VENTILATOR intended to be connected to the breathing system

---

<sup>1</sup> There exists a consolidated edition 1.1 (2000) that includes IEC 60601-1-4 (1996) and its amendment 1 (1999).

*Additional definitions:*

**2.1.101****bacterial filter**

device that removes bacteria and particulate matter from the gas stream

[ISO 4135:1995, definition 4.1.7 modified]

**2.1.102****clearly legible**

visual attribute of information displayed by the EQUIPMENT that allows the OPERATOR to discern (or identify) qualitative or quantitative values or functions under a specific set of environmental conditions

**2.1.103****emergency air intake port**

dedicated intake port through which ambient air may be drawn when the supply of FRESH GAS is insufficient or absent

[ISO 4135:1995, definition 4.2.2 modified]

**2.1.104****flow-direction-sensitive component**

VENTILATOR component through which the gas flow has to be in one direction only for its proper functioning and/or PATIENT safety

[ISO 4135:1995, definition 4.1.13]

**2.1.105****fresh gas**

gas supplied to the VENTILATOR BREATHING SYSTEM. It excludes the following:

- air drawn through the EMERGENCY AIR INTAKE PORT;
- air drawn through leaks in the VENTILATOR BREATHING SYSTEM;
- expired gas from the PATIENT

**2.1.106****fresh gas intake port**

intake port, other than the EMERGENCY AIR INTAKE PORT, through which FRESH GAS may be drawn into the VENTILATOR BREATHING SYSTEM

[ISO 4135:1995, definition 4.2.6 modified]

**2.1.107****gas exhaust port**

that port of a VENTILATOR from which gas is discharged to the atmosphere either directly or via a gas scavenging system

[ISO 4135:1995, definition 4.2.7]

**2.1.108****gas intake port**

port through which gas is drawn into the VENTILATOR BREATHING SYSTEM

### 2.1.109

#### **gas output port**

port through which gas is delivered at RESPIRATORY PRESSURES via the inspiratory limb to the PATIENT CONNECTION PORT

[ISO 4135:1995, definition 4.2.8 modified]

### 2.1.110

#### **gas return port**

port through which gas is returned at RESPIRATORY PRESSURES via the expiratory limb from the PATIENT CONNECTION PORT

[ISO 4135:1995, definition 4.2.9 modified]

### 2.1.111

#### **high pressure gas input port**

input port to which gas may be supplied at a pressure greater than 100 kPa

[ISO 4135:1995, definition 4.2.10 modified]

### 2.1.112

#### **inflating gas**

FRESH GAS that may also power the VENTILATOR

### 2.1.113

#### **inflating gas input port**

input port to which INFLATING GAS is supplied

[ISO 4135:1995, definition 4.2.11]

NOTE An input port is a port to which gas is supplied under positive pressure and through which the gas is driven by this pressure. The gas may be supplied either at a controlled pressure or at a controlled flow.

### 2.1.114

#### **inhibition (disabled)**

state in which an alarm system or part of an alarm system can not annunciate alarm signals

NOTE 1 INHIBITION may apply to an individual alarm condition, to a group of alarm conditions, or to the entire alarm system of the EQUIPMENT.

NOTE 2 INHIBITION may be invoked by the OPERATOR or by the EQUIPMENT (for instance, in a warm-up mode or when no PATIENT is connected).

NOTE 3 The duration of INHIBITION is always indefinite. Only direct action by the OPERATOR or a change in the EQUIPMENT caused by the OPERATOR (for instance, the end of a warm-up mode or when a PATIENT is connected) will revoke INHIBITION.

### 2.1.115

#### **low-pressure gas input port**

input port to which gas is supplied at a pressure not exceeding 100 kPa

[ISO 4135:1995, definition 4.2.14]

### 2.1.116

#### **manual ventilation port**

port to which a device may be connected for manual inflation of the lungs

[ISO 4135:1995, definition 4.2.15 modified]

### 2.1.117

#### **maximum limited pressure ( $P_{LIM\ max}$ )**

Highest pressure at the PATIENT CONNECTION PORT during NORMAL USE and under a SINGLE FAULT CONDITION

**2.1.118****maximum working pressure ( $P_w \text{ max}$ )**

highest pressure at the PATIENT CONNECTION PORT during NORMAL USE, irrespective of the setting of controls, other than the control intended to adjust this pressure

NOTE Even if not adjustable, this maximum is equal to or less than the MAXIMUM LIMITED PRESSURE.

**2.1.119****minimum limited pressure ( $P_{\text{LIM}} \text{ min}$ )**

lowest pressure at the PATIENT CONNECTION PORT during NORMAL USE and under a SINGLE FAULT CONDITION

NOTE This pressure may be sub-atmospheric.

**2.1.120****minute volume ( $\dot{V}$ )**

volume of gas per minute entering or leaving the lungs of the PATIENT

**2.1.121****operator's position**

intended location and orientation of the OPERATOR with respect to the EQUIPMENT for NORMAL USE according to the instructions for use

**2.1.122****patient connection port (of the VENTILATOR BREATHING SYSTEM)**

port of the VENTILATOR BREATHING SYSTEM to which the PATIENT can be connected

[ISO 4135:1995, definition 4.2.16]

NOTE Interface between the VENTILATOR BREATHING SYSTEM and the PATIENT.

**2.1.123****respiratory pressure**

pressure at the PATIENT CONNECTION PORT

<https://standards.iteh.ai/catalog/standards/sist/2fd258b7-fc51-45ce-ba60-fb2ffcdcf39/iec-60601-2-12-2001>

**2.1.124****suspended**

state of an alarm system where OPERATOR action has temporarily caused an otherwise enabled alarm system to disable all auditory or all auditory and visual alarm signals for a fixed interval

**2.1.125****ventilator**

automatic EQUIPMENT that is intended to augment or provide ventilation of the lungs of the PATIENT when connected to the airway of the PATIENT

**2.1.126****ventilator breathing system (VBS)**

breathing system bounded by the LOW PRESSURE GAS INPUT PORT(S), the GAS INTAKE PORT(S) and the PATIENT CONNECTION PORT, together with the FRESH GAS INTAKE and EXHAUST PORT(S), if these are provided

[ISO 4135:1995, definition 4.1.6 modified]