
Acoustics — Determination of sound power levels of noise sources using sound intensity —

Part 3:
Precision method for measurement by scanning

iTeh STANDARD PREVIEW
(standards.iteh.ai)

Acoustique — Détermination par intensimétrie des niveaux de puissance acoustique émis par les sources de bruit —

Partie 3: Méthode de précision pour mesurage par balayage

<https://standards.iteh.ai/catalog/standards/sis/161b375c-cba9-4012-8d19-b28456cf27c3/iso-9614-3-2002>



PDF disclaimer

This PDF file may contain embedded typefaces. In accordance with Adobe's licensing policy, this file may be printed or viewed but shall not be edited unless the typefaces which are embedded are licensed to and installed on the computer performing the editing. In downloading this file, parties accept therein the responsibility of not infringing Adobe's licensing policy. The ISO Central Secretariat accepts no liability in this area.

Adobe is a trademark of Adobe Systems Incorporated.

Details of the software products used to create this PDF file can be found in the General Info relative to the file; the PDF-creation parameters were optimized for printing. Every care has been taken to ensure that the file is suitable for use by ISO member bodies. In the unlikely event that a problem relating to it is found, please inform the Central Secretariat at the address given below.

iTeh STANDARD PREVIEW
(standards.iteh.ai)

ISO 9614-3:2002

<https://standards.iteh.ai/catalog/standards/sist/161b375c-eba9-4012-8d19-b28456cf27c3/iso-9614-3-2002>

© ISO 2002

All rights reserved. Unless otherwise specified, no part of this publication may be reproduced or utilized in any form or by any means, electronic or mechanical, including photocopying and microfilm, without permission in writing from either ISO at the address below or ISO's member body in the country of the requester.

ISO copyright office
Case postale 56 • CH-1211 Geneva 20
Tel. + 41 22 749 01 11
Fax + 41 22 749 09 47
E-mail copyright@iso.ch
Web www.iso.ch

Printed in Switzerland

Contents

| | Page |
|---|------|
| Foreword | v |
| Introduction..... | vi |
| 1 Scope..... | 1 |
| 2 Normative references..... | 1 |
| 3 Terms and definitions | 2 |
| 4 General requirements | 7 |
| 4.1 Size of sound source under test..... | 7 |
| 4.2 Character of sound radiated by the source..... | 7 |
| 4.3 Measurement uncertainty..... | 7 |
| 5 Acoustic environment..... | 9 |
| 5.1 Criteria for adequacy of the test environment | 9 |
| 5.2 Extraneous intensity | 9 |
| 5.3 Wind and gas flows..... | 9 |
| 5.4 Temperature..... | 9 |
| 5.5 Configuration of the surroundings..... | 9 |
| 5.6 Atmospheric conditions..... | 9 |
| 6 Instrumentation | 10 |
| 6.1 General | 10 |
| 6.2 Calibration and field check..... | 10 |
| 6.3 Time-series of sound intensity and sound pressure..... | 10 |
| 7 Installation and operation of the source..... | 10 |
| 7.1 General | 10 |
| 7.2 Operating conditions of the source under test..... | 11 |
| 8 Measurement of normal sound intensity component levels | 11 |
| 8.1 Determination of measurement surface | 11 |
| 8.2 Determination of scanning paths and segments..... | 11 |
| 8.3 Measurements | 12 |
| 8.4 Further actions | 14 |
| 9 Determination of sound power level | 15 |
| 9.1 Calculation of partial sound powers for each partial surface of the measurement surface | 15 |
| 9.2 Calculation of normalized sound power level..... | 15 |
| 10 Information to be recorded..... | 15 |
| Annex A (informative) List of symbols used in this part of ISO 9614 | 17 |
| Annex B (normative) Calculation of field indicators..... | 19 |
| B.1 General | 19 |
| B.2 Definition of field indicators..... | 19 |
| B.2.1 Temporal variability indicator, F_T | 19 |
| B.2.2 Unsigned pressure-intensity indicator, $F_{p I_n}$ | 19 |
| B.2.3 Signed pressure-intensity indicator, F_{pI_n} | 20 |
| B.2.4 Field non-uniformity indicator, F_S | 21 |
| Annex C (normative) Procedure for achieving the desired accuracy | 22 |
| C.1 Qualification requirements..... | 22 |
| C.1.1 General | 22 |

C.1.2 Check for the adequacy of the averaging time 22
C.1.3 Check for the repeatability of the scan on a partial surface 22
C.1.4 Check for the adequacy of the measurement equipment 22
C.1.5 Check for the presence of strong extraneous noise 23
C.1.6 Check for the field non-uniformity..... 23
C.2 Action to be taken to increase the grade of accuracy of determination 23
Annex D (informative) Effects of airflow on measurement of sound intensity..... 26
Annex E (informative) Effect of sound absorption within the measurement surface..... 27
Annex F (informative) Measurement surface and scanning procedure 28
Annex G (informative) Procedure for obtaining time-averaged intensities and squared pressures from a sequence of short-time averaged intensities and squared pressures 29
Annex H (informative) Normalization of sound power level 30
H.1 General 30
H.2 Calculation of normalized sound power level 30
Annex I (informative) Field indicators used in ISO 9614-1, -2 and -3..... 32
Bibliography..... 33

iTeh STANDARD PREVIEW
(standards.iteh.ai)

ISO 9614-3:2002
<https://standards.iteh.ai/catalog/standards/sist/161b375c-eba9-4012-8d19-b28456cf27c3/iso-9614-3-2002>

Foreword

ISO (the International Organization for Standardization) is a worldwide federation of national standards bodies (ISO member bodies). The work of preparing International Standards is normally carried out through ISO technical committees. Each member body interested in a subject for which a technical committee has been established has the right to be represented on that committee. International organizations, governmental and non-governmental, in liaison with ISO, also take part in the work. ISO collaborates closely with the International Electrotechnical Commission (IEC) on all matters of electrotechnical standardization.

International Standards are drafted in accordance with the rules given in the ISO/IEC Directives, Part 3.

The main task of technical committees is to prepare International Standards. Draft International Standards adopted by the technical committees are circulated to the member bodies for voting. Publication as an International Standard requires approval by at least 75 % of the member bodies casting a vote.

Attention is drawn to the possibility that some of the elements of this part of ISO 9614 may be the subject of patent rights other than those identified above. ISO shall not be held responsible for identifying any or all such patent rights.

ISO 9614-3 was prepared by Technical Committee ISO/TC 43, *Acoustics*, Subcommittee SC 1, *Noise*.

ISO 9614 consists of the following parts, under the general title *Acoustics — Determination of sound power levels of noise sources using sound intensity*: **(standards.iteh.ai)**

- *Part 1: Measurement at discrete points* [ISO 9614-3:2002](https://standards.iteh.ai/catalog/standards/sist/161b375c-eba9-4012-8d19-b28456cf27c3/iso-9614-3-2002)
- *Part 2: Measurement by scanning* <https://standards.iteh.ai/catalog/standards/sist/161b375c-eba9-4012-8d19-b28456cf27c3/iso-9614-3-2002>
- *Part 3: Precision method for measurement by scanning*

Annexes B and C form a normative part of this part of ISO 9614. Annexes A, D, E, F, G, H and I are for information only.

Introduction

0.1 The sound power radiated by a source is equal in value to the integral of the scalar product of the sound intensity vector and the associated elemental area vector over any surface totally enclosing the source. Other International Standards which describe methods of determination of the sound power levels of noise sources, principally ISO 3740 to ISO 3747, without exception specify sound pressure level as the primary acoustic quantity to be measured. The relationship between sound intensity level and sound pressure level at any point depends on the characteristics of the source, the characteristics of the measurement environment, and the disposition of the measurement positions with respect to the source.

The procedures specified in ISO 3740 to ISO 3747 are not always applicable, for the following reasons.

- a) Specific facilities are necessary if high precision is required. It is frequently not possible to install, and operate, large pieces of equipment in such facilities.
- b) They cannot be used in the presence of high levels of extraneous noise generated by sources other than that under investigation.

0.2 This part of ISO 9614 specifies methods of determining the sound power levels of sources, within specific ranges of uncertainty, under test conditions which are less restricted than those required by ISO 3740 to ISO 3747.

It is recommended that personnel performing sound intensity measurements according to this part of ISO 9614 are appropriately trained and experienced.

0.3 This part of ISO 9614 complements ISO 9614-1, ISO 9614-2 and the ISO 3740 to ISO 3747 series, which specify various methods for the determination of sound power levels of machines and equipment. It differs from the ISO 3740 to ISO 3747 series principally in three aspects.

- a) Measurements are made of sound intensity as well as of sound pressure.
- b) The uncertainty of the sound power level determined by the method specified in this part of ISO 9614 is classified according to the results of specified ancillary tests and calculations performed in association with the test measurements.
- c) Current limitations of intensity measurement equipment which conforms to IEC 61043 restrict measurements to the one-third octave range 50 Hz to 6,3 kHz. Octave band and band-limited A-weighted values are determined from the constituent one-third-octave band values.

0.4 The integral over any surface totally enclosing the source of the scalar product of the sound intensity vector and the associated elemental area vector provides a measure of the sound power radiated directly into the air by all sources located within the enclosing surface and excludes sound radiated by sources located outside this surface. In practice, this exclusion is effective only if the source under test and other sources of extraneous intensity on the measurement surface are stationary over time. In the presence of sound sources operating outside the measurement surface, any system lying within the surface can absorb a proportion of energy incident upon it. The total sound power absorbed within the measurement surface will appear as a negative contribution to source power, and can produce an error in the sound power determination. In order to minimize the associated error, it is therefore necessary to remove any sound-absorbing material lying within the measurement surface which is not normally present during the operation of the source under test.

This method is based on sampling of the intensity normal to the measurement surface by moving an intensity probe continuously along specified paths. The resulting sampling error is a function of the spatial variation of the normal intensity component over the measurement surface, which depends on the directivity of the source, the chosen measurement surface, the pattern and speed of the probe scanning, and the proximity of extraneous sources outside the measurement surface.

The accuracy of measurement of the normal component of sound intensity at a position is sensitive to the difference between the local sound pressure level and the local normal sound intensity level. A large difference can occur when the intensity vector at a measurement position is directed at a large angle (approaching 90°) to the local normal to the measurement surface. Alternatively, the local sound pressure level can contain strong contributions from sources outside the measurement surface, but can be associated with little net sound energy flow, as in a reverberant field in an enclosure; or the field can be strongly reactive because of the presence of the near field and/or standing waves.

The accuracy of determination of sound power level is adversely affected by a flow of sound energy into the volume enclosed by the measurement surface through a portion of that surface, even though it is, in principle, compensated by increased flow of the volume out through the remaining portion of the surface. This condition is caused by the presence of a strong extraneous source outside the measurement surface. This part of ISO 9614 limits such situations by giving relevant criteria.

iTeh STANDARD PREVIEW (standards.iteh.ai)

[ISO 9614-3:2002](https://standards.iteh.ai/catalog/standards/sist/161b375c-eba9-4012-8d19-b28456cf27c3/iso-9614-3-2002)

<https://standards.iteh.ai/catalog/standards/sist/161b375c-eba9-4012-8d19-b28456cf27c3/iso-9614-3-2002>

iTeh STANDARD PREVIEW
(standards.iteh.ai)

ISO 9614-3:2002

<https://standards.iteh.ai/catalog/standards/sist/161b375c-eba9-4012-8d19-b28456cf27c3/iso-9614-3-2002>

Acoustics — Determination of sound power levels of noise sources using sound intensity —

Part 3: Precision method for measurement by scanning

1 Scope

1.1 This part of ISO 9614 specifies a method for measuring the component of sound intensity normal to a measurement surface which is chosen so as to enclose the sound source(s) of which the sound power level is to be determined.

Surface integration of the intensity component normal to the measurement surface is approximated by subdividing the measurement surface into contiguous partial surfaces, and scanning the intensity probe over each partial surface along a continuous path which covers the extent of the partial surface. The measurement instrument determines the averaged normal intensity component and averaged squared sound pressure over the duration of each scan. The scanning operation can be performed either manually or by means of a mechanical system.

The octave band or band-limited weighted sound power level is calculated from the measured one-third-octave-band values. The method is applicable to any source for which a physically stationary measurement surface can be defined, and on which the sound generated by the source under test and by other significant extraneous sources are stationary in time. The source is defined by the choice of measurement surface. The method is applicable in specific test environments fulfilling all relevant requirements of this part of ISO 9614.

This part of ISO 9614 specifies certain ancillary procedures, described in annex C, to be followed in conjunction with the sound power determination. The results are used to indicate the quality of the determination, and hence the grade of accuracy. If the quality of the determination does not meet the requirements of this part of ISO 9614, the test procedure shall be modified in the manner indicated.

This part of ISO 9614 is not applicable to any frequency band in which the sound power of the source is found to be negative on measurement.

1.2 This part of ISO 9614 is applicable to sources situated in any environment which is neither so variable over time as to reduce the accuracy of the measurement of sound intensity to an unacceptable degree, nor subjects the intensity measurement probe to gas flows of unacceptable speed or unsteadiness (see 5.2.2, 5.3 and 5.4).

In some cases it will be found that the test conditions are too adverse to allow the requirements of this part of ISO 9614 to be met. For example, extraneous noise levels can exceed the dynamic capability of the measuring instrument or can vary to an excessive degree during the test. In such cases the method given in this part of ISO 9614 is not suitable for the determination of the sound power level of the source.

2 Normative references

The following normative documents contain provisions which, through reference in this text, constitute provisions of this part of ISO 9614. For dated references, subsequent amendments to, or revisions of, any of these publications do not apply. However, parties to agreements based on this part of ISO 9614 are encouraged to investigate the possibility of applying the most recent editions of the normative documents indicated below. For undated

references, the latest edition of the normative document referred to applies. Members of ISO and IEC maintain registers of currently valid International Standards.

IEC 60651, *Sound level meters*

IEC 60942:1998, *Electroacoustics — Sound calibrators*

IEC 61260, *Electroacoustics — Octave-band and fractional-octave-band filters*

IEC 61043:1993, *Electroacoustics — Instruments for the measurement of sound intensity — Measurements with pairs of pressure sensing microphones*

GUM:1993, *Guide to the expression of uncertainty in measurement*. BIPM, IEC, IFCC, ISO, IUPAC, IUPAP, OIML.

3 Terms and definitions

For the purposes of this part of ISO 9614, the following terms and definitions apply.

NOTE Symbols used in this part of ISO 9614 are listed in annex A. Definitions of field indicators are given in annex B.

3.1 sound pressure level

L_p
ten times the logarithm to the base 10 of the ratio of the mean-square sound pressure to the square of the reference sound pressure

NOTE 1 The reference sound pressure is 20 μPa .

NOTE 2 Sound pressure level is expressed in decibels. <https://standards.iteh.ai/catalog/standards/sist/161b375c-eba9-4012-8d19-b28456cf27c3/iso-9614-3-2002>

3.2 instantaneous sound intensity

$\vec{I}(t)$
instantaneous flow of sound energy per unit of area and per unit time in the direction of the local instantaneous acoustic particle velocity

NOTE This is a vectorial quantity which is equal to the product of the instantaneous sound pressure at a point and the associated particle velocity

$$\vec{I}(t) = p(t) \cdot \vec{u}(t) \tag{1}$$

where

$p(t)$ is the instantaneous sound pressure at a point;

$\vec{u}(t)$ is the associated instantaneous particle velocity at the same point;

t is time.

3.3 sound intensity

\bar{I}
time-averaged value of $\vec{I}(t)$ in a temporally stationary sound field

$$\bar{I} = \lim_{T \rightarrow \infty} \frac{1}{T} \int_0^T \vec{I}(t) dt \tag{2}$$

where

T is the integration period

NOTE Also

I is the signed magnitude of \vec{I} ; in this part of ISO 9614, the sign is chosen so that the energy flow going out of the sound source through the measurement surface is measured positive;

$|I|$ is the unsigned magnitude of \vec{I} .

3.4 normal sound intensity

I_n
component of sound intensity in the direction normal to a measurement surface defined by the unit normal vector \vec{n}

$$I_n = \vec{I} \cdot \vec{n} \quad (3)$$

where \vec{n} is the unit normal vector directed out of the volume enclosed by the measurement surface

3.5 normal sound intensity level

L_{I_n}
logarithmic measure of the value of the normal sound intensity, I_n , given by

$$L_{I_n} = 10 \lg \frac{|I_n|}{I_0} \text{ dB} \quad (4)$$

where I_0 is the reference sound intensity ($=10^{-12} \text{ W} \cdot \text{m}^{-2}$)

NOTE 1 It is expressed in decibels.

NOTE 2 When I_n is negative, the level is expressed as $(-)\text{XX dB}$, except when used in the evaluation of δ_{pI_0} (see 3.10).

3.6 Sound powers

3.6.1 partial sound power

P_i
time-averaged flow of sound energy per unit of time through a partial surface of a measurement surface, given by

$$P_i = \overline{I_{ni}} \cdot S_i \quad (5)$$

where

$\overline{I_{ni}}$ is the signed magnitude of the partial surface average normal sound intensity measured on the partial surface i of the measurement surface;

S_i is the area of the partial surface i .

ISO 9614-3:2002(E)

NOTE 1 When the averaged normal sound intensity level \bar{L}_{I_n} for a partial surface i is expressed as XX dB, the value of \bar{I}_{ni} is calculated from the equation.

$$\bar{I}_{ni} = I_0 10^{XX/10} \quad (6)$$

NOTE 2 When the averaged normal sound intensity level \bar{L}_{I_n} for a partial surface i is expressed as $(-)$ XX dB, the value of \bar{I}_{ni} is calculated from the equation.

$$\bar{I}_{ni} = -I_0 10^{XX/10} \quad (7)$$

3.6.2

sound power

P

total sound power generated by a source, as determined using the method given in this part of ISO 9614, given by

$$P = \sum_{i=1}^N P_i \quad (8)$$

where N is the total number of partial surfaces of the measurement surface

3.6.3

sound power level

L_W

logarithmic measure of the sound power generated by a source, as determined using this part of ISO 9614, given by

$$L_W = 10 \lg \frac{|P|}{P_0} \text{ dB} \quad (9)$$

where P_0 is the reference sound power ($= 10^{-12}$ W)

NOTE 1 It is expressed in decibels.

NOTE 2 When P is negative, the level is expressed as $(-)$ XX dB for record purposes only.

3.6.4

normalized sound power level

L_{W0}

sound power level under the reference meteorological condition (temperature $\theta_0 = 23$ °C, barometric pressure $B_0 = 101\,325$ Pa), given by

$$L_{W0} = L_W - 15 \lg \left[\frac{B}{101\,325} \times \frac{296,15}{273,15 + \theta} \right] \text{ dB} \quad (10)$$

where

θ is the air temperature, in degrees Celsius, during the actual measurement;

B is the barometric pressure, in pascals, during the actual measurement.

NOTE See annex H.

3.7 Surfaces

3.7.1

measurement surface

hypothetical surface on which intensity measurements are made, and which either completely encloses the sound source under test or, in conjunction with an acoustically rigid continuous surface, encloses the sound source under tests

NOTE In cases where the hypothetical surface is penetrated by bodies possessing solid surfaces, the measurement surface terminates at the lines of intersection between the bodies and the surface.

3.7.2

partial surface

one of a set of smaller surfaces into which a measurement surface is divided and over which a partial sound power is obtained

See Figure 1.

3.7.3

segment

one of a set of smaller surfaces into which a partial surface is divided

See Figure 2.

NOTE The idea of "segment" is introduced so that the scanning path and time are determined on a partial surface.

3.8

extraneous intensity

contribution to the sound intensity which arises from the operation of sources external to the measurement surface (source mechanisms operating outside the volume enclosed by the measurement surface)

<https://standards.iteh.ai/catalog/standards/sist/161b375c-e8a9-4012-8d19-b28456cf27c3/iso-9614-3-2002>

3.9

probe

that part of the intensity measurement system which incorporates the sensors

3.10

pressure-residual intensity index

δ_{pI_0}

difference between the indicated L_p and indicated $L_{I\delta}$ when the intensity probe is placed and oriented in a sound field such that the sound intensity is zero

$$\delta_{pI_0} = L_p - L_{I\delta} \quad (11)$$

where $L_{I\delta}$ is the level of residual intensity I_δ given by

$$L_{I\delta} = 10 \lg \frac{|I_\delta|}{I_0} \text{ dB} \quad (12)$$

NOTE 1 It is expressed in decibels.

NOTE 2 Details for determining δ_{pI_0} are given in IEC 61043.

3.11

dynamic capability index

L_d

index given by