

SLOVENSKI STANDARD

SIST EN 61094-3:2016

01-november-2016

Nadomešča:
SIST EN 61094-3:2002

Merilni mikrofoni - 3. del: Primarna metoda za kalibriranje laboratorijskih standardnih mikrofонов v prostem polju z recipročno tehniko (IEC 61094-3:2016)

Measurement microphones - Part 3: Primary method for free-field calibration of laboratory standard microphones by the reciprocity technique (IEC 61094-3:2016)

iTeh STANDARD PREVIEW
(standards.iteh.ai)

Microphones de mesure - Partie 3 : Méthode de base pour l'étalonnage en champ libre des microphones étalons de laboratoire par la méthode de réciprocité (IEC 61094-3:2016)

<https://standards.iteh.ai/catalog/standards/sist/313e2841-8531-4100-a460-f0992afd66b5/sist-en-61094-3-2016>

Ta slovenski standard je istoveten z: EN 61094-3:2016

ICS:

17.140.50	Elektroakustika	Electroacoustics
33.160.50	Pribor	Accessories

SIST EN 61094-3:2016

en

iTeh STANDARD PREVIEW
(standards.iteh.ai)

SIST EN 61094-3:2016

<https://standards.iteh.ai/catalog/standards/sist/313e2841-8531-4100-a460-f0992afd66b5/sist-en-61094-3-2016>

EUROPEAN STANDARD
NORME EUROPÉENNE
EUROPÄISCHE NORM

EN 61094-3

September 2016

ICS 17.140.50; 33.160.50

Supersedes EN 61094-3:1995

English Version

**Electroacoustics - Measurement microphones - Part 3: Primary
method for free-field calibration of laboratory standard
microphones by the reciprocity technique
(IEC 61094-3:2016)**

Électroacoustique - Microphones de mesure - Partie 3:
Méthode primaire pour l'étalonnage en champ libre des
microphones étalons de laboratoire par la méthode de
réciprocité
(IEC 61094-3:2016)

Messmikrofone - Teil 3: Primärverfahren zur Freifeld-
Kalibrierung von Laboratoriums-Normalmikrofonen nach der
Reziprozitätsmethode
(IEC 61094-3:2016)

This European Standard was approved by CENELEC on 2016-07-19. CENELEC members are bound to comply with the CEN/CENELEC Internal Regulations which stipulate the conditions for giving this European Standard the status of a national standard without any alteration.

Up-to-date lists and bibliographical references concerning such national standards may be obtained on application to the CEN-CENELEC Management Centre or to any CENELEC member.

This European Standard exists in three official versions (English, French, German). A version in any other language made by translation under the responsibility of a CENELEC member into its own language and notified to the CEN-CENELEC Management Centre has the same status as the official versions.

CENELEC members are the national electrotechnical committees of Austria, Belgium, Bulgaria, Croatia, Cyprus, the Czech Republic, Denmark, Estonia, Finland, Former Yugoslav Republic of Macedonia, France, Germany, Greece, Hungary, Iceland, Ireland, Italy, Latvia, Lithuania, Luxembourg, Malta, the Netherlands, Norway, Poland, Portugal, Romania, Slovakia, Slovenia, Spain, Sweden, Switzerland, Turkey and the United Kingdom.



European Committee for Electrotechnical Standardization
Comité Européen de Normalisation Electrotechnique
Europäisches Komitee für Elektrotechnische Normung

CEN-CENELEC Management Centre: Avenue Marnix 17, B-1000 Brussels

EN 61094-3:2016**European foreword**

The text of document 29/873/CDV, future edition 2 of IEC 61094-3, prepared by IEC TC 29, Electroacoustics, was submitted to the IEC-CENELEC parallel vote and was approved by CENELEC as EN 61094-3:2016.

The following dates are fixed:

- latest date by which the document has to be implemented at national level by publication of an identical national standard or by endorsement (dop) 2017-04-19
- latest date by which the national standards conflicting with the document have to be withdrawn (dow) 2019-07-19

This document supersedes EN 61094-3:1995.

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. CENELEC [and/or CEN] shall not be held responsible for identifying any or all such patent rights.

Endorsement notice

The text of the International Standard IEC 61094-3:2016 was approved by CENELEC as a European Standard without any modification.

In the official version, for Bibliography, the following note has to be added for the standard indicated:

SIST EN 61094-3:2016
 IEC 61094-8:2012 <https://standards.iteh.ai/standards/IEC/61094-8:2012/100-a460-f0992afd66b5/sist-en-61094-3-2016> NOTE Harmonized as EN 61094-8:2012

Annex ZA (normative)

Normative references to international publications with their corresponding European publications

The following documents, in whole or in part, are normatively referenced in this document and are indispensable for its application. For dated references, only the edition cited applies. For undated references, the latest edition of the referenced document (including any amendments) applies.

NOTE 1 When an International Publication has been modified by common modifications, indicated by (mod), the relevant EN/HD applies.

NOTE 2 Up-to-date information on the latest versions of the European Standards listed in this annex is available here: www.cenelec.eu.

<u>Publication</u>	<u>Year</u>	<u>Title</u>	<u>EN/HD</u>	<u>Year</u>
IEC 61094-1	2000	Measurement microphones -- Part 1: Specifications for laboratory standard microphones	EN 61094-1	2000
IEC 61094-2	2009	Electroacoustics - Measurement microphones -- Part 2: Primary method for the pressure calibration of laboratory standard microphones by the reciprocity technique	EN 61094-2	2009
ISO 9613-1	-	Acoustics; attenuation of sound during propagation outdoors; part 1: calculation of the absorption of sound by the atmosphere	-	-
IEC/TS 61094-7	-	Measurement microphones -- Part 7: Values for the difference between free-field and pressure sensitivity levels of laboratory standard microphones	-	-
ISO/IEC Guide 98-3	-	Uncertainty of measurement - Part 3: Guide to the expression of uncertainty in measurement (GUM:1995)	-	-

iTeh STANDARD PREVIEW
(standards.iteh.ai)

SIST EN 61094-3:2016

<https://standards.iteh.ai/catalog/standards/sist/313e2841-8531-4100-a460-f0992afd66b5/sist-en-61094-3-2016>



IEC 61094-3

Edition 2.0 2016-06

INTERNATIONAL STANDARD

NORME INTERNATIONALE



**Electroacoustics – Measurement microphones –
Part 3: Primary method for free-field calibration of laboratory standard
microphones by the reciprocity technique**

**Électroacoustique – Microphones de mesure –
Partie 3: Méthode primaire pour l'étalonnage en champ libre des microphones
étalons de laboratoire par la méthode de réciprocité**

INTERNATIONAL
ELECTROTECHNICAL
COMMISSION

COMMISSION
ELECTROTECHNIQUE
INTERNATIONALE

ICS 17.140.50; 33.160.50

ISBN 978-2-8322-3478-5

**Warning! Make sure that you obtained this publication from an authorized distributor.
Attention! Veuillez vous assurer que vous avez obtenu cette publication via un distributeur agréé.**

CONTENTS

FOREWORD.....	4
1 Scope.....	6
2 Normative references.....	6
3 Terms and definitions	6
4 Reference environmental conditions.....	7
5 Principles of free-field calibration by reciprocity.....	7
5.1 General principles	7
5.1.1 General	7
5.1.2 General principles using three microphones.....	7
5.1.3 General principles using two microphones and an auxiliary sound source	8
5.2 Basic expressions	8
5.3 Insert voltage technique	9
5.4 Free-field receiving characteristics of a microphone	9
5.5 Free-field transmitting characteristics of a microphone	10
5.6 Reciprocity procedure	11
5.7 Final expressions for the free-field sensitivity.....	11
5.7.1 Method using three microphones.....	11
5.7.2 Method using two microphones and an auxiliary sound source	12
6 Factors influencing the free-field sensitivity.....	12
6.1 General.....	12
6.2 Polarizing voltage.....	12
6.3 Shield configuration	12
6.4 Acoustic conditions	13
6.5 Position of the acoustic centre of a microphone	13
6.6 Dependence on environmental conditions	14
6.6.1 General	14
6.6.2 Static pressure.....	14
6.6.3 Temperature	14
6.6.4 Humidity	14
6.6.5 Transformation to reference environmental conditions.....	14
6.7 Considerations concerning measurement space.....	15
7 Calibration uncertainty components.....	15
7.1 General.....	15
7.2 Electrical transfer impedance	15
7.3 Deviations from ideal free-field conditions.....	15
7.4 Attenuation of sound in air.....	16
7.5 Polarizing voltage.....	16
7.6 Physical properties of air.....	16
7.7 Imperfection of theory	16
7.8 Uncertainty on free-field sensitivity level	17
Annex A (informative) Values for the position of the acoustic centre.....	19
Annex B (normative) Values of the air attenuation coefficient.....	20
B.1 General.....	20
B.2 Calculation procedure	20

Annex C (informative) Environmental influence on the sensitivity of microphones	23
C.1 General.....	23
C.2 Dependence on static pressure	23
C.3 Dependence on temperature	23
Annex D (informative) Application of time selective techniques for removal of unwanted reflections and acoustic interference between microphones	25
D.1 General.....	25
D.2 Practical considerations	25
D.2.1 Signal-to-noise ratio.....	25
D.2.2 Reflections from walls and measurement rig	25
D.3 Frequency limitations	26
D.3.1 General	26
D.3.2 Measurements based on frequency sweeps	26
D.3.3 Measurements based on pure tones	26
D.4 Generating missing portions of the frequency response previous to transforming to the time-domain.	27
D.4.1 General	27
D.4.2 Missing frequencies below the minimum measurement frequency.....	27
D.4.3 Missing frequencies above the maximum measured frequency	27
D.4.4 Filtering the extended frequency response.....	28
Bibliography	29
Figure 1 – Equivalent circuit for a receiving microphone under free-field conditions	9
Figure 2 – Equivalent circuit for a transmitting microphone under free-field conditions	10
Figure A.1 – Example of the estimated values of the acoustic centres of LS1P and LS2aP microphones given in the bibliographical references for Annex A	19
Table 1 – Uncertainty components	17
Table B.1 – Values for attenuation of sound pressure in air (in dB/m).....	22

INTERNATIONAL ELECTROTECHNICAL COMMISSION

ELECTROACOUSTICS – MEASUREMENT MICROPHONES –**Part 3: Primary method for free-field calibration of laboratory standard microphones by the reciprocity technique**

FOREWORD

- 1) The International Electrotechnical Commission (IEC) is a worldwide organization for standardization comprising all national electrotechnical committees (IEC National Committees). The object of IEC is to promote international co-operation on all questions concerning standardization in the electrical and electronic fields. To this end and in addition to other activities, IEC publishes International Standards, Technical Specifications, Technical Reports, Publicly Available Specifications (PAS) and Guides (hereafter referred to as "IEC Publication(s)"). Their preparation is entrusted to technical committees; any IEC National Committee interested in the subject dealt with may participate in this preparatory work. International, governmental and non-governmental organizations liaising with the IEC also participate in this preparation. IEC collaborates closely with the International Organization for Standardization (ISO) in accordance with conditions determined by agreement between the two organizations.
- 2) The formal decisions or agreements of IEC on technical matters express, as nearly as possible, an international consensus of opinion on the relevant subjects since each technical committee has representation from all interested IEC National Committees.
- 3) IEC Publications have the form of recommendations for international use and are accepted by IEC National Committees in that sense. While all reasonable efforts are made to ensure that the technical content of IEC Publications is accurate, IEC cannot be held responsible for the way in which they are used or for any misinterpretation by any end user.
- 4) In order to promote international uniformity, IEC National Committees undertake to apply IEC Publications transparently to the maximum extent possible in their national and regional publications. Any divergence between any IEC Publication and the corresponding national or regional publication shall be clearly indicated in the latter.
- 5) IEC itself does not provide any attestation of conformity. Independent certification bodies provide conformity assessment services and, in some areas, access to IEC marks of conformity. IEC is not responsible for any services carried out by independent certification bodies.
- 6) All users should ensure that they have the latest edition of this publication.
- 7) No liability shall attach to IEC or its directors, employees, servants or agents including individual experts and members of its technical committees and IEC National Committees for any personal injury, property damage or other damage of any nature whatsoever, whether direct or indirect, or for costs (including legal fees) and expenses arising out of the publication, use of, or reliance upon, this IEC Publication or any other IEC Publications.
- 8) Attention is drawn to the Normative references cited in this publication. Use of the referenced publications is indispensable for the correct application of this publication.
- 9) Attention is drawn to the possibility that some of the elements of this IEC Publication may be the subject of patent rights. IEC shall not be held responsible for identifying any or all such patent rights.

International Standard IEC 61094-3 has been prepared by IEC technical committee 29: Electroacoustics.

This second edition cancels and replaces the first edition published in 1995. This edition constitutes a technical revision.

This edition includes the following significant technical changes with respect to the previous edition:

- a) a new informative annex describing the use of time-selective techniques to minimize the influence of acoustic reflections from the measurement setup;
- b) provision for the calibration of microphones in driven shield configuration.

The text of this standard is based on the following documents:

CDV	Report on voting
29/873/CDV	29/892A/RVC

Full information on the voting for the approval of this standard can be found in the report on voting indicated in the above table.

This publication has been drafted in accordance with the ISO/IEC Directives, Part 2.

A list of all parts in the IEC 61094 series, published under the general title *Electroacoustics – Measurement microphones*, can be found on the IEC website.

The committee has decided that the contents of this publication will remain unchanged until the stability date indicated on the IEC website under "<http://webstore.iec.ch>" in the data related to the specific publication. At this date, the publication will be

- reconfirmed,
- withdrawn,
- replaced by a revised edition, or
- amended.

iTeh STANDARD PREVIEW

(standards.iteh.ai)

IMPORTANT – The 'colour inside' logo on the cover page of this publication indicates that it contains colours which are considered to be useful for the correct understanding of its contents. Users should therefore print this document using a colour printer.

<https://standards.iteh.ai/catalog/standards/sist/313e2841-8531-4100-a460-f0992afd66b5/sist-en-61094-3-2016>

ELECTROACOUSTICS – MEASUREMENT MICROPHONES –

Part 3: Primary method for free-field calibration of laboratory standard microphones by the reciprocity technique

1 Scope

This part of IEC 61094

- specifies a primary method of determining the complex free-field sensitivity of laboratory standard microphones so as to establish a reproducible and accurate basis for the measurement of sound pressure under free-field conditions,
- is applicable to laboratory standard microphones meeting the requirements of IEC 61094-1,
- is intended for use by laboratories with highly experienced staff and specialized equipment.

NOTE The calibration principle described in this part of IEC 61094 is also applicable to working standard microphones, preferably used without their protection grid.

2 Normative references

The following documents, in whole or in part, are normatively referenced in this document and are indispensable for its application. For dated references, only the edition cited applies. For undated references, the latest edition of the referenced document (including any amendments) applies.

IEC 61094-1:2000, *Measurement microphones – Part 1: Specifications for laboratory standard microphones*

IEC 61094-2:2009, *Electroacoustics – Measurement microphones – Part 2: Primary method for pressure calibration of laboratory standard microphones by the reciprocity technique*

IEC TS 61094-7:2006, *Measurement microphones – Part 7: Values for the difference between free-field and pressure sensitivity levels of laboratory standard microphones*

ISO 9613-1, *Acoustics – Attenuation of sound during propagation outdoors – Part 1: Calculation of the absorption of sound by the atmosphere*

ISO/IEC Guide 98-3, *Uncertainty of measurement – Part 3: Guide to the expression of uncertainty in measurement* (GUM:1995)

3 Terms and definitions

For the purposes of this document, the terms and definitions given in IEC 61094-1, IEC 61094-2, ISO/IEC Guide 98-3 and the following apply.

3.1 phase

<free-field sensitivity of a microphone> phase angle between the open-circuit voltage and the sound pressure that would exist at the position of the acoustic centre of the microphone in the absence of the microphone, for a sinusoidal plane progressive wave of given frequency and direction of sound incidence, and for given environmental conditions