



SLOVENSKI STANDARD
SIST EN 60601-2-33:2010/A1:2015
01-september-2015

Medicinska električna oprema - 2-33. del: Posebne zahteve za varnost opreme za magnetno resonanco za medicinsko diagnostiko - Dopnilo A1

Medical electrical equipment - Part 2-33: Particular requirements for the basic safety and essential performance of magnetic resonance equipment for medical diagnosis

Medizinische elektrische Geräte - Teil 2-33: Besondere Festlegungen für die Sicherheit von Magnetresonanzgeräten für die medizinische Diagnostik

Appareils électromédicaux - Partie 2-33: Exigences particulières pour la sécurité de base et les performances essentielles des appareils à résonance magnétique utilisés pour le diagnostic médical

<https://standards.iteh.ai/catalog/standards/sist/b3e5d959-c2f9-4de5-8150-f1b2dfb1fad5/sist-en-60601-2-33-2010-a1-2015>

Ta slovenski standard je istoveten z: EN 60601-2-33:2010/A1:2015

ICS:

11.040.55 Diagnostična oprema Diagnostic equipment

SIST EN 60601-2-33:2010/A1:2015 en

iTeh STANDARD PREVIEW
(standards.iteh.ai)

SIST EN 60601-2-33:2010/A1:2015

<https://standards.iteh.ai/catalog/standards/sist/b3e5d959-c2f9-4de5-8150-f1b2dfb1fad5/sist-en-60601-2-33-2010-a1-2015>

EUROPEAN STANDARD

EN 60601-2-33:2010/A1

NORME EUROPÉENNE

EUROPÄISCHE NORM

May 2015

ICS 11.040.55

English Version

Medical electrical equipment - Part 2-33: Particular requirements
for the basic safety and essential performance of magnetic
resonance equipment for medical diagnosis
(IEC 60601-2-33:2010/A1:2013)

Appareils électromédicaux - Partie 2-33: Exigences
particulières pour la sécurité de base et les performances
essentielles des appareils à résonance magnétique utilisés
pour le diagnostic médical
(IEC 60601-2-33:2010/A1:2013)

Medizinische elektrische Geräte - Teil 2-33: Besondere
Festlegungen für die Sicherheit von
Magnetresonanzgeräten für die medizinische Diagnostik
(IEC 60601-2-33:2010/A1:2013)

This amendment A1 modifies the European Standard EN 60601-2-33:2010; it was approved by CENELEC on 2015-04-14. CENELEC members are bound to comply with the CEN/CENELEC Internal Regulations which stipulate the conditions for giving this amendment the status of a national standard without any alteration.

Up-to-date lists and bibliographical references concerning such national standards may be obtained on application to the CEN-CENELEC Management Centre or to any CENELEC member.

This amendment exists in three official versions (English, French, German). A version in any other language made by translation under the responsibility of a CENELEC member into its own language and notified to the CEN-CENELEC Management Centre has the same status as the official versions.

CENELEC members are the national electrotechnical committees of Austria, Belgium, Bulgaria, Croatia, Cyprus, the Czech Republic, Denmark, Estonia, Finland, Former Yugoslav Republic of Macedonia, France, Germany, Greece, Hungary, Iceland, Ireland, Italy, Latvia, Lithuania, Luxembourg, Malta, the Netherlands, Norway, Poland, Portugal, Romania, Slovakia, Slovenia, Spain, Sweden, Switzerland, Turkey and the United Kingdom.



European Committee for Electrotechnical Standardization
Comité Européen de Normalisation Electrotechnique
Europäisches Komitee für Elektrotechnische Normung

CEN-CENELEC Management Centre: Avenue Marnix 17, B-1000 Brussels

EN 60601-2-33:2010/A1:2015**Foreword**

The text of document 62B/884/CDV, future IEC 60601-2-33:2010/A1, prepared by SC 62B "Diagnostic imaging equipment" of IEC/TC 62 "Electrical equipment in medical practice" was submitted to the IEC-CENELEC parallel vote and approved by CENELEC as EN 60601-2-33:2010/A1:2015.

The following dates are fixed:

- latest date by which the document has to be implemented at national level by publication of an identical national standard or by endorsement (dop) 2016-01-14
- latest date by which the national standards conflicting with the document have to be withdrawn (dow) 2018-04-14

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. CENELEC [and/or CEN] shall not be held responsible for identifying any or all such patent rights.

This document has been prepared under a mandate given to CENELEC by the European Commission and the European Free Trade Association, and supports essential requirements of EU Directive.

For the relationship with EU Directive 93/42/EEC, see informative Annex ZZ, included in EN 60601-2-33:2010/A1:2011.

iTeh STANDARD PREVIEW
Endorsement notice
(standards.iteh.ai)

The text of the International Standard IEC 60601-2-33:2010/A1:2013 was approved by CENELEC as a European Standard without any modification.

In the Bibliography of EN 60601-2-33:2010, the following note has to be added for the standard indicated:

ISO 7010:2011 NOTE Harmonized as EN ISO 7010:2012 (not modified).

Annex ZA (normative)

Normative references to international publications with their corresponding European publications

The following documents, in whole or in part, are normatively referenced in this document and are indispensable for its application. For dated references, only the edition cited applies. For undated references, the latest edition of the referenced document (including any amendments) applies.

NOTE 1 When an International Publication has been modified by common modifications, indicated by (mod), the relevant EN/HD applies.

NOTE 2 Up-to-date information on the latest versions of the European Standards listed in this annex is available here: www.cenelec.eu

Replacements in Annex ZA of EN 60601-2-33:2010:

<u>Publication</u>	<u>Year</u>	<u>Title</u>	<u>EN/HD</u>	<u>Year</u>
Replace the existing reference to IEC 60601-1:2005 by the following:				
IEC 60601-1	2005	Medical electrical equipment - Part 1: General requirements for basic safety and essential performance	EN 60601-1 + corr. March	2006 2010
+A1	2012		+A1 +A1/AC +A12	2013 2014 2014

Replace the existing reference to NEMA MS 4:2006 by the following:

NEMA MS 4	2010	Acoustic noise measurement procedure for - diagnostic Magnetic Resonance Imaging (MRI) devices		-
-----------	------	--	--	---

iTeh STANDARD PREVIEW
(standards.iteh.ai)

SIST EN 60601-2-33:2010/A1:2015

<https://standards.iteh.ai/catalog/standards/sist/b3e5d959-c2f9-4de5-8150-f1b2dfb1fad5/sist-en-60601-2-33-2010-a1-2015>



IEC 60601-2-33

Edition 3.0 2013-04

INTERNATIONAL STANDARD

NORME INTERNATIONALE

AMENDMENT 1
AMENDEMENT 1

Medical electrical equipment –
Part 2-33: Particular requirements for the basic safety and essential performance
of magnetic resonance equipment for medical diagnosis

Appareils électromédicaux –
Partie 2-33: Exigences particulières pour la sécurité de base et les performances
essentiels des appareils à résonance magnétique utilisés pour le diagnostic
médical

INTERNATIONAL
ELECTROTECHNICAL
COMMISSION

COMMISSION
ELECTROTECHNIQUE
INTERNATIONALE

PRICE CODE
CODE PRIX

G

ICS 11.040.55

ISBN 978-2-83220-751-2

Warning! Make sure that you obtained this publication from an authorized distributor.
Attention! Veuillez vous assurer que vous avez obtenu cette publication via un distributeur agréé.

FOREWORD

This amendment has been prepared by subcommittee 62B: Diagnostic imaging equipment, of IEC technical committee 62: Electrical equipment in medical practice.

The text of this amendment is based on the following documents:

CDV	Report on voting
62B/884/CDV	62B/904/RVC

Full information on the voting for the approval of this amendment can be found in the report on voting indicated in the above table.

The committee has decided that the contents of this amendment and the base publication will remain unchanged until the stability date indicated on the IEC web site under "<http://webstore.iec.ch>" in the data related to the specific publication. At this date, the publication will be

- reconfirmed,
- withdrawn,
- replaced by a revised edition, or
- amended.

iTeh STANDARD PREVIEW
(standards.iteh.ai)

~~SIST EN 60601-2-33:2010/A1:2015~~

<https://standards.iteh.ai/catalog/standards/sist/b3e5d959-c2f9-4de5-8150-f1b2dfb1fad5/sist-en-60601-2-33-2010-a1-2015>

INTRODUCTION

This amendment has been published to adapt IEC 60601-2-33:2010 to the technical corrections introduced by Amendment 1 (2012) to IEC 60601-1:2005.

201.1 Scope, object and related standards

In the footnote, replace IEC 60601-1:2005 by IEC 60601-1:2005+A1:2012.

201.1.1 Scope

Replace the existing fourth paragraph with the following:

The standard does not formulate specific requirements for MR EQUIPMENT or MR SYSTEMS used in INTERVENTIONAL MR EXAMINATIONS.

201.1.3 Collateral standards

Replace the second paragraph with the following:

IEC 60601-1-2:2007 applies as modified in Clause 202. IEC 60601-1-3, IEC 60601-1-10, IEC 60601-1-11 and IEC 60601-1-12 do not apply. All other published collateral standards in the IEC 60601-1 series apply as published.

201.2 Normative references

Replace the existing reference to IEC 60601-1:2005 by the following:

IEC 60601-1:2005, *Medical electrical equipment – Part 1: General requirements for basic safety and essential performance*
Amendment 1:2012

Replace the existing reference to NEMA MS 4:2006 by the following:

NEMA MS 4:2010, *Acoustic noise measurement procedure for diagnostic magnetic resonance imaging (MRI) devices*

201.3 Terms and definitions

Replace the existing reference to IEC 60601-1:2005 by IEC 60601-1:2005+A1:2012.