
**Wheat flour — Physical characteristics of
doughs —**

Part 2:

Determination of rheological properties using
an extensograph

iTeh STANDARD PREVIEW
(standards.iteh.ai)

Farines de blé tendre — Caractéristiques physiques des pâtes —

*Partie 2: Détermination des caractéristiques rhéologiques au moyen de
l'extensographe*

ISO 5530-2:1997

<https://standards.iteh.ai/catalog/standards/sist/92fa5fd8-15ef-45c5-a0ea-b63e675da2e2/iso-5530-2-1997>



Foreword

ISO (the International Organization for Standardization) is a worldwide federation of national standards bodies (ISO member bodies). The work of preparing International Standards is normally carried out through ISO technical committees. Each member body interested in a subject for which a technical committee has been established has the right to be represented on that committee. International organizations, governmental and non-governmental, in liaison with ISO, also take part in the work. ISO collaborates closely with the International Electrotechnical Commission (IEC) on all matters of electrotechnical standardization.

Draft International Standards adopted by the technical committees are circulated to the member bodies for voting. Publication as an International Standard requires approval by at least 75 % of the member bodies casting a vote.

International Standard ISO 5530-2 was prepared by Technical Committee ISO/TC 34, *Agricultural food products*, Subcommittee SC 4, *Cereals and pulses*.

This part of ISO 5530 is based on Standard No. 114 of the International Association for Cereal Science and Technology (ICC).

This second edition cancels and replaces the first edition (ISO 5530-2:1988), which has been technically revised.

ISO 5530 consists of the following parts, under the general title *Wheat flour — Physical characteristics of doughs*:

- *Part 1: Determination of water absorption and rheological properties using a farinograph*
- *Part 2: Determination of rheological properties using an extensograph*
- *Part 3: Determination of water absorption and rheological properties using a valorigraph*
- *Part 4: Determination of rheological properties using an alveograph*

Annexes A to C of this part of ISO 5530 are for information only.

© ISO 1997

All rights reserved. Unless otherwise specified, no part of this publication may be reproduced or utilized in any form or by any means, electronic or mechanical, including photocopying and microfilm, without permission in writing from the publisher.

International Organization for Standardization
Case postale 56 • CH-1211 Genève 20 • Switzerland
Internet central@iso.ch
X.400 c=ch; a=400net; p=iso; o=isocs; s=central

Printed in Switzerland

Wheat flour — Physical characteristics of doughs —

Part 2:

Determination of rheological properties using an extensograph

1 Scope

This part of ISO 5530 specifies a method, using an extensograph, for the determination of the rheological properties of a wheat flour dough in an extension test. The recorded load-extension curve is used to assess general quality of flour and its response to improving agents.

The method is applicable to flour from wheat (*Triticum aestivum* L.).

iTeh STANDARD PREVIEW
(standards.iteh.ai)

2 Normative references

The following standards contain provisions which, through reference in this text, constitute provisions of this part of ISO 5530. At the time of the publication, the editions indicated were valid. All standards are subject to revision, and parties to agreements based on this part of ISO 5530 are encouraged to investigate the possibility of applying the most recent editions of the standards indicated below. Members of IEC and ISO maintain registers of currently valid International Standards.

ISO 712:—¹⁾, *Cereals and cereal products — Determination of moisture content — Routine reference method.*

ISO 5530-1:1997, *Wheat flour — Physical characteristics of doughs — Part 1: Determination of water absorption and rheological properties using a farinograph.*

3 Definitions

For the purposes of this part of ISO 5530, the following terms and definitions apply.

3.1 extensograph water absorption

Volume of water required to produce a dough with a consistency of 500 farinograph units (FU) after 5 min mixing, under the operating conditions specified in this part of ISO 5530.

NOTE — It is expressed in millilitres per 100 g of flour at 14,0 % (*m/m*) moisture content.

3.2 stretching characteristics (of dough)

Resistance of dough to extension and the extent to which it can be stretched until breaking, under the operating conditions specified in this part of ISO 5530.

¹⁾ To be published. (Revision of ISO 712:1985)

NOTE 1 The resistance is expressed in arbitrary units (extensograph units, EU).

NOTE 2 The extent of stretching is also expressed in arbitrary units (millimetres transposition of the band of paper in the recorder).

4 Principle

Preparation of a dough from flour, water and salt in a farinograph under specified conditions. Moulding of a test piece on the balling unit and moulder of the extensograph into a standard shape. After a fixed period of time, stretching the test piece and recording the force required. Immediately after the first stretching, two repetitions, one after the other, with the same test piece, of moulding, rest period and stretching.

The size and shape of the curves obtained are a guide to the physical properties of the dough influencing the baking characteristics of the flour.

5 Reagents

Use only reagents of recognized analytical grade, and distilled or demineralized water, or water of equivalent purity.

5.1 Sodium chloride

6 Apparatus

Usual laboratory apparatus and, in particular, the following (standards.iteh.ai)

6.1 Extensograph²⁾, with a thermostat consisting of a constant temperature water bath (see annex A), with the following operating characteristics:

- rotational frequency of balling unit: $(83 \pm 3) \text{ min}^{-1}$ (rev/min);
- rotational frequency of moulder: $(15 \pm 1) \text{ min}^{-1}$ (rev/min);
- hook speed: $(1,45 \pm 0,05) \text{ cm/s}$;
- chart speed : $(0,65 \pm 0,01) \text{ cm/s}$;
- force exerted per extensograph unit: $(12,3 \pm 0,3) \text{ mN/EU}$ [$(1,25 \pm 0,03) \text{ gf/EU}$].

NOTE — Some instruments have a different calibration for force/unit deflection. The procedure described may be used with such instruments, but it is necessary for the different calibration to be taken into account when comparing the results with instruments calibrated as above.

6.2 Farinograph, connected to a similar thermostat as the extensograph, with the operating characteristics specified in ISO 5530-1, and a burette as specified in ISO 5530-1.

6.3 Balance, capable of weighing to the nearest $\pm 0,1 \text{ g}$.

²⁾ This part of ISO 5530 has been drawn up on the basis of the Brabender Extensograph. This information is given for the convenience of users of this part of ISO 5530 and does not constitute an endorsement by ISO of this apparatus. Other equipment may be used if it can be shown to give comparable results.

6.4 Spatula, made of soft plastic.

6.5 Conical flask, of 250 ml capacity.

7 Sampling

Sampling is not part of the method specified in this part of ISO 5530. A recommended sampling method is given in ISO 13690.

It is important that the laboratory receive a sample which is truly representative and has not been damaged or changed during transport and storage.

8 Procedure

8.1 Determination of the moisture content of the flour

Determine the moisture content of the flour using the method specified in ISO 712.

8.2 Preparation of apparatus

8.2.1 Turn on the thermostat of the farinograph (6.2) and circulate the water until the required temperatures are reached, prior to using the instrument. Before and during use, check the temperatures of

- the thermostats,
- the mixing bowl of the farinograph, in the hole provided for this purpose, and of
- the extensograph cabinet.

iTeh STANDARD PREVIEW
(standards.iteh.ai)

All temperatures shall be $(30 \pm 0,2)$ °C.

<https://standards.iteh.ai/catalog/standards/sist/92fa5fd8-15ef-45c5-a0ea-b63e675da2e2/iso-5530-2-1997>

8.2.2 Adjust the arm of the pen of the extensograph so as to obtain zero reading when a cradle with both its clamps plus a 150 g weight is placed in position.

8.2.3 Pour some water into the trough of each cradle support, and place the supports, cradles and clamps in the cabinet at least 15 min before use.

8.2.4 Uncouple the mixer of the farinograph from the driving shaft and adjust the position of the counterweight(s) so as to obtain zero deflection of the pointer with the motor running at the specified rotational frequency (see ISO 5530-1:1997, 6.1). Switch off the motor and then couple the mixer.

Lubricate the mixer with a drop of water between the back-plate and each of the blades. Check that the deflection of the pointer is within the range (0 ± 5) FU with the mixing blades rotating at the specified rotational frequency in the empty, clean bowl. If the deflection exceeds 5 FU, clean the mixer more thoroughly or eliminate other causes of friction.

Adjust the arm of the pen so as to obtain identical readings from the pointer and the recording pen.

Adjust the damper so that, with the motor running, the time required for the pointer to go from 1 000 FU to 100 FU is $(1,0 \pm 0,2)$ s.

8.2.5 Fill the burette of the farinograph, including the tip, with water at a temperature of $(30 \pm 0,5)$ °C.

8.3 Test portion

If necessary, bring the flour to a temperature of (25 ± 5) °C.

Weigh, to the nearest 0,1 g, the equivalent of 300 g of flour having a moisture content of 14 % (m/m). Let this mass, in grams, be m ; see table 1 in ISO 5530-1:1997, for m as a function of moisture content.

Place the flour into the farinograph mixer. Cover the mixer and keep it covered until the end of mixing (8.4.2), except for the shortest possible time when water has to be added and the dough scraped down (see ISO 5530-1:1997, A.2.2).

8.4 Preparation of the dough

8.4.1 Place $(6,0 \pm 0,1)$ g of the sodium chloride (5.1) in the conical flask (6.5). Run in approximately 135 ml of water from the burette and dissolve the salt. For flours having a low water absorption, use a smaller volume of water.

8.4.2 Mix in the farinograph mixer at the specified rotational frequency (see ISO 5530-1:1997, 6.1) for 1 min or slightly longer. Pour the salt solution (8.4.1) through a funnel into the centre hole of the bottom part of the lid, when a whole-minute line on the recorder paper passes by the pen.

In order to reduce the waiting time, the recorder paper may be moved forward during mixing of the flour. Do not move it backwards.

NOTE 1 With older farinographs, the bowl of which is covered by a single plate (see ISO 5530-1:1997, A.2.2), the salt solution is poured into the right-hand front corner of the bowl.

Add from the burette, into the right-hand front corner of the mixer, a volume of water approximately equal to that expected to obtain a consistency of 500 FU after mixing for 5 min. When the dough forms, scrape down the sides of the bowl with the spatula (6.4), adding any adhering particles to the dough without stopping the mixer. If the consistency is too high, add a little more water to obtain a consistency of 500 FU after mixing for 5 min. Stop mixing and clean the mixer.

NOTE 2 If the first dough meets the requirements of 8.4.3, test pieces from it can be moulded (8.4.4) and stretched (8.5.1).

8.4.3 Make further mixings as necessary, until a dough is obtained

- to which the salt solution and water have been added within 25 s,
- the consistency of which, measured at the centre of the curve after mixing for 5 min, is between 480 FU and 520 FU, and
- the mixing time of which is $(5 \pm 0,1)$ min.

Stop mixing after this time.

8.4.4 Take a support with two cradles from the cabinet of the extensograph (6.1); remove their clamps. Remove the dough from the mixer. Weigh a $(150 \pm 0,5)$ g test piece. Place it in the balling unit and perform 20 revolutions of the plate. Remove the dough from the balling unit and pass it once through the moulder, ensuring that the test piece enters the back centrally, base first. Roll the test piece off the moulder into the centre of a cradle, and clamp it. Set the timer for 45 min. Weigh a second test piece, and ball, mould and clamp it in the same way. Place the support with two cradles and test pieces into the cabinet.

NOTE 1 Very sticky doughs may be dusted lightly with rice flour or starch before being put into the moulder.

NOTE 2 In the case of doughs showing substantial elastic recovery, the clamps should be held down for a few seconds to ensure that they fix the dough properly.

Clean the farinograph mixer.

8.5 Determination

8.5.1 Exactly 45 min after clamping the first test piece, place the first cradle in the balance arm of the extensograph (6.1); the bridge between the two halves of the cradle shall be on the left-hand side so as not to be touched by the stretching hook when travelling. Adjust the pen to zero force. Immediately start the stretching hook. Observe the test piece (see note in 9.3). After rupture of the piece, remove the cradle.

NOTE — In recent extensographs, the hook automatically returns to its upper position. With older models it is necessary, by means of a switch, to stop the hook after breaking of the test piece, and to start it returning to its upper position.

8.5.2 Collect the dough from the cradle and the hook. Repeat the balling and moulding operations as described in 8.4.4 on this test piece. Reset the timer for 45 min.

8.5.3 Turn the recorder paper back to the same starting position as for the first test piece. Repeat the stretching operation (8.5.1) on the second test piece. Collect the dough from the cradle and the hook. Repeat the balling and moulding operations (8.4.4) on the second test piece.

8.5.4 Repeat the stretching, balling, and moulding operations described in 8.5.1 to 8.5.3, returning the moulded test pieces to the cabinet. These operations take place after slightly more than 90 min from the end of mixing.

8.5.5 Repeat the operation described in 8.5.1, stretching both test pieces in turn. This takes place after slightly more than 135 min from the end of mixing.

8.5.6 In order to carry out quick and time saving measurements, another procedure may be suitable. The difference with the standard procedure is in the rest periods. Stretching after 45 min, 90 min and 135 min after mixing are replaced by stretching after 30 min, 60 min and 90 min after mixing. The shape and the size of the curves obtained differ from those of the standard extensograms. When the quick procedure is used, it is necessary to state this in the test report.

ISO 5530-2:1997

<https://standards.iteh.ai/catalog/standards/sist/92fa5fd8-15ef-45c5-a0ea-b63e675da2e2/iso-5530-2-1997>

9 Expression of results

NOTE — To facilitate the calculations, a computer may be used. The extensograph has to be modified by adding an electrical output for transferring the data to the computer. With the appropriate software, the computer evaluates the diagram according to 9.2 to 9.5 and documents the diagram and the results.

9.1 Water absorption

Calculate the extensograph water absorption, expressed in millilitres per 100 g of flour at 14 % (*m/m*) moisture content, as specified in ISO 5530-1:1997, 9.1, for the 300 g mixer.

9.2 Resistance to stretching

9.2.1 Maximum resistance

Take as the maximum resistance to stretching, R_m , the mean of the maximum heights of the extensograph curves (see figure 1) from the two test pieces, provided that the difference between them does not exceed 15 % of their mean value.

Report each of the mean values of $R_{m,45}$, $R_{m,90}$ and $R_{m,135}$ to nearest 5 EU.

9.2.2 Resistance at constant deformation

Some workers prefer to measure the height of the curve at a fixed extension of the test piece, usually corresponding to 50 mm transposition of the recorder paper. The extension is measured from the moment that the hook touches the test piece; i.e. when the force is suddenly different from zero.

Take as the result of the resistance to stretching at constant deformation, R_{50} , the mean of the heights of the extensograph curves after 50 mm transposition of the recorder paper (see figure 1) from the two test pieces, provided that the difference between them does not exceed 15 % of their mean value.

Report each of the mean values of $R_{50,45}$, $R_{50,90}$ and $R_{50,135}$ to the nearest 5 EU.

NOTE — Owing to the greater depression of the cradle, a more resistant test piece will be extended less at 50 mm transposition of the recorder paper than a less resistant test piece. It is possible, by means of a suitable template, to read the resistances of all test pieces at the same net extension. If such a template is used, it is necessary to mention this in the test report.

9.3 Extensibility

The extensibility, E , is the distance travelled by the recorder paper from the moment that the hook touches the test piece until rupture of (one of the strings of) the test pieces. Rupture is indicated on the extensograph curve either by a smooth fall of the curve almost to zero force, or by a sharp break in the curve (see figure 1).

NOTE — Beyond the breaking point, the course, of the recording depends on the inertia of the lever system and on the time interval between the breaking of the two strings of the test piece. For measurement of the extensibility, the curve is supposed to proceed, from the breaking point, along a circular ordinate line (dashed line in figure 1) down to zero force. To properly identify the breaking point on the curve, it is necessary to observe the test piece when breaking.

Take as the result of the extensibility the mean distance on the extensograph curves from the two test pieces, provided that the difference between them does not exceed 9 % of their mean value.

Report each of the mean values of E_{45} , E_{90} and E_{135} to the nearest millimetre.




Figure 1 — Representative extensogram showing the commonly measured indices

9.4 Energy

The energy is defined as the area under the recorded curve. The energy describes the work applied when stretching a dough sample. The area is measured by a planimeter and reported in square centimetres.

9.5 Ratio (R/E)

The ratio R/E is the quotient of the resistance R_m or R_{50} and the extensibility. The ratio is an additional factor in the review of the dough behaviour.

E

10 Precision

Details of an interlaboratory test on the precision of the method are summarized in annex B. The values derived from this interlaboratory test may not be applicable to concentration ranges and matrices other than those given.

11 Test report

The test report shall specify:

- all information necessary for the complete identification of the sample;
- the sampling method used, if known;
- the test method used, with reference to this part of ISO 5530;
- all operating details not specified in this part of ISO 5530, or regarded as optional, together with details of any incidents which may have influenced the test result(s);
- the test result(s) obtained;
- if the repeatability has been checked, the final quoted result obtained.

iTeh STANDARD PREVIEW
(standards.iteh.ai)

[ISO 5530-2:1997](https://standards.iteh.ai/catalog/standards/sist/92fa5fd8-15ef-45c5-a0ea-b63e675da2e2/iso-5530-2-1997)

<https://standards.iteh.ai/catalog/standards/sist/92fa5fd8-15ef-45c5-a0ea-b63e675da2e2/iso-5530-2-1997>