

SLOVENSKI STANDARD
SIST EN 61010-2-091:2012/AC:2013
01-marec-2013

Varnostne zahteve za električno opremo za meritve, nadzorovanje in laboratorijsko uporabo - 2-091. del: Posebne zahteve za rentgenske sisteme v ohišjih

Safety requirements for electrical equipment for measurement, control and laboratory use -- Part 2-091: Particular requirements for cabinet X-ray systems

Sicherheitsbestimmungen für elektrische Mess-, Steuer-, Regel- und Laborgeräte -- Teil 2-091: Besondere Anforderungen für Röntgengeräteschränke

Règles de sécurité pour appareils électriques de mesurage, de régulation et de laboratoire -- Partie 2-091: Exigences particulières pour les équipements à rayons X montés en armoire

<https://standards.iteh.ai/catalog/standards/sist/a13380e3-b1f1-437b-951c-e0fd19054e03/sist-en-61010-2-091-2012-ac-2013>

Ta slovenski standard je istoveten z: EN 61010-2-091:2012/AC:2013

ICS:

19.080	Električno in elektronsko preskušanje	Electrical and electronic testing
71.040.10	Kemijski laboratoriji. Laboratorijska oprema	Chemical laboratories. Laboratory equipment

SIST EN 61010-2-091:2012/AC:2013 en

iTeh STANDARD PREVIEW
(standards.iteh.ai)

[SIST EN 61010-2-091:2012/AC:2013](https://standards.iteh.ai/catalog/standards/sist/a13380e3-b1f1-437b-951c-e0fd9054e03/sist-en-61010-2-091-2012-ac-2013)

<https://standards.iteh.ai/catalog/standards/sist/a13380e3-b1f1-437b-951c-e0fd9054e03/sist-en-61010-2-091-2012-ac-2013>



Corrigendum to EN 61010-2-091:2012

English version

Following BT Decision D143/C081, add the following Annex ZB:

"

Annex ZB

(informative)

A-deviations

iTeh STANDARD PREVIEW

(standards.iteh.ai)

A-deviation: National deviation due to regulations, the alteration of which is for the time being outside the competence of the CEN/CENELEC national member.

[SIST EN 61010-2-091:2012/AC:2013](#)

This European Standard falls under Directive LVD (2006/95/EC) [80e3-b1f1-437b-951c-e0fd9054e03/sist-en-61010-2-091-2012-ac-2013](#)

NOTE (from CEN/CENELEC IR Part 2:2006, 2.17) Where standards fall under EC Directives, it is the view of the Commission of the European Communities (OJ No C 59; 1982-03-09) that the effect of the decision of the Court of Justice in case 815/79 Cremonini/Vrankovich (European Court Reports 1980, p. 3583) is that compliance with A-deviations is no longer mandatory and that the free movement of products complying with such a standard should not be restricted except under the safeguard procedure provided for in the relevant Directive.

A-deviations in an EFTA-country are valid instead of the relevant provisions of the European Standard in that country until they have been removed.

Country	Divergence
France	<p>National legislation</p> <p>Section 1 of the Decree of 2 September 1991, determining technical requirements to be met by X-ray generators, used in industrial radiology, states:</p> <p><i>"The devises of X-ray generators used in industrial radiology, including accessories, put into service after the publication of this Decree shall comply with the rules set out the date of commissioning by the approved French standard NF C 74-100 concerning Radiology equipment. X-ray apparatus - Construction and tests. or any equivalent standard of a Member State of the European Economic Community "</i></p>

"