



SLOVENSKI STANDARD

SIST EN 13544-1:2007

01-oktober-2007

BUKca Yý U
SIST EN 13544-1:2002
SIST EN 13544-1:2002/A1:2005
SIST EN 13544-1:2002/AC:2005

Dihalna oprema za zdravljenje - 1. del: Razprševalni sistemi in njihovi sestavni deli

Respiratory therapy equipment - Part 1: Nebulizing systems and their components

Atemtherapiegeräte - Teil 1: Verneblersysteme und deren Bauteile

Matériel respiratoire thérapeutique - Partie 1: Systèmes de nébulisation et leurs composants

<https://standards.iteh.ai/catalog/standards/sist/2da75a13-1238-4d99-8101-79fce8146958/sist-en-13544-1-2007>

Ta slovenski standard je istoveten z: EN 13544-1:2007

ICS:

11.040.10	Anestezijska, respiratorna in reanimacijska oprema	Anaesthetic, respiratory and reanimation equipment
-----------	--	--

SIST EN 13544-1:2007

en,fr,de

iTeh STANDARD PREVIEW **(standards.iteh.ai)**

SIST EN 13544-1:2007

<https://standards.iteh.ai/catalog/standards/sist/2da75a13-1238-4d99-8101-79fce8146958/sist-en-13544-1-2007>

English Version

Respiratory therapy equipment - Part 1: Nebulizing systems and their components

Matériel respiratoire thérapeutique - Partie 1: Systèmes de nébulisation et leurs composants

Atemtherapiegeräte - Teil 1: Verneblersysteme und deren Bauteile

This European Standard was approved by CEN on 22 March 2007.

CEN members are bound to comply with the CEN/CENELEC Internal Regulations which stipulate the conditions for giving this European Standard the status of a national standard without any alteration. Up-to-date lists and bibliographical references concerning such national standards may be obtained on application to the CEN Management Centre or to any CEN member.

This European Standard exists in three official versions (English, French, German). A version in any other language made by translation under the responsibility of a CEN member into its own language and notified to the CEN Management Centre has the same status as the official versions.

CEN members are the national standards bodies of Austria, Belgium, Bulgaria, Cyprus, Czech Republic, Denmark, Estonia, Finland, France, Germany, Greece, Hungary, Iceland, Ireland, Italy, Latvia, Lithuania, Luxembourg, Malta, Netherlands, Norway, Poland, Portugal, Romania, Slovakia, Slovenia, Spain, Sweden, Switzerland and United Kingdom.

SIST EN 13544-1:2007

<https://standards.iteh.ai/catalog/standards/sist/2da75a13-1238-4d99-8101-79fce8146958/sist-en-13544-1-2007>



EUROPEAN COMMITTEE FOR STANDARDIZATION
COMITÉ EUROPÉEN DE NORMALISATION
EUROPÄISCHES KOMITEE FÜR NORMUNG

Management Centre: rue de Stassart, 36 B-1050 Brussels

Contents

Page

Foreword.....	5
Introduction	6
1 R) Scope	7
2 Normative references	7
3 Terms and definitions	9
4 General requirements and general requirements for test	10
4.1 Modifications to Clause 3 of EN 60601-1:1990	10
4.2 Clause 4 of EN 60601-1:1990	10
4.3 Alternative type-test methods	11
5 Classification.....	11
6 Identification, marking and documents.....	11
6.1 Marking on the outside of equipment or equipment parts	11
6.3 Marking of controls and instruments	12
6.4 Symbols	12
6.8.2 Instructions for use	12
6.8.3 Technical description	14
7 Power input	14
8 Basic safety categories	14
9 Removable protective means	14
10 Environmental conditions	15
11 Not used.....	15
12 Not used.....	15
13 General.....	15
14 Requirements related to classification.....	15
15 Limitation of voltage and/or energy	15
16 Enclosures and protective covers	15
17 Separation	15
18 Protective earthing, functional earthing and potential equalization	15
19 Continuous leakage currents and patient auxiliary currents	15
20 Dielectric strength	16
21 Mechanical strength	16
22 Moving parts.....	16
23 Surfaces, corners and edges.....	16
24 Stability in normal use	16
25 Expelled parts	16
26 Vibration and noise.....	16
27 Pneumatic and hydraulic power.....	17

28	Suspended masses	17
29	X-radiation	17
30	Alpha, beta, gamma, neutron radiation and other particle radiation	17
31	Microwave radiation	17
32	Light radiation (including lasers)	17
33	Infra-red radiation	17
34	Ultra-violet radiation	17
35	Acoustical energy (including ultra-sonics)	17
36	Electromagnetic compatibility	18
37	R Locations and basic requirements	18
38	R Marking, accompanying documents	18
39	R Common requirements for category AP and category APG equipment	18
40	R Requirements and tests for Category AP equipment, parts and components thereof	18
41	R Requirements and tests for Category APG equipment, parts and components thereof	18
42	Excessive temperatures	18
43	R) Fire prevention	18
44	Overflow, spillage, leakage, humidity, ingress of liquids, cleaning, sterilization and disinfection	19
45	Pressure vessels and parts subject to pressure	19
46	Human errors	19
47	Electrostatic charges	19
48	Biocompatibility	20
49	Interruption of the power supply	20
50	Accuracy of operating data	20
51	Protection against hazardous output	20
52	Abnormal operation and fault conditions	21
53	Environmental tests	21
54	General	21
55	Enclosures and covers	21
56	Components and general assembly	21
57	Mains parts, components and layout	23
58	Protective earthing - Terminals and connections	23
59	Construction and layout	23
	Annex A A (informative) Rationale	24
	Annex B B (informative) Diameters of the particles depositable fraction	27
	Annex C C (normative) Test methods for the aerosol output rate, the aerosol output and for particle sizing	28
CC.1	Method of test for the aerosol output rate	28
CC.1.1	Test conditions	28
CC.1.2	R) Principle of test	28
CC.1.3	Test equipment	28

Annex D D (normative) Mass balance checks on cascade impactor tests	36
DD.1 Aerosol output rate and aerosol output tests:	36
DD.2 Particle sizing test	36
Annex E E (informative) Environmental aspects.....	37
Annex ZA (informative) Relationship between this European Standard and the Essential Requirements of EU Directive 93/42 EEC on medical devices	39
Bibliography	41

iTeh STANDARD PREVIEW (standards.iteh.ai)

SIST EN 13544-1:2007

<https://standards.iteh.ai/catalog/standards/sist/2da75a13-1238-4d99-8101-79fce8146958/sist-en-13544-1-2007>

Foreword

This document (EN 13544-1:2007) has been prepared by Technical Committee CEN/TC 215 “Respiratory and anaesthetic equipment”, the secretariat of which is held by BSI.

This European Standard shall be given the status of a national standard, either by publication of an identical text or by endorsement, at the latest by October 2007, and conflicting national standards shall be withdrawn at the latest by October 2007.

This document supersedes EN 13544-1:2001.

This document has been prepared under a mandate given to CEN by the European Commission and the European Free Trade Association, and supports essential requirements of EU Directive(s).

For relationship with EU Directive(s), see informative Annex ZA, which is an integral part of this document.

This European Standard applies to respiratory therapy equipment and has been prepared in three parts. This Part addresses nebulizing systems; Parts 2 and 3 address respectively tubing and connectors, and air entrainment devices.

According to the CEN/CENELEC Internal Regulations, the national standards organizations of the following countries are bound to implement this European Standard: Austria, Belgium, Bulgaria, Cyprus, Czech Republic, Denmark, Estonia, Finland, France, Germany, Greece, Hungary, Iceland, Ireland, Italy, Latvia, Lithuania, Luxembourg, Malta, Netherlands, Norway, Poland, Portugal, Romania, Slovakia, Slovenia, Spain, Sweden, Switzerland and United Kingdom.

SIST EN 13544-1:2007

<https://standards.iteh.ai/catalog/standards/sist/2da75a13-1238-4d99-8101-79fce8146958/sist-en-13544-1-2007>

Introduction

This European Standard is based on EN 60601-1:1990

In EN 60601-1:1990, this type of European Standard is referred to as a “Particular Standard”. As stated in 1.3 of EN 60601-1:1990 the requirements of this European Standard take precedence over those of EN 60601-1:1990.

Clauses, subclauses, tables and figures additional to those in EN 60601-1:1990 are numbered beginning at '101'. Additional annexes are lettered beginning at 'AA' except for Annex 'ZA'.

Additional items in lettered lists are lettered beginning 'aa'.

Rationales for some of the requirements of this European Standard are given in Annex AA. Such requirements are indicated by the letter '**R**' after the clause number.

iTeh STANDARD PREVIEW (standards.iteh.ai)

SIST EN 13544-1:2007

<https://standards.iteh.ai/catalog/standards/sist/2da75a13-1238-4d99-8101-79fce8146958/sist-en-13544-1-2007>

Section one – General

1 R) Scope

The scope given in Clause 1 of EN 60601-1:1990 applies except that 1.1 is replaced by the following:

1.1 This European Standard specifies requirements for nebulizing systems used for the delivery of drugs in an aerosol form to humans through the respiratory system.

This European Standard includes gas-powered nebulizers which may be derived from e.g. compressors, pipeline systems, cylinders etc., or electrically-powered nebulizers (e.g. ultrasonic and membrane devices) or manually-powered nebulizers.

NOTE Requirements for nebulizers having also a humidification function are specified in EN ISO 8185:1997 + AC: 2002 “Humidifiers” (see 56.102).

This European Standard does not apply to nebulizers precharged with a specific medicinal product (e.g. MDI, DPI).

2 Normative references

The following referenced documents are indispensable for the application of this document. For dated references, only the edition cited applies. For undated references, the latest edition of the referenced document (including any amendments) applies.

EN 556 (all parts), *Sterilization of medical devices — Requirements for medical devices to be designated “STERILE”* <https://standards.iteh.ai/catalog/standards/sist/2da75a13-1238-4d99-8101-79fce8146958/sist-en-13544-1-2007>

EN 737-1, *Medical gas pipeline systems — Part 1: Terminal units for compressed medical gases and vacuum*

ENV 737-6, *Medical gas pipeline systems — Part 6: Dimensions and allocation of probes for terminal units for compressed medical gases and vacuum*

EN 739, *Low pressure hose assemblies for use with medical gases*

EN 980, *Graphical symbols for use in the labelling of medical devices*

EN 1041, *Information supplied by the manufacturer with medical devices*

EN 1281-2¹⁾, *Anaesthetic and respiratory equipment — Conical connectors — Part 2: Screw-threaded weight-bearing connectors (ISO 5356-2:1987 modified)*

EN 1707, *Conical fittings with a 6 % (Luer) taper for syringes, needles and certain other medical equipment — Lock fittings*

EN 20594-1, *Conical fittings with a 6 % (Luer) taper for syringes, needles and certain other medical equipment — Part 1: General requirements (ISO 594-1:1986)*

EN 60601-1:1990, *Medical electrical equipment — Part 1: General requirements for basic safety and essential performance (IEC 60601-1:1988)*

¹⁾ Will be superseded by EN ISO 5356-2, which is currently under preparation.

EN 13544-1:2007 (E)

EN 60601-1-2, *Medical electrical equipment — Part 1-2: General requirements for safety; Collateral Standard: Electromagnetic compatibility — Requirements and tests* (IEC 60601-1-2:2001)

EN 61000-4-2:1995, *Electromagnetic compatibility (EMC) — Part 4: Testing and measurement techniques — Section 2: Electrostatic discharge immunity test — Basic EMC publication* (IEC 61000-4-2:1995)

EN 61672-1, *Electroacoustics — Sound level meters — Part 1: Specifications* (IEC 61672-1:2002)

EN 61672-2, *Electroacoustics — Sound level meters — Part 2: Pattern evaluation tests* (IEC 61672-2:2003)

EN ISO 3744, *Acoustics — Determination of sound power levels of noise sources using sound pressure — Engineering method in an essentially free field over a reflecting plane* (ISO 3744:1994)

EN ISO 4135:2001, *Anaesthetic and respiratory equipment — Vocabulary* (ISO 4135:2001)

EN ISO 5356-1, *Anaesthetic and respiratory equipment — Conical connectors — Part 1: Cones and sockets* (ISO 5356-1:2004)

EN ISO 7396-1, *Medical gas pipeline systems — Part 1: Pipelines for compressed medical gases and vacuum* (ISO 7396-1:2006)

EN ISO 8185, *Humidifiers for medical use — General requirements for humidification systems* (ISO 8185:1997)

EN ISO 10524-1, *Pressure regulators for use with medical gases — Part 1: Pressure regulators and pressure regulators with flow-metering devices* (ISO 10524-1:2006)

EN ISO 10524-3, *Pressure regulators for use with medical gases — Part 3: Pressure regulators integrated with cylinder valves* (ISO 10524-3:2005)

EN ISO 11135-1, *Sterilization of health care products — Ethylene oxide — Part 1: Requirements for development, validation and routine control of a sterilization process for medical devices* (ISO 11135-1:2007)

EN ISO 11137-1, *Sterilization of health care products — Radiation — Part 1: Requirements for development, validation and routine control of a sterilization process for medical devices* (ISO 11137-1:2006)

EN ISO 11137-2, *Sterilization of health care products — Radiation — Part 2: Establishing the sterilization dose* (ISO 11137-2:2006)

EN ISO 11137-3, *Sterilization of health care products — Radiation — Part 3: Guidance on dosimetric aspects* (ISO 11137-3:2006)

EN ISO 14971, *Medical devices — Application of risk management to medical devices* (ISO 14971:2000)

EN ISO 15001, *Anaesthetic and respiratory equipment — Compatibility with oxygen* (ISO 15001:2003)

EN ISO 17665-1, *Sterilization of health care products — Moist heat — Part 1: Requirements for the development, validation and routine control of a sterilization process for medical devices* (ISO 17665-1:2006)

IEC 60079-4, *Electrical apparatus for explosive gas atmospheres — Part 4: Method of test for ignition temperature*

3 Terms and definitions

For the purposes of this document, the terms and definitions given in EN ISO 4135:2001, Clause 2 of EN 60601-1:1990 and the following apply.

2.1.5 R) applied part: Add the following item:

All parts of the nebulizer intended to be connected to the patient or to the breathing system.

3.1

aerosol

suspension of particles in gas

NOTE Particles can be liquid or solid.

3.2

aerosol output

amount of aerosol delivered by the nebulizing system for given filled volume

3.3

aerosol output rate

amount of aerosol delivered by the nebulizing system per unit of time

3.4

anatomical airways

natural pathways through which respired gases pass in either direction between the atmosphere and the alveoli (see Annex BB)

3.5

manually-powered nebulizer

nebulizer which operates by means of human power

3.6

electrically-powered nebulizer

nebulizer which operates by means of electrical power

3.7

gas-powered nebulizer (jet nebulizer)

nebulizer in which aerosol is generated by compressed gas

3.8

ultrasonic nebulizer

nebulizer in which aerosol is generated by means of ultrasound

3.9

liquid container

part of the nebulizer which contains the liquid for nebulization

3.10

maximum fill volume

maximum volume of liquid, expressed in millilitres, in the liquid container during normal operation when the nebulizer is filled to its maximum filling level

3.11

nebulizer

device which converts a liquid into an aerosol

device, including all parts, required to convert a liquid into an aerosol and make it available for inhalation

3.13

validation

confirmation, through the provision of objective evidence, that the requirements for a specific intended use or application have been fulfilled

NOTE 1 The term "validated" is used to designate the corresponding status.

NOTE 2 The use conditions for validation can be real or simulated.

4.1 Modifications to Clause 3 of EN 60601-1:1990

Clause 3 of EN 60601-1:1990 applies with the following modifications.

Replace 3.1 with the following:

Nebulizing systems and nebulizers shall, when transported, stored, installed, operated in normal use and maintained according to the instructions of the manufacturer, cause no safety hazard which could reasonably be foreseen using risk management procedures in accordance with EN ISO 14971 and which is connected with their intended application, in normal condition and in single fault condition.

In 3.6 add the following:

aa) applicable single fault conditions are:

SIST EN 13544-1:2007

<https://standards.iteh.ai/catalog/standards/sist/2da75a13-1238-4d99-8101->

- short and open-circuits of components or wiring which can:

- cause sparks to occur, or
 - increase the energy of sparks, or
 - increase temperature (see section seven);
- incorrect output resulting from software error.

NOTE See also 54.101.

bb) **R)** A situation in which a fault is not detected is considered a normal condition.

Fault conditions/hazardous situations may remain undetected over a period of time and as a consequence may lead to an unacceptable risk. In that case subsequent fault conditions need to be considered as a single fault condition.

Specific risk control measures need to be determined within the risk management process to deal with such situations.

4.2 Clause 4 of EN 60601-1:1990

Clause 4 of EN 60601-1:1990 applies.

4.3 Alternative type-test methods

The manufacturer may use type-test methods for particle sizing, different from those specified in Annex CC.

Validation of alternative type-test methods shall be performed against the reference method in Annex CC to demonstrate that an equivalent particle size distribution curve and derived results are obtained.

Demonstration shall be circulated in the technical documentation of the manufacturer.

Evidence shall be provided upon request e.g. to Regulatory Authorities.

5 Classification

Clause 5 of EN 60601-1:1990 applies.

NOTE A nebulizing system can have applied parts of different types.

6 Identification, marking and documents

Clause 6 of EN 60601-1:1990 applies with the following additions and modifications.

6.1 Marking on the outside of equipment or equipment parts

In 6.1 add the following to item e):

The name or trademark and address of the manufacturer. For devices imported into the European Union the following applies: the name and address of the person responsible or of the authorised representative of the manufacturer or the importer established within the European Community shall be provided with the device or with the accompanying document.

In 6.1 add the following additional items:

aa) all flow-direction sensitive components, breathing attachments or parts (e.g. face mask or mouth piece one-way valve) shall be either clearly and durably marked with an arrow showing direction of gas flow if operator-detachable, or shall be permanently attached (see note of 6.8.2 dd));

bb) if gas-specific, the inlet and outlet shall be identified by clear and durable marking;

cc) marking of devices, labels and packaging.

Devices, labels and/or packaging shall contain the following:

- device identification and content information;
- if appropriate, the symbol **STERILE** in accordance with EN 980 together with the method of sterilization;
- if appropriate, the batch code, preceded by the symbol **LOT** in accordance with EN 980, or serial number;
- the expiry date, if the device is sensitive to storage or shelf life in accordance with EN 980;;
- if appropriate, an indication that the device is for single use;
- any special storage and/or handling conditions;