



# SLOVENSKI STANDARD

## SIST EN 302 537-2 V1.1.2:2008

01-september-2008

9`Y\_fca U[ bYfbUnXfi y`^j cgh]b`nUXYj Y`j `nj Yn]`n`fUX]`g\_`ja `gdY\_fca `fØFAŁ!  
BUdfUj Y`fUh\_Y[ UXcgY[ UfGF8Ł!`G`i y`VYb]`g]ghYa ]i `hfUa U`b] `a c ]`nU  
a YX]V]bg\_Y`dcXUh\_Yž\_]XYi `Y`c`j `Z`Y\_j Yb b] `cVa c `] `cX`(\$%A<n`Xc`(\$&A<n`]b  
cX`(\$)`A<n`Xc`(\$`A<n!`&`XY.`<Ufa cb]n]fUb]9Bž\_]`nU`Ya UV]ghj YbY`nU H]j Y  
`YbU`"&X]fY\_hj YF/ HH9

Electromagnetic compatibility and Radio spectrum Matters (ERM) - Short Range Devices (SRD) - Ultra Low Power Medical Data Service Systems operating in the frequency range 401 MHz to 402 MHz and 405 MHz to 406 MHz - Part 2: Harmonized EN covering essential requirements of article 3.2 of the R&TTE Directive

[SIST EN 302 537-2 V1.1.2:2008](https://standards.iteh.ai/catalog/standards/sist/38fb137e-2bcc-493c-8e19-8847a724dece/sist-en-302-537-2-v1-1-2-2008)

<https://standards.iteh.ai/catalog/standards/sist/38fb137e-2bcc-493c-8e19-8847a724dece/sist-en-302-537-2-v1-1-2-2008>

Ta slovenski standard je istoveten z: EN 302 537-2 Version 1.1.2

### ICS:

33.060.99	Druga oprema za radijske komunikacije	Other equipment for radiocommunications
33.100.01	Elektromagnetna združljivost na splošno	Electromagnetic compatibility in general

SIST EN 302 537-2 V1.1.2:2008 en

**iTeh STANDARD PREVIEW**  
**(standards.iteh.ai)**

SIST EN 302 537-2 V1.1.2:2008

<https://standards.iteh.ai/catalog/standards/sist/38fb137e-2bcc-493c-8e19-8847a724deee/sist-en-302-537-2-v1-1-2-2008>

# ETSI EN 302 537-2 V1.1.2 (2007-12)

---

*Harmonized European Standard (Telecommunications series)*

**Electromagnetic compatibility  
and Radio spectrum Matters (ERM);  
Short Range Devices (SRD);  
Ultra Low Power Medical Data Service Systems  
operating in the frequency range  
401 MHz to 402 MHz and 405 MHz to 406 MHz;  
Part 2: Harmonized EN covering essential requirements  
of article 3.2 of the R&TTE Directive**

---

**iTeh STANDARD PREVIEW**  
**(standards.iteh.ai)**

[SIST EN 302 537-2 V1.1.2:2008](https://standards.iteh.ai/catalog/standards/sist/38fb137e-2bcc-493c-8e19-8847a724dece/sist-en-302-537-2-v1-1-2-2008)

<https://standards.iteh.ai/catalog/standards/sist/38fb137e-2bcc-493c-8e19-8847a724dece/sist-en-302-537-2-v1-1-2-2008>



## Reference

---

DEN/ERM-TG30-005-2

## Keywords

---

health, regulation, SRD

**ETSI**

650 Route des Lucioles  
F-06921 Sophia Antipolis Cedex - FRANCE

---

Tel.: +33 4 92 94 42 00 Fax: +33 4 93 65 47 16

Siret N° 348 623 562 00017 - NAF 742 C  
Association à but non lucratif enregistrée à la  
Sous-Préfecture de Grasse (06) N° 7803/88

**iTeh STANDARD PREVIEW**  
**(standards.iteh.ai)**

SIST EN 302 537-2 V1.1.2:2008

<https://standards.iteh.ai/catalog/standards/sist/38fb137e-2bcc-493c-8e19-8847a724d4cc/etsi-en-302-537-2-v1-1-2-2008>

**Important notice**

Individual copies of the present document can be downloaded from:

<http://www.etsi.org>

The present document may be made available in more than one electronic version or in print. In any case of existing or perceived difference in contents between such versions, the reference version is the Portable Document Format (PDF). In case of dispute, the reference shall be the printing on ETSI printers of the PDF version kept on a specific network drive within ETSI Secretariat.

Users of the present document should be aware that the document may be subject to revision or change of status. Information on the current status of this and other ETSI documents is available at

<http://portal.etsi.org/tb/status/status.asp>

If you find errors in the present document, please send your comment to one of the following services:

[http://portal.etsi.org/chaicor/ETSI\\_support.asp](http://portal.etsi.org/chaicor/ETSI_support.asp)

---

**Copyright Notification**

---

No part may be reproduced except as authorized by written permission.  
The copyright and the foregoing restriction extend to reproduction in all media.

© European Telecommunications Standards Institute 2007.  
All rights reserved.

**DECT™**, **PLUGTESTS™**, **UMTS™**, **TIPHON™**, the TIPHON logo and the ETSI logo are Trade Marks of ETSI registered for the benefit of its Members.

**3GPP™** is a Trade Mark of ETSI registered for the benefit of its Members and of the 3GPP Organizational Partners.

# Contents

Intellectual Property Rights .....	5
Foreword.....	5
1 Scope .....	6
2 References .....	7
2.1 Normative references .....	7
2.2 Informative references.....	7
3 Definitions and abbreviations.....	8
3.1 Definitions.....	8
3.2 Abbreviations .....	8
4 Technical requirements and specifications.....	8
4.1 Environmental profile.....	8
4.2 Conformance requirements .....	8
4.2.1 Mechanical and electrical design .....	8
4.2.1.1 General .....	8
4.2.1.2 Antennas .....	8
4.2.1.3 Controls.....	8
4.2.1.4 Transmitter shut-off facility .....	8
4.2.2 Frequency error.....	8
4.2.2.1 Definition .....	8
4.2.2.2 Limits .....	9
4.2.2.3 Conformance.....	9
4.2.3 Emission bandwidth.....	9
4.2.3.1 Definition .....	9
4.2.3.2 Limits .....	9
4.2.3.3 Conformance.....	9
4.2.4 Effective radiated power of the fundamental emission.....	9
4.2.4.1 Definition .....	9
4.2.4.2 Limits .....	9
4.2.4.3 Conformance.....	9
4.2.5 Spurious emissions .....	9
4.2.5.1 Definition .....	9
4.2.5.2 Limits .....	9
4.2.5.3 Conformance.....	9
4.2.6 Frequency stability under low voltage conditions .....	10
4.2.6.1 Definition .....	10
4.2.6.2 Limits .....	10
4.2.6.3 Conformance.....	10
4.2.7 Spurious radiation of receivers .....	10
4.2.7.1 Definition .....	10
4.2.7.2 Limits .....	10
4.2.7.3 Conformance.....	10
4.2.8 Spectrum access.....	10
4.2.8.1 LBT/AFA spectrum access .....	10
4.2.8.1.1 Definition.....	10
4.2.8.1.2 Limits .....	10
4.2.8.1.3 Conformance .....	10
4.2.8.2 Low Power Low Duty Cycle spectrum access .....	11
4.2.8.2.1 Definition.....	11
4.2.8.2.2 Limits for low duty cycle low power spectrum access .....	11
4.2.8.2.3 Conformance .....	11
5 Testing for compliance with technical requirements.....	11
5.1 Environmental conditions for testing .....	11
5.2 Interpretation of the measurement results .....	11

5.3	Essential radio test suites.....	12
5.3.1	Frequency error.....	12
5.3.2	Emission bandwidth.....	12
5.3.3	Effective radiated power of the fundamental emission.....	12
5.3.4	Spurious emissions .....	12
5.3.5	Frequency stability under low voltage conditions .....	12
5.3.6	Spurious radiation of receivers .....	12
5.3.7	Spectrum access for systems using LBT/AFA.....	13
5.3.8	Spectrum access for devices using low power and low duty cycle.....	13
5.3.9	Normal and extreme test-conditions .....	13
5.3.10	Test power source .....	13
5.3.11	Choice of samples for test suites.....	13
<b>Annex A (normative):</b>	<b>HS Requirements and conformance Test specifications Table (HS-RTT).....</b>	<b>14</b>
<b>Annex B (informative):</b>	<b>The EN title in the official languages .....</b>	<b>17</b>
<b>Annex C (informative):</b>	<b>Bibliography.....</b>	<b>19</b>
History .....		20

## iTeh STANDARD PREVIEW (standards.iteh.ai)

[SIST EN 302 537-2 V1.1.2:2008](https://standards.iteh.ai/catalog/standards/sist/38fb137e-2bcc-493c-8e19-8847a724deee/sist-en-302-537-2-v1-1-2-2008)

<https://standards.iteh.ai/catalog/standards/sist/38fb137e-2bcc-493c-8e19-8847a724deee/sist-en-302-537-2-v1-1-2-2008>

## Intellectual Property Rights

IPRs essential or potentially essential to the present document may have been declared to ETSI. The information pertaining to these essential IPRs, if any, is publicly available for **ETSI members and non-members**, and can be found in ETSI SR 000 314: *"Intellectual Property Rights (IPRs); Essential, or potentially Essential, IPRs notified to ETSI in respect of ETSI standards"*, which is available from the ETSI Secretariat. Latest updates are available on the ETSI Web server (<http://webapp.etsi.org/IPR/home.asp>).

Pursuant to the ETSI IPR Policy, no investigation, including IPR searches, has been carried out by ETSI. No guarantee can be given as to the existence of other IPRs not referenced in ETSI SR 000 314 (or the updates on the ETSI Web server) which are, or may be, or may become, essential to the present document.

## Foreword

This Harmonized European Standard (Telecommunications series) has been produced by ETSI Technical Committee Electromagnetic compatibility and Radio spectrum Matters (ERM).

The present document has been produced by ETSI in response to a mandate from the European Commission issued under Council Directive 98/34/EC [3] (as amended) laying down a procedure for the provision of information in the field of technical standards and regulations.

The present document is intended to become a Harmonized Standard, the reference of which will be published in the Official Journal of the European Communities referencing the Directive 1999/5/EC of the European Parliament and of the Council of 9 March 1999 on radio equipment and telecommunications terminal equipment and the mutual recognition of their conformity (R&TTE Directive) [6].

Technical specifications relevant to Directive 1999/5/EC are given in annex A.

The present document is part 2 of a multi-part deliverable covering radio equipment in the frequency range 401 MHz to 402 MHz and 405 MHz to 406 MHz for Ultra Low Power Active Medical Devices and implants. This includes Body Worn, Hand-Held, Data systems, etc., the medical section of which is regulated under the Medical Device Directive [4] and Active Medical Implants and Peripherals the medical section of which is regulated under the Active Implantable Medical Device Directive [4], as identified below:

- Part 1: "Technical characteristics and test methods";
- **Part 2: "Harmonized EN covering essential requirements of article 3.2 of the R&TTE Directive".**

### National transposition dates

Date of adoption of this EN:	14 December 2007
Date of latest announcement of this EN (doa):	31 March 2008
Date of latest publication of new National Standard or endorsement of this EN (dop/e):	30 September 2008
Date of withdrawal of any conflicting National Standard (dow):	30 September 2009

# 1 Scope

The present document covers various individual devices which when operating together form a system operating as a Medical Data Service (MEDS) that provides medical practitioners with therapeutic and/or diagnostic information used to provide improved medical treatment of a patient and/or to provide an interactive system for patient control of therapeutic devices. MEDS is intended only for transmission of non-time critical data, the loss of which will not compromise the health and/or safety of the patient.

Devices covered by the present document are an evolving new technology to be made available worldwide by the medical equipment industry that will provide high speed communications capability between devices associated with an individual patient that are part of a complete MEDS system as defined in clause 3.1. Examples of MEDS devices falling under the scope of the present document are portable body worn physiological sensors that allow ambulatory monitoring, implanted devices and external system devices used to transfer data collected by a MEDS system to medical practitioners that will use the data to diagnose and treat a patient.

These devices utilize ultra low power radio transmitters in combination with medical devices, the medical portion of which is regulated by the Medical Device Directive [4] (MDD) or the Active Implantable Medical Device Directive (AIMD [7]). The radio part of medical devices regulated by the MDD is hereafter referred to as ULP-AMD, ULP-AMD-P for peripheral devices, and ULP-BWD for body worn devices. ULP-BWD are devices, such as a physiological parameter sensor or handheld devices that are intended to operate in very close proximity to the human body, including touching the body, whose radio antenna is external to the body and is used to communicate with a device that is part of a MEDS system. The radio part of medical devices regulated under the AIMD is hereafter referred to as Ultra Low Power-Active Medical Implants (ULP-AMI) and peripherals (ULP-AMI-P) used in a Medical Data Service (MEDS).

The present document contains required characteristics considered necessary for the radio sections to meet in order to efficiently use the available spectrum for the purpose of transferring data that is used in diagnosing and delivering therapies to individuals with various illnesses. Of particular importance is the inclusion of spectrum monitoring and access requirements (listen before talk protocol) designed to significantly reduce any interference potential between MEDS systems operating in the band or between a MEDS system and the primary users of the band.

The present document is a specific product standard applicable to ultra low power transmitters that are part of a system used in the MEDS operating in spectrum within the frequency bands 401 MHz to 402 MHz and 405 MHz to 406 MHz. The present document contains the technical characteristics for ultra low power radio equipment and is addressed by CEPT/ERC/REC 70-03 [5] and annex 12 to that document. It does not necessarily include all the characteristics, which may be required by a user, nor does it necessarily represent the optimum performance achievable.

It applies to ultra low power systems and accessories operating in spectrum within the bands 401 MHz to 402 MHz and 405 MHz to 406 MHz that operate in the MEDS service:

- for telecommand and telemetry between any devices that are part of a MEDS (see definition of MEDS);
- with or without an integral antenna; and/or
- with an antenna connection provided only for the purpose of connecting an external dedicated antenna.

In addition to the present document, other ENs that specify technical requirements in respect of essential requirements under other parts of article 3 of the R&TTE Directive [6] will apply to equipment within the scope of the present document.



## 2 References

References are either specific (identified by date of publication and/or edition number or version number) or non-specific.

- For a specific reference, subsequent revisions do not apply.
- Non-specific reference may be made only to a complete document or a part thereof and only in the following cases:
  - if it is accepted that it will be possible to use all future changes of the referenced document for the purposes of the referring document;
  - for informative references.

Referenced documents which are not found to be publicly available in the expected location might be found at <http://docbox.etsi.org/Reference>.

For online referenced documents, information sufficient to identify and locate the source shall be provided. Preferably, the primary source of the referenced document should be cited, in order to ensure traceability. Furthermore, the reference should, as far as possible, remain valid for the expected life of the document. The reference shall include the method of access to the referenced document and the full network address, with the same punctuation and use of upper case and lower case letters.

NOTE: While any hyperlinks included in this clause were valid at the time of publication ETSI cannot guarantee their long term validity.

### 2.1 Normative references

The following referenced documents are indispensable for the application of the present document. For dated references, only the edition cited applies. For non-specific references, the latest edition of the referenced document (including any amendments) applies.

- [1] ETSI EN 302 537-1 (V1.2.1): "Electromagnetic compatibility and Radio spectrum Matters (ERM); Short Range Devices (SRD); Ultra Low Power Medical Data Service Systems operating in the frequency range 401 MHz to 402 MHz and 405 MHz to 406 MHz; Part 1: Technical characteristics and test methods".
- [2] ETSI TR 100 028 (V1.4.1): "ElectroMagnetic Compatibility and Radio Spectrum Matters (ERM); Uncertainties in the measurement of mobile radio equipment characteristics".

### 2.2 Informative references

- [3] Directive 98/34/EC of the European Parliament and of the Council of 22 June 1998 laying down a procedure for the provision of information in the field of technical standards and regulations.
- [4] Council Directive 93/42/EEC of 14 June 1993 concerning medical devices (MD Directive).
- [5] CEPT/ERC/REC 70-03 (2006): "Relating to the use of Short Range Devices (SRD)".
- [6] Directive 1999/5/EC of the European Parliament and of the Council of 9 March 1999 on radio equipment and telecommunications terminal equipment and the mutual recognition of their conformity (R&TTE Directive).
- [7] Council Directive 90/385/EEC of 20 June 1990 on the approximation of the laws of the Member States relating to active implantable medical devices.

---

## 3 Definitions and abbreviations

### 3.1 Definitions

For the purposes of the present document, the terms and definitions given in EN 302 537-1 [1], clause 3.1 apply.

### 3.2 Abbreviations

For the purposes of the present document, the abbreviations given in EN 302 537-1 [1], clause 3.3 apply.

---

## 4 Technical requirements and specifications

### 4.1 Environmental profile

The technical requirements of the present document apply under the environmental profile for operation of the equipment as described in the user's manual and declared by the provider. The equipment shall comply with all the technical requirements of the present document at all times when operating within the boundary limits of the operational environmental profile as described above. The provider shall declare that interruption of the communications link for his MEDS system shall not result in compromising the health and safety of the patient.

### 4.2 Conformance requirements

#### 4.2.1 Mechanical and electrical design

##### 4.2.1.1 General

The equipment shall be designed, constructed and manufactured in accordance with sound engineering practice and with the aim of minimizing harmful disturbance to other equipment and services. It should not be disturbed by harmful interference from other electronic devices and users of the band. Transmitters and receivers may be individual or combination units.

##### 4.2.1.2 Antennas

Equipment operating in the MEDS service shall have an integral antenna, an external dedicated antenna or both. If provision for an external antenna connection is made, the connector shall be a unique type to prevent use of an antenna other than a dedicated antenna supplied by the provider.

##### 4.2.1.3 Controls

Those controls which, if maladjusted, might increase the interfering potentialities of the equipment, shall not be accessible to the user.

##### 4.2.1.4 Transmitter shut-off facility

If the transmitter is equipped with an automatic transmitter shut-off facility or battery-saving feature and it interferes with testing of the device, it shall be capable of being made inoperative for the purpose of testing.

### 4.2.2 Frequency error

#### 4.2.2.1 Definition

The frequency error shall be as defined in EN 302 537-1 [1], clause 8.1.1.

#### 4.2.2.2 Limits

The frequency error limits shall be as defined in EN 302 537-1 [1], clause 8.1.2.

#### 4.2.2.3 Conformance

Conformance tests as defined in clause 5.3.1 shall be carried out.

### 4.2.3 Emission bandwidth

#### 4.2.3.1 Definition

The emission bandwidth shall be as defined in EN 302 537-1 [1], clause 8.2.1.

#### 4.2.3.2 Limits

The emission bandwidth limits shall be as defined in EN 302 537-1 [1], clause 8.2.2.

#### 4.2.3.3 Conformance

Conformance tests as defined in clause 5.3.2 shall be carried out.

### 4.2.4 Effective radiated power of the fundamental emission

#### 4.2.4.1 Definition

The effective radiated power shall be as defined in EN 302 537-1 [1], clause 8.3.1.

#### 4.2.4.2 Limits

- The e.r.p. limit shall be as defined in EN 302 537-1 [1], clause 8.3.2.1 for systems using LBT and AFA for spectrum access.
- The e.r.p. limit shall be as defined in EN 302 537-1 [1], clause 8.3.2.2 for systems using low duty cycle and low power for spectrum access.

#### 4.2.4.3 Conformance

Conformance tests as defined in clause 5.3.3 shall be carried out.

### 4.2.5 Spurious emissions

#### 4.2.5.1 Definition

The spurious emissions shall be as defined in EN 302 537-1 [1], clause 8.4.1.

#### 4.2.5.2 Limits

The spurious emissions limits shall be as defined in EN 302 537-1 [1], clause 8.4.2.

#### 4.2.5.3 Conformance

Conformance tests as defined in clause 5.3.4 shall be carried out.