



SLOVENSKI STANDARD

SIST EN 15216:2008

01-februar-2008

Karakterizacija odpadkov - Določevanje celotnih raztopljenih snovi (TDS) v vodi in izlužkih

Characterization of waste - Determination of total dissolved solids (TDS) in water and eluates

Charakterisierung von Abfällen - Bestimmung des Gesamtgehaltes an gelösten Feststoffen (TDS) in Wasser und Eluaten

Caractérisation des déchets - Détermination de la concentration en matières solides dissoutes totales (TDS) de l'eau et des éluats

<https://standards.iteh.ai/catalog/standards/sist/78555912-150f-450c-a238-1d34497c7cab/sist-en-15216-2008>

Ta slovenski standard je istoveten z: **EN 15216:2007**

ICS:

13.030.20 Tekoči odpadki. Blato Liquid wastes. Sludge

SIST EN 15216:2008

en,fr,de

iTeh STANDARD PREVIEW
(standards.iteh.ai)

SIST EN 15216:2008

<https://standards.iteh.ai/catalog/standards/sist/78555912-150f-450c-a238-1d34497c7cab/sist-en-15216-2008>

EUROPEAN STANDARD

EN 15216

NORME EUROPÉENNE

EUROPÄISCHE NORM

October 2007

ICS 13.030.20

English Version

Characterization of waste - Determination of total dissolved solids (TDS) in water and eluates

Caractérisation des déchets - Détermination de la concentration en matières solides dissoutes totales (TDS) de l'eau et des éluats

Charakterisierung von Abfällen - Bestimmung des Gesamtgehaltes an gelösten Feststoffen (TDS) in Wasser und Eluaten

This European Standard was approved by CEN on 1 September 2007.

CEN members are bound to comply with the CEN/CENELEC Internal Regulations which stipulate the conditions for giving this European Standard the status of a national standard without any alteration. Up-to-date lists and bibliographical references concerning such national standards may be obtained on application to the CEN Management Centre or to any CEN member.

This European Standard exists in three official versions (English, French, German). A version in any other language made by translation under the responsibility of a CEN member into its own language and notified to the CEN Management Centre has the same status as the official versions.

CEN members are the national standards bodies of Austria, Belgium, Bulgaria, Cyprus, Czech Republic, Denmark, Estonia, Finland, France, Germany, Greece, Hungary, Iceland, Ireland, Italy, Latvia, Lithuania, Luxembourg, Malta, Netherlands, Norway, Poland, Portugal, Romania, Slovakia, Slovenia, Spain, Sweden, Switzerland and United Kingdom.

[SIST EN 15216:2008](https://standards.iteh.ai/catalog/standards/sist/78555912-150f-450c-a238-1d34497c7cab/sist-en-15216-2008)

<https://standards.iteh.ai/catalog/standards/sist/78555912-150f-450c-a238-1d34497c7cab/sist-en-15216-2008>



EUROPEAN COMMITTEE FOR STANDARDIZATION
COMITÉ EUROPÉEN DE NORMALISATION
EUROPÄISCHES KOMITEE FÜR NORMUNG

Management Centre: rue de Stassart, 36 B-1050 Brussels

Contents

Page

Foreword.....	3
Introduction	4
1 Scope	5
2 Terms and definitions	5
3 Principle.....	5
4 Sample storage and preparation.....	5
5 Equipment	5
6 Procedure	6
7 Calculation of results	6
8 Quality control.....	6
9 Performance characteristics	7
10 Test report	7
Annex A (informative) Summary of general requirements and recommendations.....	8
Bibliography	9

[SIST EN 15216:2008](https://standards.itech.ai/catalog/standards/sist/78555912-150f-450c-a238-1d34497c7cab/sist-en-15216-2008)
<https://standards.itech.ai/catalog/standards/sist/78555912-150f-450c-a238-1d34497c7cab/sist-en-15216-2008>

Foreword

This document (EN 15216:2007) has been prepared by Technical Committee CEN/TC 292 “Characterization of waste”, the secretariat of which is held by NEN, in cooperation with CEN/TC 230.

This European Standard shall be given the status of a national standard, either by publication of an identical text or by endorsement, at the latest by April 2008, and conflicting national standards shall be withdrawn at the latest by April 2008.

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. CEN [and/or CENELEC] shall not be held responsible for identifying any or all such patent rights.

According to the CEN/CENELEC Internal Regulations, the national standards organizations of the following countries are bound to implement this European Standard: Austria, Belgium, Bulgaria, Cyprus, Czech Republic, Denmark, Estonia, Finland, France, Germany, Greece, Hungary, Iceland, Ireland, Italy, Latvia, Lithuania, Luxembourg, Malta, Netherlands, Norway, Poland, Portugal, Romania, Slovakia, Slovenia, Spain, Sweden, Switzerland and the United Kingdom.

iTeh STANDARD PREVIEW (standards.iteh.ai)

[SIST EN 15216:2008](https://standards.iteh.ai/catalog/standards/sist/78555912-150f-450c-a238-1d34497c7cab/sist-en-15216-2008)

<https://standards.iteh.ai/catalog/standards/sist/78555912-150f-450c-a238-1d34497c7cab/sist-en-15216-2008>

EN 15216:2007 (E)

Introduction

Total dissolved solids are a common sum parameter which is in use in water analysis and in analysis of eluates. TC 292 and TC 230 cooperated on this item to create one horizontal standard which is of most practical use for laboratories.

iTeh STANDARD PREVIEW (standards.iteh.ai)

[SIST EN 15216:2008](https://standards.iteh.ai/catalog/standards/sist/78555912-150f-450c-a238-1d34497c7cab/sist-en-15216-2008)

<https://standards.iteh.ai/catalog/standards/sist/78555912-150f-450c-a238-1d34497c7cab/sist-en-15216-2008>

1 Scope

This European standard specifies a method for the determination of total dissolved solids (TDS) in water and eluates, provided they are not volatile under the conditions specified or that they do not release water molecules from hydration. It applies to water and eluates containing more than 200 mg/l of total dissolved solids. Samples with lower amounts of dissolved solids can be analyzed by repetition of the drying step.

2 Terms and definitions

For the purposes of this document, the following term and definition applies.

2.1

total dissolved solids

TDS

ρ_{TDS}

mass of dissolved constituents per unit volume of water remaining after a specified filtering and drying process

NOTE TDS is expressed in milligrams per litre.

3 Principle

The test sample is filtered and then dried in a dish to a constant mass in an oven at $(105 \pm 3) ^\circ\text{C}$.

4 Sample storage and preparation

During storage, the laboratory sample may be subject to changes which are liable to influence the analysis result. If storage is necessary, samples shall be kept at $(4 \pm 2) ^\circ\text{C}$.

NOTE Freezing should be avoided due to risk of formation of precipitation.

It is recommended to carry out the test immediately after sampling of the water sample or preparation of the eluate.

5 Equipment

5.1 Drying system thermostatically controlled and capable of maintaining a temperature of $(105 \pm 3) ^\circ\text{C}$; e.g. drying oven, infrared system.

5.2 Desiccator with an active drying agent such as silica gel.

5.3 Analytical balance with an accuracy of 1 mg or better.

5.4 Membrane filter (porosity of $0,45 \mu\text{m}$).

5.5 Filtration apparatus.

5.6 Evaporation dish or crucible. Temperature tolerant laboratory vessel withstanding $105 ^\circ\text{C}$. Suitable materials are glass, ceramic, porcelain, aluminium and other materials inert to the test portion.

EN 15216:2007 (E)**6 Procedure**

Filter the test sample through the membrane filter as specified in 5.4 unless it has been membrane filtered before (e. g. eluates prepared according to EN 12457).

Dry the dish, for example in the oven, at $(105 \pm 3) ^\circ\text{C}$ and weigh to the nearest 1 mg (to obtain m_a) after cooling to ambient temperature in the desiccator.

Homogenise and transfer a proper aliquot (V_{FB}) if necessary in stages, into the dish to ensure that not less than 20 mg and not more than 1 000 mg of dry mass, m_{D} , remains after evaporation to dryness.

Evaporate the sample to dryness by a drying system according to 5.1 and weigh to the nearest 1 mg (to obtain m_b) after cooling to ambient temperature in the desiccator.

The mass of the dry matter, m_{D} , shall be considered constant if, after drying for a further half-hour period, the mass obtained does not differ by more than 0,5 % of the previous value or 2 mg whichever is the greater. Otherwise, drying shall be repeated until constant mass has been reached.

NOTE If a constant value is not obtained even after drying for a third time, record the value determined last and note this in the test report.

7 Calculation of results

Calculate the total dissolved solids from equation (1):

$$\rho_{\text{TDS}} = \frac{m_{\text{D}}}{V_{\text{FB}}} \quad (1)$$

with $m_{\text{D}} = m_{\text{b}} - m_{\text{a}}$

iTech STANDARD PREVIEW
(standards.itech.ai)
SIST EN 15216:2008
<https://standards.itech.ai/catalog/standards/sist/78555912-150f-450c-a238-1d34497c7cab/sist-en-15216-2008>

where

ρ_{TDS} is the concentration of total dissolved solids, in mg/l;

m_{D} is the mass of the dry matter, in mg;

m_{a} is the mass of the empty dish, in mg;

m_{b} is the mass of the dish containing the dry matter, in mg;

V_{FB} is the volume of the filtered test sample taken, in l.

Values shall be rounded to the nearest 1 mg/l.

8 Quality control

Where uncertainty exists about the homogeneity or behaviour of the sample, it is recommended that the analysis is carried out in duplicate.

9 Performance characteristics

The method performance characteristics given in Table 1 have been established in a European intercomparison study on two surface water samples, two waste water samples and two eluates, carried out in 2006.

Table 1 — Performance characteristics for the determination of total dissolved solids in water and eluates

Material	p	N	O %	Mean mg/l	S_r %	S_R %
Surface water A	17	34	0	478	2,15	4,49
Surface water B	17	34	0	3810	0,39	2,53
Waste water A	17	34	0	27300	1,34	4,33
Waste water B	17	34	0	4840	1,28	5,55
Eluate A	17	34	0	4090	1,74	6,27
Eluate B	17	34	0	3440	0,71	3,16

p is number of participating laboratories
 N is the number of outlier-free individual analytical results
 O is the percentage of outliers
 S_r is the estimate of the relative repeatability standard deviation
 S_R is the estimate of the relative reproducibility standard deviation

SIST EN 15216:2008

10 Test report <https://standards.iteh.ai/catalog/standards/sist/78555912-150f-450c-a238-1d34497c7cab/sist-en-15216-2008>

The test report shall contain at least the following details:

- reference to this European standard;
- all data which are necessary to identify the sample;
- calculated concentration of total dissolved solids;
- details of all procedural steps which deviate from this standard together with all circumstances that may have influenced the results;
- type of drying procedure.