



Designation: C 1287 – 95 (Reapproved 2001)

## Standard Test Method for Determination of Impurities in Uranium Dioxide by Inductively Coupled Plasma Mass Spectrometry<sup>1</sup>

This standard is issued under the fixed designation C 1287; the number immediately following the designation indicates the year of original adoption or, in the case of revision, the year of last revision. A number in parentheses indicates the year of last reapproval. A superscript epsilon ( $\epsilon$ ) indicates an editorial change since the last revision or reapproval.

### 1. Scope

1.1 This test method covers the determination of 61 elements in uranium dioxide samples by inductively coupled plasma mass spectrometry (ICP-MS). The elements are listed in Table 1 along with their lower reporting limits.

1.2 Similar levels of these elements in other uranic compounds can also be determined if they are treated and converted to the same uranium concentration solution.

1.3 *This standard does not purport to address all of the safety concerns, if any, associated with its use. It is the responsibility of the user of this standard to establish appropriate safety and health practices and determine the applicability of regulatory limitations prior to use.* For a specific warning statement, see Note 1.

NOTE 1—**Warning:** The ICP-MS is a source of intense ultra-violet radiation from the radio frequency induced plasma. Protection from radio frequency radiation and UV radiation is provided by the instrument under normal operation.

1.4 The test method for the additional elements boron, sodium, silicon, phosphorus, potassium, and calcium is given in Appendix X1.

1.5 The test method for technetium-99 is given in Appendix X2.

### 2. Referenced Documents

#### 2.1 ASTM Standards:

C 753 Specification for Nuclear-Grade, Sinterable Uranium Dioxide Powder<sup>2</sup>

C 776 Specification for Sintered Uranium Dioxide Pellets<sup>2</sup>

C 787 Specification for Uranium Hexafluoride for Enrichment<sup>2</sup>

C 788 Specification for Nuclear-Grade Uranyl Nitrate Solution<sup>2</sup>

C 967 Specification for Uranium Ore Concentrate<sup>2</sup>

C 996 Specification for Uranium Hexafluoride Enriched to

<sup>1</sup> This test method is under the jurisdiction of ASTM Committee C26 on Nuclear Fuel Cycle and is the direct responsibility of Subcommittee C26.05 on Methods of Test.

Current edition approved Sept. 10, 1995. Published November 1995. Originally published as C 1287 – 94. Last previous edition C 1287 – 94.

<sup>2</sup> *Annual Book of ASTM Standards*, Vol 12.01.

Less Than 5 % <sup>235</sup>U<sup>2</sup>

D 1193 Specification for Reagent Water<sup>3</sup>

### 3. Summary of Test Method

3.1 The sample is dissolved in acid. A fixed quantity of internal standard is added to monitor and correct for signal instability. The level of impurities in the solution is measured by ICP-MS. Customized software calculates the concentration of each element.

3.2 Uranium-concentration-matched standard solutions are used to calibrate the ICP-MS instrument. The calibration is linear up to at least 0.2  $\mu\text{g/ml}$  (100  $\mu\text{g/g}$  U) for each analyte.<sup>4,5</sup>

### 4. Significance and Use

4.1 This test method is capable of measuring the elements in Table 1, some of which are required by Specifications C 753, C 776, C 787, C 788, C 967 and C 996.

### 5. Apparatus

5.1 *ICP-MS*, controlled by computer and fitted with the associated software and peripherals.

5.2 *Autosampler*, with tube racks and disposable plastic sample tubes compatible with 5.1 (optional).

5.3 *Variable Micropipettes:*

5.3.1 10  $\mu\text{L}$  to 100  $\mu\text{L}$  capacity.

5.3.2 100  $\mu\text{L}$  to 1000  $\mu\text{L}$  capacity.

5.3.3 1000  $\mu\text{L}$  to 10.00 mL capacity.

5.4 *Volumetric Flasks:*

5.4.1 50 mL capacity—polypropylene.

5.4.2 100 mL capacity—polypropylene.

5.4.3 1 L capacity—glass.

5.5 *Platinum Dish*—100 mL capacity.

5.6 *Silica Beaker*—250 mL capacity.

<sup>3</sup> *Annual Book of ASTM Standards*, Vol 11.01.

<sup>4</sup> "ICP-MS Versus Conventional Methods for the Analysis of Trace Impurities in Nuclear Fuel," by Allenby, P., Clarkson, A. S., Makinson, P. R., presented at 2nd Surrey Conference on Plasma Source Mass Spectrometry, Guildford, UK, July 1987.

<sup>5</sup> "Trace Metals in NBL Uranium Standard CRM 124 Using ICP-MS," by Aldridge, A. J., Clarkson, A. S., Makinson, P. R., Dawson, K. W., presented at 1st Durham International Conference on Plasma Source Mass Spectrometry, Durham, UK, September 1988.

**TABLE 1 Reporting Limits of Impurity Elements in Uranium Dioxide**

NOTE 1—Acquisition time = 10 s/isotope using peak jump mode.

NOTE 2—103 Rh was used as an internal standard.

NOTE 3—The LRL is based on the within run standard deviation ( $S_p$ ) of 20 uranium-matched blank determinations for each analyte. This limit equals  $4 \times S_p$ , rounded up to a preferred value in the series 1, 1.5, 2, 3, 4, 6, multiplied or divided by the appropriate integer power of ten.

NOTE 4—The upper reporting limit can be increased by extending the calibration to 10 µg/mL (5000 µg/g U) if the ICP-MS used has an extended dynamic range (EDR) accessory.

Analyte	Mass Used	Analyte Group	Lower Reporting Limit (LRL), µg/g U	Upper Reporting Limit (URL), µg/g U
Lithium	7	A	0.01	100
Beryllium	9	A	0.04	100
Magnesium	24	A	4	100
Aluminum	27	D	2	1000
Scandium	45	A	4	100
Titanium	48	B	0.2	100
Vanadium	51	B	0.04	100
Chromium	52	B	0.1	100
Manganese	55	A	0.1	100
Iron	56	A	15	100
Cobalt	59	A	0.02	100
Nickel	60	A	0.4	100
Copper	65	A	0.2	100
Zinc	66	A	0.3	100
Gallium	69	A	0.04	100
Germanium	74	A	0.2	100
Arsenic	75	A	0.2	100
Selenium	82	A	3	100
Rubidium	85	A	0.06	100
Strontium	88	A	0.06	100
Yttrium	89	A	0.04	100
Zirconium	90	B	0.02	100
Niobium	93	B	0.01	100
Molybdenum	95	B	0.04	100
Ruthenium	102	B	0.02	100
Palladium	106	B	0.2	100
Silver	107	A	0.1	100
Cadmium	111	A	0.03	100
Indium	115	A	0.04	100
Tin	116	B	0.04	100
Antimony	121	B	0.02	100
Tellurium	130	B	0.4	100
Caesium	133	A	0.06	100
Barium	138	A	0.02	100
Lanthanum	139	C	0.1	100
Cerium	140	C	0.01	100
Praseodymium	141	C	0.01	100
Neodymium	146	C	0.01	100
Samarium	149	C	0.01	100
Europium	151	C	0.01	100
Gadolinium	158	C	0.01	100
Terbium	159	C	0.01	100
Dysprosium	163	C	0.01	100
Holmium	165	C	0.01	100
Erbium	166	C	0.01	100
Thulium	169	C	0.01	100
Ytterbium	174	C	0.01	100
Lutetium	175	C	0.01	100
Hafnium	178	B	0.01	100
Tantalum	181	B	0.01	100
Tungsten	184	B	0.01	100
Rhenium	187	A	0.02	100
Osmium	190	B	0.2	100
Iridium	193	B	0.2	100
Platinum	195	B	0.2	100
Gold	197	B	0.06	100
Mercury	202	A	0.4	100
Thallium	205	A	0.02	100
Lead	208	A	0.02	100
Bismuth	209	A	0.03	100
Thorium	232	B	0.01	100

5.7 Watch Glasses—75 mm diameter.

## 6. Reagents

6.1 The sensitivity of the ICP-MS technique requires the use of ultra high purity reagents in order to be able to obtain the low levels of detection. All the reagents below are ultra high purity grade unless otherwise stated:

6.1.1 Element stock standards at 1000 µg/mL for all the elements in Table 1.

6.1.2 Hydrofluoric acid (40 g/100 g).

6.1.3 Nitric acid (specific gravity 1.42)—Concentrated nitric acid (HNO<sub>3</sub>).

6.1.4 Orthophosphoric acid (specific gravity 1.70).

6.1.5 Rhodium Stock Solution (1000 µg/mL Rh)—Commercially available solution (see Note 2).

NOTE 2—Rhodium stock solution is commercially available supplied with a certificate of analysis for the element and a full range of trace impurities. The solutions are prepared by the manufacturer using a variety of media designed to keep each element in solution for a minimum of one year.

6.1.6 Sulfuric acid (specific gravity 1.84)—Concentrated sulfuric acid (H<sub>2</sub>SO<sub>4</sub>).

6.1.7 Uranium Standard Base Solution—Uranyl nitrate solution to Specification C 788, of known uranium (100 g/L) and aluminum content ( $\leq 2$  µg/g U). The total metallic impurity (TMI) content must not exceed 50 µg/g U and no individual analyte must exceed 10 µg/g U.

6.1.8 Purity of Water—Unless otherwise indicated, references to water shall be understood to mean reagent water conforming to Specification D 1193, Type I.

## 7. Standards

7.1 Three separate mixed standard solutions (A, B and C) are prepared to prevent the precipitation of some elements (as insoluble chlorides, fluorides etc.; see Table 1 for details of the analyte groups). Analyte group A contains element stock solutions prepared in HNO<sub>3</sub> or HNO<sub>3</sub>/HF, analyte group B contains element stock solutions prepared in HCl or HCl/HF, and analyte group C contains the rare earth element stock solutions. The mixed standard solutions should be prepared to contain only the analytes of interest. Other combinations of mixed standard solutions may be prepared to minimize the precipitation of the analytes.

7.1.1 Mixed standard solution A is prepared from stock solutions of each element from analyte group A. Transfer 1000 µL of the stock solution (1000 µg/mL) of each element into a 50 mL polypropylene volumetric flask and add 500 µL of concentrated nitric acid. Dilute to 50 mL with water and mix. This multi-element standard contains 20 µg/mL of each analyte in 1 % nitric acid. This solution must be used on the day of preparation.

7.1.2 Mixed standard solution B is prepared from stock solutions of each element from analyte group B. Transfer 1000 µL of the stock solution (1000 µg/mL) of each element into a 50 mL polypropylene volumetric flask and add 500 µL of concentrated nitric acid. Dilute to 50 mL with water and mix. This multi-element standard contains 2 µg/mL of each analyte in 1 % nitric acid. This solution must be used within one week of preparation.



7.1.3 Mixed standard solution C is prepared from stock solutions of each element from analyte group C. Transfer 1000  $\mu\text{L}$  of the stock solution (1000  $\mu\text{g}/\text{mL}$ ) of each element into a 50 mL polypropylene volumetric flask and add 500  $\mu\text{L}$  of concentrated nitric acid. Dilute to 50 mL with water and mix. This multi-element standard contains 20  $\mu\text{g}/\text{mL}$  of each analyte in 1 % nitric acid. This solution must be used within one week of preparation.

7.2 Standard solution D is prepared from the stock solution of aluminum from analyte group D. Transfer 1000  $\mu\text{L}$  of the stock solution (1000  $\mu\text{g}/\text{mL}$  Al) into a 50 mL polypropylene volumetric flask and add 500  $\mu\text{L}$  of concentrated nitric acid. Dilute to 50 mL with water and mix. This standard contains 20  $\mu\text{g}/\text{mL}$  of aluminum in 1 % nitric acid. This solution must be used within one week of preparation.

7.3 Rhodium internal standard solution is prepared from the stock solution. Transfer 1000  $\mu\text{L}$  of the stock solution (1000  $\mu\text{g}/\text{mL}$  Rh) into a 100 mL polypropylene volumetric flask and add 1000  $\mu\text{L}$  of concentrated nitric acid. Dilute to 100 mL with water and mix. This internal standard solution contains 10  $\mu\text{g}/\text{mL}$  Rh in a 1 % nitric acid solution. Other internal standards may be used.

7.4 Diluent solution is prepared from rhodium stock standard solution. Transfer 1000  $\mu\text{L}$  of the stock solution (1000  $\mu\text{g}/\text{mL}$  Rh) into a 1 L volumetric flask and add 10.00 mL of concentrated nitric acid. Dilute to 1 L with water and mix. This diluent solution contains 0.1  $\mu\text{g}/\text{mL}$  Rh in 1 % nitric acid solution. Other internal standard diluent solutions may be used.

## 8. Procedure

NOTE 3—A uranium-free reagent blank is used to eliminate bias due to the analyte concentrations in the uranium standard base solution. However, a uranium-free reagent blank for the determination of aluminum cannot be prepared. Small variations in the concentration of the orthophosphoric acid/sulfuric acid mixture cause large variations in aluminum and rhodium signals. This leads to large errors in the reagent blank correction. A uranium-matched reagent blank is necessary to provide a constant acid concentration in the nebulized solution.

8.1 *Sample Preparation for the Determination of All Elements Except Aluminum:*

8.1.1 Weigh a portion of uranium dioxide containing between 2.45 and 2.55 g of uranium into a platinum dish. Record the weight to the nearest 0.001 g.

8.1.2 Add 10 mL of water and 12.5 mL of concentrated nitric acid. Heat on a hotplate to assist dissolution.

8.1.3 Add 2.5 mL of hydrofluoric acid (40 g/100 g) and warm at about 80°C for 5 min.

8.1.4 Allow the solution to cool and transfer quantitatively to a 50 mL polypropylene volumetric flask. Dilute to 50 mL with water and mix. This solution contains 50 g of uranium per litre in 25 % nitric acid/5 % hydrofluoric acid.

8.1.5 Transfer 4.00 mL of the solution in 8.1.4 and 1.00 mL of the rhodium internal standard solution (see 7.3) into a 100 mL polypropylene volumetric flask. Dilute to 100 mL with water and mix. This solution contains 2 g of uranium per litre and 0.1  $\mu\text{g}/\text{mL}$  Rh in 1 % nitric acid/0.2 % hydrofluoric acid.

8.1.6 A uranium-free reagent blank (see 8.3.1) and a control or recovery sample must be prepared with every run of samples.

8.1.7 Analyze these solutions as in 8.4 using the calibration solutions prepared in 8.3.1. The solutions must be analyzed within 8 h of preparation to minimize the effects of analyte precipitation.

8.2 *Sample Preparation for the Determination of Aluminum:*

8.2.1 Weigh a portion of uranium dioxide powder or pellet sample equivalent to  $1.00 \pm 0.05$  g of uranium into a silica beaker. Record the weight to the nearest 0.001 g.

8.2.2 Add 50 mL of concentrated nitric acid and heat the mixture on a hotplate to assist dissolution.

8.2.3 Add further 10 mL portions of concentrated nitric acid until dissolution is complete.

8.2.4 Evaporate the solution until fumes of nitric acid are no longer evolved.

8.2.5 Add 5 mL of concentrated sulfuric acid and 1 mL of orthophosphoric acid (specific gravity 1.70) and evaporate the mixture until fumes are no longer evolved. This acid mixture aids the dissolution of alumina which has been heated above 1000°C.

8.2.6 Add 5 mL of concentrated nitric acid and evaporate the mixture until fumes are no longer evolved.

8.2.7 Add 20 mL of water and heat gently to assist dissolution. Allow the solution to cool and quantitatively transfer it into a 50 mL volumetric flask. Dilute to 50 mL with water and mix.

8.2.8 Dispense 1.00 mL of the solution in 8.2.7 and mix with 9.00 mL of the diluent solution (see 7.4). This solution contains 2 g of uranium per litre and 0.09  $\mu\text{g}/\text{mL}$  Rh.

8.2.9 A uranium-matched reagent blank (see 8.3.2) and a control or recovery sample must be prepared with every run of samples.

8.2.10 Analyze these solutions as in 8.4 using the calibration solutions prepared in 8.3.2. The solutions must be analyzed within 8 h of preparation to minimize the effects of analyte precipitation.

8.3 *Preparation of Blanks and Calibration Standard Solutions:*

8.3.1 *For the Determination of All Elements Except Aluminum:*

8.3.1.1 *Uranium-free Reagent Blank*—Transfer 12.5 mL of concentrated nitric acid and 2.5 mL of hydrofluoric acid (40 g/100 g) into a 50 mL polypropylene volumetric flask. Continue as instructed from 8.1.5 onwards.

8.3.1.2 *Uranium-matched Calibration Blank*—Transfer 2.00 mL of the uranium standard base solution (see 6.1.8; this is equivalent to 0.20 g of uranium) into a 100 mL polypropylene volumetric flask. Add 1000  $\mu\text{L}$  of concentrated nitric acid, 200  $\mu\text{L}$  of hydrofluoric acid (40 g/100 g) and 1000  $\mu\text{L}$  of rhodium internal standard solution (see 7.3). Dilute to 100 mL with water and mix. This solution contains 2 g of uranium per litre and 0.1  $\mu\text{g}/\text{mL}$  Rh in 1 % nitric acid/0.2 % hydrofluoric acid.

8.3.1.3 *Uranium-matched Calibration Standard*—Transfer 2.00 mL of the uranium standard base solution (see 6.1.8; this is equivalent to 0.20 g of uranium) into a 100 mL polypropylene volumetric flask. Add 1000  $\mu\text{L}$  of concentrated nitric acid, 200  $\mu\text{L}$  of hydrofluoric acid (40 g/100 g), 1000  $\mu\text{L}$  of each mixed standard solution (see 7.1.1, 7.1.2 and 7.1.3) and 1000



µL of rhodium internal standard solution (see 7.3). Dilute to 100 mL with water and mix. This solution contains 2 g of uranium per litre, 0.2 µg/mL of each analyte (equivalent to 100 µg/g U) and 0.1 µg/mL Rh in 1 % nitric acid/0.2 % hydrofluoric acid.

### 8.3.2 For the Determination of Aluminum:

**8.3.2.1 Uranium-matched Reagent/Calibration Blank**—Transfer 10.00 mL of the uranium standard base solution (see 6.1.8; this is equivalent to 1.0 g of uranium) into a silica beaker. Continue as instructed from 8.2.2 onwards. The final solution contains 2 g of uranium per litre and 0.09 µg/mL Rh.

**8.3.2.2 Uranium-matched Calibration Standard**—Transfer 10.00 mL of the uranium standard base solution (see 6.1.8; this is equivalent to 1.0 g of uranium) into a silica beaker. Add 100 µL of aluminum standard solution (see 7.2) and continue as instructed from 8.2.2 onwards. The final solution contains 2 g of uranium per litre, 0.2 µg/mL Al (equivalent to 100 µg/g U) and 0.09 µg/mL Rh.

### 8.4 Measurement of Elements by ICP-MS:

**8.4.1** To avoid contamination problems when nebulizing the samples, which contain hydrofluoric acid, the nebulizer system (that is, spray chamber and nebulizer) must be made from fluorinated plastic materials (for example, TFE-fluorocarbon or polychlorotrifluoroethylene).

**8.4.1.1** Set up the ICP-MS for the analysis using the parameters given in the manufacturer's operating manual. Nebulize the uranium-matched reagent/calibration blank solution to optimize conditions using the 103 Rh internal standard.

#### INSTRUMENT OPERATING CONDITIONS

Solution Pumping Rate	Sample solution IN: 1.25 mL/min
ICP Incident Power	1400 watts
ICP Reflected Power	<10 watts
Plasma Argon Coolant	14 L/min at 70 psig
Plasma Argon Auxiliary	0.7 L/min at 70 psig
Plasma Argon Nebulizer	0.93 L/min at 40 psig
Integration Method	Valley Int.
Integration Area	0.80000 daltons (Atomic Mass Units—AMU)
Background Counts	10.0000 counts/s
Dead Time	100.000 µs
Safe Resting Mass	129.253 daltons (AMU)

**8.4.1.2** Acquire the data for all blank, calibration standard, control/recovery and sample solutions using the ICP-MS for the analytes required using the masses specified by the element menu and given in Table 1. The element menu must also contain the mass for the internal standard (normally 103 Rh). Uranium-matched calibration solutions are run at the start and end of each run.

## 9. Calculation

**9.1** The use of a uranium-free reagent blank allows the analyte concentrations in the uranium-matched blank to be ignored. For aluminum, however, a uranium-matched reagent/calibration blank must be used. The aluminum concentration of the uranium standard base solution (*Z*) must be added to the sample concentration to avoid reporting biased results (that is, if the aluminum content of the uranium standard base solution is 2 µg/g U, the uranium-matched reagent/calibration blank contains 2 µg/g U and the uranium-matched calibration standard contains 102 µg/g U. The calculation software sets these values at zero and 100 µg/g U, respectively.)

NOTE 4—Analyte counts are normalized using the internal standard count ratio (ISCR).

$$ISCR = \frac{103 \text{ Rh counts for first solution nebulized}}{103 \text{ Rh counts for each subsequent solution}} \quad (1)$$

The first solution nebulized is usually the uranium-matched reagent/calibration blank solution. The normalization and calculation of analyte concentrations is performed by the ICP-MS software.

**9.2** The element concentration, *M* (expressed as µg/g U), for the elements in 8.1 and 8.2 is calculated from:

$$M = \frac{(A_s - A_b)}{(A_c - A_u)} \times C_c \times 500 \quad (2)$$

for elements in 8.1, and

$$M = \left[ \frac{(A_s - A_u)}{(A_c - A_u)} \times C_c \times 500 \right] + Z \quad (3)$$

for aluminum in 8.2.

where:

*C<sub>c</sub>* = the concentration of the element in the uranium-matched calibration standard solution (µg/mL),

*A<sub>s</sub>* = the normalized peak signal of each element in the sample,

*A<sub>b</sub>* = the normalized peak signal of each element in the uranium-free reagent blank solution,

*A<sub>c</sub>* = the normalized peak signal of each element in the uranium-matched calibration standard solution,

*A<sub>u</sub>* = the normalized peak signal of each element in the uranium-matched reagent/calibration blank solution, and

*Z* = the aluminum content of the uranium standard base solution (µg/g U),

**9.3** Corrections for isobaric effects are not needed when impurities are at or below the upper reporting limit. The isotopes listed in Table 1 can be measured without significant isobaric interference except 48-titanium which suffers an interference from 48-calcium. Four hundred µg Ca/g U is equivalent to about 1 µg Ti/g U. Other titanium isotopes (mass 47 or mass 49) can be used but the lower reporting limit is increased to 1.5 and 2 µg/g U, respectively.

**9.4** There is a correction for the molecular interference of ArCl<sup>+</sup> on arsenic at mass 75. This can be corrected by monitoring ArCl<sup>+</sup> at mass 77 and correcting by direct proportion the contribution of ArCl<sup>+</sup> at mass 75 from the known ratio of ArCl<sup>+</sup> 77:75. This correction is only required if concentrations of chloride are greater than 10 µg/mL in the nebulized solutions.

## 10. Precision and Bias

### 10.1 Precision:

**10.1.1** The primary control samples (PCS) used were prepared in-house. They were prepared by adding a known amount of each analyte, as a solution, to characterized, high purity UO<sub>3</sub> hydrate (UOH). The UOH was then dried, blended, and ignited to the octoxide (U<sub>3</sub>O<sub>8</sub>). The U<sub>3</sub>O<sub>8</sub> was then thoroughly blended and checked for homogeneity by replicate analysis. To aid the validation of results, the standard solutions