



INTERNATIONAL STANDARD ISO 9680:1993
TECHNICAL CORRIGENDUM 1

Published 1995-03-15

INTERNATIONAL ORGANIZATION FOR STANDARDIZATION • МЕЖДУНАРОДНАЯ ОРГАНИЗАЦИЯ ПО СТАНДАРТИЗАЦИИ • ORGANISATION INTERNATIONALE DE NORMALISATION

Dental operating light

TECHNICAL CORRIGENDUM 1

Appareil d'éclairage opératoire dentaire

RECTIFICATIF TECHNIQUE 1

Technical corrigendum 1 to International Standard ISO 9680:1993 was prepared by Technical Committee ISO/TC 106, *Dentistry*, Subcommittee SC 6, *Dental equipment*.

iTeh Standards
(<https://standards.iteh.ai>)
Document Preview

Page 1, clause 2

Delete reference to ISO 4211:1979.

Page 2, subclause 5.2.2

[ISO 9680:1993/Cor 1:1995](https://standards.iteh.ai/catalog/standards/iso/e240744e-c8a5-4a89-9014-a2bfc0915fd8/iso-9680-1993-cor-1-1995)

<https://standards.iteh.ai/catalog/standards/iso/e240744e-c8a5-4a89-9014-a2bfc0915fd8/iso-9680-1993-cor-1-1995>

Insert the word "maximum" in the first sentence, to read "The area of maximum illuminance shall lie within. . .".

Page 5, subclause 5.4

Change title to "Cleaning and disinfection".

Invert the order of the first two paragraphs.

Page 10, subclause 7.3.13

Change title to "Cleaning and disinfection"

Delete text, and insert "Test according to IEC 601-1:1988, clause 44.7".

UDC 616.314-7:628.94

Ref. No. ISO 9680:1993/Cor.1:1995(E)

Descriptors: dentistry, dental equipment, lighting equipment, classification, specifications, tests, marking, packaging, technical data sheets.

© ISO 1995

Printed in Switzerland