



# SLOVENSKI STANDARD SIST EN ISO 6143:2006

01-november-2006

---

5 bU]nUd`]bcj `!`Df]a Yf`UbY`a YrcXY`nUXc`c Yj Ub`Y`]b`dfYj Yf`Ub`Y`gYghUj Y  
\_U]Vf]fb]`d`]bg\_]`na Yg]f]GC`\*`%`'`.&\$`\$%`

Gas analysis - Comparison methods for determining and checking the composition of calibration gas mixtures (ISO 6143:2001)

Gasanalyse - Vergleichsverfahren zur Bestimmung und Überprüfung der Zusammensetzung von Kalibriergasgemischen (ISO 6143:2001)

Analyse des gaz - Méthodes comparatives pour la détermination et la vérification de la composition des mélanges de gaz pour étalonnage (ISO 6143:2001)

**Ta slovenski standard je istoveten z: EN ISO 6143:2006**

---

**ICS:**

71.040.40      Kemijska analiza      Chemical analysis

**SIST EN ISO 6143:2006**      en

**iTeh STANDARD PREVIEW**  
**(standards.iteh.ai)**

[SIST EN ISO 6143:2006](https://standards.iteh.ai/catalog/standards/sist/c8ca766d-2277-417e-816c-57c8858bfe49/sist-en-iso-6143-2006)

<https://standards.iteh.ai/catalog/standards/sist/c8ca766d-2277-417e-816c-57c8858bfe49/sist-en-iso-6143-2006>

EUROPEAN STANDARD  
NORME EUROPÉENNE  
EUROPÄISCHE NORM

**EN ISO 6143**

August 2006

ICS 71.040.40

English Version

**Gas analysis - Comparison methods for determining and checking the composition of calibration gas mixtures (ISO 6143:2001)**

Analyse des gaz - Méthodes comparatives pour la détermination et la vérification de la composition des mélanges de gaz pour étalonnage (ISO 6143:2001)

Gasanalyse - Vergleichsverfahren zur Bestimmung und Überprüfung der Zusammensetzung von Kalibriergasgemischen (ISO 6143:2001)

This European Standard was approved by CEN on 21 July 2006.

CEN members are bound to comply with the CEN/CENELEC Internal Regulations which stipulate the conditions for giving this European Standard the status of a national standard without any alteration. Up-to-date lists and bibliographical references concerning such national standards may be obtained on application to the Central Secretariat or to any CEN member.

This European Standard exists in three official versions (English, French, German). A version in any other language made by translation under the responsibility of a CEN member into its own language and notified to the Central Secretariat has the same status as the official versions.

CEN members are the national standards bodies of Austria, Belgium, Cyprus, Czech Republic, Denmark, Estonia, Finland, France, Germany, Greece, Hungary, Iceland, Ireland, Italy, Latvia, Lithuania, Luxembourg, Malta, Netherlands, Norway, Poland, Portugal, Romania, Slovakia, Slovenia, Spain, Sweden, Switzerland and United Kingdom.

<https://standards.iteh.ai/catalog/standards/sist/c8ca766d-2277-417e-816c-57c8858bfe49/sist-en-iso-6143-2006>



EUROPEAN COMMITTEE FOR STANDARDIZATION  
COMITÉ EUROPÉEN DE NORMALISATION  
EUROPÄISCHES KOMITEE FÜR NORMUNG

**Management Centre: rue de Stassart, 36 B-1050 Brussels**

**EN ISO 6143:2006 (E)****Foreword**

The text of ISO 6143:2001 has been prepared by Technical Committee ISO/TC 158 "Analysis of gases" of the International Organization for Standardization (ISO) and has been taken over as EN ISO 6143:2006 by Technical Committee CEN/SS N21 "Gaseous fuels and combustible gas", the secretariat of which is held by CMC.

This European Standard shall be given the status of a national standard, either by publication of an identical text or by endorsement, at the latest by February 2007, and conflicting national standards shall be withdrawn at the latest by February 2007.

According to the CEN/CENELEC Internal Regulations, the national standards organizations of the following countries are bound to implement this European Standard: Austria, Belgium, Cyprus, Czech Republic, Denmark, Estonia, Finland, France, Germany, Greece, Hungary, Iceland, Ireland, Italy, Latvia, Lithuania, Luxembourg, Malta, Netherlands, Norway, Poland, Portugal, Romania, Slovakia, Slovenia, Spain, Sweden, Switzerland and United Kingdom.

**Endorsement notice**

The text of ISO 6143:2001 has been approved by CEN as EN ISO 6143:2006 without any modifications.

**iTeh STANDARD PREVIEW**  
**(standards.iteh.ai)**

[SIST EN ISO 6143:2006](https://standards.iteh.ai/catalog/standards/sist/c8ca766d-2277-417e-816c-57c8858bfe49/sist-en-iso-6143-2006)

<https://standards.iteh.ai/catalog/standards/sist/c8ca766d-2277-417e-816c-57c8858bfe49/sist-en-iso-6143-2006>

# INTERNATIONAL STANDARD

**ISO  
6143**

Second edition  
2001-05-01

---

---

## Gas analysis — Comparison methods for determining and checking the composition of calibration gas mixtures

*Analyse des gaz — Méthodes comparatives pour la détermination et la  
vérification de la composition des mélanges de gaz pour étalonnage*

**iTeh STANDARD PREVIEW**  
**(standards.iteh.ai)**

[SIST EN ISO 6143:2006](https://standards.iteh.ai/catalog/standards/sist/c8ca766d-2277-417e-816c-57c8858bfe49/sist-en-iso-6143-2006)

<https://standards.iteh.ai/catalog/standards/sist/c8ca766d-2277-417e-816c-57c8858bfe49/sist-en-iso-6143-2006>



Reference number  
ISO 6143:2001(E)

© ISO 2001

## ISO 6143:2001(E)

**PDF disclaimer**

This PDF file may contain embedded typefaces. In accordance with Adobe's licensing policy, this file may be printed or viewed but shall not be edited unless the typefaces which are embedded are licensed to and installed on the computer performing the editing. In downloading this file, parties accept therein the responsibility of not infringing Adobe's licensing policy. The ISO Central Secretariat accepts no liability in this area.

Adobe is a trademark of Adobe Systems Incorporated.

Details of the software products used to create this PDF file can be found in the General Info relative to the file; the PDF-creation parameters were optimized for printing. Every care has been taken to ensure that the file is suitable for use by ISO member bodies. In the unlikely event that a problem relating to it is found, please inform the Central Secretariat at the address given below.

## iTeh STANDARD PREVIEW (standards.iteh.ai)

SIST EN ISO 6143:2006

<https://standards.iteh.ai/catalog/standards/sist/c8ca766d-2277-417e-816c-57c8858bfe49/sist-en-iso-6143-2006>

© ISO 2001

All rights reserved. Unless otherwise specified, no part of this publication may be reproduced or utilized in any form or by any means, electronic or mechanical, including photocopying and microfilm, without permission in writing from either ISO at the address below or ISO's member body in the country of the requester.

ISO copyright office  
Case postale 56 • CH-1211 Geneva 20  
Tel. + 41 22 749 01 11  
Fax + 41 22 749 09 47  
E-mail [copyright@iso.ch](mailto:copyright@iso.ch)  
Web [www.iso.ch](http://www.iso.ch)

Printed in Switzerland

## Contents

Page

Foreword.....	iv
Introduction.....	v
1 Scope .....	1
2 Terms and definitions .....	1
3 Symbols and abbreviated terms .....	3
4 Principle.....	4
5 General procedure.....	6
6 Special procedures.....	14
7 Test report .....	14
Annex A (normative) Procedures for data evaluation.....	16
Annex B (informative) Examples .....	23
Annex C (informative) Computer implementation of recommended methods.....	31
Bibliography.....	33

**iTeh STANDARD PREVIEW**

**(standards.iteh.ai)**

SIST EN ISO 6143:2006

<https://standards.iteh.ai/catalog/standards/sist/c8ca766d-2277-417e-816c-57c8858bfe49/sist-en-iso-6143-2006>

## ISO 6143:2001(E)

## Foreword

ISO (the International Organization for Standardization) is a worldwide federation of national standards bodies (ISO member bodies). The work of preparing International Standards is normally carried out through ISO technical committees. Each member body interested in a subject for which a technical committee has been established has the right to be represented on that committee. International organizations, governmental and non-governmental, in liaison with ISO, also take part in the work. ISO collaborates closely with the International Electrotechnical Commission (IEC) on all matters of electrotechnical standardization.

International Standards are drafted in accordance with the rules given in the ISO/IEC Directives, Part 3.

Draft International Standards adopted by the technical committees are circulated to the member bodies for voting. Publication as an International Standard requires approval by at least 75 % of the member bodies casting a vote.

Attention is drawn to the possibility that some of the elements of this International Standard may be the subject of patent rights. ISO shall not be held responsible for identifying any or all such patent rights.

International Standard ISO 6143 was prepared by Technical Committee ISO/TC 158, *Analysis of gases*, to cancel and replace the first edition (ISO 6143:1981), of which the methods for the design and evaluation of calibrations of analytical systems have been updated and a method for estimating the uncertainty of the composition of calibration gas mixtures has been added. It also cancels and replaces ISO 6711:1981, of which entirely new methods for checking the composition of calibration gases have been specified, thus replacing the method which is no longer in use.

Annex A forms a normative part of ISO 6143. Annexes B and C are for information only.

[SIST EN ISO 6143:2006  
https://standards.iteh.ai/catalog/standards/sist/c8ca766d-2277-417e-816c-57c8858bfe49/sist-en-iso-6143-2006](https://standards.iteh.ai/catalog/standards/sist/c8ca766d-2277-417e-816c-57c8858bfe49/sist-en-iso-6143-2006)



## Introduction

In gas analysis, calibration of analytical systems, as specified in the first edition of ISO 6143, has largely been confined to the determination of a straight line through the origin, or of a straight-line segment, using only the minimum number of calibration standards (one for a straight line through the origin, two for a line segment). The approach adopted in the revision, relating to calibration as well as to uncertainty evaluation, goes far beyond this simple scheme by

- including non-linear response curves and/or functions,
- replacing interpolation by regression,
- taking into account the uncertainty on the calibration standards,
- including validation of calculated response curves and/or functions,
- calculating uncertainties by uncertainty propagation.

As a consequence of adopting non-linear response models, advanced regression techniques (errors in both variables) and uncertainty propagation, the main calculation procedures can only be performed on a computer, using a specific program. Such a program is available (see annex C). As an alternative, sufficient information is given in the document to enable the user to develop a program on his own.

[SIST EN ISO 6143:2006  
https://standards.iteh.ai/catalog/standards/sist/c8ca766d-2277-417e-816c-57c8858bfe49/sist-en-iso-6143-2006](https://standards.iteh.ai/catalog/standards/sist/c8ca766d-2277-417e-816c-57c8858bfe49/sist-en-iso-6143-2006)

**iTeh STANDARD PREVIEW**  
**(standards.iteh.ai)**

SIST EN ISO 6143:2006

<https://standards.iteh.ai/catalog/standards/sist/c8ca766d-2277-417e-816c-57c8858bfe49/sist-en-iso-6143-2006>

# Gas analysis — Comparison methods for determining and checking the composition of calibration gas mixtures

## 1 Scope

This International Standard provides methods for

- determining the composition of a calibration gas mixture by comparison with appropriate reference gas mixtures,
- calculating the uncertainty of the composition of a calibration gas mixture in relation to the known uncertainty of the composition of the reference gas mixtures with which it was compared,
- checking the composition attributed to a calibration gas mixture by comparison with appropriate reference gas mixtures,
- comparing the composition of several calibration gas mixtures, e.g. for the purpose of comparing different methods of gas mixture preparation, or for testing consistency among gas mixtures of closely related composition.

**NOTE** In principle, the method described in this document is also applicable to the analysis of (largely) unknown samples instead of prospective calibration gas mixtures (i.e. gas mixtures which are intended for use as calibration gas mixtures). Such applications, however, require appropriate care and consideration of additional uncertainty components, for example concerning the effect of matrix differences between the reference gases used for calibration and the analysed sample.

## 2 Terms and definitions

For the purposes of this International Standard, the following terms and definitions apply.

### 2.1

#### **composition**

characteristic of a gas mixture given by the kind and content of each specified mixture component (analyte) and the composition of the complementary gas (matrix)

**NOTE** In this International Standard, the analyte content is specified as a mole fraction, exclusively. Mole fractions have the advantage of being perfectly independent of the pressure and the temperature of the gas mixture. Therefore their use is recommended. However, for specific measuring systems, other composition measures (e.g. mass concentrations) may be more appropriate. Their use then requires due care concerning the dependence on pressure and temperature.

### 2.2

#### **comparison method**

method for determining the content of a specified gas mixture component (analyte) by measuring an instrumental response

**NOTE** Comparison of measuring systems requires calibration, in which the relationship between response and analyte content is established. This is achieved by measuring the response to known values of analyte content provided by reference gas mixtures.

**ISO 6143:2001(E)****2.3****calibration**

set of operations that establish, under specified conditions, the relationship between values of quantities indicated by a measuring instrument or measuring system, or values represented by a material measure or reference material, and the corresponding values realized by standards

[VIM]

**2.4****response function**

functional relationship between instrumental response and analyte content

NOTE 1 The response function can be expressed in two different ways as a calibration function or an analysis function, depending on the choice of the dependent and the independent variable.

NOTE 2 The response function is conceptual and cannot be determined exactly. It is determined approximately through calibration.

**2.4.1****calibration function**

instrumental response expressed as a function of analyte content

**2.4.2****analysis function**

analyte content expressed as a function of instrumental response

**2.5****uncertainty of measurement**

parameter, associated with the result of a measurement, that characterizes the dispersion of the values that can reasonably be attributed to the measurand

**iTeh STANDARD PREVIEW**  
**(standards.iteh.ai)**

[SIST EN ISO 6143:2006](https://standards.iteh.ai/catalog/standards/sist/c8ca766d-2277-417e-816c-57c8858bfe49/sist-en-iso-6143-2006)

[GUM]

<https://standards.iteh.ai/catalog/standards/sist/c8ca766d-2277-417e-816c-57c8858bfe49/sist-en-iso-6143-2006>

NOTE In keeping with the GUM, in this International Standard the uncertainty of the composition of a gas mixture is expressed as a standard uncertainty, i.e. as a single standard deviation.

**2.6****traceability**

property of the result of a measurement or the value attributed to a standard whereby it can be related to stated references, usually national or international standards, through an unbroken chain of comparisons all having stated uncertainties

[VIM]

**2.7****measurement standard**

material measure, measuring instrument, reference material, or measuring system, intended to define, realize, conserve, or reproduce a unit or one or more values of a quantity to serve as a reference

[VIM]

**2.8****reference standard**

standard, generally having the highest metrological quality available at a given location or in a given organization, from which measurements made there are derived

[VIM]

**2.9****working standard**

standard that is used routinely to calibrate or check material measures, measuring instruments or reference materials

[VIM]

NOTE A working standard is usually calibrated against a reference standard.

**2.10****reference material**

material or substance one or more of whose property values are sufficiently homogeneous and well established to be used for the calibration of an apparatus, the assessment of a measuring method, or for assigning values to materials

[ISO Guide 30]

**2.11****calibration gas mixture**

gas mixture whose composition is sufficiently well established and stable to be used as a working standard of composition

**2.12****reference gas mixture**

gas mixture whose composition is sufficiently well established and stable to be used as a reference standard of composition

iTeh STANDARD PREVIEW  
(standards.iteh.ai)

**3 Symbols and abbreviated terms**

$a_j$	parameters of the calibration function $F$ ( $j = 0, 1, \dots, N$ )
$b_j$	parameters of the analysis function $G$ ( $j = 0, 1, \dots, N$ )
$D$	sensitivity matrix
$F$	calibration function, $y = F(x)$ , for the specified analyte
$G$	analysis function, $x = G(y)$ , for the specified analyte
$k$	coverage factor
$L$	limit of detection
$M_{\text{cal}}$	(sample of) calibration gas mixture
$M_{\text{ref}}$	(sample of) reference gas mixture
$Q$	transform matrix
$S$	sum of weighted squared deviations
$S_{\text{res}}$	residual sum of weighted squared deviations
$t$	Student's $t$ -factor
$U(q)$	expanded uncertainty of an estimated quantity $q$ , $U(q) = ku(q)$