

SLOVENSKI STANDARD

SIST EN ISO 10360-5:2011

01-december-2011

Nadomešča:

SIST EN ISO 10360-5:2001

SIST ISO 10360-5:2002

Specifikacija geometrijskih veličin izdelka - Preskusi za sprejemljivost in ponovno overjanje koordinatnih merilnih strojev (KMS) - 5. del: Koordinatni merilni stroji, uporabljeni kot enotipalni in večtipalni sondirni sistemi (ISO 10360-5:2010)

Geometrical product specifications (GPS) - Acceptance and reverification tests for coordinate measuring machines (CMM) - Part 5: CMMs using single and multiple stylus contacting probing systems (ISO 10360-5:2010)

(standards.iteh.ai)

Geometrische Produktspezifikation (GPS) - Annahmeprüfung und Bestätigungsprüfung für Koordinatenmessgeräte (KMG) - Teil 5: Prüfung der Antastabweichungen von KMG mit berührendem Messkopfsystem (ISO 10360-5:2010)

Spécification géométrique des produits (GPS) - Essais de réception et de vérification périodique des machines à mesurer tridimensionnelles (MMT) - Partie 5: MMT utilisant des systèmes de palpage à stylet simple et à stylets multiples (ISO 10360-5:2010)

Ta slovenski standard je istoveten z: EN ISO 10360-5:2010

ICS:

| | | |
|-----------|---|---|
| 17.040.30 | Merila | Measuring instruments |
| 17.040.40 | Specifikacija geometrijskih veličin izdelka (GPS) | Geometrical Product Specification (GPS) |

SIST EN ISO 10360-5:2011

en,fr,de

iTeh STANDARD PREVIEW
(standards.iteh.ai)

[SIST EN ISO 10360-5:2011](#)

<https://standards.iteh.ai/catalog/standards/sist/3b524726-1753-4cc9-8904-fbed0693c7f5/sist-en-iso-10360-5-2011>

EUROPEAN STANDARD

EN ISO 10360-5

NORME EUROPÉENNE

EUROPÄISCHE NORM

September 2010

ICS 17.040.30

Supersedes EN ISO 10360-5:2000

English Version

Geometrical product specifications (GPS) - Acceptance and
reverification tests for coordinate measuring machines (CMM) -
Part 5: CMMs using single and multiple stylus contacting probing
systems (ISO 10360-5:2010)

Spécification géométrique des produits (GPS) - Essais de
réception et de vérification périodique des machines à
mesurer tridimensionnelles (MMT) - Partie 5: MMT utilisant
des systèmes de palpé à stylet simple ou à stylets
multiples (ISO 10360-5:2010)

Geometrische Produktspezifikation (GPS) -
Annahmeprüfung und Bestätigungsprüfung für
Koordinatenmessgeräte (KMG) - Teil 5: Prüfung der
Antastabweichungen von KMG mit berührendem
Messkopfsystem (ISO 10360-5:2010)

This European Standard was approved by CEN on 12 June 2010.

CEN members are bound to comply with the CEN/CENELEC Internal Regulations which stipulate the conditions for giving this European Standard the status of a national standard without any alteration. Up-to-date lists and bibliographical references concerning such national standards may be obtained on application to the CEN Management Centre or to any CEN member.

This European Standard exists in three official versions (English, French, German). A version in any other language made by translation under the responsibility of a CEN member into its own language and notified to the CEN Management Centre has the same status as the official versions.

<https://standards.iteh.ai/catalog/standards/sist/3b524726-1753-4cc9-8904->

CEN members are the national standards bodies of Austria, Belgium, Bulgaria, Croatia, Cyprus, Czech Republic, Denmark, Estonia, Finland, France, Germany, Greece, Hungary, Iceland, Ireland, Italy, Latvia, Lithuania, Luxembourg, Malta, Netherlands, Norway, Poland, Portugal, Romania, Slovakia, Slovenia, Spain, Sweden, Switzerland and United Kingdom.



EUROPEAN COMMITTEE FOR STANDARDIZATION
COMITÉ EUROPÉEN DE NORMALISATION
EUROPÄISCHES KOMITEE FÜR NORMUNG

Management Centre: Avenue Marnix 17, B-1000 Brussels

Contents

Page

Foreword.....3

**iTeh STANDARD PREVIEW
(standards.iteh.ai)**

[SIST EN ISO 10360-5:2011](https://standards.iteh.ai/catalog/standards/sist/3b524726-1753-4cc9-8904-fbed0693c7f5/sist-en-iso-10360-5-2011)

<https://standards.iteh.ai/catalog/standards/sist/3b524726-1753-4cc9-8904-fbed0693c7f5/sist-en-iso-10360-5-2011>

Foreword

The text of ISO 10360-5:2010 has been prepared by Technical Committee ISO/TC 213 “Dimensional and geometrical product specifications and verification” of the International Organization for Standardization (ISO) and has been taken over as EN ISO 10360-5:2010 by Technical Committee CEN/TC 290 “Dimensional and geometrical product specification and verification” the secretariat of which is held by AFNOR.

This European Standard shall be given the status of a national standard, either by publication of an identical text or by endorsement, at the latest by March 2011, and conflicting national standards shall be withdrawn at the latest by March 2011.

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. CEN [and/or CENELEC] shall not be held responsible for identifying any or all such patent rights.

This document supersedes EN ISO 10360-5:2000.

According to the CEN/CENELEC Internal Regulations, the national standards organizations of the following countries are bound to implement this European Standard: Austria, Belgium, Bulgaria, Croatia, Cyprus, Czech Republic, Denmark, Estonia, Finland, France, Germany, Greece, Hungary, Iceland, Ireland, Italy, Latvia, Lithuania, Luxembourg, Malta, Netherlands, Norway, Poland, Portugal, Romania, Slovakia, Slovenia, Spain, Sweden, Switzerland and the United Kingdom.

ITEH STANDARD PREVIEW
(standards.iteh.ai)
Endorsement notice

The text of ISO 10360-5:2010 has been approved by CEN as a EN ISO 10360-5:2010 without any modification.

SIST EN ISO 10360-5:2011
<https://standards.iteh.ai/catalog/standards/sist/3b524726-1753-4cc9-8904-fbed0693c7f5/sist-en-iso-10360-5-2011>

iTeh STANDARD PREVIEW
(standards.iteh.ai)

[SIST EN ISO 10360-5:2011](#)

<https://standards.iteh.ai/catalog/standards/sist/3b524726-1753-4cc9-8904-fbed0693c7f5/sist-en-iso-10360-5-2011>

INTERNATIONAL
STANDARDISO
10360-5Second edition
2010-09-15

**Geometrical product specifications
(GPS) — Acceptance and reverification
tests for coordinate measuring machines
(CMM) —**

Part 5:

**CMMs using single and multiple stylus
contacting probing systems****(standards.iteh.ai)**

Spécification géométrique des produits (GPS) — Essais de réception et de vérification périodique des machines à mesurer tridimensionnelles (MMT)

<https://standards.iteh.ai/catalog/standards/sist/3b524726-1753-4cc9-8904->

Partie 5: MMT utilisant des systèmes de palpation à stylet simple et à stylets multiples

Reference number
ISO 10360-5:2010(E)

© ISO 2010

ISO 10360-5:2010(E)**PDF disclaimer**

This PDF file may contain embedded typefaces. In accordance with Adobe's licensing policy, this file may be printed or viewed but shall not be edited unless the typefaces which are embedded are licensed to and installed on the computer performing the editing. In downloading this file, parties accept therein the responsibility of not infringing Adobe's licensing policy. The ISO Central Secretariat accepts no liability in this area.

Adobe is a trademark of Adobe Systems Incorporated.

Details of the software products used to create this PDF file can be found in the General Info relative to the file; the PDF-creation parameters were optimized for printing. Every care has been taken to ensure that the file is suitable for use by ISO member bodies. In the unlikely event that a problem relating to it is found, please inform the Central Secretariat at the address given below.

iTeh STANDARD PREVIEW (standards.iteh.ai)

[SIST EN ISO 10360-5:2011](https://standards.iteh.ai/catalog/standards/sist/3b524726-1753-4cc9-8904-fbed0693c7f5/sist-en-iso-10360-5-2011)

<https://standards.iteh.ai/catalog/standards/sist/3b524726-1753-4cc9-8904-fbed0693c7f5/sist-en-iso-10360-5-2011>

**COPYRIGHT PROTECTED DOCUMENT**

© ISO 2010

All rights reserved. Unless otherwise specified, no part of this publication may be reproduced or utilized in any form or by any means, electronic or mechanical, including photocopying and microfilm, without permission in writing from either ISO at the address below or ISO's member body in the country of the requester.

ISO copyright office
Case postale 56 • CH-1211 Geneva 20
Tel. + 41 22 749 01 11
Fax + 41 22 749 09 47
E-mail copyright@iso.org
Web www.iso.org

Published in Switzerland

Contents

Page

| | |
|--|-----------|
| Foreword | v |
| Introduction..... | vi |
| 1 Scope | 1 |
| 2 Normative references | 2 |
| 3 Terms and definitions | 2 |
| 4 Symbols..... | 6 |
| 5 Requirements for metrological characteristics..... | 7 |
| 5.1 Single-stylus probing error | 7 |
| 5.2 Single-stylus probing configuration..... | 8 |
| 5.3 Multi-stylus probing errors and values | 8 |
| 5.4 Multi-stylus probing configurations | 9 |
| 5.5 Styli | 9 |
| 5.6 Environmental conditions | 9 |
| 5.7 Operating conditions | 9 |
| 6 Acceptance tests and reverification tests | 10 |
| 6.1 General | 10 |
| 6.2 Single-stylus probing configuration..... | 10 |
| 6.2.1 Application | 10 |
| 6.2.2 Principle..... | 10 |
| 6.2.3 Measuring equipment | 10 |
| 6.2.4 Procedure | 11 |
| 6.2.5 Derivation of test results | 12 |
| 6.3 Fixed multi-probe and multi-stylus probing systems..... | 12 |
| 6.3.1 Principle..... | 12 |
| 6.3.2 Measuring equipment | 13 |
| 6.3.3 Procedure | 14 |
| 6.3.4 Data analysis..... | 15 |
| 6.4 Articulating probing systems..... | 16 |
| 6.4.1 Principle..... | 16 |
| 6.4.2 Measuring equipment | 17 |
| 6.4.3 Procedure | 17 |
| 6.4.4 Data analysis..... | 19 |
| 7 Compliance with specification..... | 19 |
| 7.1 Acceptance tests | 19 |
| 7.2 Reverification tests | 20 |
| 8 Applications | 20 |
| 8.1 Acceptance tests | 20 |
| 8.2 Reverification tests | 20 |
| 8.3 Interim checks | 20 |
| 9 Indication in product documentation and data sheets..... | 21 |
| Annex A (informative) Symbols and subscripts | 23 |
| Annex B (informative) Checking the probing system prior to the ISO 10360-2 test | 24 |
| Annex C (informative) Interpretation of multi-stylus test results..... | 25 |
| Annex D (normative) Maximum permissible error/limit figures | 27 |

ISO 10360-5:2010(E)

| | |
|---|-----------|
| Annex E (informative) Relation to the GPS matrix model | 28 |
| Bibliography | 30 |

**iTeh STANDARD PREVIEW
(standards.iteh.ai)**

[SIST EN ISO 10360-5:2011](https://standards.iteh.ai/catalog/standards/sist/3b524726-1753-4cc9-8904-fbed0693c7f5/sist-en-iso-10360-5-2011)

<https://standards.iteh.ai/catalog/standards/sist/3b524726-1753-4cc9-8904-fbed0693c7f5/sist-en-iso-10360-5-2011>

Foreword

ISO (the International Organization for Standardization) is a worldwide federation of national standards bodies (ISO member bodies). The work of preparing International Standards is normally carried out through ISO technical committees. Each member body interested in a subject for which a technical committee has been established has the right to be represented on that committee. International organizations, governmental and non-governmental, in liaison with ISO, also take part in the work. ISO collaborates closely with the International Electrotechnical Commission (IEC) on all matters of electrotechnical standardization.

International Standards are drafted in accordance with the rules given in the ISO/IEC Directives, Part 2.

The main task of technical committees is to prepare International Standards. Draft International Standards adopted by the technical committees are circulated to the member bodies for voting. Publication as an International Standard requires approval by at least 75 % of the member bodies casting a vote.

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. ISO shall not be held responsible for identifying any or all such patent rights.

ISO 10360-5 was prepared by Technical Committee ISO/TC 213, *Dimensional and geometrical product specifications and verification*.

This second edition cancels and replaces the first edition (ISO 10360-5:2000), which has been technically revised, and ISO/PAS 12868:2009.

ISO 10360 consists of the following parts, under the general title *Geometrical Product Specifications (GPS) — Acceptance and reverification tests for coordinate measuring machines (CMM)*:

- Part 1: Vocabulary
- Part 2: CMMs used for measuring linear dimensions
- Part 3: CMMs with the axis of a rotary table as the fourth axis
- Part 4: CMMs used in scanning measuring mode
- Part 5: CMMs using single and multiple stylus contacting probing systems
- Part 6: Estimation of errors in computing Gaussian associated features
- Part 7: CMMs equipped with video probing systems
- Part 9: CMMs with multiple probing systems

The following parts are under preparation:

- Part 8: CMMs with optical distance sensors
- Part 10: Laser trackers for measuring point-to-point distances

ISO 10360-5:2010(E)

Introduction

This part of ISO 10360 is a geometrical product specification (GPS) standard and is to be regarded as a general GPS standard (see ISO/TR 14638). It influences chain link 5 of the chains of standards of size, distance, radius, angle, form, orientation, location, run-out and datums.

For more detailed information on the relation of this part of ISO 10360 to other standards and the GPS matrix model, see Annex E.

The acceptance and reverification tests described in this part of ISO 10360 are applicable to coordinate measuring machines (CMMs) that use contacting probes, with or without multiple styli or multiple articulated-probe positions, when measuring a workpiece.

Experience has shown that the multi-stylus errors calculated using this part of ISO 10360 are significant and, at times, the dominant errors in the CMM. Owing to the virtually infinite variety of modern CMM probing system configurations, the tests specified by this part of ISO 10360 have been limited to providing a testing format only. The tests are intended to provide information on the ability of a CMM to measure a feature or features, using a contacting probe and, when relevant, using multiple styli, multiple probes or multiple articulated-probe positions.

The situations to which they are applicable include

- single-stylus probing systems,
- multiple styli connected to the CMM probe (e.g. a star),
- installations using an articulating probing system (motorized or manual) that can be prequalified,
- installations using a repeatable probe-changing system,
- installations using a repeatable stylus-changing system, and
- multi-probe installations.

It is believed that the procedures given in this part of ISO 10360 will be helpful in identifying CMM system uncertainty components for specific measurement tasks, and that the user will be able to reduce errors by removing contributing elements such as long probe extensions and styli, then retesting the new configuration set.

The tests in this part of ISO 10360 are sensitive to many errors attributable to both the CMM and the probing system, and are to be performed in addition to the length-measuring tests given in ISO 10360-2.

The primary objective is to determine the practical performance of the complete CMM and probing system. Therefore, the tests are designed to reveal measuring errors which are likely to occur when such a combined system is used on real workpieces, e.g. errors generated by the interaction between large probe-tip-offset lengths and uncorrected CMM rotation errors. The errors found here differ from those found in the E_L tests in ISO 10360-2:2009, 6.5, because with multiple styli the net CMM travel may be very different from the measured length. See Annex C for more information.

Geometrical product specifications (GPS) — Acceptance and reverification tests for coordinate measuring machines (CMM) —

Part 5: CMMs using single and multiple stylus contacting probing systems

1 Scope

This part of ISO 10360 specifies acceptance and periodic reverification tests of CMM performance with contacting probing systems and is only applicable to CMMs using

- any type of contacting probing system,
- a discrete point probing mode, and
- spherical or hemispherical stylus tip(s).

It complements ISO 10360-7, which is the module for CMMs with video probing systems, and ISO 10360-2, which is universal, i.e. not probe-type specific.

NOTE It is the CMM probing performance tests which are specified by the maximum permissible errors (MPEs), due to the impracticality of isolating the performance of the probing system from that of the CMM, even on a small artefact such as a test sphere.

This part of ISO 10360 applies to CMMs supplied with any of the following:

- a) single-stylus probing system;
- b) multi-stylus probing systems with fixed multiple styli attached to a single probe (e.g. “star” stylus);
- c) multiple probing systems such as those with a stylus for each of their probes;
- d) systems with articulating probing;
- e) stylus and probe changing systems;
- f) manual (non-driven) CMMs.

This part of ISO 10360 is not applicable to non-contacting probing systems, which require different testing procedures.

The terms “multi-stylus size error”, etc., should strictly be written “combined CMM and multi-stylus probing-system size error”, etc. For convenience, the wording has been truncated.

If it is desired to isolate the probing-system performance as far as is practical, the influence of the CMM can be minimized. See Annex C for more information.