



# SLOVENSKI STANDARD

## SIST EN 16168:2013

01-februar-2013

---

**Blato, obdelani biološki odpadki in tla - Določevanje celotnega dušika z metodo suhega sežiga**

Sludge, treated biowaste and soil - Determination of total nitrogen using dry combustion method

Schlamm, behandelter Bioabfall und Boden - Bestimmung des Gesamt-Stickstoffgehalts mittels trockener Verbrennung

Boue, biodéchet traité et sol - Détermination de la teneur totale en azote par combustion sèche

[SIST EN 16168:2013](https://standards.iteh.ai/catalog/standards/sist/c41199da-e8cb-42b0-b06b-3b82c6308767/sist-en-16168-2013)

[https://standards.iteh.ai/catalog/standards/sist/c41199da-e8cb-42b0-b06b-](https://standards.iteh.ai/catalog/standards/sist/c41199da-e8cb-42b0-b06b-3b82c6308767/sist-en-16168-2013)

[3b82c6308767/sist-en-16168-2013](https://standards.iteh.ai/catalog/standards/sist/c41199da-e8cb-42b0-b06b-3b82c6308767/sist-en-16168-2013)

**Ta slovenski standard je istoveten z: EN 16168:2012**

---

**ICS:**

13.030.20	Tekoči odpadki. Blato	Liquid wastes. Sludge
13.080.10	Kemijske značilnosti tal	Chemical characteristics of soils

**SIST EN 16168:2013**

**en,fr,de**

**iTeh STANDARD PREVIEW**  
**(standards.iteh.ai)**

SIST EN 16168:2013

<https://standards.iteh.ai/catalog/standards/sist/c41199da-e8cb-42b0-b06b-3b82c6308767/sist-en-16168-2013>

EUROPEAN STANDARD

EN 16168

NORME EUROPÉENNE

EUROPÄISCHE NORM

August 2012

ICS 13.030.01; 13.080.10

English Version

## Sludge, treated biowaste and soil - Determination of total nitrogen using dry combustion method

Boues, biodéchets traités et sols - Détermination de la teneur totale en azote par combustion sèche

Schlamm, behandelter Bioabfall und Boden - Bestimmung des Gesamt-Stickstoffgehalts mittels trockener Verbrennung

This European Standard was approved by CEN on 24 May 2012.

CEN members are bound to comply with the CEN/CENELEC Internal Regulations which stipulate the conditions for giving this European Standard the status of a national standard without any alteration. Up-to-date lists and bibliographical references concerning such national standards may be obtained on application to the CEN-CENELEC Management Centre or to any CEN member.

This European Standard exists in three official versions (English, French, German). A version in any other language made by translation under the responsibility of a CEN member into its own language and notified to the CEN-CENELEC Management Centre has the same status as the official versions.

CEN members are the national standards bodies of Austria, Belgium, Bulgaria, Croatia, Cyprus, Czech Republic, Denmark, Estonia, Finland, Former Yugoslav Republic of Macedonia, France, Germany, Greece, Hungary, Iceland, Ireland, Italy, Latvia, Lithuania, Luxembourg, Malta, Netherlands, Norway, Poland, Portugal, Romania, Slovakia, Slovenia, Spain, Sweden, Switzerland, Turkey and United Kingdom.

<https://standards.iteh.ai/catalog/standards/sist/c41199da-e8cb-42b0-b06b-3b82c6308767/sist-en-16168-2013>



EUROPEAN COMMITTEE FOR STANDARDIZATION  
COMITÉ EUROPÉEN DE NORMALISATION  
EUROPÄISCHES KOMITEE FÜR NORMUNG

Management Centre: Avenue Marnix 17, B-1000 Brussels

<b>Contents</b>	<b>Page</b>
Foreword.....	3
Introduction .....	4
1 Scope .....	5
2 Normative references .....	5
3 Terms and definitions .....	5
4 Principle.....	5
5 Interferences and sources of errors .....	5
6 Reagents.....	6
7 Apparatus .....	6
8 Sample storage and sample pretreatment .....	6
9 Procedure .....	6
9.1 General.....	6
9.2 Confirmation of calibration.....	7
9.3 Blank test.....	7
9.4 Determination of total nitrogen content.....	7
10 Calculation.....	7
11 Expression of results .....	8
12 Precision.....	8
13 Test report .....	8
Annex A (informative) Repeatability and reproducibility data.....	9
A.1 Materials used in the interlaboratory comparison study .....	9
A.2 Interlaboratory comparison results .....	10
Bibliography .....	11

iTeH STANDARD PREVIEW  
(standards.iteh.ai)

[SIST EN 16168:2013](https://standards.iteh.ai/catalog/standards/sist/c41199da-e8cb-42b0-b06b-3b82c6308767/sist-en-16168-2013)

<https://standards.iteh.ai/catalog/standards/sist/c41199da-e8cb-42b0-b06b-3b82c6308767/sist-en-16168-2013>

## Foreword

This document (EN 16168:2012) has been prepared by Technical Committee CEN/TC 400 "Project Committee - Horizontal standards in the fields of sludge, biowaste and soil", the secretariat of which is held by DIN.

This European Standard shall be given the status of a national standard, either by publication of an identical text or by endorsement, at the latest by February 2013, and conflicting national standards shall be withdrawn at the latest by February 2013.

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. CEN [and/or CENELEC] shall not be held responsible for identifying any or all such patent rights.

This document has been prepared under a mandate given to CEN by the European Commission and the European Free Trade Association.

The preparation of this document by CEN is based on a mandate by the European Commission (Mandate M/330), which assigned the development of standards on sampling and analytical methods for hygienic and biological parameters as well as inorganic and organic determinants, aiming to make these standards applicable to sludge, treated biowaste and soil as far as this is technically feasible.

According to the CEN/CENELEC Internal Regulations, the national standards organizations of the following countries are bound to implement this European Standard: Austria, Belgium, Bulgaria, Croatia, Cyprus, Czech Republic, Denmark, Estonia, Finland, Former Yugoslav Republic of Macedonia, France, Germany, Greece, Hungary, Iceland, Ireland, Italy, Latvia, Lithuania, Luxembourg, Malta, Netherlands, Norway, Poland, Portugal, Romania, Slovakia, Slovenia, Spain, Sweden, Switzerland, Turkey and the United Kingdom.

<https://standards.iteh.ai/catalog/standards/sist/c41199da-e8cb-42b0-b06b-3b82c6308767/sist-en-16168-2013>

## Introduction

This European Standard is applicable and validated for several types of matrices as indicated in Table 1 (see also Annex A for the results of the validation).

**Table 1 — Matrices for which this European Standard is applicable and validated**

Matrix	Materials used for validation
Sludge	Municipal sludge
Biowaste	Fresh compost Compost
Soil	Sludge amended soil Agricultural soil

**WARNING** — Persons using this European Standard should be familiar with usual laboratory practice. This European Standard does not purport to address all of the safety problems, if any, associated with its use. It is the responsibility of the user to establish appropriate safety and health practices and to ensure compliance with any national regulatory conditions.

**IMPORTANT** — It is absolutely essential that tests conducted according to this European Standard be carried out by suitably trained staff.

<https://standards.iteh.ai/catalog/standards/sist/c41199da-e8cb-42b0-b06b-3b82c6308767/sist-en-16168-2013>

## 1 Scope

This European Standard specifies the determination of total nitrogen (organic and inorganic) according to the procedure of Dumas in sludge, treated biowaste and soil. A typical limit of detection is 0,02 % nitrogen, and a typical limit of quantification is 0,08 % nitrogen.

## 2 Normative references

The following documents, in whole or in part, are normatively referenced in this document and are indispensable for its application. For dated references, only the edition cited applies. For undated references, the latest edition of the referenced document (including any amendments) applies.

EN 15934, *Sludge, treated biowaste, soil and waste — Calculation of dry matter fraction after determination of dry residue or water content*

EN 16179, *Sludge, treated biowaste and soil — Guidance for sample pretreatment*

EN ISO 3696, *Water for analytical laboratory use — Specification and test methods (ISO 3696)*

EN ISO 5667-15, *Water quality — Sampling — Part 15: Guidance on the preservation and handling of sludge and sediment samples (ISO 5667-15)*

ISO 18512, *Soil quality — Guidance on long and short term storage of soil samples*

## 3 Terms and definitions

For the purposes of this document, the following terms and definitions apply.

### 3.1

#### **total nitrogen**

amount of nitrogen that is released after Dumas combustion of the sample

## 4 Principle

The total nitrogen content of the material is determined by heating it to a temperature of at least 850 °C in the presence of oxygen. Mineral (inorganic) and organic nitrogen compounds are oxidized and/or volatilized. The combustion products are oxides of nitrogen (NO<sub>x</sub>) and molecular nitrogen (N<sub>2</sub>). After transforming all nitrogen into molecular nitrogen, the content of the nitrogen gas is measured using thermal conductivity or other device specific detectors.

## 5 Interferences and sources of errors

Pores in the material to be analysed are filled with air and therefore with nitrogen. Nitrogen also enters the combustion cell when it is opened to exchange the sample. Purging of the cell and the material to be analysed by inert gas is sufficient in leaving no nitrogen gas behind. Moist samples shall be used only in special cases and handled with care, as they can leach out of sample vials during the process or contaminate the device by spattering.

Fluctuations of total nitrogen may be caused by differences in nitrogen content of carrier gases used. Therefore a blank determination shall be performed after changing gas bottles and each day before starting the analytical series.

**EN 16168:2012 (E)****6 Reagents**

Use only reagents of recognized analytical grade, unless otherwise specified.

**6.1 Water**, grade 2 as specified in EN ISO 3696.

**6.2 Combustion gas (oxygen)**, free of nitrogen.

**6.3 Inert gas**, carrier gas, free of nitrogen, e.g. helium.

**6.4 Ethylenediaminetetraacetic acid (EDTA)**, or other calibration substances with known content of nitrogen, e.g. acetanilide ( $C_8H_9NO$ ), L-aspartic acid ( $C_4H_7NO_4$ ), sulfanilic acid or other amino acids or reference standards (e.g. NIST) which contain the matrix materials of investigations.

**7 Apparatus**

Usual laboratory apparatus, and in particular the following:

**7.1 Dumas apparatus**, suitable and calibrated for determination of total nitrogen by combustion of the sample at a temperature of at least 850 °C, and equipped with a detection device for measurement of nitrogen gas.

**7.2 Crucibles**, adapted to the Dumas apparatus (7.1) of variable sizes, e.g. 1 ml to 20 ml of nominal volume or special foil.

iTeh STANDARD PREVIEW  
(standards.iteh.ai)

**8 Sample storage and sample pretreatment**

Store soil samples according to ISO 18512 and sludge samples according to EN ISO 5667-15.

For the purpose of this European Standard biowaste may be stored like soil.

Pretreat the sample according to EN 16179, if not otherwise specified.

During the drying procedure or a milling process, care shall be taken to avoid losses of ammonium-N and/or nitrate-N. Prolonged drying at 40 °C or at room temperature may cause losses of nitrogen due to microbial activity within the sample. Therefore rapid drying methods should be used.

If necessary, moist samples can be used. In this case determine the dry mass on a special sample, so that the result can be referred to dry mass.

Determine the dry mass of the sample according to EN 15934.

**9 Procedure****9.1 General**

The total nitrogen contribution from water (6.1), reagents and gases (6.2, 6.3) shall be significantly less than the lowest total nitrogen content to be determined. The overall total nitrogen content of water (6.1), reagents and gases (6.2, 6.3) shall be checked by performing a blank test (see 9.3).

Ascertain homogeneity of the laboratory sample and the air-dried test sample.

In special cases the use of moist samples is necessary. Take care that the sample is homogeneous and avoid spattering during the combustion process in the Dumas apparatus.



The Dumas apparatus (7.1) is used to combust the material at a minimum temperature of 850 °C in the presence of oxygen (6.2), to reduce the nitrogen oxides, to eliminate the interfering gases and to detect the content of molecular nitrogen gas formed.

## 9.2 Confirmation of calibration

Confirm the calibration of the apparatus by analyzing calibration substances (6.4) to control the combustion and the apparatus on the day of use as follows:

Weigh an adequate amount of EDTA (6.4) or any calibrating substance and measure the amount of nitrogen. If necessary, check the linearity of the analyser with different amounts of EDTA (6.4) or the calibrating substance used.

If the confirmation value deviates by more than 3 %, the calibration should be updated.

## 9.3 Blank test

Carry out blank determinations (empty crucibles) in each series.

If a reproducible blank value is measured, it shall be considered during calibration. Otherwise, the reason shall be determined and removed.

NOTE Fingerprints on sample containers may lead to blank values.

## 9.4 Determination of total nitrogen content

Weigh a portion of the dried or moist sample to be analyzed to the nearest of 0,1 % accuracy into the crucible (7.2). The amount depends on the expected total nitrogen content and on the size of the crucible.

Carry out the analysis in accordance with the manufacturer's instructions. Use oxygen (6.2) as combustion gas. For reduction, oxidation, removal and/or fixing of combustion gases that interfere with the analysis, refer to the manufacturer's instructions.

## 10 Calculation

Calculate the content of nitrogen  $w_{\text{N}}$ , in milligrams per kilogram, using Formula (1):

$$w_{\text{N}} = \frac{X \times 100}{m \times w_{\text{dm}}} \quad (1)$$

where

- $w_{\text{N}}$  is the content of nitrogen on the basis of dry matter, expressed in milligrams per kilogram (mg/kg);
- $X$  is the primary result in milligram (mg) nitrogen;
- $m$  is the mass of the test sample in the crucible, expressed in kilograms (kg);
- $w_{\text{dm}}$  is the dry matter fraction, expressed in percent (%), determined according to EN 15934.