



SLOVENSKI STANDARD

SIST EN 15934:2012

01-december-2012

Blato, obdelani biološki odpadki, tla in odpadki - Izračun suhe snovi z določevanjem suhega ostanka ali vsebnosti vode

Sludge, treated biowaste, soil and waste - Calculation of dry matter fraction after determination of dry residue or water content

Schlamm, behandelter Bioabfall, Boden und Abfall - Berechnung des Trockenmassenanteils nach Bestimmung des Trockenrückstands oder des Wassergehalts

(standards.iteh.ai)

Boues, bio-déchets traités, sols et déchets - Calcul de la teneur en matière sèche par détermination du résidu sec ou de la teneur en eau

<https://standards.iteh.ai/catalog/standards/sist/7e0473bb-75f6-480b-9b66-8442c7320af8/sist-en-15934-2012>

Ta slovenski standard je istoveten z: EN 15934:2012

ICS:

13.030.20 Tekoči odpadki. Blato Liquid wastes. Sludge

SIST EN 15934:2012

en,fr,de

iTeh STANDARD PREVIEW
(standards.iteh.ai)

SIST EN 15934:2012

<https://standards.iteh.ai/catalog/standards/sist/7e0473bb-75f6-480b-9b66-8442c7320af8/sist-en-15934-2012>

EUROPEAN STANDARD

EN 15934

NORME EUROPÉENNE

EUROPÄISCHE NORM

August 2012

ICS 13.030.01

English Version

Sludge, treated biowaste, soil and waste - Calculation of dry matter fraction after determination of dry residue or water content

Boues, bio-déchets traités, sols et déchets - Calcul de la teneur en matière sèche par détermination du résidu sec ou de la teneur en eau

Schlamm, behandelter Bioabfall, Boden und Abfall - Berechnung des Trockenmassenanteils nach Bestimmung des Trockenrückstands oder des Wassergehalts

This European Standard was approved by CEN on 24 May 2012.

CEN members are bound to comply with the CEN/CENELEC Internal Regulations which stipulate the conditions for giving this European Standard the status of a national standard without any alteration. Up-to-date lists and bibliographical references concerning such national standards may be obtained on application to the CEN-CENELEC Management Centre or to any CEN member.

This European Standard exists in three official versions (English, French, German). A version in any other language made by translation under the responsibility of a CEN member into its own language and notified to the CEN-CENELEC Management Centre has the same status as the official versions.

CEN members are the national standards bodies of Austria, Belgium, Bulgaria, Croatia, Cyprus, Czech Republic, Denmark, Estonia, Finland, Former Yugoslav Republic of Macedonia, France, Germany, Greece, Hungary, Iceland, Ireland, Italy, Latvia, Lithuania, Luxembourg, Malta, Netherlands, Norway, Poland, Portugal, Romania, Slovakia, Slovenia, Spain, Sweden, Switzerland, Turkey and United Kingdom.



EUROPEAN COMMITTEE FOR STANDARDIZATION
COMITÉ EUROPÉEN DE NORMALISATION
EUROPÄISCHES KOMITEE FÜR NORMUNG

Management Centre: Avenue Marnix 17, B-1000 Brussels

Contents

	Page
Foreword.....	3
Introduction	4
1 Scope	5
2 Normative references	5
3 Terms and definitions	5
4 Principle	5
4.1 General	5
4.2 Principle of Method A — Drying at 105 °C (dry residue)	6
4.3 Principle of Method B — Direct Karl-Fischer-titration (water content)	6
5 Sample preparation	6
6 Method A — Drying at 105 °C	6
6.1 General	6
6.2 Interferences	6
6.3 Hazards	6
6.4 Apparatus	6
6.5 Procedure	7
6.6 Calculation of results	7
7 Method B — Direct Karl-Fischer-titration (volumetric/coulometric detection)	8
7.1 Interferences	8
7.2 Reagents	8
7.3 Apparatus	8
7.4 Procedure	8
7.4.1 Determination of the equivalence factor	8
7.4.2 Analysis of liquid samples	9
7.4.3 Analysis of solid samples	9
8 Calculation of dry matter fraction	10
9 Precision	10
10 Test report	10
Annex A (informative) Repeatability and reproducibility	11
A.1 Performance characteristics	11
A.2 Types of samples and sample preparation	12
A.3 Homogeneity and stability	13
A.4 Dry matter fraction results by different methods	14
Annex B (informative) Azeotropic distillation with toluene	15
B.1 General	15
B.2 Interferences	15
B.3 Hazards	15
B.4 Reagents	15
B.5 Apparatus	15
B.6 Procedure	16
B.7 Calculation of results	16
B.8 System performance check	16
B.9 Performance data	17
Annex C (informative) Summary of general requirements and recommendations	19

Foreword

This document (EN 15934:2012) has been prepared by Technical Committee CEN/TC 400 "Project Committee - Horizontal standards in the fields of sludge, biowaste and soil", the secretariat of which is held by DIN.

This European Standard shall be given the status of a national standard, either by publication of an identical text or by endorsement, at the latest by February 2013, and conflicting national standards shall be withdrawn at the latest by February 2013.

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. CEN [and/or CENELEC] shall not be held responsible for identifying any or all such patent rights.

This document has been prepared under a mandate given to CEN by the European Commission and the European Free Trade Association.

The preparation of this document by CEN is based on a mandate by the European Commission (Mandate M/330), which assigned the development of standards on sampling and analytical methods for hygienic and biological parameters as well as inorganic and organic determinants, aiming to make these standards applicable to sludge, treated biowaste and soil as far as this is technically feasible.

According to the CEN/CENELEC Internal Regulations, the national standards organizations of the following countries are bound to implement this European Standard: Austria, Belgium, Bulgaria, Croatia, Cyprus, Czech Republic, Denmark, Estonia, Finland, Former Yugoslav Republic of Macedonia, France, Germany, Greece, Hungary, Iceland, Ireland, Italy, Latvia, Lithuania, Luxembourg, Malta, Netherlands, Norway, Poland, Portugal, Romania, Slovakia, Slovenia, Spain, Sweden, Switzerland, Turkey and the United Kingdom.

<https://standards.iteh.ai/catalog/standards/sist/7e0473bb-75f6-480b-9b66-8442c7320af8/sist-en-15934-2012>

Introduction

In case of analysis of solids and sludge water is usually not considered as part of the sample and results are generally related to dry matter, which can be calculated by determination of the dry residue (dry matter fraction). For this purpose, and for the determination of the water content, two methods are described in this European Standard. The choice of the method depends on the type of sample and its content of volatile substances excluding water.

As a result of the validation study, the determination of water content by azeotropic distillation has been replaced by Karl-Fischer-titration. Nevertheless, the distillation may be useful in certain cases. This method is described in Annex B (informative).

This European Standard is applicable and validated for several types of matrices as indicated in Table 1 (see also Annex A for the results of the validation).

Table 1 — Matrices for which this European Standard is applicable and validated

Matrix	Materials used for validation
Sludge (only method A)	Municipal sludge
Biowaste (only method A)	Fresh compost
Soil (only method A)	Sludge amended soil
Waste (method A and B)	Contaminated soil, Dredged sludge, Nickel sludge, Filter cake, Distillation residue, Drilling emulsion

WARNING — Persons using this European Standard should be familiar with usual laboratory practice. This European Standard does not purport to address all of the safety problems, if any, associated with its use. It is the responsibility of the user to establish appropriate safety and health practices and to ensure compliance with any national regulatory conditions.

IMPORTANT — It is absolutely essential that tests conducted according to this European Standard be carried out by suitably trained staff.

1 Scope

This European Standard specifies methods for the calculation of the dry matter fraction of sludge, treated biowaste, soil and waste for which the results of performed analysis are to be calculated to the dry matter basis. Depending on the nature and origin of the sample, the calculation is based on a determination of the dry residue (Method A) or a determination of the water content (Method B). It applies to samples containing more than 1 % (mass fraction) of dry residue or more than 1 % (mass fraction) of water.

Method A applies to sludge, treated biowaste, soil and solid waste. Method B applies to liquid waste and to samples which are suspected or known to contain volatiles except for water.

2 Normative references

The following documents, in whole or in part, are normatively referenced in this document and are indispensable for its application. For dated references, only the edition cited applies. For undated references, the latest edition of the referenced document (including any amendments) applies.

EN 15002, *Characterization of waste — Preparation of test portions from the laboratory sample*

EN 16179, *Sludge, treated biowaste and soil — Guidance for sample pretreatment*

3 Terms and definitions

For the purposes of this document, the following terms and definitions apply.

3.1

dry residue

remaining mass fraction of a sample after a drying process at 105 °C under specified conditions

3.2

water content

mass fraction of water in a sample determined by the method after drying at 105 °C or by Karl-Fischer-titration under specified conditions

3.3

dry matter fraction

mass fraction of a sample excluding water expressed as mass fraction calculated by determination of dry residue or water content

4 Principle

4.1 General

Depending on the origin (sludge, biowaste, soil, waste) and the nature of the sample (liquid, solid or multiphase) either the water content or the dry residue is to be determined. The results from the determination of water content or dry residue are used to calculate the dry matter fraction. In case of multiphase (waste) samples these samples shall be homogenized. If homogenization is not possible, a phase separation according to EN 15002 shall be applicable and the phases are analysed separately.

EN 15934:2012 (E)

4.2 Principle of Method A — Drying at 105 °C (dry residue)

The sample is dried to a constant mass at 105 °C. This method applies to solid samples and samples which become solid during the drying process. The method generally applies to soil, sludge and treated biowaste. Solid samples containing amounts of volatiles which are expected to affect the results shall be analysed by Method B.

4.3 Principle of Method B — Direct Karl-Fischer-titration (water content)

The method usually applies to liquid samples and homogenised multiphase samples. It can also be used for solid samples in case the “water content” is of interest and in case volatiles are expected to interfere with the determination according to Method A. The water content of a sample is determined by direct Karl-Fischer-titration and either volumetric or coulometric detection. Liquid samples are directly added to the cell, while solid samples are extracted by use of methanol.

5 Sample preparation

Pretreat the samples according to EN 16179 or EN 15002, if not otherwise specified.

6 Method A — Drying at 105 °C**6.1 General**

When performing any kind of analysis on samples, the dry matter fraction shall be determined simultaneously on identical test portions, in order to base all other test results on the determined dry matter content of each individual sample.

iTeh STANDARD PREVIEW
(standards.iteh.ai)

6.2 Interferences

[SIST EN 15934:2012
https://standards.iteh.ai/catalog/standards/sist/7e0473bb-75f6-480b-9b66-8442c7320af8/sist-en-15934-2012](https://standards.iteh.ai/catalog/standards/sist/7e0473bb-75f6-480b-9b66-8442c7320af8/sist-en-15934-2012)

The samples can change during the drying process, e.g. by absorption of carbon dioxide in the case of alkaline samples, or of oxygen by reducing substances. Volatile compounds evaporating at 105 °C are measured as water using this procedure.

NOTE For alkaline samples or samples susceptible to oxidation the drying can be carried out in a nitrogen atmosphere.

6.3 Hazards

Flammable or explosive gases may be released in the drying process.

6.4 Apparatus

6.4.1 Drying system thermostatically controlled and capable of maintaining a temperature of (105 ± 5) °C; e.g. drying oven, infrared system, halogen lamp system.

6.4.2 Desiccator with an active drying agent such as silica gel.

6.4.3 Precision balance with an accuracy of 1 mg or better.

6.4.4 Evaporating dish or crucible. Temperature tolerant laboratory vessel withstanding 105 °C. Suitable materials are metallic, ceramic, borosilicate glass or quartz.

NOTE 1 If the sample is to be ashed following this procedure, it should be ensured that the material will withstand the elevated temperatures.

NOTE 2 The material of the evaporating dish or crucible should be compatible with the nature of the sample to avoid interferences upon weighing due to possible oxidations.

6.5 Procedure

Place an evaporating dish or crucible (6.4.4) in the drying system (6.4.1) set at (105 ± 5) °C for a minimum of 30 min. After cooling in the desiccator (6.4.2) to ambient temperature, weigh the basin to the nearest 1 mg.

Depending on the expected dry residue, weigh into the evaporating dish or crucible (6.4.4) a suitable amount of material so that the residue obtained has a weight of at least 0,5 g. Weigh the loaded basin to the nearest 1 mg.

Place the evaporating dish or crucible (6.4.4) containing the sample in the drying system (6.4.1) set at (105 ± 5) °C until the residue appears dry, typically overnight.

NOTE For some samples e.g. sludge, there is a risk of formation of a cake surface, which hinders an even drying. To avoid this, a glass rod can be weighed along with the dish. If cake formation occurs during drying, the glass rod is used to stir the sample, to break up the cake. This should be repeated as necessary. The rod should be reweighed afterwards.

After cooling in the desiccator (6.4.2) weigh the evaporating dish or crucible and contents for the first time.

The dry residue shall be regarded as constant if the mass obtained after further 1 h of drying does not differ by more than 0,5 % of the previous value or 2 mg whichever is the greater. Otherwise repeat the drying process.

In case of weight inconstancy after three cycles the drying process may be stopped (after at least 16 h). The result of the last weighing shall be recorded in the test report. A drying period of 16 h is sufficient.

In case of infrared or halogen lamp drying systems the corresponding manufacturer's instructions apply. The technique of choice shall be recorded in the test report.

Other techniques than oven drying, infrared or halogen lamp drying are allowed, provided they are proven to give comparable results.

6.6 Calculation of results

The dry residue is calculated according to Formula (1):

$$w_{\text{dr}} = \frac{m_{\text{c}} - m_{\text{a}}}{m_{\text{b}} - m_{\text{a}}} \cdot f \quad (1)$$

where

w_{dr} is the dry residue of the sample, expressed as mass fraction in percent (%) or in grams per kilogram (g/kg);

m_{a} is the mass of the empty dish or crucible, expressed in grams (g);

m_{b} is the mass of the dish or crucible containing the sample, expressed in grams (g);

m_{c} is the mass of the dish or crucible containing the dried sample, expressed in grams (g);

f is the conversion factor $f = 100$ for expression of results as mass fraction in percent (%) and $f = 1\,000$ for expression of results in grams per kilogram (g/kg).

EN 15934:2012 (E)

7 Method B — Direct Karl-Fischer-titration (volumetric/coulometric detection)**7.1 Interferences**

Substances like peroxides, oxides, hydroxides, ammonia, carbonates, hydrogen carbonates, disulfites, nitrite, sulfite, thiosulfate, hydrazine and derivatives, iron (II), copper (I), tin (II), silanols, arsenite, arsenate, selenite, tellurite, phenols, boric compounds, aldehydes, acetone and other reactive ketones may cause interferences. Especially all reductive or oxidative reagents are potentially interfering. In case of concern about interferences results can be verified by azeotropic distillation (Annex B).

Most of the interferences are reduced in case of extraction with methanol before analysis.

7.2 Reagents

All reagents used shall be at least of analytical grade and suitable for their specific purposes. Hygroscopic substances shall be stored in a desiccator.

7.2.1 Karl-Fischer reagent as recommended by provider of KF-system.

7.2.2 Methanol. The methanol used shall not contain more than a mass fraction of 0,005 % water.

7.2.3 Toluene if necessary to solve a sample.

7.3 Apparatus

Usual laboratory glassware and the following equipment shall be used.

7.3.1 Karl-Fischer-titrator.

7.3.2 Shaking device. <https://standards.iteh.ai/catalog/standards/sist/7e0473bb-75f6-480b-9b66-8442c7320af8/sist-en-15934-2012>

7.3.3 Weighing pipette.

7.3.4 Conical flasks, 250 ml with ground glass neck and stopper.

7.4 Procedure**7.4.1 Determination of the equivalence factor**

Fill the titration vessel (7.3.1) with methanol (7.2.2). Stir without splattering. Titrate with Karl-Fischer reagent (7.2.1) to the end point. Using a weighing pipette (7.3.3) add approximately 80 mg of water (W) with an accuracy of 0,1 %. Titrate to the end point and record the consumption (V_1).

The equivalence factor F is calculated according to Formula (2):

$$F = \frac{W}{V_1} \quad (2)$$

where

F is the equivalence factor, expressed in milligrams of water per millilitre (mg/ml);

W is the quantity of water, expressed in milligrams (mg);

V_1 is the volume of the titration fluid, expressed in millilitres (ml).

The equivalence factor shall be determined weekly.

Instead of water other suitable standards with defined water content are applicable. In case of coulometric detection no titer adjustment is required.

7.4.2 Analysis of liquid samples

Fill the titration vessel (7.3.1) with methanol (7.2.2). Stir without splattering. Titrate with Karl-Fischer reagent (7.2.1) to the end point. Add an appropriate amount of the liquid sample M with an accuracy of 0,1 % (in case of water contents less than 10 % about 1 g of sample is appropriate).

Record the added volume of the titration liquid (V) for the sample.

The water content of the liquid sample is calculated according to Formula (3):

$$w_w = \frac{V \times F}{M \times 1000} \times 100 \quad (3)$$

where

w_w is the water content, expressed as mass fraction in percent (%);

V is the added volume of the titration liquid, expressed in millilitres (ml);

F is the equivalence factor (according to Formula (2)), expressed in milligrams water per millilitre (mg/ml);

M is the mass of an aliquot of the sample, expressed in grams (g).

The dry matter fraction is calculated by Formula (6).
<https://standards.iteh.ai/catalog/standards/sist/7e0473bb-75f6-480b-9b66-8442c7320af8/sist-en-15934-2012>

7.4.3 Analysis of solid samples

Weigh approximately 10 g solid sample M to 0,1 % accurately into a conical flask (7.3.4). Pipette 50,0 ml methanol (7.2.2). Shake for 10 min in a shaking machine (7.3.2). Fill the titration vessel with methanol (7.2.2). Stir without splattering. Titrate with Karl-Fischer reagent (7.2.1) to the end point. Pipette an appropriate volume (V_E) of the extract into the sample vessel (in case of water contents less than 5 % about 10 ml of extract is appropriate).

Note the added volume of titration liquid (V) for the extract. Also perform a blank run. Note the added volume of the titration liquid V_0 for the blank.

The water content of the solid sample, expressed as mass fraction in percent, is calculated according to Formula (4):

$$w_w = \frac{(V - V_0) \times F \times 50}{M \times 1000 \times V_E} \times 100 \quad (4)$$

where

w_w is the water content of phase i , expressed as mass fraction in percent (%);

V is the added volume of the titration liquid, expressed in millilitres (ml);

V_0 is the added volume of the titration liquid in blank run, expressed in millilitres (ml);