

First edition  
2003-12-01

Corrected version  
2004-06-15

---

---

**Solid mineral fuels — Determination of  
total cadmium content of coal**

*Combustibles minéraux solides — Dosage du cadmium total dans le  
charbon*

**iTeh STANDARD PREVIEW**  
**(standards.iteh.ai)**

[SIST ISO 15238:2005](https://standards.iteh.ai/catalog/standards/sist/ed45e1f9-a6e3-4a96-800f-b667cbc26254/sist-iso-15238-2005)

[https://standards.iteh.ai/catalog/standards/sist/ed45e1f9-a6e3-4a96-800f-  
b667cbc26254/sist-iso-15238-2005](https://standards.iteh.ai/catalog/standards/sist/ed45e1f9-a6e3-4a96-800f-b667cbc26254/sist-iso-15238-2005)



Reference number  
ISO 15238:2003(E)

© ISO 2003

**PDF disclaimer**

This PDF file may contain embedded typefaces. In accordance with Adobe's licensing policy, this file may be printed or viewed but shall not be edited unless the typefaces which are embedded are licensed to and installed on the computer performing the editing. In downloading this file, parties accept therein the responsibility of not infringing Adobe's licensing policy. The ISO Central Secretariat accepts no liability in this area.

Adobe is a trademark of Adobe Systems Incorporated.

Details of the software products used to create this PDF file can be found in the General Info relative to the file; the PDF-creation parameters were optimized for printing. Every care has been taken to ensure that the file is suitable for use by ISO member bodies. In the unlikely event that a problem relating to it is found, please inform the Central Secretariat at the address given below.

**iTeh STANDARD PREVIEW**  
**(standards.iteh.ai)**

[SIST ISO 15238:2005](https://standards.iteh.ai/catalog/standards/sist/ed45e1f9-a6e3-4a96-800f-b667cbc26254/sist-iso-15238-2005)

<https://standards.iteh.ai/catalog/standards/sist/ed45e1f9-a6e3-4a96-800f-b667cbc26254/sist-iso-15238-2005>

© ISO 2003

All rights reserved. Unless otherwise specified, no part of this publication may be reproduced or utilized in any form or by any means, electronic or mechanical, including photocopying and microfilm, without permission in writing from either ISO at the address below or ISO's member body in the country of the requester.

ISO copyright office  
Case postale 56 • CH-1211 Geneva 20  
Tel. + 41 22 749 01 11  
Fax + 41 22 749 09 47  
E-mail [copyright@iso.org](mailto:copyright@iso.org)  
Web [www.iso.org](http://www.iso.org)

Published in Switzerland

## Foreword

ISO (the International Organization for Standardization) is a worldwide federation of national standards bodies (ISO member bodies). The work of preparing International Standards is normally carried out through ISO technical committees. Each member body interested in a subject for which a technical committee has been established has the right to be represented on that committee. International organizations, governmental and non-governmental, in liaison with ISO, also take part in the work. ISO collaborates closely with the International Electrotechnical Commission (IEC) on all matters of electrotechnical standardization.

International Standards are drafted in accordance with the rules given in the ISO/IEC Directives, Part 2.

The main task of technical committees is to prepare International Standards. Draft International Standards adopted by the technical committees are circulated to the member bodies for voting. Publication as an International Standard requires approval by at least 75 % of the member bodies casting a vote.

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. ISO shall not be held responsible for identifying any or all such patent rights.

ISO 15238 was prepared by Technical Committee ISO/TC 27, *Solid mineral fuels*, Subcommittee SC 5, *Methods of analysis*.

In this corrected version of ISO 15238:2003, Table 1 has been amended by the replacement of the published value of  $0,6\omega_{Cd} - 1$  for the Repeatability limit with the correct value of  $0,16\omega_{Cd} - 1$ .

[SIST ISO 15238:2005](https://standards.iteh.ai/catalog/standards/sist/ed45e1f9-a6e3-4a96-800f-b667cbc26254/sist-iso-15238-2005)

<https://standards.iteh.ai/catalog/standards/sist/ed45e1f9-a6e3-4a96-800f-b667cbc26254/sist-iso-15238-2005>

## Introduction

Cadmium occurs naturally in coal. It is an element that can potentially be released during the combustion process.

Quantitative recovery can be achieved by strict adherence to the procedure set out in this International Standard.

The method may also be suitable for other elements, including lead and arsenic, although the choice of matrix modifier may vary.

## iTeh STANDARD PREVIEW (standards.iteh.ai)

[SIST ISO 15238:2005](https://standards.iteh.ai/catalog/standards/sist/ed45e1f9-a6e3-4a96-800f-b667cbc26254/sist-iso-15238-2005)

<https://standards.iteh.ai/catalog/standards/sist/ed45e1f9-a6e3-4a96-800f-b667cbc26254/sist-iso-15238-2005>

# Solid mineral fuels — Determination of total cadmium content of coal

## 1 Scope

This International Standard specifies a procedure for the determination of the total cadmium content of coal.

This procedure has not been validated with coals that spontaneously ignite. Prior to use with such sample types, users should validate the method.

## 2 Normative references

The following referenced documents are indispensable for the application of this document. For dated references, only the edition cited applies. For undated references, the latest edition of the referenced document (including any amendments) applies.

ISO 331, *Coal — Determination of moisture in the analysis sample — Direct gravimetric method*

ISO 1170, *Coal and coke — Calculation of analyses to different bases*

ISO 3696, *Water for analytical laboratory use — Specifications and test methods*

ISO 5068, *Brown coals and lignites — Determination of moisture content — Indirect gravimetric method*

ISO 5069-2, *Brown coals and lignites — Principles of sampling — Part 2: Sample preparation for determination of moisture content and for general analysis*

ISO 13909-2, *Hard coal and coke — Mechanical sampling — Part 2: Coal — Sampling from moving streams*

## 3 Principle

The coal sample is ashed, followed by dissolution of the ash in a mixture of hydrochloric, nitric and hydrofluoric acids. The cadmium species present are quantified by graphite furnace atomic absorption spectroscopy.

## 4 Reagents

**WARNING — Care should be exercised when handling the reagents, many of which are toxic and corrosive.**

During the analysis, unless otherwise stated, use only reagents of recognized analytical grade.

**4.1 Water**, conforming to the requirements of Grade 1 of ISO 3696.

**4.2 Aqua regia solution.** Mix 1 part by volume of concentrated nitric acid solution (relative density 1,42), 3 parts of concentrated hydrochloric acid solution (relative density 1,19), and 1 part water.

**4.3 Concentrated hydrofluoric acid solution** (relative density 1,15).

**WARNING — Hydrofluoric acid is an extremely aggressive chemical which shall be handled with care.**

**4.4 Boric acid solution**, saturated. Dissolve 60 g of boric acid in 1 litre of hot water, cool and allow to stand for three days before decanting the clear solution.

**4.5 Cadmium standard stock solution**, 10 µg/ml. Prepare the stock solution from a high purity metal oxide or salt, having a purity greater than 99,9 %. Dilute to volume with 1 % (V/V) nitric acid solution.

NOTE The cadmium standard stock solution may also be prepared from commercially available certified cadmium solution.

**4.6 Cadmium standard solution**, 0,1 µg/ml. Dilute 10,0 ml of cadmium standard stock solution (4.5) to 1,0 litre with 1 % (V/V) nitric acid solution.

## 5 Apparatus

**5.1 Balance**, capable of weighing to the nearest 0,1 mg.

**5.2 Silica or platinum combustion crucible** (45 × 35 × 14 mm).

**5.3 Muffle furnace**, with a temperature control.

**5.4 Plastics bottles**, made of high density polyethylene (HDPE) or fluorinated ethylene propylene (FEP), of 125 ml capacity, with screw-cap lids, capable of withstanding 130 °C when sealed and containing liquid.

**5.5 Volumetric flasks**, made of high density polyethylene (HDPE) or fluorinated ethylene propylene (FEP), of 100 ml capacity.

**5.6 Boiling-water bath.**

**5.7 Graphite-furnace atomic absorption spectrometer**, with background correction.

## 6 Preparation of sample

The test sample is the general analysis test sample prepared in accordance with ISO 5069-2 or ISO 13909-2 as appropriate. Ensure that the moisture content of the sample is in equilibrium with the laboratory atmosphere, exposing it if necessary, in a thin layer for the minimum time required to achieve equilibrium.

Before commencing the determination, mix the equilibrated sample for at least 1 min, preferably by mechanical means.

If the results are to be calculated other than on the "air-dried" basis (see Clause 8), then, after weighing the test portion, determine the moisture content of a further portion of the test sample by the method described in ISO 331 or ISO 5068 as appropriate.

## 7 Procedure

### 7.1 Ashing of coal sample

Weigh, to the nearest 0,1 mg, the amount of sample necessary to obtain a minimum of 0,2 g ash in one or more open 50 ml silica combustion crucibles (5.2). The coal layer in the crucible shall not exceed a thickness of 0,15 g/cm<sup>2</sup>. Place the crucibles in a cold muffle furnace (5.3), set to attain (450 ± 10) °C in about 1 h, and

maintain at that temperature for a further 2 h. After cooling and weighing, calculate the percentage of ash in the coal. Carefully mix the ash in the crucibles before transferring it to a sample bottle.

NOTE The ashing time of 3 h is a minimum. Ashing overnight (18 h) is allowed.

## 7.2 Preparation of test solution

Place approximately 0,2 g ( $\pm 0,1$  mg) of ash in a 125 ml plastics bottle (5.4). Add 3 ml of aqua regia (4.2), 5 ml of concentrated hydrofluoric acid solution (4.3), and tighten the screw cap. Place the bottle on a boiling water bath for 2 h. Remove the bottle and allow it to cool to ambient temperature, cautiously open, then add 50 ml of saturated boric acid solution (4.4). Warm again for approximately 1 h; if a residue remains at this stage it may be ignored. Remove the bottle from the water bath and cool again to ambient temperature before transferring the solution quantitatively to a 100 ml plastics volumetric flask, and diluting to 100 ml with water.

Care should be exercised when adding the acid mixture to the ash since the 450 °C ashing temperature may not have decomposed all carbonates present in the coal.

NOTE 1 If desired, it is permitted to prepare the solutions gravimetrically, adjusting the calculations in the appropriate manner.

NOTE 2 As an alternative method for the dissolution of the prepared ash, microwave dissolution or acid dissolution in a pressurized vessel can be used, if quantitative recovery of cadmium can be demonstrated.

## 7.3 Preparation of blank solution

Prepare a blank solution by following exactly the dissolution procedure described above (7.2) but omitting the coal (ash) sample.

## 7.4 Graphite-furnace atomic absorption analysis

[SIST ISO 15238:2005](https://standards.iteh.ai/catalog/standards/sist/ed45e1f9-a6e3-4a96-800f-b667cbc26254/sist-iso-15238-2005)

### 7.4.1 Calibration

<https://standards.iteh.ai/catalog/standards/sist/ed45e1f9-a6e3-4a96-800f-b667cbc26254/sist-iso-15238-2005>

Prepare standards by diluting aliquot portions of the cadmium standard stock solution (4.5), 3 ml of aqua regia solution (4.2), 5 ml of concentrated hydrofluoric acid solution (4.3) and 50 ml of saturated boric acid solution (4.4) to 100 ml with water.

Standards containing 0,0, 1,0, 2,5 and 5,0 µg/l have been found suitable. The standards used should be verified for each instrument.

### 7.4.2 Measurement of standards

Measure the absorbance of each standard solution using the graphite-furnace atomic absorption spectrometer (5.7). Plot the corresponding absorbance response for each matrix-matched standard against concentration, producing a calibration curve for the instrument. Sample responses are then compared directly with the calibration.

The performance of the instrument shall be verified by the analysis of certified reference materials.

NOTE 1 No additional matrix modifier is used for cadmium determination and the instrument parameters allow for this by using a low char temperature.

NOTE 2 An alternative quantification procedure is the use of standard analyte additions to the final solution.

### 7.4.3 Measurement of test solution

Repeat the procedure described in 7.4.2 for the test solution (7.2). Determine the cadmium concentration in the test solution using the calibration graph (7.4.2).

#### 7.4.4 Measurement of blank solution

Repeat the procedure described in 7.4.2 for the blank (7.3). Determine the cadmium concentration in the blank solution using the calibration graph (7.4.2)

### 8 Expression of results

#### 8.1 Ash content

Calculate the ash content,  $\omega_A$ , as a percentage by mass of the coal sample, using the equation:

$$\omega_A = \frac{(m_c - m_a)}{(m_b - m_a)} \times 100$$

where

$m_a$  is the mass of the ashing combustion crucible;

$m_b$  is the mass of the ashing combustion crucible plus coal before ashing;

$m_c$  is the mass of the ashing combustion crucible plus coal after ashing.

#### 8.2 Cadmium content

**iTeh STANDARD PREVIEW**  
(standards.iteh.ai)

Calculate the cadmium content,  $\omega_{Cd}$ , in ng/g, in the coal test sample using the equation:

$$\omega_{Cd} = \omega_A \frac{(\rho_t - \rho_b)}{1000 m} \times 1000$$

SIST ISO 15238:2005  
<https://standards.iteh.ai/catalog/standards/sist/ed45e1f9-a6e3-4a96-800f-b667cbc26254/sist-iso-15238-2005>

where

$\omega_A$  is the ash content of the sample as analysed, in percentage by mass;

$\rho_t$  is the cadmium concentration of the test solution, in  $\mu\text{g/l}$ ;

$\rho_b$  is the cadmium concentration of the blank solution, in  $\mu\text{g/l}$ ;

$m$  is the mass of the sample as analysed, in grams.

Report the result as the mean of duplicate determinations to the nearest 20 ng/g on the “as analysed” basis.

For calculation of the results to bases other than “as analysed”, see ISO 1170.

### 9 Precision

#### 9.1 Repeatability limit

The means of the results of duplicate determinations, carried out at different times, in the same laboratory, by the same operator, with the same apparatus, on representative portions taken from the same general analysis sample, should not differ by more than the value given in Table 1.



## 9.2 Reproducibility critical difference

The results of duplicate determinations, carried out in each of two laboratories, on representative portions taken from the same analysis sample after the last stage of sample preparation, should not differ by more than the value given in Table 1.

Table 1

	Maximum acceptable differences between results obtained (calculated to the same moisture content)	
	Repeatability limit	Reproducibility critical difference
Cadmium content (ng/g)	$0,16\omega_{\text{Cd}} - 1$	$0,22\omega_{\text{Cd}} + 3$

## 10 Test report

The test report shall include the following information:

- identification of the sample tested;
- the method used by reference to this International Standard, i.e. ISO 15238:2003;
- the date of the determination;
- the results and the method of expression used.

[SIST ISO 15238:2005](https://standards.iteh.ai/catalog/standards/sist/ed45e1f9-a6e3-4a96-800f-b667cbc26254/sist-iso-15238-2005)  
<https://standards.iteh.ai/catalog/standards/sist/ed45e1f9-a6e3-4a96-800f-b667cbc26254/sist-iso-15238-2005>