

# SLOVENSKI STANDARD SIST EN 15692:2009

01-julij-2009

Etanol kot komponenta za dodajanje motornemu bencinu - Določevanje vode - Karl Fischerjeva metoda potenciometrične titracije

Ethanol as a blending component for gasoline - Determination of water content - Karl Fischer potentiometric titration method

Ethanol zu Verwendung als Blendkomponente in Ottokraftstoff - Bestimmung des Wassergehaltes - Potentiometrische Titration nach Karl Fischer VV

Ethanol pour essence au plomb - Détermination de la teneur de l'eau - Méthode de Karl Fischer titrage par potentiométrie

https://standards.iteh.ai/catalog/standards/sist/724d92a2-b420-4634-9cc0-

Ta slovenski standard je istoveten z: EN 15692-2009

ICS:

71.080.60 Alkoholi. Etri Alcohols. Ethers 75.160.20 Tekoča goriva Liquid fuels

SIST EN 15692:2009 en,fr,de

**SIST EN 15692:2009** 

# iTeh STANDARD PREVIEW (standards.iteh.ai)

SIST EN 15692:2009

https://standards.iteh.ai/catalog/standards/sist/724d92a2-b420-4634-9cc0-ad2c6bc38782/sist-en-15692-2009

**EUROPEAN STANDARD** 

EN 15692

NORME EUROPÉENNE

**EUROPÄISCHE NORM** 

**April 2009** 

ICS 75.160.20

#### **English Version**

# Ethanol as a blending component for petrol - Determination of water content - Karl Fischer potentiometric titration method

Ethanol comme base de mélange à l'essence -Détermination de la teneur en eau - Méthode de titrage potentiométrique Karl Fischer

Ethanol zu Verwendung als Blendkomponente in Ottokraftstoff - Bestimmung des Wassergehaltes -Potentiometrische Titration nach Karl Fischer

This European Standard was approved by CEN on 19 March 2009.

CEN members are bound to comply with the CEN/CENELEC Internal Regulations which stipulate the conditions for giving this European Standard the status of a national standard without any alteration. Up-to-date lists and bibliographical references concerning such national standards may be obtained on application to the CEN Management Centre or to any CEN member.

This European Standard exists in three official versions (English, French, German). A version in any other language made by translation under the responsibility of a CEN member into its own language and notified to the CEN Management Centre has the same status as the official versions.

CEN members are the national standards bodies of Austria, Belgium, Bulgaria, Cyprus, Czech Republic, Denmark, Estonia, Finland, France, Germany, Greece, Hungary, Iceland, Ireland, Italy, Latvia, Lithuania, Luxembourg, Malta, Netherlands, Norway, Poland, Portugal, Romania, Slovakia, Slovenia, Spain, Sweden, Switzerland and United Kingdom.

https://standards.iteh.ai/catalog/standards/sist/724d92a2-b420-4634-9cc0ad2c6bc38782/sist-en-15692-2009



EUROPEAN COMMITTEE FOR STANDARDIZATION COMITÉ EUROPÉEN DE NORMALISATION EUROPÄISCHES KOMITEE FÜR NORMUNG

Management Centre: Avenue Marnix 17, B-1000 Brussels

Con	tents P	Page
Foreword		3
1	Scope	4
2	Normative references	4
3	Terms and definitions	4
4	Principle	4
5	Reagents and materials	4
6	Apparatus	5
7	Sampling and sample handling	5
8 8.1 8.2	Procedure	6
9	Calculation	7
10	Expression of results	7
11 11.1 11.2	Precision of results Teh STANDARD PREVIEW  General Standards.iteh.ai  Repeatability, r	7 8
11.3	Reproducibility, RSIST EN 15692-2009	8
12	Test report	8
Biblio	graphy	9

#### **Foreword**

This document (EN 15692:2009) has been prepared by Technical Committee CEN/TC 19 "Gaseous and liquid fuels, lubricants and related products of petroleum, synthetic and biological origin", the secretariat of which is held by NEN.

This European Standard shall be given the status of a national standard, either by publication of an identical text or by endorsement, at the latest by October 2009, and conflicting national standards shall be withdrawn at the latest by October 2009.

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. CEN [and/or CENELEC] shall not be held responsible for identifying any or all such patent rights.

This document was prepared by CEN/TC 19's Ethanol Task Force under its Working Group 21 and is based on ISO 760 [1]. It is developed as an alternative to EN 15489 [2], delivering a method more widely used in the alcohol and beverage industry environment.

According to the CEN/CENELEC Internal Regulations, the national standards organizations of the following countries are bound to implement this European Standard: Austria, Belgium, Bulgaria, Cyprus, Czech Republic, Denmark, Estonia, Finland, France, Germany, Greece, Hungary, Iceland, Ireland, Italy, Latvia, Lithuania, Luxembourg, Malta, Netherlands, Norway, Poland, Portugal, Romania, Slovakia, Slovenia, Spain, Sweden, Switzerland and the United Kingdom.

(standards.iteh.ai)

SIST EN 15692:2009 https://standards.iteh.ai/catalog/standards/sist/724d92a2-b420-4634-9cc0-ad2c6bc38782/sist-en-15692-2009

#### 1 Scope

This European standard specifies a method for the direct determination of water in ethanol to be used in gasoline blends. It is applicable in the range 0.05 % (m/m) to 0.54 % (m/m).

NOTE For the purposes of this European Standard, the term "(m/m)" is used to represent the mass fraction.

WARNING — Use of this standard may involve hazardous materials, operations and equipment. This standard does not purport to address all of the safety problems associated with its use. It is the responsibility of the user of this standard to establish appropriate safety and health practices and determine the applicability of regulatory limitations prior to use.

#### 2 Normative references

The following referenced documents are indispensable for the application of this document. For dated references, only the edition cited applies. For undated references, the latest edition of the referenced document (including any amendments) applies.

EN ISO 3170, Petroleum liquids – Manual sampling (ISO 3170:2004)

EN ISO 3696, Water for analytical laboratory use – Specification and test methods (ISO 3696:1987)

# Terms and definitions iTeh STANDARD PREVIEW

(standards.iteh.ai)

For the purposes of this European Standard, the following term and definition applies.

3.1 SIST EN 15692:2009

water content

https://standards.iteh.ai/catalog/standards/sist/724d92a2-b420-4634-9cc0-

content of water determined by potentiometric Karl Fischer procedure as given in this document

#### 4 Principle

A weighed test portion is injected into the titration vessel of a potentiometric Karl Fischer apparatus. The water present is titrated to a potentiometric end point using Karl Fischer reagent. Iodine  $(I_2)$ , with presence of anhydride sulfur  $(SO_2)$ , of methanol  $(CH_3OH)$  and of an appropriate nitrogen base (RN), is introduced for the Karl Fischer reaction. Based on the stoichiometry of the reaction, one mole of iodine reacts with one mole of water.

The reaction can be expressed as follows:

CH3OH + SO2 + RN → [RNH]SO3CH3

 $H2O + I2 + [RNH]SO3CH3 + 2RN \rightarrow [RNH]SO4CH3 + 2[RNH]I$ 

#### 5 Reagents and materials

Use only chemicals and reagents of recognized analytical grade.

**5.1 Karl Fischer reagent**, pyridine-free Karl Fischer reagent, containing iodine, sulfur dioxide and a odourless amine and with a nominal water equivalent content of either 2 mg or 5 mg water per ml equivalent.

The Karl Fisher reagent shall be standardized daily before use (see 8.1).

- **5.2 Titration solvent methanol,** anhydrous methanol with a water content of less than 0,05 % (*m/m*).
- **5.3** Water, complying with grade 3 of EN ISO 3696.

#### 6 Apparatus

**6.1 Karl Fischer titrator**, using a potentiometric end-point.

NOTE Karl Fischer titrators are commercially available and some of them automatically stop the titration at the endpoint. Instructions for operation of these devices are provided by the manufacturer and are not described here. The method described in this document is using a titrator with a 5 ml burette.

#### 6.2 Electrode platinum/platinum

#### 6.3 Syringes

NOTE Needles with bores between 0,5 mm and 0,8 mm have been found suitable, such as syringe of single use with Luer connection, 5 ml or 10 ml capacity.

- **6.3.1 Syringe**, of approximately 10  $\mu$ l capacity, fitted with a needle of sufficient length to enable the tip to reach under the surface of the liquid in the titration vessel when inserted through the inlet-port septum.
- **6.3.2 Syringe**, of approximately 10 ml capacity, fitted with a needle of sufficient length to enable the tip to reach under the surface of the liquid in the titration vessel when inserted through the inlet-port septum. **Teh STANDARD PREVIEW**
- 6.4 Analytical balance, capable of weighing with an accuracy of 0,1 mg.
- **6.5 Automatic burette**, of 5 ml or 10 ml capacity, protected from humidity by the use of a molecular sieve (6.6) at the top.

  SIST EN 15692:2009

standards.iteh.ai

NOTE Although it is recommended that an automatic burette connected to a reservoir containing the Karl Fischer reagent is used, a burette, of approximately 5 ml capacity, fitted with a guard tube filled with molecular sieve (6.6) to prevent the ingress of moisture, may also be used.

**6.6** Molecular sieve, granulometry proximally 1,6 mm to 2,5 mm, pore size close to 1 nm.

If required to be dried it shall be placed in an oven at proximally 140 °C for about 8 hours, then cooled in a dessicator to room temperature.

#### 7 Sampling and sample handling

- **7.1** Samples shall be taken as described in EN ISO 3170.
- **7.2** Take care to minimise the uptake of atmospheric moisture during sampling and sample handling.
- **7.3** Samples shall be shaken before used.
- **7.4** Samples shall be stored capped at room temperature in a dry place to avoid modification of water content.

NOTE The use of a glass bottle that can be sealed with a septum has been found suitable for sampling and sample handling. A test portion of the sample can be taken through the septum with a syringe fitted with a needle.

#### 8 Procedure

#### 8.1 Standardization of the Karl Fischer reagent

**8.1.1** Add sufficient of the titration solvent (5.2) to the clean, dry titration vessel (6.1) to cover the electrodes. Seal all openings to the vessel, start the magnetic stirrer and adjust for smooth stirring action. Turn on the indication circuit and add KF reagent (5.1) from the burette until the end point is reached.

Swirl the titration vessel to dry inside walls. Add more KF reagent if needed until a steady end point is reached and maintained for at least 15 s.

Repeat these swirling and titration steps until the vessels walls are dry.

- **8.1.2** Fill a 10  $\mu$ l syringe (6.3.1) with pure water, taking care to eliminate air bubbles. Wipe the needle with a tissue to remove any residual water from its surface. Using the balance (6.4) weigh the syringe and the water and record the mass in mg (W).
- **8.1.3** Insert the needle of the syringe into the titration vessel via the inlet port septum. Ensure that the tip of the needle is below the surface of the titration solvent. Add the contents of the syringe to the titration solvent in the vessel which has been adjusted to its end point. Titrate the water with KF reagent until a steady end point is reached and maintained at least 15 s.

After adding water, do not shake the vessel.

**8.1.4** Record, to the nearest 0,01 ml, the volume of titrant needed to reach the end point (7).

Reweigh the empty syringe and record the mass in mg.

**8.1.5** To check the conformance of the KF reagent, calculate the water equivalence of KF reagent, use the following equation: https://standards.itch.ai/catalog/standards/sist/724d92a2-b420-4634-9cc0-ad2c6bc38782/sist-en-15692-2009

$$F = W/T \tag{1}$$

where,

- F is the water equivalence of the KF reagent, expressed in mg/ml;
- W is the mass of water added in mg (see 8.1.2);
- T is the volume of titrant needed to reach the end point, expressed in ml (see 8.1.4).
- **8.1.6** Repeat the procedure specified in 8.1.1 to 8.1.5 to give a duplicate value.
- **8.1.7** If the variation between the two titrations is greater than 2 % relative, discard the contents of the titration vessel. Introduce a further portion of appropriate titration solvent into the vessel and repeat the standardization procedure starting from 8.1.1.
- **8.1.8** If the titrations for two further portions of water still vary by more than 2 % relative, it is likely that either the Karl Fischer reagent and/or the titration solvent have aged. Replace these with fresh reagents and repeat the procedure starting from 8.1.1
- **8.1.9** Record the mean water equivalence value.

#### 8.2 Analysis

**8.2.1** Add fresh titration solvent (5.2) to the clean, dry titration vessel (6.1) and titrate the solvent to the end point conditions as described in 8.1.1.

**8.2.2** Dry the inside of a 10 ml syringe (6.3.2) by drawing the titration solvent up into the syringe and discharging back into the titration vessel. If the vessel contents become wet, add KF reagent (5.1) until the end point is maintained for at least 15 s without further addition of titrant.

Repeat this drying procedure until no further addition of KF reagent is necessary to maintain the end point state for at least 15 s (alternatively, oven-dried syringes, cooled in a dessicator may be used).

- **8.2.3** Immediately after mixing, use the dry syringe to withdraw at least three portions of the sample and discard as waste.
- **8.2.4** Immediately withdraw a 4 ml to 8 ml (depending on the water content) test portion of the sample, clean the needle with a paper tissue and weight the syringe and contents to the nearest 0,1 mg.

Insert the syringe into the titration vessel via the inlet port septum, such that the needle is above the surface of the solvent, and discharge its contents. Withdraw the syringe, reseal the vessel, reweigh to the nearest 0,1 mg and record the mass  $m_1$  of the sample test portion taken.

- **8.2.5** Titrate to the end point state, which shall be stable for at least 15 s. Do not shake the cell after addition of the sample.
- **8.2.6** Record, to the nearest 0,01 ml, the volume of titrant needed to reach the end point (V) and the water equivalence of KF reagent ( $m_2$ ).

NOTE The solvent should be changed when the test portion content exceeds 2 g of sample per 15 ml of solvent or when 4 ml of titrant per 15 ml of solvent has been added to the titration vessel.

### iTeh STANDARD PREVIEW

#### 9 Calculation

(standards.iteh.ai)

Calculate the water content of the sample,  $C_{sr}$  expressed in % (m/m), using the following equation:

$$C_{s} = \frac{m_{2} \times V}{m_{1} \times 10}$$
 https://standards.iteh.ai/catalog/standards/sist/724d92a2-b420-4634-9cc0-ad2c6bc38782/sist-en-15692-2009 (2)

where

 $m_1$  is the mass of the test portion, expressed in grams (g) (see 8.2.4);

m<sub>2</sub> is the mass of water in 1 ml of Karl Fischer reagent, expressed in mg/ml (see 8.2.6);

V is the volume of titrant required to reach the end point titration, expressed in ml (see 8.2.6).

#### 10 Expression of results

Report the water content of the sample to the nearest 0,01 % (m/m).

#### 11 Precision

#### 11.1 General

The precision given was derived from statistical analysis by EN ISO 4259 [3] of the results of interlaboratory testing of a matrix of ethanol samples produced in Europe from bio materials such as raw wine, molasses, pulp and corn.