



Designation: D 511 – 93 (Reapproved 1998)

AMERICAN SOCIETY FOR TESTING AND MATERIALS
100 Barr Harbor Dr., West Conshohocken, PA 19428
Reprinted from the Annual Book of ASTM Standards. Copyright ASTM

Standard Test Methods for Calcium and Magnesium In Water¹

This standard is issued under the fixed designation D 511; the number immediately following the designation indicates the year of original adoption or, in the case of revision, the year of last revision. A number in parentheses indicates the year of last reapproval. A superscript epsilon (ϵ) indicates an editorial change since the last revision or reapproval.

This standard has been approved for use by agencies of the Department of Defense.

1. Scope

1.1 These test methods cover the determination of calcium and magnesium in water by complexometric titration and atomic absorption spectrophotometric procedures. Two test methods are included, as follows:

	Sections
Test Method A—Complexometric Titration	7 to 14
Test Method B—Atomic Absorption Spectrophotometric	15 to 23

1.2 *This standard does not purport to address all of the safety concerns, if any, associated with its use. It is the responsibility of the user of this standard to establish appropriate safety and health practices and determine the applicability of regulatory limitations prior to use.* Specific hazard statements are given in Note 2 and Note 7.

2. Referenced Documents

2.1 ASTM Standards:

- D 1129 Terminology Relating to Water²
- D 1193 Specification for Reagent Water²
- D 2777 Practice for Determination of Precision and Bias of Applicable Methods of Committee D-19 on Water²
- D 3370 Practices for Sampling Water from Closed Conduits²
- D 4691 Practice for Measuring Elements in Water by Flame Atomic Absorption Spectrophotometry²
- D 4841 Practice for Estimation of Holding Times for Water Samples Containing Organic and Inorganic Constituents²

3. Terminology

3.1 *Definitions*—For definitions of terms used in these test methods, refer to Terminology D 1129.

4. Significance and Use

4.1 Calcium and magnesium salts in water are the primary components of water hardness which can cause pipe or tube scaling.

5. Purity of Reagents

5.1 Reagent grade chemicals shall be used in all tests. Unless otherwise indicated, it is intended that all reagents shall conform to the specifications of the Committee on Analytical Reagents of the American Chemical Society.³ Other grades may be used, provided it is first ascertained that the reagent is of sufficiently high purity to permit its use without lessening the accuracy of the determination.

5.2 *Purity of Water*—Unless otherwise indicated, references to water shall be understood to mean reagent water conforming to Specification D 1193, Type I, II, or III water. Type I is preferred and more commonly used. Type II water was specified at the time of round-robin testing of these test methods.

NOTE 1—The user must ensure the type of reagent water is sufficiently free of interferences. The water should be analyzed using this test method.

6. Sampling

6.1 Collect the sample in accordance with Practices D 3370.

6.2 If total recoverable calcium and magnesium concentrations are being determined, acidify the water sample with HNO₃ (sp gr 1.42) to a pH of 2 or less immediately at the time of collection; normally about 2 mL/L are required. The holding time for the samples may be calculated in accordance with Practice D 4841.

6.3 If dissolved calcium and magnesium concentrations are being determined, filter the samples through a 0.45- μ m membrane filter and acidify with HNO₃ (sp gr 1.42) to a pH of 2 or less immediately at time of collection; normally about 2 mL/L are required.

6.4 A number of sources of calcium contamination have been encountered in laboratories. Among the most common are plastic ware, paper towels, and dust. Rinsing plastic ware with sample prior to use, avoiding contact of apparatus with paper towels, and keeping exposure to the air to a minimum will limit the possibility of contamination.

¹ These test methods are under the jurisdiction of ASTM Committee D-19 on Water and are the direct responsibility of Subcommittee D19.05 on Inorganic Constituents in Water.

Current edition approved April 15, 1993. Published June 1993. Originally published as D 511 – 37. Last previous edition D 511 – 92.

² *Annual Book of ASTM Standards*, Vol 11.01.

³ *Reagent Chemicals, American Chemical Society Specifications*, American Chemical Society, Washington, DC. For suggestions on the testing of reagents not listed by the American Chemical Society, see *Analar Standards for Laboratory Chemicals*, BDH Ltd., Poole, Dorset, U.K., and the *United States Pharmacopeia and National Formulary*, U.S. Pharmaceutical Convention, Inc. (USPC), Rockville, MD.

TEST METHOD A—COMPLEXOMETRIC TITRATION

7. Scope

7.1 This test method is applicable to most waters in a range from 1 to 1000 mg/L of calcium plus magnesium expressed as calcium, but may fail in the analysis of highly colored waters, brines, or waters that contain excessive amounts of metals. The upper and lower limits may be extended by either dilution or use of micro apparatus.

7.2 Data are not available to determine which matrices were used to obtain the precision and bias data, and it is the responsibility of the analyst to determine the acceptability of this test method for the matrix being analyzed.

8. Summary of Test Method

8.1 EDTA (ethylenediamine tetraacetic acid or its salts) is added to a sample containing calcium and magnesium ions after the pH of the solution is adjusted to 10 for the determination of calcium and magnesium or from pH 12 to 13 for the determination of calcium alone. The EDTA initially complexes the calcium and then the magnesium. The end point is observed by the use of a suitable indicator. At a pH of 12 to 13 magnesium is precipitated. Magnesium is determined by the difference between an aliquot titrated at pH 10 and one titrated at pH 12 to 13.

9. Interferences

9.1 EDTA reacts with iron, manganese, copper, zinc, lead, cobalt, nickel, barium, strontium, calcium, magnesium, and several other metals. The interference of heavy metals is minimized by the addition of hydroxylamine and cyanide, which reduce or complex the metals, or both. Metal concentrations as high as 5 mg/L of iron, 10 mg/L of manganese, 10 mg/L of copper, 10 mg/L of zinc, and 10 mg/L of lead can be tolerated when hydroxylamine and cyanide are added.

9.2 In the titration of calcium plus magnesium, the higher oxidation states of manganese above Mn^{+2} react rapidly with the indicator to form discolored oxidation products. Hydroxylamine hydrochloride reagent is used to reduce manganese to the divalent state. The divalent manganese interference can be eliminated by addition of one or two small crystals of potassium ferrocyanide.

9.2.1 Orthophosphate and sulfate ions interfere at concentrations in excess of 500 and 10 000 mg/L, respectively.

9.2.2 In the presence of aluminum concentrations in excess of 10 mg/L, the blue color that indicates that the end point has been reached will appear and then, on short standing, will revert to red. The reversion should not be confused with the gradual change that normally takes place in the titrated sample several minutes after the titration has been completed.

9.3 In the titration of calcium, ammonium purpurate reacts with strontium but not with magnesium or barium. However, the end point in the presence of strontium is sluggish, and the titration is not strictly stoichiometric. Barium does not titrate as calcium, but affects the indicator in some unknown way so that no end point, or at best a poor end point, is obtained. Barium can be removed by prior precipitation with sulfuric acid, but care must be exercised to prevent precipitation of calcium.

Orthophosphate will precipitate calcium at the pH of the test.

9.4 A possible interference from the commonly used polyphosphates, organic phosphonates, and EDTA/NTA compounds in water treatment should be recognized.

10. Apparatus

10.1 *Titration Assembly*—Some analysts prefer to use conventional lighting and hand stirring. Others report better results by using a visual-titration assembly consisting of a motor-driven stirrer, 25-mL buret, white-porcelain-base buret holder, and shaded incandescent lamp. The sample beaker is placed near the front of the porcelain base and the reaction is viewed diagonally downward through the side of the beaker and against the white background. Illumination is from behind the beaker. The capacity of the buret, type lighting, and background color may be varied depending on the ionic concentrations normally encountered and the indicator chosen.

11. Reagents

11.1 *Buffer Solution, Ammonium Chloride-Ammonium Hydroxide*—Dissolve 67.6 g of ammonium chloride (NH_4Cl) in 200 mL of water. Add 570 mL of concentrated ammonium hydroxide (NH_4OH , sp gr 0.900). Add 5.00 g of magnesium salt of EDTA and dilute to 1000 mL. Store in a tightly stoppered plastic bottle to prevent the loss of ammonia. Discard the solution when 1 mL added to a neutralized sample fails to produce a pH of 10.0 ± 0.1 at the titration end point. To attain highest accuracy, adjust the magnesium level to exact equivalence through the appropriate addition of a small amount of either disodium EDTA or magnesium sulfate ($MgSO_4$).

11.2 *Calcium Indicator Solution*:

11.2.1 *Ammonium Purpurate*—Mix thoroughly 1.0 g of ammonium purpurate with 200 g of sucrose. Place in a bottle provided with a dispensing spoon of 0.2-g capacity.

11.2.2 *Fluorescein Methylene Iminodiacetic Acid*⁴—Grind 0.2 g of fluorescein methylene iminodiacetic acid and 0.12 g of thymolphthalein with 20 g of potassium chloride to 40 to 50 mesh size. Place in a bottle provided with a dispensing spoon of 0.2-g capacity.

11.3 *Calcium Solution, Standard* (1.00 mL = 0.400 mg calcium)—Suspend 1.000 g of calcium carbonate ($CaCO_3$), dried at 180°C for 1.0 h before weighing, in approximately 600 mL of water and dissolve cautiously with a minimum of dilute HCl. Dilute to 1000 mL with water in a volumetric flask.

11.4 *Chrome Black T Solution* (4.0 g/L)—Dissolve 0.4 g of Chrome Black T in 100 mL of water. This solution has a shelf life of approximately 1 week. Alternatively, a dry powder mixture composed of 0.5 g of dye and 100 g of powdered sodium chloride may be used. Store this in a dark-colored bottle provided with a dispensing spoon of approximately 0.2-g capacity. The shelf life is at least 1 year.

11.5 *EDTA Solution, Standard* (0.01 M, 1 mL = 0.401 mg calcium or 0.243 mg manganese)—Dissolve 3.72 g of Na_2EDTA dihydrate, which has been dried overnight over H_2SO_4 in a desiccator, in water and dilute to 1000 mL in a

⁴ Calcein, W., Fluorescein Complexon, and Fluorexon, supplied by various commercial firms, have been found satisfactory for this purpose.

 D 511

volumetric flask. The reagent is stable for several weeks. Check the titer of the reagent by titrating 25.00 mL of CaCO₃ standard solution as described in the procedure for sample analysis.

11.6 *Hydroxylamine Hydrochloride Solution* (30 g/L)—Dissolve 30 g of hydroxylamine hydrochloride (NH₂OH·HCl) in water and dilute to 1000 mL.

11.7 *Potassium Ferrocyanide*—(K₄Fe(CN)₆·3H₂O).

11.8 *Sodium Cyanide Solution* (25 g/L)—Dissolve 25 g of sodium cyanide (NaCN) in water and dilute to 1000 mL.

NOTE 2—**Warning:** Sodium cyanide is a deadly poison. Do not add NaCN to any acid solution or acidify any solution containing it. Use this reagent in a fume hood.

11.9 *Sodium Hydroxide Solution* (80 g/L)—Dissolve 80 g of sodium hydroxide (NaOH) in 800 mL of water. Cool and dilute to 1000 mL.

12. Procedure

12.1 Calcium Plus Magnesium:

12.1.1 Measure 100.0 mL of a well-mixed acidified sample into a 125-mL beaker or flask.

NOTE 3—If only dissolved calcium plus magnesium is to be determined, omit 12.1.1 and proceed to 12.1.5.

12.1.2 Add 5 mL of hydrochloric acid (HCl, sp gr 1.19) to each sample.

12.1.3 Heat the samples on a steambath or hot plate until the volume has been reduced to 15 to 20 mL, making certain that the samples do not boil.

NOTE 4—For samples with high levels of dissolved or suspended matter, the amount of reduction in volume is left to the discretion of the analyst.

12.1.4 Cool and filter the samples through a suitable filter (such as fine-textured, acid-washed, ashless paper) into 100-mL volumetric flasks. Wash the paper 2 or 3 times with water and bring to the volume.

12.1.5 Pipet the filtered sample (50.00-mL maximum) into a 150-mL beaker and adjust the volume to approximately 50 mL. Adjust the pH to 7 to 10 by the dropwise addition of ammonium hydroxide (NH₄OH, sp gr 0.900).

NOTE 5—For analysis of brines an appropriate aliquot size often can be determined from knowledge of the specific gravity, for example:

1.000 to 1.025, use 25 mL
1.025 to 1.050, use 10 mL
1.050 to 1.090, use 5 mL
1.090 to 1.120, use 1 mL
1.120 to 1.180, use 0.1 mL

12.1.6 Insert the beaker in the titration assembly and start the stirrer.

12.1.7 Add 1 mL of NH₂OH·HCl solution.

12.1.8 Add 1 mL of buffer solution. Measure the pH and verify that it falls in the pH range from 10.0 + 0.1.

12.1.9 Add 2 mL of NaCN solution (**Warning:** see Note 2.) unless interfering metals are known to be absent.

12.1.10 If manganese is present, add one or two small crystals K₄Fe(CN)₆·3H₂O.

12.1.11 Add 4 to 5 drops of Chrome Black T indicator solution. If powdered indicator is used add approximately 0.2 g.

12.1.12 Titrate with standard EDTA solution until blue or purple swirls begin to show. The end point is reached when all traces of red and purple have disappeared and the solution is clear blue in color. The titration should be completed within 5 min of the buffer addition. If more than 15 mL of titrant is required, take a smaller sample aliquot and repeat the test.

12.1.13 Record the volume of EDTA solution required to titrate calcium plus magnesium.

12.1.14 Determine a reagent blank correction by similarly titrating 50 mL of water including all added reagents.

12.2 Calcium:

12.2.1 Refill the buret with EDTA standard solution.

12.2.2 Pipet another aliquot of the same sample (50.00 mL maximum) into a 150-mL beaker and adjust the volume to approximately 50 mL (see Note 5).

12.2.3 Insert the beaker in the titration assembly and start the stirrer.

12.2.4 Add 1 mL of NH₂OH·HCl.

12.2.5 Add 1 mL of NaOH solution. The pH should range from 12 to 13.

12.2.6 Add 1 mL of NaCN (**Warning**, see Note 2).

12.2.7 Add 0.2 g of calcium indicator solution and proceed immediately with the titration.

12.2.8 Titrate with standard EDTA solution to the appropriate end point. If ammonium purpurate is used, the end point will be indicated by color change from salmon pink to orchid purple. If fluorescein methylene iminodiacetic acid is used, the end point will be indicated by a color change from deep green to purple. The titration should be completed within 5 min of the addition of NaOH solution. If more than 15 mL of titrant is required, take a smaller sample aliquot and repeat the test.

12.2.9 Record the volume of EDTA solution required to titrate the calcium.

12.2.10 Determine a reagent blank correction by similarly titrating 50 mL of water including all added reagents.

13. Calculation

13.1 Calculate the concentration of calcium and magnesium in milligrams per litre using Eq 1 and Eq 2:

$$\text{Calcium, mg/L} = (A \times B/D) \times 40\ 100 \quad (1)$$

$$\text{Magnesium, mg/L} = ((C \times B/E) - (A \times B/D)) \times 24\ 300 \quad (2)$$

where:

A = EDTA standard solution required to titrate calcium in 12.2.9 minus the blank correction determined in 12.2.10, mL,

B = molarity of EDTA standard solution,

C = EDTA standard solution required to titrate calcium plus magnesium in 12.1.13 minus the blank correction determined in 12.1.14, mL,

D = sample taken in 12.2.2, mL, and

E = sample taken in 12.1.5, mL.

13.2 If the concentration of strontium is determined to be significant, make a correction for strontium concentration using Eq 3:

$$\begin{aligned} \text{Corrected mg/L calcium} &= \text{mg/L calcium} \\ &- (\text{mg/L strontium} \times 0.46) \end{aligned} \quad (3)$$