
**Water quality — Evaluation of ultimate
aerobic biodegradability of organic
compounds in aqueous medium — Static
test (Zahn-Wellens method)**

*Qualité de l'eau — Évaluation, en milieu aqueux, de la biodégradabilité
aérobie ultime des composés organiques — Essai statique
(méthode Zahn-Wellens)*

iTeh STANDARD PREVIEW
(standards.iteh.ai)

ISO 9888:1999

<https://standards.iteh.ai/catalog/standards/sist/e18bf63-e3a8-4953-aa7d-3bca2eac95ad/iso-9888-1999>



Contents

1 Scope	1
2 Normative references	1
3 Terms and definitions	2
4 Principle.....	3
5 Test environment.....	3
6 Reagents.....	3
7 Apparatus	5
8 Procedure	5
9 Calculation and expression of results.....	7
10 Validity of results.....	8
11 Test report	8
Annex A (informative) Example of a biodegradation curve	10
Bibliography	11

iTeh STANDARD PREVIEW
(standards.iteh.ai)

ISO 9888:1999
<https://standards.iteh.ai/catalog/standards/sist/e18bf63-e3a8-4953-aa7d-3bca2eac95ad/iso-9888-1999>

© ISO 1999

All rights reserved. Unless otherwise specified, no part of this publication may be reproduced or utilized in any form or by any means, electronic or mechanical, including photocopying and microfilm, without permission in writing from the publisher.

International Organization for Standardization
Case postale 56 • CH-1211 Genève 20 • Switzerland
Internet iso@iso.ch

Printed in Switzerland

Foreword

ISO (the International Organization for Standardization) is a worldwide federation of national standards bodies (ISO member bodies). The work of preparing International Standards is normally carried out through ISO technical committees. Each member body interested in a subject for which a technical committee has been established has the right to be represented on that committee. International organizations, governmental and non-governmental, in liaison with ISO, also take part in the work. ISO collaborates closely with the International Electrotechnical Commission (IEC) on all matters of electrotechnical standardization.

International Standards are drafted in accordance with the rules given in the ISO/IEC Directives, Part 3.

Draft International Standards adopted by the technical committees are circulated to the member bodies for voting. Publication as an International Standard requires approval by at least 75 % of the member bodies casting a vote.

International Standard ISO 9888 was prepared by Technical Committee ISO/TC 147, *Water quality*, Subcommittee SC 5, *Biological methods*.

This second edition cancels and replaces the first edition (ISO 9888:1991), which has been technically revised.

Annex A of this International Standard is for information only.

iTeh STANDARD PREVIEW (standards.iteh.ai)

[ISO 9888:1999](https://standards.iteh.ai/catalog/standards/sist/e18bfe63-e3a8-4953-aa7d-3bca2eac95ad/iso-9888-1999)

<https://standards.iteh.ai/catalog/standards/sist/e18bfe63-e3a8-4953-aa7d-3bca2eac95ad/iso-9888-1999>

iTeh STANDARD PREVIEW
(standards.iteh.ai)

ISO 9888:1999

<https://standards.iteh.ai/catalog/standards/sist/e18bf63-e3a8-4953-aa7d-3bca2eac95ad/iso-9888-1999>

Water quality — Evaluation of ultimate aerobic biodegradability of organic compounds in aqueous medium — Static test (Zahn-Wellens method)

WARNING — Activated sludge and sewage may contain potentially pathogenic organisms. Take appropriate precautions when handling them. Handle with care toxic test compounds and those with unknown properties.

1 Scope

This International Standard specifies a method for the evaluation in aqueous medium of the ultimate biodegradability and, as additional information, the primary biodegradability and the total elimination from water, of organic compounds at a given concentration by aerobic microorganisms.

The conditions described in this International Standard normally correspond to optimal conditions for allowing the maximum value of biodegradation to occur with the chosen inoculum in the test time. These conditions may even be more favourable than in full-scale wastewater treatment plants, especially if their hydraulic retention time, sludge age or the adaptation of the activated sludge is not optimal.

The method applies to organic compounds which are

- a) water-soluble at the concentration used under the test conditions and not expected to be transformed to insoluble metabolites if biodegradation and not elimination only shall be determined;
- b) nonvolatile, or which have a negligible vapour pressure under the test conditions;
- c) not lost by foaming from the test solution;
- d) not inhibitory to the test microorganisms at the concentration chosen for the test. Inhibitory effects can be determined using a suitable test method (e.g. see ISO 8192). If the test compound is toxic, the test concentration must be lowered, or a pre-exposed inoculum can be used.

This International Standard is also applicable to the measurement of biodegradation and elimination of dissolved organic compounds in wastewater (also called "test compound" in the method).

NOTE If more information is required to predict the behaviour of test compounds or wastewater in a treatment plant, a simulation test (e.g. the activated sludge simulation test ISO 11733) should be performed. For appropriate use of this method and for alternative biodegradation methods, see ISO 15462.

2 Normative references

The following normative documents contain provisions which, through reference in this text, constitute provisions of this International Standard. For dated references, subsequent amendments to, or revisions of, any of these publications do not apply. However, parties to agreements based on this International Standard are encouraged to investigate the possibility of applying the most recent editions of the normative documents indicated below. For undated references, the latest edition of the normative document referred to applies. Members of ISO and IEC maintain registers of currently valid International Standards.

ISO 6060:1989, *Water quality — Determination of the chemical oxygen demand*.

ISO 8245, *Water quality — Guidelines for the determination of total organic carbon (TOC) and dissolved organic carbon (DOC)*.

3 Terms and definitions

For the purposes of this International Standard, the following terms and definitions apply.

3.1

ultimate aerobic biodegradation

breakdown of a chemical compound or organic matter by microorganisms in the presence of oxygen to carbon dioxide, water and mineral salts of any other elements present (mineralization) and the production of new biomass

3.2

primary biodegradation

structural change (transformation) of a chemical compound by microorganisms resulting in the loss of a specific property

3.3

activated sludge

biomass produced in the aerobic treatment of wastewater by the growth of bacteria and other microorganisms in the presence of dissolved oxygen

3.4

concentration of suspended solids of an activated sludge

amount of solids obtained by filtration or centrifugation of a known volume of activated sludge and drying at about 105 °C to constant mass

3.5

total organic carbon

TOC

all that carbon present in organic matter which is dissolved and suspended in the water

3.6

dissolved organic carbon

DOC

that part of the organic carbon in the water which cannot be removed by specified phase separation

NOTE Phase separation may be specified for example by centrifugation at 40 000 m·s⁻² for 15 min or by membrane filtration using membranes with pores of 0,2 µm to 0,45 µm diameter.

3.7

chemical oxygen demand

COD

mass concentration of oxygen equivalent to the amount of a specified oxidant consumed by a chemical compound or organic matter when a water sample is treated with that oxidant under defined conditions

NOTE COD is expressed in this case as milligrams oxygen consumed per milligram or gram of test compound.

3.8

lag phase

time from the start of a test until adaptation and/or selection of the degrading microorganisms are achieved and the biodegradation degree of a chemical compound or organic matter has increased to about 10 % of the maximum level of biodegradation

NOTE Lag phase is recorded in days.

3.9

maximum level of biodegradation

maximum degree of biodegradation of a chemical compound or organic matter in a test, above which no further biodegradation takes place during the test

NOTE Maximum level of biodegradation is recorded in percent.

3.10

biodegradation phase

time from the end of the lag phase of a test until about 90 % of the maximum level of biodegradation has been reached

NOTE Biodegradation phase is recorded in days.

3.11

plateau phase

time from the end of the biodegradation phase until the end of the test

NOTE Plateau phase is recorded in days.

3.12

pre-exposure

pre-incubation of an inoculum in the presence of the test chemical compound or organic matter, with the aim of enhancing the ability of this inoculum to biodegrade the test material by adaptation and/or selection of the microorganisms

3.13

preconditioning

pre-incubation of an inoculum under the conditions of the subsequent test in the absence of the test chemical compound or organic matter, with the aim of improving the performance of the test by acclimatization of the microorganisms to the test conditions

4 Principle

iTeh STANDARD PREVIEW

The biodegradation or elimination of water-soluble organic compounds or wastewater ingredients by aerobic microorganisms is determined using a static aqueous test system. The test mixture contains an inorganic medium, activated sludge as a mixed inoculum and an organic test compound as the sole source of carbon and energy other than the sludge. The amount of test compound added is chosen to result in an initial concentration of dissolved organic carbon (DOC) between 50 mg/l and 400 mg/l, or of chemical oxygen demand (COD) between 100 mg/l and 1000 mg/l, depending on its water solubility and on its toxicity to the bacteria in the inoculum.

Measurement of the concentration of DOC (or COD) is made at the beginning and end of the test (normally 28 d) and at intermediate time intervals, as required. To allow for any significant adsorption of the test compound onto the sludge, samples are also taken 3 h after the beginning of the test. Values obtained at this time are used as the basis for calculating the percentage of ultimate biodegradability at each sampling time. As additional information, the total elimination of DOC or COD from the aqueous phase may be obtained by calculating the removal based on a measured value before addition of the salts of the test medium and the inoculum. Possible other abiotic elimination processes, such as stripping into the air, may be determined by an abiotic elimination control without inoculum.

Use of specific analysis may give additional information on primary biodegradability of test compounds.

5 Test environment

Incubation shall take place in the dark or in diffuse light, at a temperature within the range 20 °C to 25 °C which shall not vary by more than 2 °C during the test and in an enclosure which shall be free from gases and vapours which are toxic to microorganisms.

6 Reagents

Use only reagents of recognized analytical grade, including the following.

6.1 Water, distilled or deionized, containing less than 1 mg/l DOC.

6.2 Test medium

6.2.1 Composition

6.2.1.1 Solution a)

Anhydrous potassium dihydrogenphosphate (KH_2PO_4)	8,5 g
Anhydrous dipotassium hydrogenphosphate (K_2HPO_4)	21,75 g
Disodium hydrogenphosphate dihydrate ($\text{Na}_2\text{HPO}_4 \cdot 2\text{H}_2\text{O}$)	33,4 g
Ammonium chloride (NH_4Cl)	0,5 g
Dissolve in water (6.1) to make up to	1000 ml

NOTE In order to check this buffer solution, it is recommended that the pH is measured. If it is not at about 7,4 a fresh solution should be prepared.

6.2.1.2 Solution b)

Dissolve 22,5 g magnesium sulfate heptahydrate ($\text{MgSO}_4 \cdot 7\text{H}_2\text{O}$) in water (6.1) to make up to 1000 ml.

6.2.1.3 Solution c)

Dissolve 36,4 g calcium chloride dihydrate ($\text{CaCl}_2 \cdot 2\text{H}_2\text{O}$) in water (6.1) to make up to 1000 ml.

6.2.1.4 Solution d)

Dissolve 0,25 g iron(III) chloride hexahydrate ($\text{FeCl}_3 \cdot 6\text{H}_2\text{O}$) in water (6.1) to make up to 1000 ml. Prepare this solution freshly before use or add a drop of concentrated hydrochloric acid (HCl) to avoid precipitation.

6.2.2 Preparation of the test medium

1000 ml test medium shall contain 10 ml of solution a) and 1 ml of each of solutions b) to d). Add the correct amounts of the stock solutions as described in 8.3. Prepare the test medium freshly before use. The solutions a) to c) may be stored up to 6 months in the dark at room temperature.

NOTE If a test compound influences the pH value of the mixture at the chosen test concentration, an increase in the buffer capacity of the test medium may be required. At high test concentrations, a lack of nitrogen may also occur. In these cases it is recommended that the nutrient composition of the test medium be improved by adding, for example, 100 ml of solution a) instead of 10 ml or just the phosphate amount of solution a). The C:N:P ratio for a test concentration of 400 mg/l DOC is then changed from 100:0,3:30 to 100:3:300. As at higher nitrogen concentrations nitrification may occur, higher phosphate buffer concentrations are required to keep the pH stable.

6.3 Sodium hydroxide solution

Dissolve sodium hydroxide (NaOH) in the water (6.1) to obtain a solution of concentration 0,1 mol/l to 0,5 mol/l.

6.4 Hydrochloric acid solution

Dissolve hydrochloric acid (HCl) in the water (6.1) to obtain a solution of concentration 0,1 mol/l to 0,5 mol/l.

6.5 Mercury chloride solution

Dissolve 1 g of mercury(II) chloride (HgCl_2) in 100 ml of the water (6.1).

7 Apparatus

Ordinary laboratory equipment and the following shall be used

7.1 Glass vessels, of capacity 1 litre to 5 litres equipped with agitators with glass or metal stirrers; rotate to ensure adequate mixing.

Each vessel shall be fitted with 2 mm to 4 mm inner diameter glass tubes or glass frits to introduce air. The air shall be free from organic carbon and toxic vapours and shall be presaturated with water vapour to reduce losses by evaporation.

The glassware shall be carefully cleaned and, in particular, be free from traces of organic or toxic matter.

7.2 Measuring equipment, of sufficient sensitivity for the measurement of dissolved organic carbon (see ISO 8245) or the measurement of chemical oxygen demand (see ISO 6060) and, if required, for substance-specific analysis.

7.3 Centrifuge or device for filtration, with paper filters or membrane filters (of pore size/nominal aperture diameter 0,2 to 0,45 μm) which adsorb or release organic carbon to a minimum degree.

7.4 pH-meter (usual laboratory equipment).

8 Procedure

8.1 Preparation of the test solutions

8.1.1 Solution of the test compound in the water (6.1) or in the test medium (6.2.2), at a suitable concentration (e.g. 3 000 mg/l).

In the case of wastewater, determine the TOC, DOC and pH. The test is acceptable to the wastewater if the DOC is >90 % of the TOC, which indicates sufficiently water-soluble compounds for this method based on DOC measurement. Add the wastewater directly or dilute it with water (6.1) to obtain the desired test concentration. If the wastewater has an extreme pH value (<3 or >10), neutralize to $\text{pH } 7 \pm 1$ with sodium hydroxide (6.3) or hydrochloric acid (6.4) solution.

8.1.2 Solution of the reference compound in the water (6.1) or in the test medium (6.2.2), at a suitable concentration (e.g. 3 000 mg/l). Use a known biodegradable water-soluble organic compound (e.g. diethylene glycol, ethylene glycol, sodium benzoate or aniline) which have degradation degrees >90 %).

8.2 Preparation of the inoculum

Take a sample of activated sludge from the aeration tank of a biological wastewater treatment plant. Mix the sample well and wash the activated sludge by repeatedly (e.g. two to three times) adding tap water or the test medium (6.2), centrifuging or settling, and decanting of the supernatant. Before use, determine the concentration of suspended solids (use e.g. ISO 11923). If necessary, concentrate the sludge by settling so that the volume of sludge added to obtain the desired concentration of suspended solids is minimal (about 10 % of the total volume). Keep the inoculum aerated at room temperature until just before use.

NOTE Depending on the purpose of the test, the wastewater treatment plant should receive wastewater which is predominantly municipal. Activated sludge may also be taken from a laboratory treatment plant. To get as many different species or strains of bacteria as possible, it may be preferable in special cases to make a mixture from various sources. Pre-exposed inocula may be used in certain circumstances. When such inocula are used, this should be clearly stated in the test results (e.g. percentage biodegradation using pre-exposed inocula) and the method of pre-exposure detailed in the test report. Pre-exposed inocula can be obtained from laboratory biodegradation tests conducted under a variety of conditions as appropriate. Suitable tests are e.g. a Zahn-Wellens test (this test, ISO 9888) already performed with the test compound (see 8.3) or the SCAS test (ISO 9887). Inocula may also be collected from locations where relevant environmental conditions exist (e.g. wastewater treatment plants dealing with similar compounds).