



Designation: D5901 – 03



Designation: 434/98

Standard Test Method for Freezing Point of Aviation Fuels (Automated Optical Method)¹

This standard is issued under the fixed designation D5901; the number immediately following the designation indicates the year of original adoption or, in the case of revision, the year of last revision. A number in parentheses indicates the year of last reapproval. A superscript epsilon (ϵ) indicates an editorial change since the last revision or reapproval.

1. Scope*

1.1 This test method covers the determination of the temperature below which solid hydrocarbon crystals may form in aviation turbine fuels.

NOTE 1—This test method describes an alternative procedure and automated apparatus which mimics the apparatus and procedure described in Test Method [D2386](#).

1.2 The measuring range of the apparatus is from 0 to -70°C . The precision statements were derived from samples with freezing point temperatures from -45 to -65°C .

NOTE 2—Typical aviation fuel has freezing point temperatures in the -40 to -65°C range. Samples with higher freezing points were not available for the current interlaboratory program.

1.3 The values stated in SI units are to be regarded as the standard. The values given in parentheses are for information only.

1.4 *This standard does not purport to address all of the safety concerns, if any, associated with its use. It is the responsibility of the user of this standard to establish appropriate safety and health practices and determine the applicability of regulatory limitations prior to use. For specific warning statements, see Section 7.*

2. Referenced Documents

2.1 *ASTM Standards:*²

[D2386](#) Test Method for Freezing Point of Aviation Fuels³

[D4057](#) Practice for Manual Sampling of Petroleum and Petroleum Products³

[D4177](#) Practice for Automatic Sampling of Petroleum and Petroleum Products⁴

[E1](#) Specification for ASTM Liquid-in-Glass Thermometers

3. Terminology

3.1 *Definitions:*

3.1.1 *freezing point, n*—in aviation fuels, the fuel temperature at which solid hydrocarbon crystals, formed on cooling, disappear when the temperature of the fuel is allowed to rise under specified conditions of test.

3.2 *Definitions of Terms Specific to This Standard:*

3.2.1 *automated method, n*—the robotic automation of a manual procedure and apparatus.

4. Summary of Test Method

4.1 After insertion of 25 mL of the test specimen into a test chamber, the test specimen is cooled while being continuously stirred and monitored by an optical system. The temperature of the specimen is measured with an electronic temperature measuring device. When crystal formation is detected in the specimen, the temperature is recorded and the specimen in the test chamber is allowed to warm, while being continuously stirred and monitored, until the crystals in the specimen completely disappear. The temperature of the specimen when the last crystals disappear is recorded as the freezing point (automated method).

5. Significance and Use

5.1 The freezing point of an aviation fuel is an index of the lowest temperature of its utility for certain applications. Solid hydrocarbon crystals can restrict the flow of fuel in the fuel system of the aircraft. The temperature of the fuel in the aircraft tank normally decreases during flight depending on aircraft speed, altitude, and flight duration. The freezing point of the fuel must always be lower than the minimum operational fuel temperature.

¹ This test method is under the jurisdiction of ASTM Committee D02 on Petroleum Products and Lubricants and is the direct responsibility of Subcommittee D02.07 on Flow Properties.

Current edition approved May 10, 2003. Published July 2003. Originally approved in 1996. Last previous edition approved in 1999 as D5901-99. DOI: 10.1520/D5901-03.

² *Annual Book of ASTM Standards*, Vol 05.01.

³ *Annual Book of ASTM Standards*, Vol 05.02.

⁴ *Annual Book of ASTM Standards*, Vol 14.03.

*A Summary of Changes section appears at the end of this standard.

5.2 Petroleum blending operations require precise measurement of the freezing point.

5.3 This test method expresses results with a resolution of 0.1°C.

5.4 This test method provides results that have been found to be equivalent to the results from Test Method D2386 on identical samples. When the specification requires the use of Test Method D2386, do not substitute this test method or any other method.

6. Apparatus (see Annex A1)

6.1 *Automated Apparatus*⁵—The apparatus as described in Annex A1 shall consist of a test chamber comprising a jacketed test tube supported in a jacketed enclosure configuration that is capable of cooling and heating the test specimen to the temperatures required in the test. The apparatus shall have a nitrogen purge collar as part of the closure assembly for the test chamber, which prevents moisture from combining with the test specimen. The apparatus shall be capable of measuring the temperature of the test specimen, continuously stirring the test specimen at the prescribed rate, automatically cooling and then heating the test specimen, monitoring the test specimen with an electronic optical system for appearance and disappearance of the crystals in the test specimen under the conditions of the test, and recording the appearance and disappearance temperatures.

6.2 *Circulating Bath*, refrigeration unit equipped with a circulating pump capable of maintaining the temperature of a quantity of methyl alcohol at least 20°C lower than the minimum test specimen temperature expected.

NOTE 3—To achieve a typical test chamber cooling condition of –75°C, the circulating bath should be capable of achieving –85 to –90°C, since approximately 5 to 10°C is consumed in the circulation lines and insulation.

7. Reagents and Materials

7.1 *Cooling Medium, Methyl Alcohol*—A commercial or technical grade of anhydrous methanol is suitable for use as the cooling medium. (**Warning**—Extremely flammable.) (**Warning**—Toxic.)

7.2 *Nitrogen Gas*, dry nitrogen gas which has a dew point below the lowest temperature expected to be attained by the test specimen under the conditions of the test. (**Warning**—Compressed gas under high pressure.) (**Warning**—Inert gas can be an asphyxiant when inhaled.)

7.3 *Cleaning Solvents*, suitable for cleaning and drying the test chamber, such as petroleum naphtha or methyl alcohol. (**Warning**—Flammable. Liquid causes eye burns. Vapor harmful. May be fatal or cause blindness if swallowed or inhaled.)

8. Sampling

8.1 Obtain a sample in accordance with Practices D4057 or D4177.

8.2 At least 25 mL of sample is required for each test. Refer to Practice D4057.

9. Preparation of Apparatus

9.1 Prepare the apparatus for operation in accordance with the manufacturer's instructions.

9.2 Clean and dry the test chamber with petroleum naphtha to rinse out any previous specimen followed by a second rinse of alcohol to remove naphtha. Dry with moisture-free air or gas. Ensure that moisture does not remain inside the test chamber.

9.3 Prepare the refrigerated circulating bath for operation in accordance with the manufacturer's instructions and allow it to attain a temperature lower than –75°C. The temperature of the alcohol, at the test chamber, shall not be below –80°C unless the expected freezing point is below –60°C.

9.4 Confirm that the supply of nitrogen purge gas is connected and regulated in accordance with the manufacturer's instructions.

10. Calibration and Standardization

10.1 Ensure that all of the manufacturer's instructions for calibrating, checking, and operating the apparatus are followed including calibration of the temperature measuring system against a certified standard temperature device.

10.2 A sample with a mutually agreed upon freezing point such as one from an interlaboratory test program, Test Method D2386 or equivalent, can be used to verify performance of the apparatus within the precisions of this test method.

11. Procedure

11.1 Measure out 25 ± 1 mL of the fuel and transfer it to the clean, dry, test chamber. Support the test chamber in the position recommended by the manufacturer, enclosing the top of the test chamber with a closure assembly supporting the stirrer, temperature measuring device, optical system, and nitrogen purge collar. Adjust the temperature measuring device position, if necessary, so that it is positioned in the center of the test chamber. Ensure that the bottom of the temperature measuring device is between 10 to 15 mm from the bottom of the test chamber. Connect the cooling medium inlet and outlet hoses to the respective connections on the test chamber according to the manufacturer's instructions.

11.2 Start the operation of the apparatus according to the manufacturer's instructions. This shall enable the flow of the cooling medium for cooling of the specimen, the flow of the purge gas and the stirring of the specimen continuously and without interruption. The stirrer shall move up and down vertically at the rate of 1 to 1.5 cycles per second, taking care that the stirrer loops approach the bottom of the test chamber on the downstroke and remain below the specimen surface on the upstroke.

11.3 The optical system shall monitor the specimen for the appearance of hydrocarbon crystals. The apparatus shall disregard any cloud-like formation, due to water, that appears in the test specimen at approximately –10°C which does not increase in intensity as the specimen temperature decreases.

11.4 After the crystals are detected, the apparatus shall discontinue the flow of the cooling medium allowing the test

⁵ The sole source of supply of the apparatus known to the committee at this time is Herzog Model SC 860, available from Walter Herzog, Lauda, Germany. If you are aware of alternative suppliers, please provide this information to ASTM International Headquarters. Your comments will receive careful consideration at a meeting of the responsible technical committee¹, which you may attend.