



# SLOVENSKI STANDARD

## SIST ISO 6145-6:2004

01-oktober-2004

---

5 bU]nUd`]bcj `!`Df]dfUj U\_U]Vf]fb\ `d`]bg\_\] `na Yg]`n`X]bUa ] b]a ]]j c`i a Yff]`g\_]a ]  
a YrcXUa ]!\* "XY.`? f]h] bY`cXdfh]bY

Gas analysis -- Preparation of calibration gas mixtures using dynamic volumetric methods -- Part 6: Critical orifices

### iTeh STANDARD PREVIEW

Analyse des gaz -- Préparation des mélanges de gaz pour étalonnage à l'aide de méthodes volumétriques dynamiques -- Partie 6: Orifices critiques

[SIST ISO 6145-6:2004](https://standards.iteh.ai/catalog/standards/sist/e9b3c7b8-111a-4c9e-bc33-20766568ff84/sist-iso-6145-6-2004)

Ta slovenski standard je istoveten z: [ISO 6145-6:2003](https://standards.iteh.ai/catalog/standards/sist/e9b3c7b8-111a-4c9e-bc33-20766568ff84/sist-iso-6145-6-2004)

---

#### ICS:

71.040.40

Kemijska analiza

Chemical analysis

**SIST ISO 6145-6:2004**

**en**

**iTeh STANDARD PREVIEW**  
**(standards.iteh.ai)**

[SIST ISO 6145-6:2004](#)

<https://standards.iteh.ai/catalog/standards/sist/e9b3c7b8-111a-4c9e-bc33-26766588ff84/sist-iso-6145-6-2004>

# INTERNATIONAL STANDARD

**ISO**  
**6145-6**

Second edition  
2003-05-15

---

---

## Gas analysis — Preparation of calibration gas mixtures using dynamic volumetric methods —

### Part 6: Critical orifices

**iTeh STANDARD PREVIEW**  
**(standards.iteh.ai)**  
*Analyse des gaz — Préparation des mélanges de gaz pour étalonnage à  
l'aide de méthodes volumétriques dynamiques —  
Partie 6: Orifices critiques*

[SIST ISO 6145-6:2004](https://standards.iteh.ai/catalog/standards/sist/e9b3c7b8-111a-4c9e-bc33-26766588ff84/sist-iso-6145-6-2004)

[https://standards.iteh.ai/catalog/standards/sist/e9b3c7b8-111a-4c9e-bc33-  
26766588ff84/sist-iso-6145-6-2004](https://standards.iteh.ai/catalog/standards/sist/e9b3c7b8-111a-4c9e-bc33-26766588ff84/sist-iso-6145-6-2004)



Reference number  
ISO 6145-6:2003(E)

© ISO 2003

**ISO 6145-6:2003(E)****PDF disclaimer**

This PDF file may contain embedded typefaces. In accordance with Adobe's licensing policy, this file may be printed or viewed but shall not be edited unless the typefaces which are embedded are licensed to and installed on the computer performing the editing. In downloading this file, parties accept therein the responsibility of not infringing Adobe's licensing policy. The ISO Central Secretariat accepts no liability in this area.

Adobe is a trademark of Adobe Systems Incorporated.

Details of the software products used to create this PDF file can be found in the General Info relative to the file; the PDF-creation parameters were optimized for printing. Every care has been taken to ensure that the file is suitable for use by ISO member bodies. In the unlikely event that a problem relating to it is found, please inform the Central Secretariat at the address given below.

## iTeh STANDARD PREVIEW (standards.iteh.ai)

[SIST ISO 6145-6:2004](https://standards.iteh.ai/catalog/standards/sist/e9b3c7b8-111a-4c9e-bc33-26766588ff84/sist-iso-6145-6-2004)

<https://standards.iteh.ai/catalog/standards/sist/e9b3c7b8-111a-4c9e-bc33-26766588ff84/sist-iso-6145-6-2004>

© ISO 2003

All rights reserved. Unless otherwise specified, no part of this publication may be reproduced or utilized in any form or by any means, electronic or mechanical, including photocopying and microfilm, without permission in writing from either ISO at the address below or ISO's member body in the country of the requester.

ISO copyright office  
Case postale 56 • CH-1211 Geneva 20  
Tel. + 41 22 749 01 11  
Fax + 41 22 749 09 47  
E-mail [copyright@iso.org](mailto:copyright@iso.org)  
Web [www.iso.org](http://www.iso.org)

Published in Switzerland

<b>Contents</b>	<b>Page</b>
<b>1 Scope</b> .....	<b>1</b>
<b>2 Normative references</b> .....	<b>1</b>
<b>3 Principle</b> .....	<b>1</b>
<b>4 Application to preparation of gas mixtures</b> .....	<b>2</b>
<b>4.1 Description of the orifice system and the experimental procedure</b> .....	<b>2</b>
<b>4.2 Area of validity</b> .....	<b>4</b>
<b>4.3 Operating conditions</b> .....	<b>4</b>
<b>5 Calculation of operating parameters and results</b> .....	<b>4</b>
<b>5.1 Selection of suitable orifice systems</b> .....	<b>4</b>
<b>5.2 Calculation of volume fraction</b> .....	<b>5</b>
<b>5.3 Sources of uncertainty</b> .....	<b>5</b>
<b>5.4 Uncertainty of volume fraction</b> .....	<b>6</b>
<b>Annex A (informative) Premixed gases for preparation of mixtures of high dilution</b> .....	<b>7</b>
<b>Annex B (informative) Practical hints</b> .....	<b>9</b>
<b>Bibliography</b> .....	<b>11</b>

**ITeH STANDARD PREVIEW**  
 (standards.iteh.ai)

SIST ISO 6145-6:2004

<https://standards.iteh.ai/catalog/standards/sist/e9b3c7b8-111a-4c9e-bc33-26766588ff84/sist-iso-6145-6-2004>

## ISO 6145-6:2003(E)

## Foreword

ISO (the International Organization for Standardization) is a worldwide federation of national standards bodies (ISO member bodies). The work of preparing International Standards is normally carried out through ISO technical committees. Each member body interested in a subject for which a technical committee has been established has the right to be represented on that committee. International organizations, governmental and non-governmental, in liaison with ISO, also take part in the work. ISO collaborates closely with the International Electrotechnical Commission (IEC) on all matters of electrotechnical standardization.

International Standards are drafted in accordance with the rules given in the ISO/IEC Directives, Part 2.

The main task of technical committees is to prepare International Standards. Draft International Standards adopted by the technical committees are circulated to the member bodies for voting. Publication as an International Standard requires approval by at least 75 % of the member bodies casting a vote.

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. ISO shall not be held responsible for identifying any or all such patent rights.

ISO 6145-6 was prepared by Technical Committee ISO/TC 158, *Analysis of gases*.

This second edition cancels and replaces the first edition (ISO 6145-6:1986), which has been technically revised.

ISO 6145 consists of the following parts, under the general title *Gas analysis — Preparation of calibration gas mixtures using dynamic volumetric methods*:

- *Part 1: Methods of calibration*
- *Part 2: Volumetric pumps*
- *Part 4: Continuous injection method*
- *Part 5: Capillary calibration devices*
- *Part 6: Critical orifices*
- *Part 7: Thermal mass-flow controllers*
- *Part 9: Saturation method*
- *Part 10: Permeation method*

Diffusion will be the subject of a future Part 8 to ISO 6145. Part 3 to ISO 6145, entitled *Periodic injections into a flowing gas*, has been withdrawn.

## Introduction

This part of ISO 6145 is one of a series of standards that present various dynamic volumetric methods used for the preparation of calibration gas mixtures.

## iTeh STANDARD PREVIEW (standards.iteh.ai)

[SIST ISO 6145-6:2004](https://standards.iteh.ai/catalog/standards/sist/e9b3c7b8-111a-4c9e-bc33-26766588ff84/sist-iso-6145-6-2004)

<https://standards.iteh.ai/catalog/standards/sist/e9b3c7b8-111a-4c9e-bc33-26766588ff84/sist-iso-6145-6-2004>

**iTeh STANDARD PREVIEW**  
**(standards.iteh.ai)**

SIST ISO 6145-6:2004

<https://standards.iteh.ai/catalog/standards/sist/e9b3c7b8-111a-4c9e-bc33-26766588ff84/sist-iso-6145-6-2004>



# Gas analysis — Preparation of calibration gas mixtures using dynamic volumetric methods —

## Part 6: Critical orifices

### 1 Scope

This part of ISO 6145 specifies a method for the continuous production of calibration gas mixtures, containing two or more components, from pure gases or other gas mixtures by use of critical orifice systems. By selection of appropriate combinations of orifices and with the use of pure gases, the volume fraction of the calibration component in the calibration gas mixture can be varied by a factor of  $10^4$ . Additionally, it can be changed by a factor of  $10^2$  by changing the initial pressures in the orifice systems. The uncertainty of the method depends mainly upon the flow calibration method and the variations in temperature and outlet pressure. The relative expanded uncertainty in the volume fraction obtainable for a binary mixture (at a coverage factor of 2) is 3 %.

If pre-mixed gases are used instead of pure gases, much lower volume fractions can be obtained (see Annex A). The mass flow rates or volume flow rates, from which the mass or volume fractions are determined, can be calculated and can be independently measured by a suitable method given in ISO 6145-1.

The merits of the method are that multi-component mixtures can be prepared as readily as binary mixtures if the appropriate number of orifices is utilized, and that a large quantity of calibration gas mixture can be prepared on a continuous basis. The range of flow rates can be from several millilitres per minute to approximately 10 l/min.

Although particularly applicable to preparation of gas mixtures at barometric pressure, the method also provides a means of preparation of calibration gas mixtures at pressures above barometric pressure.

Annex B gives practical hints on the use of the method.

### 2 Normative references

The following referenced documents are indispensable for the application of this document. For dated references, only the edition cited applies. For undated references, the latest edition of the referenced document (including any amendments) applies.

ISO 6143, *Gas analysis — Comparison methods for determining and checking the composition of calibration gas mixtures*

ISO 6145-1, *Gas analysis — Preparation of calibration gas mixtures using dynamic volumetric methods — Part 1: Methods of calibration*

### 3 Principle

When passed through a critical orifice at increasing upstream pressure  $p_1$ , the volume flow rate of gas passing through the orifice will increase. When the ratio of the gas pressure upstream  $p_1$  and the gas pressure downstream of the orifice  $p_2$  has reached a critical value, on further increase of  $p_1$  the volume flow rate of the gas becomes independent with respect to  $p_2$ .