
**Water quality — Determination of
chromium — Atomic absorption
spectrometric methods**

*Qualité de l'eau — Dosage du chrome — Méthodes par spectrométrie
d'absorption atomique*

iTeh STANDARD PREVIEW
(standards.iteh.ai)

[ISO 9174:1998](https://standards.iteh.ai/catalog/standards/sist/60ae9f29-c72e-44bd-b3fa-95093459e5db/iso-9174-1998)

<https://standards.iteh.ai/catalog/standards/sist/60ae9f29-c72e-44bd-b3fa-95093459e5db/iso-9174-1998>



Contents

Page

1	Scope	1
2	Normative references	1
3	Determination of chromium by flame atomic absorption spectrometry	1
4	Determination of chromium by electrothermal atomization atomic absorption spectrometry	5
	Annex A (informative) Pretreatment and digestion of sludge and sediment samples	8
	Annex B (informative) Precision data	9

iTeh STANDARD PREVIEW (standards.iteh.ai)

[ISO 9174:1998](https://standards.iteh.ai/catalog/standards/sist/60ae9f29-c72e-44bd-b3fa-95093459e5db/iso-9174-1998)

<https://standards.iteh.ai/catalog/standards/sist/60ae9f29-c72e-44bd-b3fa-95093459e5db/iso-9174-1998>

© ISO 1998

All rights reserved. Unless otherwise specified, no part of this publication may be reproduced or utilized in any form or by any means, electronic or mechanical, including photocopying and microfilm, without permission in writing from the publisher.

International Organization for Standardization
Case postale 56 • CH-1211 Genève 20 • Switzerland
Internet iso@iso.ch

Printed in Switzerland

Foreword

ISO (the International Organization for Standardization) is a worldwide federation of national standards bodies (ISO member bodies). The work of preparing International Standards is normally carried out through ISO technical committees. Each member body interested in a subject for which a technical committee has been established has the right to be represented on that committee. International organizations, governmental and non-governmental, in liaison with ISO, also take part in the work. ISO collaborates closely with the International Electrotechnical Commission (IEC) on all matters of electrotechnical standardization.

Draft International Standards adopted by the technical committees are circulated to the member bodies for voting. Publication as an International Standard requires approval by at least 75 % of the member bodies casting a vote.

iTeh STANDARD PREVIEW

International Standards are drafted in accordance with the rules given in the ISO/IEC Directives, Part 3.

International Standard ISO 9174 was prepared by Technical Committee ISO/TC 147, *Water quality*, Subcommittee SC 2, *Physical, chemical, biochemical methods*.

This second edition cancels and replaces the first edition (ISO 9174:1990), which has been technically revised.

Annexes A and B of this International Standard are for information only.

Introduction

Chromium occurs in water in the oxidation states III and VI. The two methods described in this International Standard determine chromium in both oxidation states, either as acid-soluble chromium or as water-soluble chromium, depending on the sample pretreatment. The method chosen depends on the concentration of chromium in the water to be examined.

Annex A provides information on pretreatment and digestion of chromium from sludges and sediments.

iTeh STANDARD PREVIEW (standards.iteh.ai)

[ISO 9174:1998](https://standards.iteh.ai/catalog/standards/sist/60ae9f29-c72e-44bd-b3fa-95093459e5db/iso-9174-1998)

<https://standards.iteh.ai/catalog/standards/sist/60ae9f29-c72e-44bd-b3fa-95093459e5db/iso-9174-1998>

Water quality — Determination of chromium — Atomic absorption spectrometric methods

1 Scope

This International Standard specifies two methods for the determination of chromium in water by atomic absorption spectrometry. The two methods are covered in separate clauses as follows:

- Clause 3: Determination of chromium by flame atomic absorption spectrometry;
- Clause 4: Determination of chromium by electrothermal atomization atomic absorption spectrometry.

Clause 3 is applicable to the analysis of water and waste water when the concentration range is between 0,5 mg/l and 20 mg/l of chromium. When the concentration is below 0,5 mg/l, the determination can be carried out after carefully evaporating an acidified sample to small volume, taking care to avoid the formation of a precipitate.

NOTE The use of evaporation will increase the effect of interfering substances and therefore for concentrations below 0,1 mg/l the method in clause 4 is given.

Clause 4 is applicable to the analysis of water and waste water when the concentration range is between 5 µg/l and 100 µg/l of chromium by injecting a sample volume of 20 µl. It is applicable to the determination of higher concentrations by using a smaller sample volume.

ISO 9174:1998

<https://standards.iteh.ai/catalog/standards/sist/60ae9f29-c72e-44bd-b3fa-95093459e5db/iso-9174-1998>

2 Normative references

The following normative documents contain provisions which, through reference in this text, constitute provisions of this International Standard. For dated references, subsequent amendments to, or revisions of, any of these publications do not apply. However, parties to agreements based on this International Standard are encouraged to investigate the possibility of applying the most recent editions of the normative documents indicated below. For undated references, the latest edition of the normative document referred to applies. Members of ISO and IEC maintain registers of currently valid International Standards.

ISO 5667-2 : 1991, *Water quality — Sampling — Part 2: Guidance on sampling techniques*.

ISO 5667-3 : 1994, *Water quality — Sampling — Part 3: Guidance on the preservation and handling of samples*.

3 Determination of chromium by flame atomic absorption spectrometry

3.1 Principle

The method is based on the atomic absorption spectrometric measurement of the chromium content of the acidified sample in a nitrous oxide/acetylene flame. Measurement at a wavelength of 357,9 nm. Addition of lanthanum salt to reduce matrix interferences is necessary.

3.2 Reagents

All reagents shall be of recognized analytical grade. Use deionized water or water distilled from an all-glass apparatus. The water used for blank tests and for preparing reagents and standard solutions shall have a chromium content that is negligible compared with the smallest concentrations to be determined in the samples.

3.2.1 Hydrochloric acid, HCl, $\rho \approx 1,18$ g/ml.

3.2.2 Nitric acid, HNO₃, $\rho \approx 1,42$ g/ml.

3.2.3 Nitric acid, $c(\text{HNO}_3) = 1,5$ mol/l.

Introduce about 500 ml of water into a 1 000 ml volumetric flask, add 100 ml of nitric acid (see 3.2.2) and dilute to the mark with water.

3.2.4 Hydrogen peroxide, H₂O₂, 30 % mass fraction solution.

3.2.5 Lanthanum chloride, (LaCl₃), solution with $\rho(\text{La}) = 20$ g/l.

Dissolve 23,5 g of lanthanum oxide La₂O₃ in 200 ml of hydrochloric acid (see 3.2.1), dilute to 1 000 ml with water and mix.

WARNING: Observe appropriate precautions. The reaction of La₂O₃ with hydrochloric acid is strongly exothermic.

3.2.6 Chromium solutions.

3.2.6.1 Chromium, stock solution, with $\rho(\text{Cr}) = 1,000$ g/l.

Dry a portion of potassium dichromate (K₂Cr₂O₇) at 105 °C ± 2 °C for about 2 h. Cool and dissolve 2,825 g ± 0,001 g of the dried potassium dichromate in water. Add 5 ml ± 1 ml of nitric acid (see 3.2.2) and dilute to 1 000 ml with water in a volumetric flask.

1 ml of this stock solution contains 1,00 mg of chromium.

Store this solution in either polyethene or borosilicate glass containers at room temperature. The solution is stable at room temperature for about year if it is stored in the dark and at a pH between 1 and 2.

NOTE Chromium stock solutions are commercially available.

3.2.6.2 Chromium, standard solution, with $\rho(\text{Cr}) = 50$ mg/l.

Introduce 50,00 ml ± 0,01 ml of the chromium stock solution (see 3.2.6.1) into a 1 000 ml volumetric flask. Add 1 ml of nitric acid (see 3.2.2), make up to the mark with water and mix.

This solution is stable for at least 1 month.

3.3 Apparatus

3.3.1 Atomic absorption spectrometer, equipped with a chromium hollow cathode lamp and a nitrous oxide/acetylene burner, and operated in accordance with the manufacturer's instructions. It is essential that the manufacturer's safety recommendations are strictly observed when using the nitrous oxide/acetylene flame.

3.3.2 Glassware.

Before use carefully soak all glassware for about 24 h in nitric acid (see 3.2.3), then rinse thoroughly with water.

If low concentrations are expected (for example in ground water), the glassware should be kept under nitric acid (see 3.2.3) until use.

Do not use glassware which has been cleaned with chromic acid.

3.3.3 Membrane filters, of nominal pore diameter 0,45 µm, washed thoroughly with nitric acid (see 3.2.3) and rinsed with water (see 3.2).

3.4 Sampling and preparation of test portions

3.4.1 General

Collect and preserve samples in accordance with ISO 5667-2 and ISO 5667-3. See also annex A.

Collect samples in high density polyethene or borosilicate glass containers which have been previously cleaned with nitric acid (see 3.2.3) and then rinsed with water (see 3.2).

3.4.2 Acid-soluble chromium

Treat the samples by addition of sufficient nitric acid (see 3.2.2), immediately after collection, to adjust the pH to between 1 and 2.

To 90 ml of the acidified sample, or another suitable aliquot in accordance with the expected concentration, add 1 ml of hydrogen peroxide (see 3.2.4) and 2 ml of nitric acid (see 3.2.2). Boil and evaporate to a volume of approximately 50 ml.

The sample shall not be reduced to dryness.

In the case of samples with low chromium concentrations, this digestion method should be carried out in an autoclave or a microwave oven.

It is important to follow the manufacturer's instructions.

Add 10 ml of nitric acid (see 3.2.2) to the evaporated solution. Transfer the solution to a 100 ml volumetric flask. Add 10 ml of lanthanum chloride solution (see 3.2.5) into the flask, dilute to the mark with water and mix.

3.4.3 Water-soluble chromium

Filter the sample through a membrane filter (see 3.3), as soon as possible after collection, and acidify the filtrate immediately with nitric acid (see 3.2.2) to give a pH between 1 and 2.

Add 10 ml of lanthanum chloride solution (see 3.2.5) to a 100 ml volumetric flask and make up to the mark with the acidified filtrate, or use another suitable aliquot in accordance with the expected concentration, and mix.

3.5 Procedure

3.5.1 Blank test

Carry out a blank test in parallel with the determination by the same procedure and using the same quantities of all the reagents as in the sampling and determination, but replacing the test portion by water.

3.5.2 Preparation of the calibration solutions

Before each set of determinations, prepare from the chromium standard solution (see 3.2.6.2) at least five calibration solutions covering the range of concentrations to be determined; for example:

pipette 1,0 ml, 2,5 ml, 5,0 ml, 10,0 ml and 20,0 ml of chromium standard solution (see 3.2.6.2) into a series of 100 ml volumetric flasks. Add to each flask depending on the pretreatment (see 3.4.2 and 3.4.3) 10 ml or 2 ml of nitric acid (see 3.2.2) respectively and 10 ml of lanthanum chloride solution (see 3.2.5), dilute to the mark with water and mix. These solutions correspond to chromium concentrations of 0,50 mg/l; 1,25 mg/l; 2,50 mg/l; 5,00 mg/l and 10,0 mg/l respectively. Proceed accordingly for other concentration ranges of interest.

3.5.3 Calibration

Set up the instrument in accordance with the manufacturer's instructions (at wavelength $\lambda = 357,9$ nm) using a nitrous oxide/acetylene flame.

Aspirate a calibration solution (see 3.5.2) and optimize the aspiration, the burner height and the flame conditions. Adjust the response of the instrument to zero absorbance with water.

In turn, aspirate the set of calibration solutions (see 3.5.2) and the blank solution (see 3.5.1). Plot a graph having the chromium concentrations of the calibration solutions, in milligrams per litre, as abscissa and the corresponding absorbance values as ordinate. Alternatively, data-processing facilities can be used to equate the concentration against the absorbance.

3.5.4 Test portion measurement

Aspirate the prepared test portion (see 3.4.2 or 3.4.3) into the flame and measure the absorbance for chromium. After each measurement aspirate water and readjust the zero, if necessary.

3.6 Expression of results

By reference to the calibration graph obtained (see 3.5.3) or data-processing facilities used, determine the concentration of chromium corresponding to the absorbances of the test portion (see 3.5.4) and of the blank solution (see 3.5.1).

Calculate the concentration of chromium of the sample using the equation:

$$\rho = \frac{(A_s - A_{so}) \cdot V_w}{V_p \cdot b}$$

where:

iTeh STANDARD PREVIEW
(standards.iteh.ai)

- ρ is the chromium concentration of the sample, in milligrams per litre;
- A_s is the absorbance of the test portion;
- A_{so} is the absorbance of the blank;
- V_p is the volume of the acidified sample in accordance with 3.4.2 or 3.4.3, in millilitres;
- b is the sensitivity (slope of the calibration graph), in litres per milligram;
- V_w is the volume of the test portion in accordance with 3.4.2 or 3.4.3, in millilitres.

Precision data for this method are given in Table B.1.

3.7 Interferences

The following ions have been tested and shown not to cause interference below the concentration given in Table 1.

This information was taken from practical tests carried out in a working group in 1984.

Table 1 — Interfering ions

Ion	Concentration mg/l	Ion	Concentration mg/l
Sulfate	10 000	Iron	500
Chloride	12 000	Nickel	100
Sodium	9 000	Copper	100
Potassium	9 000	Cobalt	100
Magnesium	2 000	Aluminium	100
Calcium	2 000	Zinc	100

The total dry residue of the sample shall not exceed 15 g/l and the conductivity should be lower than 2000 mS/m.

NOTE It is possible to determine the effect of possible interferences by analysing samples spiked with chromium and various concentrations of the potential interfering substance using the method of standard additions.

3.8 Test report

The test report shall contain at least the following information:

- a) a reference to this International Standard;
- b) a reference to the procedure used;
- c) complete identification of the sample;
- d) the results of the analysis (see 3.6);
- e) any other information relevant to the procedure.

4 Determination of chromium by electrothermal atomization atomic absorption spectrometry

4.1 Principle

Chromium is determined directly in acidified samples by electrothermal atomization atomic absorption spectrometry. The sample is introduced into an electrically heated graphite tube and the absorbance is measured at a wavelength of 357,9 nm. If necessary, the standard additions technique is employed.

4.2 Reagents

All reagents shall be of recognized analytical grade. Use deionized water or water distilled from an all-glass apparatus. The water used for blank tests and for preparing reagents and standard solutions shall have a chromium content that is negligible compared with the smallest concentrations to be determined in the samples.

4.2.1 Chromium standard solution, with $\rho(\text{Cr}) = 0,5 \text{ mg/l}$.

Add 10,00 ml \pm 0,01 ml of the chromium standard solution (see 3.2.6.2) to a 1 000 ml volumetric flask. Add 10 ml of nitric acid (see 3.2.2) and dilute to the mark with water.

This solution is stable for at least 1 month.

4.3 Apparatus

4.3.1 Atomic absorption spectrometer, equipped with an electrothermal atomizer, chromium hollow cathode lamp and a suitable device for correction of non-specific absorbants.

4.3.2 Atomizer tubes.

The atomizer tubes will vary according to the instrument used but shall be pyrolytically coated and preconditioned before use in accordance with the manufacturer's instructions.

NOTE It is permissible to use a L'vov platform.

4.3.3 Automatic sample injector (autosampler) or microsyringe, of appropriate volume.

For high precision, an automatic sample injector is necessary.

4.4 Sampling and preparation of test portions

Follow the procedures described in 3.4 but omit the lanthanum chloride solution.