



Designation: F526 – 97 (Reapproved 2003)

Standard Test Method for Measuring Dose for Use in Linear Accelerator Pulsed Radiation Effects Tests¹

This standard is issued under the fixed designation F526; the number immediately following the designation indicates the year of original adoption or, in the case of revision, the year of last revision. A number in parentheses indicates the year of last reapproval. A superscript epsilon (ϵ) indicates an editorial change since the last revision or reapproval.

This standard has been approved for use by agencies of the Department of Defense.

1. Scope

1.1 This test method covers a calorimetric measurement of the dose delivered in a single pulse of electrons from an electron linear accelerator used as an ionizing source in radiation-effects testing. The test method is designed for use with pulses of electrons in the energy range from 10 to 50 MeV and is only valid for cases in which both the calorimeter and the test specimen to be irradiated are “thin” compared to the range of these electrons in the materials of which they are constructed.

1.2 The procedure described can be used in those cases in which (1) the dose delivered in a single pulse is 5 Gy² (500 rad) or greater, or (2) multiple pulses of a lower dose can be delivered in a time short compared to the thermal time constant of the calorimeter. The minimum dose per pulse that can be acceptably monitored depends on the variables of the particular test, including pulse rate, pulse uniformity, and the thermal time constant of the calorimeter.

1.3 A determination of the dose is made directly for the material of which the calorimeter block is made. The dose in other materials can be calculated from this measured value by formulas presented in this test method. The need for such calculations and the choice of materials for which calculations are to be made shall be subject to agreement by the parties to the test.

1.4 The values stated in SI units are to be regarded as the standard. The values in parenthesis are provided for information only.

1.5 *This standard does not purport to address the safety concerns, if any, associated with its use. It is the responsibility of the user of this standard to establish appropriate safety and health practices and determine the applicability of regulatory limitations prior to use.*

¹ This test method is under the jurisdiction of ASTM Committee F01 on Electronics and is the direct responsibility of Subcommittee F01.11 on Nuclear & Space Radiation Effects.

Current edition approved June 10, 2003. Published March 1998. Originally published as F526 – 77 T. Last previous edition F526 – 91. DOI: 10.1520/F0526-97R03.

² In 1975 the General Conference on Weights and Measures adopted the unit gray (symbol—Gy) for absorbed dose; 1 Gy = 100 rad.

2. Referenced Documents

2.1 *ASTM Standards*:³

E230 Specification and Temperature-Electromotive Force (EMF) Tables for Standardized Thermocouples

3. Terminology

3.1 *Definitions*:

3.1.1 *thermal time constant of a calorimeter*—the time for the temperature excursion of the calorimeter resulting from a radiation pulse to drop to 1/e of its initial maximum value.

4. Summary of Test Method

4.1 *Single-Pulse Method*—This method consists of (1) irradiating, with a single pulse of high-energy electrons from an electron linear accelerator (linac), a small block of material to which either a thermistor or a thermocouple made from small-diameter wire is attached; (2) recording and measuring the resulting signal from a bridge circuit or directly from the thermocouple; (3) calculating the dose deposited in the block based on the temperature rise and the specific heat of the material; and (4) if required, calculating the equivalent dose in other specified materials.

4.2 *Multiple-Pulse Method*—If the dose available in a single pulse is not large enough to give measurable results, the linac is pulsed repeatedly within a time short compared to the thermal time constant of the calorimeter. This method is similar to the single-pulse method except that the average dose delivered in each pulse is calculated from the measured cumulative dose of all the pulses.

5. Significance and Use

5.1 An accurate measure of the dose during radiation-effects testing is necessary to ensure the validity of the data taken, to enable comparison to be made of data taken at different

³ For referenced ASTM standards, visit the ASTM website, www.astm.org, or contact ASTM Customer Service at service@astm.org. For *Annual Book of ASTM Standards* volume information, refer to the standard's Document Summary page on the ASTM website.

facilities, and to verify that components or circuits are tested to the radiation specification applied to the system for which they are to be used.

5.2 The primary value of a calorimetric method for measuring dose is that the results are absolute. They are based only on physical properties of materials, that is, the specific heat of the calorimeter-block material and the Seebeck emf of the thermocouple used or the temperature coefficient of resistance (α) of the thermistor used, all of which can be established with non-radiation measurements.

5.3 The method permits repeated measurements to be made during a radiation effects test without requiring entry into the radiation cell between measurements.

6. Interferences

6.1 *Thermal Isolation*—If the thermal isolation of the calorimeter is not sufficient, the thermal time constant of the calorimeter response will be too short for it to be useful.

NOTE 1—This condition can be caused by insufficient insulation material or by heat loss through the thermocouple wires themselves.

6.2 *Thermal Equilibrium*—The initial value of the transient temperature change following a radiation pulse may not reflect the true temperature change of the calorimeter-block material.

NOTE 2—This situation can be brought about by a temperature rise occurring in the materials at the point of attachment of the thermocouple or the thermistor different from that in the calorimeter-block material. As long as the calorimeter block comprises the great bulk of the calorimeter

material, the temperature will quickly equilibrate to that of the block, and the subsequent temperature record will be that of the calorimeter-block material (see Appendix X1).

6.3 *Pulse Reproducibility*—If pulse-to-pulse reproducibility of the radiation source varies more than $\pm 20\%$, a good measure of the dose per pulse may not be attainable from the average value calculated in the multiple-pulse method.

7. Apparatus

7.1 *Linac*—Electron linear accelerator and associated instrumentation and controls suitable for use as an ionizing source in radiation-effects testing.

7.2 *Calorimeter*—Special instrument suitable for measuring the dose delivered by the linac and constructed in accordance with any of several designs utilizing any of several materials as indicated in Appendix X1. Although measurement differences resulting from the use of different designs should not be significant, all parties to the test shall agree to a single design utilizing a single calorimeter-block material and a specific thermocouple or thermistor. The calorimeter design shall be such that the surface density in the beam path is less than or equal to no more than 20% of the range of the beam-energy electrons (see Fig. 1).

7.3 *D-C Low Noise Amplifier (LNA)*, with a gain of 1000 to 10 000 (see Fig. 2).

NOTE 3—An analog nanovoltmeter with a recorder output can also be used as a low noise amplifier. These devices produce a 1-V output for a full scale reading.





FIG. 1 Block Diagram of Calorimeter Dosimeter Circuit



S1 for Auto Zero

FIG. 2 FIG. Recommended Low Noise Amplifier Schematic Diagram

7.3.1 Response time no greater than 0.1 s for the amplifier output to reach 90 % of its final reading,

7.3.2 Noise level less than 10 mV rms referred to the output,

7.3.3 Measurement accuracy of 2 % of full scale or better, and

7.3.4 Normal-mode rejection capability such that a-c voltages of 50 Hz and above and 60 dB greater than the range setting shall affect the instrument reading by less than 2 %.

NOTE 4—If the meter does not have an internal nulling circuit, it may be necessary to use a simple bucking circuit to null out thermal emfs in the measuring circuit to keep the meter on scale at the high-gain positions used in this measurement (see Fig. 1).

7.4 Data Recorder—Linear-response recorder meeting the following specifications:

7.4.1 Recording speed sufficient to capture 5 to 10 s of calorimeter response.

7.4.2 Response time for full-scale deflection of 0.1 s or less.

7.4.3 Deviation of response from linearity of no more than ± 2 %, and

7.4.4 Sensitivity compatible with the recorder output of the d-c LNA. Typically 2mV full scale.

7.5 Voltage Calibration Source—Voltage source capable of meeting the following specifications:

7.5.1 Output voltages including 1.5, 3.0, 5.0, 10.0, 15, 30, and 50 μ V, and 100 μ V,

7.5.2 Accuracy of ± 1 % of the selected voltage, or better,

7.5.3 Thermally generated voltages of less than 100 nV with the source stabilized, and

7.5.4 Source resistance of 100 Ω or less.

7.6 Wheatstone Bridge Circuit, designed so that the thermistor forms one leg of the bridge, and so that the adjustable resistor of the bridge will be equal to the resistance of the thermistor at balance (see Fig. 1B).

8. Sampling

8.1 The number of measurements shall be subject to agreement by the parties to the test.

9. Calibration

9.1 The LNA and data recorder should be calibrated to be within ± 2 % of full scale.

10. Procedure

10.1 Single-Pulse Method:

10.1.1 Position the calorimeter at the location where the dose measurement is desired.

10.1.2 Connect all components of the calorimetric dosimeter system in accordance with the circuit shown in Fig. 1.

10.1.3 Set the LNA for a gain of 10 000 (1000), if using the thermistor circuit).

10.1.4 For the thermocouple measurements, adjust either the internal nulling circuit of the LNA or the external bucking circuit so that the meter deflection caused by the quiescent level of the calorimeter output is less than full scale. For thermistor measurements adjust the bridge for a null. Use the zero-adjust capability of the data recorder to position the recorder trace near the center of the recorder chart.

NOTE 5—With either system, there will likely be a drift as the temperature of the calorimeter equilibrates. This drift is compensated for in data reduction and may be neglected if the rate of change is much less than that caused by the radiation pulse.

10.1.5 With the data recorder sweep speed set within the range from 0.5 to 2.0 cm/s, inclusive, trigger the recorder and pulse the linac.

10.1.6 If the transient deflection of the recorder is less than 10 % of full scale, set the recorder range to the next lower range and repeat 10.1.5.

10.1.7 Repeat 10.1.5 and 10.1.6 until a range is found for which the greater-than-10 % criterion is met, or until there are no more ranges to try.

10.1.7.1 When a range is found for which this greater-than-10 % criterion is met, annotate the data recorder output beside the recorded transient with the shot number, date, LNA gain, calorimeter identification, and description of irradiation geometry (including scatterer thickness and distance of the calorimeter from the scatterer) as shown in Fig. 3 and Fig. 4.

10.1.7.2 If no range is found for which a 10 % deflection is obtained which is easily distinguishable from noise, use the multiple-pulse method beginning with 10.2.2.

10.1.7.3 Otherwise, repeat 10.1.7.1 four more times.

10.2 Multiple-Pulse Method:

10.2.1 Carry out 10.1.1 through 10.1.4.

10.2.2 With the recorder chart speed set within the range from 0.5 to 2.0 cm/s, inclusive, pulse the Linac repeatedly within a time that is short compared to the thermal time constant of the calorimeter to give a recorder deflection greater than 10 % of full scale.




FIG. 3 Typical Chart Record of Calorimeter Dosimetry Using Single-Pulse Method




FIG. 4 Typical Digital Oscilloscope Recording of the Calorimeter Response

10.2.2.1 From the data, measure the voltage rise resulting from this series of pulses.

10.2.2.2 For the time interval beginning with the cessation of the radiation and equal in duration to the total time during which the radiation dose was accumulated, measure the thermocouple voltage drop.

10.2.2.3 Calculate the ratio of the voltage from 10.2.2.2 to that of 10.2.2.1.

10.2.2.4 If this ratio is less than 0.15, continue with 10.2.3 (the thermal time constant of the calorimeter is sufficiently greater than the radiation time for the dose to be determined accurately).

10.2.2.5 If this ratio is equal to or greater than 0.15, repeat 10.2.2 through 10.2.2.5 using a higher pulse repetition rate for a shorter radiation time period.

10.2.3 Annotate the data recorder output, as well as the number of pulses used (see Fig. 5, Fig. 6, and Fig. 7).

10.2.4 Repeat 10.2.2 and 10.2.3 four more times, omitting the time constant determination (10.2.2.1 through 10.2.2.5).

11. Calculation and Interpretation of Results

11.1 Single-Pulse Method:

11.1.1 On the recorder output, determine the perpendicular to the time axis at the start of each transient, as shown in Fig. 3.

11.1.2 Determine whether a period of time was required for the temperature to equilibrate after the pulse, as indicated by the presence of a spike (Fig. 5a) or a flat portion (Fig. 5b) of the data recorder trace at the end of the transient.

11.1.2.1 If no such feature is present, draw a line extrapolating the steepest part of the cooling curve following each radiation pulse back to intersect the perpendicular line (see 11.1.1). When using digital storage oscilloscopes, built in cursors usually can be used.

NOTE 6—These lines are dashed in Fig. 3.

11.1.2.2 If such a feature is present, draw a line extrapolat-

TABLE 1 Physical Properties of Some Calorimeter-Block Materials

Material	Energy Loss ^A dE/dx (10 ⁻¹⁴ J·m ² /kg)	Specific Heat, c _p ^B (J/kg·K)	Density, ρ ^B (10 ³ kg/m ³)
C	2.92	711	2.10
Al	2.74	900	2.70
Si	2.84	711	2.33
Fe	2.52	452	7.87
Cu	2.42	385	8.96
Ge	2.45	322	5.32
W	2.08	134	19.3
Au	2.06	130	19.3
Pb	2.07	128	11.4

^AThe data are given for 20-MeV electrons, but ratios based on these values are good to better than 2 % over the energy range from 10 to 50 MeV, inclusive. These values have been converted to SI units from data given in the tables of Berger and Seltzer: *Tables of Energy Losses and Ranges of Electrons and Positrons*, NASA SP-3012 (1964); and *Additional Stopping Powers and Range Tables for Protons, Mesons, and Electrons*, NASA SP-3036 (1966).

^BThese values have been converted to SI units from data given in the *Handbook of Tables for Applied Engineering Science*, 2nd ed., CRC Press, Cleveland, OH (1973). (The specific heat values are applicable in the range from 18 to 30°C, inclusive.)

ing from the slope of the curve where a smooth cooling trend resumes. Do this for each pulse.


NOTE 7—These lines are dashed in Fig. 5.

11.1.3 Measure along each perpendicular line the length from the start of each transient to the intersection of the perpendicular line with the extrapolated line.


11.1.4 Convert these measurements to output voltage level.

11.1.5 For each pulse calculate and record the dose in Gy (calorimeter-block material) producing the transient, using for a thermocouple measurement, the relation:

$$\text{Dose} = 100 V_{c_p} / PG \quad (1)$$



(a) Spike Indicating Initial Thermocouple Junction Temperature Higher than that of the Calorimeter Block.



(b) Flat Portion Indicating Initial Thermocouple Junction Temperature Lower than that of the Calorimeter Block.

FIG. 5 Possible Aberrations Observed in Strip-Chart Recorder Transient Signals

where:

V = deflection caused by irradiation pulse, in micrometres,

c_p = specific heat of calorimeter-block material, J/kg·K,

P = temperature coefficient of the calorimeter thermocouple in the vicinity of room temperature, $\mu\text{V}/\text{K}$,

G = gain of low noise amplifier, and,

100 = numerical conversion factor, rad·kg/J.

11.1.6 For a thermistor measurement, use the equation (Appendix X2):

$$\text{Dose} = \frac{(R_A + R_B)^2 k c_p}{R_A R_B \alpha E} V \quad (2)$$

where:

R_A = value of the fixed bridge resistors, Ω ,

R_B = value of the variable bridge resistor, Ω ,

k = numerical conversion constant = 10^{-2} J/kg·rad,

α = thermistor temperature coefficient of resistance, K^{-1} ,

E = bridge voltage, V, and

V and c_p have the same meaning as above.

11.1.7 Average and record the results obtained from the above calculation for each of the five radiation pulses,

11.2 Multiple-Pulse Method:

11.2.1 Draw a line perpendicular to the time axis at the time midway between the start and end of the sets of multiple radiation pulses, as shown in Fig. 6.

11.2.2 For each multiple-pulse transient, draw a linear extrapolation of the cooling curve immediately preceding the radiation, and extend it to intercept the perpendicular line (see 11.2.1).

NOTE 8—These lines are dashed in Fig. 6.

11.2.3 For each transient, draw a line extrapolating back the cooling curve, following the transient, to intercept the perpendicular line drawn for that transient.

NOTE 9—These lines are also dashed in Fig. 6.

11.2.4 For each transient, measure the length along the perpendicular line between the intersections with the extended and extrapolated lines.

11.2.5 Convert these measurements to fractions of full-scale width.

11.2.6 Calculate and record the dose delivered in each burst of multiple pulses in accordance with 11.1.5.

11.2.7 Divide the dose calculated for each set of pulses by the number of pulses in the set to obtain the average dose per pulse for that set. Record these figures.

11.2.8 Average the five values obtained. Record this figure.

NOTE 10—This figure provides the best estimate of the average dose per pulse. However, this average value is seldom useful if the pulse-to-pulse reproducibility is not within $\pm 20\%$ of a median value.

11.3 Dose Conversion:

11.3.1 To convert the dose measured in 11.1 or 11.2 to dose in a material other than that of the calorimeter block, use the equation:

$$\text{Dose B} = \frac{dE / dx_{(B)}}{dE / dx_{(A)}} \text{Dose A} \quad (3)$$