



# Sterilization of health care products — Requirements for validation and routine control — Radiation sterilization

## TECHNICAL CORRIGENDUM 1

*Stérilisation des dispositifs médicaux — Prescriptions pour la validation et le contrôle de routine — Stérilisation par irradiation*

RECTIFICATIF TECHNIQUE 1

**iTeh STANDARD PREVIEW**  
(standards.iteh.ai)

Technical Corrigendum 1 to International Standard ISO 11137:1995 was prepared by Technical Committee ISO/TC 198, *Sterilization of health care products*.

[ISO 11137:1995/Cor 1:1997](https://standards.iteh.ai/catalog/standards/sist/4ba60c05-1aeb-4686-87cb-d0e1daeb7c7/iso-11137-1995-cor-1-1997)

<https://standards.iteh.ai/catalog/standards/sist/4ba60c05-1aeb-4686-87cb-d0e1daeb7c7/iso-11137-1995-cor-1-1997>

Page 25

### **B.3.4.2.2.3 Stage 3: Perform verification dose experiment**

(For Method 2A) Replace the last two sentences of the first paragraph with the following text:

“The actual dose may vary from the D\* kGy dose by + 1,0 kGy or + 10 %, whichever is greater. If the delivered dose is less than 90 % of the D\* kGy dose, the verification dose experiment may be repeated.”

Page 28

### **B.3.4.2.4.3 Stage 3: Perform verification dose experiment**

(For Method 2B) Replace the last two sentences of the first paragraph with the following text.

“The actual dose may vary from the D\* kGy dose by + 1,0 kGy or + 10 %, whichever is greater. If the delivered dose is less than 90 % of the D\* kGy dose, the verification dose experiment may be repeated.”

**iTeh STANDARD PREVIEW**  
This page intentionally left blank  
**(standards.iteh.ai)**

[ISO 11137:1995/Cor 1:1997](https://standards.iteh.ai/catalog/standards/sist/4ba60c05-1aeb-4686-87cb-d0e1daeb7c7/iso-11137-1995-cor-1-1997)

<https://standards.iteh.ai/catalog/standards/sist/4ba60c05-1aeb-4686-87cb-d0e1daeb7c7/iso-11137-1995-cor-1-1997>