

TECHNICAL SPECIFICATION
SPÉCIFICATION TECHNIQUE
TECHNISCHE SPEZIFIKATION

FINAL DRAFT
FprCEN/TS 16317

September 2011

ICS 65.080

English Version

Fertilizers - Determination of trace elements - Determination of arsenic by inductively coupled plasma-atomic emission spectrometry (ICP-AES) after aqua regia dissolution

Düngemittel - Bestimmung von Elementspuren -
Bestimmung von Arsen mit Atomemissionsspektrometrie
mit induktiv gekoppeltem Plasma (ICP-AES) nach
Königswasserauflösung

This draft Technical Specification is submitted to CEN members for formal vote. It has been drawn up by the Technical Committee CEN/TC 260.

CEN members are the national standards bodies of Austria, Belgium, Bulgaria, Croatia, Cyprus, Czech Republic, Denmark, Estonia, Finland, France, Germany, Greece, Hungary, Iceland, Ireland, Italy, Latvia, Lithuania, Luxembourg, Malta, Netherlands, Norway, Poland, Portugal, Romania, Slovakia, Slovenia, Spain, Sweden, Switzerland and United Kingdom.

Recipients of this draft are invited to submit, with their comments, notification of any relevant patent rights of which they are aware and to provide supporting documentation.

Warning : This document is not a Technical Specification. It is distributed for review and comments. It is subject to change without notice and shall not be referred to as a Technical Specification.



EUROPEAN COMMITTEE FOR STANDARDIZATION
COMITÉ EUROPÉEN DE NORMALISATION
EUROPÄISCHES KOMITEE FÜR NORMUNG

Management Centre: Avenue Marnix 17, B-1000 Brussels

Contents

Page

Foreword.....	3
1 Scope	4
2 Normative references	4
3 Terms and definitions	4
4 Principle.....	4
5 Sampling and sample preparation	4
6 Reagents.....	4
7 Apparatus	5
8 Procedure	6
9 Calculation and expression of the results	9
10 Precision.....	11
11 Test report	12
Annex A (informative) Results of the inter-laboratory test	13
A.1 Inter-laboratory tests.....	13
A.2 Statistical results for the determination of arsenic by ICP-AES.....	13
Bibliography	14

Foreword

This document (FprCEN/TS 16317:2011) has been prepared by Technical Committee CEN/TC 260 "Fertilizers and liming materials", the secretariat of which is held by DIN.

This document is currently submitted to the Formal Vote.

This document has been prepared under a mandate given to CEN by the European Commission and the European Free Trade Association.

FprCEN/TS 16317:2011 (E)

1 Scope

This document specifies a method for the determination of the content of arsenic in fertilizers using inductively coupled plasma-atomic emission spectrometry (ICP-AES) after aqua regia dissolution. Limits of quantification are dependant on the sample matrix as well as on the instrument, but can roughly be expected to be 1,5 mg/kg for As.

2 Normative references

The following referenced documents are indispensable for the application of this document. For dated references, only the edition cited applies. For undated references, the latest edition of the referenced document (including any amendments) applies.

EN 1482-2, *Fertilizers and liming materials — Sampling and sample preparation — Part 2: Sample preparation*

EN 12944-1:1999, *Fertilizers and liming materials and soil improvers — Vocabulary — Part 1: General terms*

EN 12944-2:1999, *Fertilizers and liming materials and soil improvers — Vocabulary — Part 2: Terms relating to fertilizers*

EN ISO 3696, *Water for analytical laboratory use — Specification and test methods (ISO 3696:1987)*

3 Terms and definitions

For the purposes of this document, the terms and definitions given in EN 12944-1:1999 and EN 12944-2:1999 apply.

4 Principle

Arsenic is extracted from the sample with aqua regia and conventional boiling. The concentration in the extract is measured by inductively coupled plasma-atomic emission spectrometry (ICP-AES) with axial or radial viewing.

5 Sampling and sample preparation

Sampling is not part of the methods specified in this Technical Specification. A recommended sampling method is given in EN 1482-1.

Sample preparation shall be carried out in accordance with EN 1482-2.

6 Reagents

Use only reagents of recognized analytical grade.

Commercially available stock solutions shall be replaced according to the specifications from supplier or after one year if prepared in the laboratory from available salts. Standard solutions shall be renewed monthly as a general rule of thumb.

- 6.1 Water**, conforming to grade 2 according to EN ISO 3696.
- 6.2 Hydrochloric acid**, $c(\text{HCl}) = 12 \text{ mol/l}$; 37 % mass/volume; $\rho \approx 1,18 \text{ g/ml}$.
- 6.3 Nitric acid**, $c(\text{HNO}_3) = 16 \text{ mol/l}$; not less than 65 % mass/volume, $\rho \approx 1,42 \text{ g/ml}$.
- 6.4 Mixed solution of 0,8 mol/l nitric acid and 1,8 mol/l hydrochloric acid.**

Mix 150 ml of hydrochloric acid (6.2) and 50 ml nitric acid (6.3) to 1,0 l of water (6.1).

- 6.5 Standard stock solution**, arsenic standard stock solution, e.g. $\rho = 1\,000 \text{ mg/l}$ in arsenic.

Use suitable stock solutions. Single-element stock solutions with adequate specification stating the acid used and the preparation technique are commercially available. It is recommended to use a commercially available standard stock solution for arsenic.

6.6 Working standard solutions.

Depending on the scope, different working standard solutions may be necessary.

- 6.6.1 Working standard solution I**, $\rho = 100 \text{ mg/l}$ for arsenic.

Dilute 10,0 ml of the stock solution of arsenic (6.5) to 100,0 ml with the mixed acid solution (6.4) in the same 100 ml flask. This solution is used to prepare spiked test solutions and standard and calibration solutions.

- 6.6.2 Working standard solution II**, $\rho = 10 \text{ mg/l}$ for arsenic.

Dilute 10,0 ml of the working standard solution I (6.6.1) to 100,0 ml with the mixed acid solution (6.4) in a 100 ml flask. This solution is used to prepare spiked test solutions and calibration solutions.

7 Apparatus

- 7.1 Common laboratory glassware.**

- 7.2 Analytical balance**, capable of weighing to an accuracy of 1 mg.

- 7.3 Inductively coupled plasma-atomic emission spectrometer**, with axial or radial viewing of the plasma and with suitable background correction.

The settings of the working conditions (e.g. gas flows, RF or plasma power, sample uptake rate, integration time, number of replicates) shall be optimised according the manufacturer's instructions. Radial viewing of the plasma may be used if it can be shown that the limit of quantification for arsenic is below the required legal limit values.

The use of axial orientation of the viewing optics requires good control on the matrix effects coming from „easily ionisable elements“ (i.e. the influence of easily ionisable elements in varying concentrations on the signal intensities of the analytes). For alkali-elements, this can be done by adding caesium-chloride solution (CsCl). In general, matrix matching of calibration solutions or calibration by standard additions with several calibration standards will correct accurately for these matrix effects. Spike recovery of one known standard combined with external calibration can, if used properly, also correct sufficiently for matrix effects (see 8.1). Correction by internal standardisation is also a good option, but the accuracy of the measurements after internal standard correction should be validated properly prior to use on unknown fertilizer samples.

FprCEN/TS 16317:2011 (E)

7.4 Dilutor.

Instrument used for automated volumetric dilutions or other appropriate equipment (e.g. pipettes and volumetric glassware) to perform dilutions. The precision and accuracy of this type of equipment for volumetric dilutions shall be established, and controlled and documented regularly.

7.5 Ash-free filter paper, i.e. Whatman 589/2 or equivalent quality.

8 Procedure

8.1 General

Calibrations by standard additions with several standards or by matrix matching are very powerful calibration techniques and can be used to accurately correct for matrix effects from easy-ionisable elements (multiplicative matrix effects). Additive matrix effects (i.e. spectral interferences) are not corrected for with standard additions calibration. For matrix matching, additive matrix effects will be corrected for when the added matrix is the cause of the matrix effect. The main drawback of calibration by standard addition with several standards is the requirement for a calibration function for each sample type, which is a time consuming process. For matrix matching a profound knowledge of the sample matrix is needed, which is not necessarily available. These two techniques may thus not be practical to use in routine fertilizer laboratories.

Correction by internal standardisation is also a good option, but the accuracy of the measurements after internal standard correction should be validated properly prior to use on unknown fertilizer samples.

It is therefore suggested that calibration are to be performed by means of external calibration and correction of matrix effects by addition of one known spike of a standard solution (spike recovery). The method of external calibration and correction for spike recovery allows for the analysis of fertilisers with unknown matrix composition or with a matrix that cannot be synthetically imitated easily. This calibration technique may not be as precise as calibration by standard additions with several standards but the increased uncertainty is small compared to the total uncertainty of the method, if the total analyte concentration is in the linear working range after the spike and the added spike corresponds to at least a doubling of the analyte concentration. Many matrix errors can be compensated for by this procedure, if they are not additive (e.g. spectral interferences). Aliquots of the sample solution are analysed by the mean of external calibration and then one aliquot is spiked with known concentrations of the analytes without changing the matrix of the sample solution. The calculated spike recovery is then used to correct the concentration calculated from the external calibration function. The concentration of the spikes shall be in the linear working range of the ICP-AES.

8.2 Preparation of the test solution

8.2.1 General

The following extraction procedure leads, in most cases, for mineral fertilisers, to trace element results which correspond to the total contents of these elements.

Calibration with several standard additions and external calibration after matrix matching or by correction for matrix effects with internal standardization may also be used.

8.2.2 Preparation

8.2.2.1 Weigh $3 \text{ g} \pm 0,003 \text{ g}$ of the prepared sample and transfer to a suitable reaction vessel (action 1).

8.2.2.2 Moisten the sample with about 0,5 ml to 1,0 ml of water (6.1) and add, with mixing, $21 \text{ ml} \pm 0,1 \text{ ml}$ of hydrochloric acid (6.2) followed by $7 \text{ ml} \pm 0,1 \text{ ml}$ of nitric acid (6.3) drop wise if necessary to reduce foaming. Connect a condenser to the reaction vessel and let the mixture stand at room temperature until any effervescence almost ceases to allow for slow oxidation of any organic mass in the sample (action 2).

8.2.2.3 Transfer to the heating device and raise the temperature of the reaction mixture slowly to reflux conditions and maintain for 2 h ensuring that the condensation zone is lower than 1/3 of the height of the condenser, then allow to cool. Rinse the condenser further with 10 ml of water (6.1) (action 3).

If the digested sample contains particulates which can clog nebulisers or interfere with the injection of the sample, the sample should be centrifuged, allowed to settle, or filtered before transferring into a suitable sized volumetric flask. For example, the solution should be allowed to pass through the filter paper and then the insoluble residue washed onto the filter paper with a minimum of water (6.1). The method used has to be reported in the test report. Filter paper may cause contaminations (e.g. lead) and it may be necessary to use ash-free filter paper (7.5).

8.2.2.4 Transfer the digested sample into a 150 ml volumetric flask and dilute to the mark with water (6.1). This yields an acid concentration approximately equal to the mixed acid solution (6.4). This test solution corresponds to a 50 times dilution of the solid sample (action 4).

8.2.2.5 Test solutions are diluted with the mixed acid solution (6.4) using a dilutor (7.6) to obtain a concentration of arsenic between 0,03 mg/l and 5 mg/l (action 5).

NOTE 1 It is important that the total dilution of the test solution is equal to the dilution of the spiked test solution (see 8.3 on how to prepare the spiked test solution).

NOTE 2 The concentration of arsenic in the solution in action 5 (8.2.2.5) is chosen such that it is above the typical limits of quantification, and that the concentration fall within the linear working range of the analytical technique.

8.3 Preparation of the test solution for the correction of matrix effects by spike recovery

For each test solution analysed, a spiked test solution with a known addition of a standard solution is required to correct for matrix effects by correcting for the spike recovery measured under close to identical measurement conditions. The addition of a spike of the standard solution shall increase the analyte concentrations by at least 100 % without changing the matrix of the test solution (from 8.2.2.4, action 4) or the diluted test solution (from 8.2.2.5, action 5). See Example 1 and Example 2 below for suggestions on how to spike a diluted and an undiluted test solution when determining arsenic.

EXAMPLE 1 If the test solution (from 8.2.2.4, action 4 or from 8.2.2.5, action 5) contains 1,0 mg/l to 5 mg/l of arsenic, spike addition corresponding to 1,0 mg/l of a multi-element standard solution containing arsenic can be done while diluting the sample 5 times. Thus, taking 2,00 ml test solution and adding 1,00 ml 10 mg/l working standard solution II (6.6.2) and 7,00 ml mixed acid solution (6.4). The test solution (from 8.2.2.4, action 4 or from 8.2.2.5, action 5) shall also be 5 times diluted with the mixed acid solution (6.4) prior to analysis by ICP-AES.

EXAMPLE 2 If the test solution (from 8.2.2.4, action 4) contains 0,03 mg/l to 1,0 mg/l arsenic it should be analysed without further dilution. Add 0,100 ml of a suitable standard solution (e.g. 100 mg/l working standard solution I (6.6.1) corresponding to an addition of 1,0 mg/l of arsenic) to 9,90 ml test solution, thus preparing a spiked test solution of 10,0 ml without changing the matrix of the test solution significantly. The test solution (from 8.2.2.4, action 4) is measured using the same dilution (9,90 ml test solution and 0,100 ml mixed acid solution (6.4)) by ICP-AES.

8.4 Preparation of the blank test solution

Carry out a blank test at the same time as the extraction, with only the reagents and follow the same procedure as for the samples. The blank test solutions should be analysed without further dilution to achieve best possible detection capability. Contaminations from arsenic in the mixed acid solution used for further dilutions of the sample test solutions should be checked before each analysis (i.e. by observing the corresponding analyte signals in the calibration blank solutions or acid blanks).