

SLOVENSKI STANDARD

SIST-TS CEN/TS 16450:2013

01-december-2013

Zunanji zrak - Avtomatski merilni sistemi za merjenje koncentracije delcev (PM10; PM2,5)

Ambient air - Automated measuring systems for the measurement of the concentration of particulate matter (PM10; PM2,5)

Außenluft - Automatische Messeinrichtungen zur Bestimmung der Staubkonzentration (PM10; PM2,5)

Air ambiant - Systèmes automatisés de mesure de la concentration de matière particulaire (PM10; PM2,5)

iTeh STANDARD PREVIEW
(standards.iteh.ai)
<https://standards.iteh.ai/catalog/standards/sist/97a03c44-4564-4220-a992-68b51d02c91a/sist-ts-cen-ts-16450-2013>

Ta slovenski standard je istoveten z: CEN/TS 16450:2013

ICS:

13.040.20 Kakovost okoljskega zraka Ambient atmospheres

SIST-TS CEN/TS 16450:2013

en,fr,de

iTeh STANDARD PREVIEW
(standards.iteh.ai)

SIST-TS CEN/TS 16450:2013

<https://standards.iteh.ai/catalog/standards/sist/97a03c44-4564-4220-a992-68b51d02c91a/sist-ts-cen-ts-16450-2013>

TECHNICAL SPECIFICATION
SPÉCIFICATION TECHNIQUE
TECHNISCHE SPEZIFIKATION

CEN/TS 16450

May 2013

ICS 13.040.20

English Version

**Ambient air - Automated measuring systems for the
measurement of the concentration of particulate matter (PM10;
PM2,5)**

Air ambiant - Systèmes automatisés de mesurage de la
concentration de matière particulaire (PM10 ; PM2,5)

Außenluft - Automatische Messeinrichtungen zur
Bestimmung der Staubkonzentration (PM10; PM2,5)

This Technical Specification (CEN/TS) was approved by CEN on 6 November 2012 for provisional application.

The period of validity of this CEN/TS is limited initially to three years. After two years the members of CEN will be requested to submit their comments, particularly on the question whether the CEN/TS can be converted into a European Standard.

CEN members are required to announce the existence of this CEN/TS in the same way as for an EN and to make the CEN/TS available promptly at national level in an appropriate form. It is permissible to keep conflicting national standards in force (in parallel to the CEN/TS) until the final decision about the possible conversion of the CEN/TS into an EN is reached.

CEN members are the national standards bodies of Austria, Belgium, Bulgaria, Croatia, Cyprus, Czech Republic, Denmark, Estonia, Finland, Former Yugoslav Republic of Macedonia, France, Germany, Greece, Hungary, Iceland, Ireland, Italy, Latvia, Lithuania, Luxembourg, Malta, Netherlands, Norway, Poland, Portugal, Romania, Slovakia, Slovenia, Spain, Sweden, Switzerland, Turkey and United Kingdom.

[SIST-TS CEN/TS 16450:2013](https://standards.iteh.ai/catalog/standards/sist/97a03c44-4564-4220-a992-68b51d02c91a/sist-ts-cen-ts-16450-2013)

<https://standards.iteh.ai/catalog/standards/sist/97a03c44-4564-4220-a992-68b51d02c91a/sist-ts-cen-ts-16450-2013>



EUROPEAN COMMITTEE FOR STANDARDIZATION
COMITÉ EUROPÉEN DE NORMALISATION
EUROPÄISCHES KOMITEE FÜR NORMUNG

Management Centre: Avenue Marnix 17, B-1000 Brussels

Contents

Page

Foreword.....	3
1 Scope	4
2 Normative references	5
3 Terms and definitions	5
4 Symbols and abbreviated terms	9
5 Principle.....	10
6 Sampling	11
7 Performance criteria and test procedures	12
8 Field operation and ongoing quality control	27
9 Data handling, validation and data reports	33
10 Expression of results	34
11 Test reports and documentation.....	35
Annex A (informative) Examples of principles of AMS for monitoring of particulate matter	36
Annex B (normative) Orthogonal regression algorithms.....	39
Annex C (normative) Performing calibrations of the AMS.....	41
Annex D (normative) Elements of type approval report.....	42
Annex E (informative) Elements of suitability evaluation report.....	44
Annex F (informative) Relationship with EU Directives	45
Bibliography.....	46

Foreword

This document (CEN/TS 16450:2013) has been prepared by Technical Committee CEN/TC 264 “Air quality”, the secretariat of which is held by DIN.

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. CEN [and/or CENELEC] shall not be held responsible for identifying any or all such patent rights.

According to the CEN-CENELEC Internal Regulations, the national standards organisations of the following countries are bound to announce this Technical Specification: Austria, Belgium, Bulgaria, Croatia, Cyprus, Czech Republic, Denmark, Estonia, Finland, Former Yugoslav Republic of Macedonia, France, Germany, Greece, Hungary, Iceland, Ireland, Italy, Latvia, Lithuania, Luxembourg, Malta, Netherlands, Norway, Poland, Portugal, Romania, Slovakia, Slovenia, Spain, Sweden, Switzerland, Turkey and the United Kingdom.

iTeh STANDARD PREVIEW (standards.iteh.ai)

[SIST-TS CEN/TS 16450:2013](https://standards.iteh.ai/catalog/standards/sist/97a03c44-4564-4220-a992-68b51d02c91a/sist-ts-cen-ts-16450-2013)

<https://standards.iteh.ai/catalog/standards/sist/97a03c44-4564-4220-a992-68b51d02c91a/sist-ts-cen-ts-16450-2013>

CEN/TS 16450:2013 (E)

1 Scope

In order to be in compliance with EU Air Quality Directive requirements [1], the reference methods given in the Directive for the measurement of mass concentrations of particulate matter are not commonly used for operation in routine monitoring networks. These networks usually apply automated continuous measurement systems (AMS), such as those based on the use of oscillating microbalances or β -ray attenuation, and on in-situ optical methods. Such AMS are typically capable of producing 24-hour average measurement values over a measurement range up to 1 000 $\mu\text{g}/\text{m}^3$ and 1-hour average measurement values up to 10 000 $\mu\text{g}/\text{m}^3$, if applicable, where the volume of air is the volume at ambient conditions near the inlet at the time of sampling.

The 1-hour average values may be used for:

- direct information of the public;
- aggregation to produce daily or yearly average concentration values for regulatory reporting purposes.

EU Air Quality Directive 2008/50/EC [1] allows the use of such systems after demonstration of equivalence with the reference method, i.e., after demonstration that these systems meet the Data Quality Objectives for continuous measurements. Guidelines for the demonstration of equivalence are given in Reference [2].

This Technical Specification lays down the minimum performance requirements and test procedures for the selection of appropriate AMS for particulate matter (type approval). This includes the evaluation of its equivalence with the reference method.

Further, this Technical Specification describes minimum requirements for ongoing quality assurance – quality control (QA/QC) of AMS deployed in the field. These requirements are necessary to ensure that uncertainties of measured concentrations are kept within the required limits during extended periods of continuous monitoring in the field, and include procedures for maintenance, calibration and control checks.

Additional procedures are described that determine whether an instrument's equivalence to the reference method is maintained through possible pollution climate changes, over periods longer than five years.

Lastly, this Technical Specification describes requirements and procedures for the treatment and validation of raw measurement data that are to be used for the assembly of daily or yearly average concentration values. Experiences with existing methods for data treatment and validation – for similar AMS – have learned that the different ways of data treatment and validation applied may lead to significant differences in reported results for similar datasets [3].

When the Technical Specification is used for other purposes than the EU Directive, the range and uncertainty requirements may not apply.

This Technical Specification contains information for different groups of users.

Clauses 5 and 6 and Annex A contain general information about the principles of automated continuous measurement systems for particulate matter, and relevant equipment.

Clause 7 and Annexes B and C are specifically directed towards test houses and laboratories that perform type-approval testing of automated continuous measurement systems for particulate matter. These clauses contain information about:

- type-approval test conditions, test procedures and test requirements;
- system performance requirements;
- evaluation of the type-approval test results;
- evaluation of the uncertainty of the measurement results of the automated continuous measurement systems for particulate matter based on the type-approval test results.

Clauses 8 to 11 are directed towards monitoring networks performing the practical measurements of particulate matter in ambient air. These clauses contain information about:

- initial installation of the system in the monitoring network and acceptance testing;
- ongoing quality assurance/quality control;
- verification of equivalence;
- treatment, validation and reporting of measurement results.

2 Normative references

The following documents, in whole or in part, are normatively referenced in this document and are indispensable for its application. For dated references, only the edition cited applies. For undated references, the latest edition of the referenced document (including any amendments) applies.

EN 12341, *Air quality — Determination of the PM_{10} fraction of suspended particulate matter — Reference method and field test procedure to demonstrate reference equivalence of measurement methods*

EN 15267-1, *Air quality — Certification of automated measuring systems — Part 1: General principles*

EN 15267-2, *Air quality — Certification of automated measuring systems — Part 2: Initial assessment of the AMS manufacturer's quality management system and post certification surveillance for the manufacturing process*

iTeh STANDARD PREVIEW

3 Terms and definitions (standards.iteh.ai)

For the purpose of this document, the following terms and definitions apply.

3.1

ambient air

outdoor air in the troposphere, excluding workplaces as defined by Directive 89/654/EEC [5] where provisions concerning health and safety at work apply and to which members of the public do not have regular access

[SOURCE: Directive 2008/50/EC [1]]

3.2

automated measuring system

AMS

entirety of all measuring instruments and additional devices necessary for obtaining a measurement result

3.3

availability of the AMS

fraction of the time period for which valid measuring data of the ambient air concentration is available from an AMS

[SOURCE: EN 14211 [6]]

CEN/TS 16450:2013 (E)

3.4

calibration

operation that, under specified conditions, in a first step, establishes a relation between the quantity values with measurement uncertainties provided by measurement standards and corresponding indications with associated measurement uncertainties and, in a second step, uses this information to establish a relation for obtaining a measurement result from an indication

Note 1 to entry: A calibration may be expressed by a statement, calibration function, calibration diagram, calibration curve, or calibration table. In some cases, it may consist of an additive or multiplicative correction of the indication with associated measurement uncertainty.

Note 2 to entry: Calibration should not be confused with adjustment of a measuring system, often mistakenly called "self-calibration", nor with verification of a calibration.

[SOURCE: JCGM 200:2012 (VIM) [7]]

3.5

combined standard uncertainty

standard uncertainty of the result of a measurement when that result is obtained from the values of a number of other quantities, equal to the positive square root of a sum of terms, the terms being the variances or covariances of these other quantities weighted according to how the measurement result varies with changes in these quantities

[SOURCE: ENV 13005:1999 [8]]

3.6

coverage factor

numerical factor used as a multiplier of the combined standard uncertainty in order to obtain an expanded uncertainty

[SOURCE: ENV 13005:1999 [8]]

iTeh STANDARD PREVIEW
(standards.iteh.ai)

[SIST-TS CEN/TS 16450:2013](https://standards.iteh.ai/catalog/standards/sist/97a03c44-4564-4220-a992-68b51d02c91a/sist-ts-cen-ts-16450-2013)

<https://standards.iteh.ai/catalog/standards/sist/97a03c44-4564-4220-a992-68b51d02c91a/sist-ts-cen-ts-16450-2013>

3.7

data capture

percentage of the time for which the AMS has produced valid data to the time for which the aggregated value is to be calculated, excluding periods of regular calibration or normal maintenance

[SOURCE: Directive 2008/50/EC [1]]

3.8

designated body

body which has been designated for a specific task (type approval tests and/or QA/QC activities in the field) by the competent authority in the Member States

Note 1 to entry: It is recommended that the designated body is accredited for the specific task according to EN ISO/IEC 17025 [9].

3.9

detection limit

smallest concentration of a measurand that can be reliably detected by a specific measurement process

3.10

equivalent method

method other than the reference method for the measurement of a specified air pollutant meeting the data quality objectives for fixed measurements specified in the relevant Air Quality Directive [1]

Note 1 to entry: Equivalence is granted for defined (regional) situations within a Member State, but may be granted for situations encompassing more than one region or Member State.

3.11**expanded uncertainty**

quantity defining an interval about the result of a measurement that may be expected to encompass a large fraction of the distribution of values that could reasonably be attributed to the measurand

Note 1 to entry: The fraction may be viewed as the coverage probability or level of confidence of the interval.

Note 2 to entry: To associate a specific level of confidence with the interval defined by the expanded uncertainty requires explicit or implicit assumptions regarding the probability distribution characterised by the measurement result and its combined standard uncertainty. The level of confidence that may be attributed to this interval can be known only to the extent to which such assumptions may be justified.

[SOURCE: ENV 13005:1999 [8]]

3.12**influence quantity**

quantity that is not the measurand but that affects the result of the measurement

[SOURCE: ENV 13005:1999 [8]]

3.13**interferent**

component of the air sample, excluding the measured constituent, that affects the output signal

3.14**limit value**

level fixed on the basis of scientific knowledge, with the aim of avoiding, preventing or reducing harmful effects on human health and/or the environment as a whole, to be attained within a given period and not to be exceeded once attained

[SOURCE: Directive 2008/50/EC [1]] [SIST-TS CEN/TS 16450:2013](https://standards.iteh.ai/catalog/standards/sist/97a03c44-4564-4220-a992-68b51d02c91a/sist-ts-cen-ts-16450-2013)

<https://standards.iteh.ai/catalog/standards/sist/97a03c44-4564-4220-a992-68b51d02c91a/sist-ts-cen-ts-16450-2013>

3.15**monitoring station**

enclosure located in the field in which an AMS has been installed to measure particulate matter in such a way that its performance and operation comply with the prescribed requirements

3.16**parallel measurement**

measurement from measuring systems, sampling from the same air over the same time period

3.17**performance characteristic**

one of the parameters assigned to an AMS in order to define its performance

3.18**performance criterion**

limiting quantitative numerical value assigned to a performance characteristic, to which conformance is tested

3.19**period of unattended operation**

time period over which the drift is within the performance criterion for long term drift

CEN/TS 16450:2013 (E)

3.20**PM_x**

particulate matter suspended in air which passes through a size-selective inlet with a 50 % efficiency cut-off at x µm aerodynamic diameter

Note 1 to entry: By convention, the size-selective standard inlet designs prescribed in EN 12341 – used at the prescribed flow rates – possess the required characteristics to sample the relevant PM fraction suspended in ambient air.

Note 2 to entry: The efficiency of the size selectiveness of other inlets used may have a significant effect on the fraction of PM surrounding the cut-off, and, consequently on the mass concentration of PM_x determined.

3.21**reference method**

RM

measurement method(ology) which, by convention, gives the accepted reference value of the measurand

3.22**sampled air**

ambient air that has been sampled through the sampling inlet and sampling system

3.23**sampling inlet**

entrance to the sampling system where ambient air is collected from the atmosphere

3.24**standard uncertainty**

uncertainty of the result of a measurement expressed as a standard deviation

[SOURCE: ENV 13005:1999 [8]]

3.25**surrounding temperature**

temperature of the air directly surrounding the AMS (temperature inside the monitoring station or laboratory)

3.26**time coverage**

percentage of the reference period of the relevant limit value for which valid data for aggregation have been collected

3.27**type approval**

decision taken by a designated body that the pattern of an AMS conforms to the requirements as laid down in this document

3.28**type approval test**

examination of two or more AMS of the same pattern which are submitted by a manufacturer to a designated body including the tests necessary for approval of the pattern

3.29**uncertainty (of measurement)**

parameter associated with the result of a measurement that characterises the dispersion of the values that could reasonably be attributed to the measurand

[SOURCE: ENV 13005:1999 [8]]

3.30**zero air**

air containing particulate matter at a level $\leq 1,0 \mu\text{g}/\text{m}^3$

4 Symbols and abbreviated terms

For the purposes of this document, the following symbols and abbreviated terms apply:

a, c	intercept of orthogonal regression of results of AMS vs. reference results
A	availability of the AMS
b, d	slope of orthogonal regression of results of AMS vs. reference results
k	coverage factor
ΔP	pressure difference determined for the time interval Δt (leak test)
P_0	pressure at $t=0$ (leak test)
T_s	is the surrounding air temperature
$T_{s,lab}$	is the surrounding air temperature at the laboratory
Δt	time interval needed for the pressure rise (leak test)
t_{valid}	time during which valid data have been collected (field test)
$t_{cal,maint}$	time spent for scheduled calibrations and maintenance (field test)
t_{field}	total duration of the field test
V_L	volume leak rate (leak test)
V_{sys}	estimated total volume of the system (dead volume)
u	standard uncertainty
u_a	uncertainty of the intercept of the regression formula
u_b	uncertainty of the slope of the regression formula
u_{bs}	between-AMS uncertainty
u_{RM}	random uncertainty of reference method results
w	relative uncertainty
W	expanded relative uncertainty
x_i	individual measurement result of AMS
y_i	individual reference measurement result
AMS	Automated Measuring System
EU	European Union
GDE	Guide to the Demonstration of Equivalence of Ambient Air Monitoring Methods
GUM	Guide to the Expression of Uncertainty in Measurement
JCGM	Joint Committee for Guides in Metrology
LOD	Limit of Detection
PM	Particulate Matter
QA/QC	Quality Assurance / Quality Control
RM	Reference Method(ology)
RSS	Residual Sum of Squares

CEN/TS 16450:2013 (E)

5 Principle

5.1 General

A number of measuring principles may be used to measure the mass concentration of particulate matter in ambient air. This Technical Specification is not limited to the application of a single system for the automated continuous measurement. In general (but not necessarily), the measuring system will consist of:

- a size-selective inlet for PM₁₀ or PM_{2.5} (when using an optical system for size classification of particulate matter a size-selective inlet is not required);
- a sample tube of a length needed to meet the specific sampling height requirements given in Reference [1];
- a measuring section;
- a vacuum pump;
- flow meters;
- temperature and pressure sensors;
- hardware and software for data collection, storage and calculation of measurement results.

Auxiliary equipment may include:

- sample tube heaters;
- systems for (partial) drying of the sampled air;
- humidity sensors;
- hardware/software for performing compensation measurements, i.e., measurements to compensate for unwanted effects of interferences or random variations in the PM mass determination.

5.2 Measuring principles

A number of measuring principles is currently applied in routine monitoring practice. Descriptions of the most common principles – which do not preclude other principles – are given in Annex A.

5.3 Type approval

The type approval of an AMS according to Clause 7 and subsequent QA/QC and verification procedures according to Clause 8 provide evidence that the defined requirements concerning data quality laid out in relevant EU Directives can be satisfied. The AMS manufacturer should fulfil the requirements laid down in EN 15267-1 and EN 15267-2. A designated body should perform the type approval tests. The type approval should be awarded by or on behalf of the national competent authority of a Member State.

The type approval test is based on the evaluation of performance characteristics determined under a prescribed series of tests. In this Technical Specification test procedures are described for the determination of the actual values of the performance characteristics for at least two AMS in a laboratory and the same two AMS in the field. The evaluation for type approval of an AMS includes the calculation of the expanded uncertainty of the measuring result based on the numerical values of the tested performance characteristics and comparison with a prescribed maximum uncertainty.

Iteh STANDARD PREVIEW
(standards.iteh.ai)

SIST-TS CEN/TS 16450:2013

<https://standards.iteh.ai/catalog/standards/sist/97a03c44-4564-4220-a992-68b51d02c91a/sist-ts-cen-ts-16450-2013>

Appropriate experimental evidence should be provided by

- type approval tests performed under conditions of intended use of the specified method of measurement, and
- calculation of expanded uncertainty of results of measurement.

5.4 Suitability testing

Before putting a type-approved AMS in to operation, the body responsible for the field operation should test its suitability for the specific field conditions by performing suitability tests at a minimum of two locations representative for these conditions.

5.5 Field operation and quality control

After the initial installation of the approved AMS at the monitoring station its correct functioning should be tested.

Requirements for quality assurance and quality control are given for the operation and maintenance of the AMS, to ensure that the uncertainty of subsequent measurement results obtained in the field is not compromised.

5.6 Data handling and validation

Practical experience with existing methods for data handling and validation – for similar AMS – has shown that the different ways of data treatment and validation applied may lead to significant differences in reported results for similar datasets [3].

Hence, requirements and recommendations are given for the treatment and validation of raw measurement data collected by the AMS.

[SIST-TS CEN/TS 16450:2013](https://standards.iteh.ai/catalog/standards/sist/97a03c44-4564-4220-a992-68b51d02c91a/sist-ts-cen-ts-16450-2013)

<https://standards.iteh.ai/catalog/standards/sist/97a03c44-4564-4220-a992-68b51d02c91a/sist-ts-cen-ts-16450-2013>

6 Sampling

6.1 General

Conditions and layout of the sampling equipment will contribute to the uncertainty of the measurement; to minimise this contribution to the measurement uncertainty, performance criteria for the sampling equipment are given in the following subclauses.

NOTE In Annex A examples of equipment are schematically presented.

6.2 Sampling location

The location where the ambient air should be sampled and analysed is not specified as this depends strongly on the category of a monitoring station (such as measurements in e.g. a traffic or urban background area).

NOTE For guidance on sampling points on a micro scale, cf. Reference [1].

6.3 Sample inlet and sampling line

Each AMS is equipped with its own sample inlet and sampling line. Sampling inlets may be – but not necessarily are – size-selective inlets for PM₁₀ or PM_{2,5}.

NOTE Examples of designs of size-selective inlets for PM₁₀ or PM_{2,5} can be found in EN 12341.

The sample inlet and sampling line should be made of an inert, non-corroding, electrically conducting material, preferably stainless steel, or anodised aluminium or aluminium alloy. The inlet should be constructed in such a way that ingress of rainwater into the sampling line (or system) is prevented.