



Designation: D 5808 – 03

# Standard Test Method for Determining Organic Chloride in Aromatic Hydrocarbons and Related Chemicals by Microcoulometry<sup>1</sup>

This standard is issued under the fixed designation D 5808; the number immediately following the designation indicates the year of original adoption or, in the case of revision, the year of last revision. A number in parentheses indicates the year of last reapproval. A superscript epsilon ( $\epsilon$ ) indicates an editorial change since the last revision or reapproval.

## 1. Scope

1.1 This test method covers the organic chlorides in aromatic hydrocarbons, their derivatives, and related chemicals.

1.2 This test method is applicable to samples with chloride concentrations from 1 to 25 mg/kg.

1.3 This test method is preferred over Test Method D 5194 for products, such as styrene, that are polymerized by the sodium biphenyl reagent.

1.4 The following applies to all specified limits in this standard: for purposes of determining conformance with this standard, an observed value or a calculated value shall be rounded off “to the nearest unit” in the last right-hand digit used in expressing the specification limit, in accordance with the rounding-off method of Practice E 29.

1.5 *This standard does not purport to address all of the safety concerns, if any, associated with its use. It is the responsibility of the user of this standard to establish appropriate safety and health practices and determine the applicability of regulatory limitations prior to use.* For specific hazard statements, see 7.3 and Section 9.

## 2. Referenced Documents

- 2.1 *ASTM Standards:*  
D 1193 Specification for Reagent Water<sup>2</sup>  
D 3437 Practice for Sampling and Handling Liquid Cyclic Products<sup>3</sup>  
D 5194 Test Method for Trace Chloride in Liquid Aromatic Hydrocarbons<sup>3</sup>  
E 29 Practice for Using Significant Digits in Test Data to Determine Conformance with Specifications<sup>4</sup>  
E 691 Practice for Conducting an Interlaboratory Study to

<sup>1</sup> This test method is under the jurisdiction of ASTM Committee D16 on Aromatic Hydrocarbons and Related Chemicals is the direct responsibility of Subcommittee D16.04 on Instrumental Analysis.

Current edition approved Aug. 10, 2003. Published August 2003. Originally approved in 1995. Last previous edition approved in 1995 as D 5808 - 95.

<sup>2</sup> *Annual Book of ASTM Standards*, Vol 11.01.

<sup>3</sup> *Annual Book of ASTM Standards*, Vol 06.04.

<sup>4</sup> *Annual Book of ASTM Standards*, Vol 14.02.

Determine the Precision of a Test Method<sup>4</sup>

## 2.2 Other Document:

OSHA Regulations—29CFR paragraphs 1910.1000 and 1910.1200<sup>5</sup>

## 3. Terminology

### 3.1 Definitions:

3.1.1 *dehydration tube*—chamber containing concentrated sulfuric acid that scrubs the effluent gases from combustion to remove water vapor.

3.1.2 *oxidative pyrolysis*—a process in which a sample is combusted in an oxygen-rich atmosphere at high temperature to break down the components of the sample into elemental oxides.

3.1.3 *recovery factor*—an indication of the efficiency of the measurement computed by dividing the measured value of a standard by its theoretical value.

3.1.4 *reference sensor pair*—detects changes in silver ion concentration.

3.1.5 *test titration*—a process that allows the coulometer to set the endpoint and gain values to be used for sample analysis.

3.1.6 *titration parameters*—various instrumental conditions that can be changed for different types of analysis.

3.1.7 *working electrode (generator electrode)*—an electrode consisting of an anode and a cathode separated by a salt bridge; maintains a constant silver ion concentration.

## 4. Summary of Test Method

4.1 A liquid specimen is injected into a combustion tube maintained at 900°C having a flowing stream of oxygen and argon carrier gas. Oxidative pyrolysis converts the organic halides to hydrogen halides that then flow into a titration cell where it reacts with silver ions present in the electrolyte. The silver ion thus consumed is coulometrically replaced and the total electrical work to replace it is a measure of the organic halides in the specimen injected (see Annex A1).

<sup>5</sup> Available from U.S. Government Printing Office Superintendent of Documents, 732 N. Capitol St., NW, Mail Stop: SDE, Washington, DC 20401.

## 5. Significance and Use

5.1 Organic as well as inorganic chlorine compounds can prove harmful to equipment and reactions in processes involving hydrocarbons.

5.2 Maximum chloride levels are often specified for process streams and for hydrocarbon products.

5.3 Organic chloride species are potentially damaging to refinery processes. Hydrochloric acid can be produced in hydrotreating or reforming reactors and this acid accumulates in condensing regions of the refinery.

## 6. Interferences

6.1 Both nitrogen and sulfur interfere at concentrations greater than approximately 0.1 %.

NOTE 1—To ensure reliable detectability, all sources of chloride contamination must be eliminated.

6.2 Bromides and iodides, if present, will be calculated as chlorides. However, fluorides are not detected by this test method.

6.3 Organic chloride values of samples containing inorganic chlorides will be biased high due to partial recovery of inorganic species during combustion. Interference from inorganic species can be reduced by water washing the sample before analysis. This does not apply to water soluble samples.

## 7. Apparatus

7.1 *Pyrolysis Furnace*, which can maintain a temperature sufficient to pyrolyze the organic matrix and convert all chlorine present in the sample to hydrogen chloride.

7.2 *Pyrolysis Tube*, made of quartz and constructed so that when a sample is volatilized in the front of the furnace, it is swept into the pyrolysis zone by an inert gas, where it combusts when in the presence of oxygen. The inlet end of the tube must have a sample inlet port with a septum through which the sample can be injected by syringe. The inlet end must also have side arms for the introduction of oxygen and inert carrier gas. The pyrolysis tube must be of ample volume, so that complete pyrolysis of the sample is ensured.

7.3 *Titration Cell*, containing a reference and sensor pair of electrodes and a generator anode/cathode pair of electrodes to maintain constant chloride ion concentration. An inlet from the pyrolysis tube and magnetic stirring is also required. (**Warning**—Excessive stirring speed will decouple the stirring bar and cause it to rise in the titration cell and possibly damage the electrodes. A slight vortex in the cell will be adequate.)

7.4 *Microcoulometer*, capable of measuring the potential of the sensing-reference electrode pair, and comparing this potential with a bias potential, and amplifying the difference to the working electrode pair to generate a current. The microcoulometer output voltage signal should be proportional to the generating current.

7.5 *Automatic Boat Drive*, having variable stops, such that the sample boat may be driven into the furnace, and stopped at various points as it enters the furnace.

7.6 *Controller*, with connections for the reference, working, and sensor electrodes. The controller is used for setting of operating parameters and integration of data.

7.7 *Dehydration Tube*, positioned at the end of the pyrolysis tube so that effluent gases are bubbled through a sulfuric acid solution, and water vapor is subsequently trapped, while all other gases are allowed to flow into the titration cell.

7.8 *Gas-Tight Sampling Syringe*, having a 50  $\mu\text{l}$  capacity, capable of accurately delivering 10 to 40  $\mu\text{l}$  of sample.

7.9 *Quartz Boats*.

## 8. Reagents and Materials

8.1 *Purity of Reagents*—Reagent grade chemicals shall be used in all tests. Unless otherwise indicated, it is intended that all reagents shall conform to the specifications of the Committee on Analytical Reagents of the American Chemical Society, where such specifications are available.<sup>6</sup> Other grades may be used, provided that the reagent is of sufficiently high purity to permit its use without lessening the accuracy of the determination.

8.2 *Purity of Water*—Unless otherwise indicated, references to water shall be understood to mean reagent water conforming to Specification D 1193, Type II or III.

8.3 *Acetic Acid*—Glacial acetic acid ( $\text{CH}_3\text{COOH}$ ).

8.4 *Argon or Helium*, 99.9 % minimum purity required as carrier gas.

8.5 *Amidosulfonic Acid* ( $\text{H}_2\text{NSO}_3\text{H}$ ), minimum purity 99.3–100.3 %.

8.6 *Sodium Acetate*, anhydrous, ( $\text{NaCH}_3\text{CO}_2$ ), fine granular.

8.7 *Cell Electrolyte Solution*—Dissolve 1.35 g sodium acetate ( $\text{NaCH}_3\text{CO}_2$ ) in 850 mL of acetic acid ( $\text{CH}_3\text{COOH}$ ), and dilute to 1000 mL with water or follow manufacturer's recommendations.

NOTE 2—Bulk quantities of the electrolyte should be stored in a dark bottle or in a dark place and be prepared fresh at least every two weeks.

8.8 *Oxygen*, 99.6 % minimum purity is required as the reactant gas.

8.9 *Gas Regulators*, two-stage gas regulators must be used for the reactant and carrier gas.

8.10 *Potassium Nitrate* ( $\text{KNO}_3$ ), fine granular.

8.11 *Potassium Chloride* ( $\text{KCl}$ ), fine granular.

8.12 *Potassium Sulfate* ( $\text{K}_2\text{SO}_4$ ), crystalline.

8.13 *Working Electrode Solution* (10 %  $\text{KNO}_3$ )—Dissolve 50 g potassium nitrate ( $\text{KNO}_3$ ) in 500 mL of distilled water.

8.14 *Inner Chamber Reference Electrode Solution* (1 M  $\text{KCl}$ )—Dissolve 7.46 g potassium chloride ( $\text{KCl}$ ) in 100 mL of distilled water.

8.15 *Outer Chamber Reference Electrode Solution* (1 M  $\text{KNO}_3$ )—Dissolve 10.1 g potassium nitrate ( $\text{KNO}_3$ ) in 100 mL of distilled water.

8.16 *Sodium Chloride* ( $\text{NaCl}$ ), fine granular.

8.17 *Sodium Perchlorate* ( $\text{NaClO}_4$ ), crystalline.

8.18 *Sulfuric Acid*, (sp gr 1.84), ( $\text{H}_2\text{SO}_4$ ) concentrated.

8.19 *2,4,6-Trichlorophenol* ( $\text{TCP}$ ) ( $\text{C}_6\text{H}_3\text{OCl}_3$ ), fine granular.

<sup>6</sup> *Reagent Chemicals, American Chemical Society Specifications*, American Chemical Society, Washington, DC. For suggestions on the testing of reagents not listed by the American Chemical Society, see *Analar Standards for Laboratory Chemicals*, BDH Ltd., Poole, Dorset, U.K., and the *United States Pharmacopeia and National Formulary*, U.S. Pharmaceutical Convention, Inc. (USPC), Rockville, MD.