

# SLOVENSKI STANDARD oSIST prEN 10351:2009

01-september-2009

Kemična analiza železovih zlitin - Analiza z direktno spektrometrijo optične emisije z induktivno sklopljeno plazmo malolegiranih jekel - Določevanje Mn, P, Cu, Ni, Cr, Mo, V, Co, Al (skupni) in Sn [rutinska metoda]

Chemical analysis of ferrous materials - Inductively coupled plasma optical emission spectrometric analysis of low alloyed steels - Determination of Mn, P, Cu, Ni, Cr, Mo, V, Co, Al (total) and Sn [Routine method]

Analyse chimique des matériaux sidérurgiques - Analyse des aciers faiblement alliés par spectrométrie d'émission optique avec source à plasma induit - Détermination de Mn, P, Cu, Ni, Cr, Mo, V, Co, Al (total) et Sn [Méthode de routine]

Ta slovenski standard je istoveten z: prEN 10351

ICS:

77.040.30 Kemijska analiza kovin Chemical analysis of metals

oSIST prEN 10351:2009 en,fr

**oSIST** prEN 10351:2009

oSIST prEN 10351:2009

# EUROPEAN STANDARD NORME EUROPÉENNE EUROPÄISCHE NORM

# DRAFT prEN 10351

June 2009

ICS 77.140.10; 77.140.45

#### **English Version**

Chemical analysis of ferrous materials - Inductively coupled plasma optical emission spectrometric analysis of low alloyed steels - Determination of Mn, P, Cu, Ni, Cr, Mo, V, Co, Al (total) and Sn [Routine method]

Analyse chimique des matériaux sidérurgiques - Analyse des aciers faiblement alliés par spectrométrie d'émission optique avec source à plasma induit - Détermination de Mn, P, Cu, Ni, Cr, Mo, V, Co, Al (total) et Sn [Méthode de routine]

This draft European Standard is submitted to CEN members for enquiry. It has been drawn up by the Technical Committee ECISS/TC 20.

If this draft becomes a European Standard, CEN members are bound to comply with the CEN/CENELEC Internal Regulations which stipulate the conditions for giving this European Standard the status of a national standard without any alteration.

This draft European Standard was established by CEN in three official versions (English, French, German). A version in any other language made by translation under the responsibility of a CEN member into its own language and notified to the CEN Management Centre has the same status as the official versions.

CEN members are the national standards bodies of Austria, Belgium, Bulgaria, Cyprus, Czech Republic, Denmark, Estonia, Finland, France, Germany, Greece, Hungary, Iceland, Ireland, Italy, Latvia, Lithuania, Luxembourg, Malta, Netherlands, Norway, Poland, Portugal, Romania, Slovakia, Slovenia, Spain, Sweden, Switzerland and United Kingdom.

Recipients of this draft are invited to submit, with their comments, notification of any relevant patent rights of which they are aware and to provide supporting documentation.

**Warning**: This document is not a European Standard. It is distributed for review and comments. It is subject to change without notice and shall not be referred to as a European Standard.



EUROPEAN COMMITTEE FOR STANDARDIZATION COMITÉ EUROPÉEN DE NORMALISATION EUROPÄISCHES KOMITEE FÜR NORMUNG

Management Centre: Avenue Marnix 17, B-1000 Brussels

Contents		Page	
Forev	Foreword4		
1	Scope	5	
2	Normative references		
3	Principle		
4	Reagents		
4.1	Hydrochloric acid, HCI ( $\rho_{20}$ = 1,19 g/ml)		
4.2	Hydrochloric acid, solution 1 + 1		
4.3 4.4	Nitric acid, HNO <sub>3</sub> (ρ <sub>20</sub> = 1,33 g/ml)		
4.4 4.5	Nitric acid, solution 1 + 1		
4.5 4.6	Hydrofluoric acid, HF ( $\rho_{20}$ = 1,13 g/ml)		
4.6 4.7	Sulphuric acid, solution 1 + 1		
4.7 4.8	Fusion mixture		
4.9	Aluminium, 1 g/l standard solution		
4.10	Chromium 1 g/l standard solution		
4.11	Cobalt 1 g/l standard solution		
4.12	Copper 1 g/l standard solution		
4.13	Manganese 1 g/l standard solution		
4.14	Molybdenum 1 g/l standard solution		
4.15	Nickel 1 g/l standard solution		
4.16 4.17	Phosphorus 1 g/l standard solution Tin 1 g/l standard solution [freshly prepared]		
4.17 4.18	Vanadium 1 g/l standard solution		
4.19	Internal standard solution, 1 g/l		
4.20	Pure iron		
_	Apparatus		
5 5.1	Close texture filter paper and/or 0,45 µm cellulose nitrate filter	وع ه	
5.2	Filter funnel suitable for 0,45 µm filters		
5.3	Vacuum vessel		
5.4	Platinum crucibles and platinum covers		
5.5	Atomic emission spectrometer, equipped with inductively coupled plasma (ICP)		
6	Sampling	۵	
_	. •		
7	Sample solution preparation		
7.1 7.2	Test portion		
1.2	Preparation of the test solution		
8	Calibration process		
8.1	Preparation of the multi-elemental calibration solutions		
8.2	Standardization or re-calibration	12	
9	Determination	13	
9.1	Adjustment of the apparatus		
9.2	Spectrometric measurements of the calibration solutions		
9.3	Analysis	14	
10	Expression of results	15	
11	Precision		
12	Tost raport	10	

# oSIST prEN 10351:2009

# prEN 10351:2009 (E)

Annex A (informative) Optical emission spectrometer (OES) — Suggested performance criteria to be checked	20
Annex B (normative) Synoptic of the operations related to § 9	23
Annex C (informative) Test samples used for the precision test	24
Annex D (informative) Detailed results obtained from the precision test	25
Annex E (informative) Graphical representation of the precision data	34
Bibliography	39

# **Foreword**

This document (prEN 10351:2009) has been prepared by Technical Committee ECISS/TC 20 "Methods of chemical analysis of ferrous products", the secretariat of which is held by SIS.

This document is currently submitted to the CEN Enquiry.

### 1 Scope

This document specifies an inductively coupled plasma emission spectrometry routine method for the analysis of unalloyed and low alloyed steels, whose iron content should be at least 95 %.

This method is applicable to the elements listed in Table 1 within the ranges shown.

Mass fraction % Element min. max. Mn 0.005 2.00 Ρ 0.005 0.05 Cu 0.005 0.80 Ni 0.010 2.00 Cr 0.010 1.60 0.005 0.80 Mo V 0.002 0.40 Co 0.002 0.10 Al (total) 0.020 0.30 Sn 0.001 0.10

Table 1 — Application ranges

In all cases, the ranges specified can be extended or adapted (after validation) for the determination of other mass fractions, provided that the iron content in the samples under concern still is above 95 %.

Other elements may be included. However such elements and their mass fractions should be carefully checked, taking into account the eventual interferences, the sensitivity, the resolution and the linearity criteria for each instrument and each wavelength.

Depending also on the sensitivity of each instrument, suitable dilutions of the calibration and the test sample solutions may be necessary.

Moreover, even if the method described is a "multi elemental" method <u>it is not</u> absolutely necessary to carry out the determination of all the elements of its scope simultaneously: the measurement conditions have to be optimised by each laboratory, depending on the performances of each apparatus available.

- NOTE 1 The accuracy of the method is unsatisfactory for phosphorus contents from 0,05 to 0,1 %.
- NOTE 2 The trueness of the method couldn't be checked for vanadium contents below 0,05 %.
- NOTE 3 The precision of the method is unsatisfactory for aluminium (total) contents below 0,02 %.

#### 2 Normative references

This working draft incorporates by dated or undated reference, provisions from other publications. These normative references are cited at the appropriate places in the text and the publications are listed hereafter. For dated references, subsequent amendments to, or revisions of any of these publications apply to this working draft only when incorporated in it by amendment or revision. For undated references, the latest edition of the publication referred applies (including amendments)

EN ISO 14284, Steel and iron — sampling and preparation of samples for the determination of chemical composition.

ISO 648, Laboratory glassware — One mark pipettes.

ISO 1042, Laboratory glassware — One mark volumetric flasks.

ISO 3696, Water for analytical laboratory use — Specification and test methods.

ISO 5725-1:1994, Accuracy (trueness and precision) of measurement methods and results — Part 1: General principles and definitions.

ISO 5725-2:1994, Accuracy (trueness and precision) of measurement methods and results — Part 2: Basic method for the determination of repeatability and reproducibility of a standard measurement method.

ISO 5725-3:1994, Accuracy (trueness and precision) of measurement methods and results — Part 3: Intermediate measures of the precision of a standard measurement method.

CEN/TR 10345:2008, Guideline for statistical data treatment of inter laboratory tests for validation of analytical methods.

### 3 Principle

Dissolution of a test portion with nitric and hydrochloric acids. Filtration and ignition of the acid insoluble residue. Removal of silica with hydrofluoric acid. Fusion of the residue with a mixture of orthoboric acid and potassium carbonate, dissolution of the melt with acid and addition of this solution to the reserved filtrate.

After suitable dilution and, if necessary, addition of an internal reference element, nebulisation of the solution into an inductively coupled plasma emission spectrometer and measurement of the intensity of the emitted light (including, eventually, that of the internal reference element).

## 4 Reagents

During the analysis, use only reagents of recognised analytical grade and only distilled water or water of equivalent purity.

The same reagents should be used for the preparation of calibration solutions and of sample solutions.

- 4.1 Hydrochloric acid, HCI ( $\rho_{20}$  = 1,19 g/ml)
- 4.2 Hydrochloric acid, solution 1 + 1

Add 500 ml of hydrochloric acid (4.1) to 500 ml of water.

- 4.3 Nitric acid, HNO<sub>3</sub> ( $\rho_{20} = 1,33 \text{ g/ml}$ )
- 4.4 Nitric acid, solution 1 + 1

Add 500 ml of nitric acid (4.3) to 500 ml of water.

4.5 Hydrofluoric acid, HF ( $\rho_{20}$  = 1,13 g/ml)

WARNING — Hydrofluoric acid is extremely irritating and corrosive to skin and mucous membranes producing severe skin burns which are slow to heal. In the case of contact with skin, wash well with water, apply a topical gel containing 2,5 % (mass fraction) calcium gluconate, and seek immediate medical treatment.