

# INTERNATIONAL STANDARD

ISO  
9345-2

First edition  
2003-03-15

---

---

## Optics and optical instruments — Microscopes: Imaging distances related to mechanical reference planes —

### Part 2: Infinity-corrected optical systems

iTeh Standards  
(<https://standards.iteh.ai>)  
Document Preview

[ISO 9345-2:2003](#)

<https://standards.iteh.ai/catalog/standards/iso/b583d68-9f90-4c81-b687-84c38fe67e0b/iso-9345-2-2003>



Reference number  
ISO 9345-2:2003(E)

© ISO 2003

**PDF disclaimer**

This PDF file may contain embedded typefaces. In accordance with Adobe's licensing policy, this file may be printed or viewed but shall not be edited unless the typefaces which are embedded are licensed to and installed on the computer performing the editing. In downloading this file, parties accept therein the responsibility of not infringing Adobe's licensing policy. The ISO Central Secretariat accepts no liability in this area.

Adobe is a trademark of Adobe Systems Incorporated.

Details of the software products used to create this PDF file can be found in the General Info relative to the file; the PDF-creation parameters were optimized for printing. Every care has been taken to ensure that the file is suitable for use by ISO member bodies. In the unlikely event that a problem relating to it is found, please inform the Central Secretariat at the address given below.

**iTeh Standards**  
**(<https://standards.iteh.ai>)**  
**Document Preview**

[ISO 9345-2:2003](#)

<https://standards.iteh.ai/catalog/standards/iso/b583d68-9f90-4c81-b687-84c38fe67e0b/iso-9345-2-2003>

© ISO 2003

All rights reserved. Unless otherwise specified, no part of this publication may be reproduced or utilized in any form or by any means, electronic or mechanical, including photocopying and microfilm, without permission in writing from either ISO at the address below or ISO's member body in the country of the requester.

ISO copyright office  
Case postale 56 • CH-1211 Geneva 20  
Tel. + 41 22 749 01 11  
Fax + 41 22 749 09 47  
E-mail [copyright@iso.org](mailto:copyright@iso.org)  
Web [www.iso.org](http://www.iso.org)

Published in Switzerland

## Foreword

ISO (the International Organization for Standardization) is a worldwide federation of national standards bodies (ISO member bodies). The work of preparing International Standards is normally carried out through ISO technical committees. Each member body interested in a subject for which a technical committee has been established has the right to be represented on that committee. International organizations, governmental and non-governmental, in liaison with ISO, also take part in the work. ISO collaborates closely with the International Electrotechnical Commission (IEC) on all matters of electrotechnical standardization.

International Standards are drafted in accordance with the rules given in the ISO/IEC Directives, Part 2.

The main task of technical committees is to prepare International Standards. Draft International Standards adopted by the technical committees are circulated to the member bodies for voting. Publication as an International Standard requires approval by at least 75 % of the member bodies casting a vote.

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. ISO shall not be held responsible for identifying any or all such patent rights.

ISO 9345-2 was prepared by Technical Committee ISO/TC 172, *Optics and optical instruments*, Subcommittee SC 5, *Microscopes and endoscopes*.

ISO 9345 consists of the following parts, under the general title *Optics and optical instruments — Microscopes: Imaging distances related to mechanical reference planes*:

- *Part 1: Tube length 160 mm*
- *Part 2: Infinity-corrected optical systems*

[ISO 9345-2:2003](#)

<https://standards.iteh.ai/catalog/standards/iso/b583d68-9f90-4c81-b687-84c38fe67e0b/iso-9345-2-2003>



# Optics and optical instruments — Microscopes: Imaging distances related to mechanical reference planes —

## Part 2: Infinity-corrected optical systems

### 1 Scope

This part of ISO 9345 specifies the imaging distances of objectives, eyepieces and the focal length of "normal" tube lenses of microscopes with infinity-corrected optical systems.

NOTE A specific combination of eyepiece, objective and tube lens is frequently used to correct aberrations. Therefore the combination of an objective from one manufacturer and the tube lens or eyepiece from another manufacturer, although conforming to this part of ISO 9345, may cause errors in magnification and/or in optical performance.

## iTeh Standards

### 2 Normative references

The following referenced documents are indispensable for the application of this document. For dated references, only the edition cited applies. For undated references, the latest edition of the referenced document (including any amendments) applies.

ISO 8039:1997, *Optics and optical instruments — Microscopes — Magnification*

<https://standards.iteh.ai/catalog/standards/iso/b583d68-9f90-4c81-b687-84c38fe67e0b/iso-9345-2-2003>

ISO 9345-1:1996, *Optics and optical instruments — Microscopes — Imaging distances related to mechanical reference planes — Part 1: Tube length 160 mm*

### 3 Terms and definitions

For the purposes of this document, the following terms and definitions apply.

#### 3.1

##### parfocalizing distance of the objective

$l_1$

distance between the object plane (i.e. the uncovered surface of the object) and the locating flange of the objective when the microscope is in its (focused) working condition

NOTE See Figure 1, Figure 2 and footnote b in Table 1.

#### 3.2

##### image distance of the objective

$l_2$

distance in air between the objective locating surface and the primary image plane

NOTE An infinity-corrected objective alone produces a primary image at infinity. In combination with an infinity-corrected tube lens, the primary image is produced in the back focal plane of this tube lens (see Figure 1).

## 3.3

## parfocalizing distance of the eyepiece

 $l_3$ 

distance between the locating flange of the eyepiece and the plane upon which the eyepiece is focused

NOTE This plane is coincident with the primary image plane of the microscope when the eyepiece is mounted in the viewing tube (see Figure 1).

## 3.4

## focal length of the “normal” tube lens

 $f_{NTL}$ 

focal length related to the magnification and the focal length of the objectives which are designed to operate with this tube lens

[ISO 8039]

## 4 Requirements

## 4.1 Nominal dimensions and tolerances

The nominal dimensions shall be as given in Table 1 and as illustrated in Figure 1.

Table 1 — Nominal dimensions and tolerances

Feature	Symbol	Nominal values/range mm	Numerical aperture	Tolerance mm
Parfocalizing distance of objective <sup>a, b</sup>	$l_1$	$45 + 15 n$ ( $n = -1, 0, 1, 2, 3, 4$ )	$\leq 0,1$ $> 0,1$ to $\leq 0,25$ $> 0,25$ to $\leq 0,45$ $> 0,45$	$\pm 0,2$ <sup>c</sup> $\pm 0,06$ $\pm 0,03$ $\pm 0,01$
Image distance of objective <sup>d</sup>	$l_2$	$150 \leq f_{NTL} \leq 250$		
Parfocalizing distance of eyepiece	$l_3$	10		$\pm 0,2$
Focal length of “normal” tube lens <sup>e</sup>	$f_{NTL}$	$150 \leq f_{NTL} \leq 250$		

<sup>a</sup> The choice of a parfocalizing distance for an objective depends on the design concept of the microscope as a whole. The parfocalizing distance,  $l_1 = 45$  mm of objectives has become the standard value for microscopes with tube length 160 mm (see ISO 9345-1) and also has been adopted for various existing infinity-corrected microscope systems. Examples of common values in use are given in Annex A.

<sup>b</sup> The parfocalizing distance,  $l_1$ , shown in Figure 1 and Table 1, is intended to apply to objectives when used with uncovered objects (specimens). Objectives for use with objects covered by a cover glass shall have the following parfocalizing distance, to allow for the virtual displacement of the object by the cover glass (see also Figure 2):

$$l_1 + t \frac{n-1}{n} \text{ mm}$$

where

$t$  is the thickness of cover glass;  
 $n$  is the refractive index of the cover glass.

<sup>c</sup> The tolerance  $\pm 0,2$  mm for the parfocalizing distance of objectives with numerical aperture  $\leq 0,1$  does not necessarily apply to objectives with magnifications lower than 4  $\times$ .

<sup>d</sup> In infinity-corrected optical systems the primary image is always produced by the objective in combination with a tube lens. The distance between the locating flange of the objective and the tube lens depends on the design of the microscope. The microscope shall have such a design that, in combination with objectives and tube lenses in accordance with this part of ISO 9345, the primary image is produced 10 mm below the eyepiece-locating surface of the viewing tube.

<sup>e</sup> The choice of focal length for a “normal” tube lens depends on the design concept of the microscope system. Its value shall be in the range of  $150 \text{ mm} \leq f_{NTL} \leq 250 \text{ mm}$ . Examples of common values in use are given in Annex A.