
**General requirements for the competence
of testing and calibration laboratories**

*Prescriptions générales concernant la compétence des laboratoires
d'étalonnages et d'essais*

iTeh STANDARD PREVIEW
(standards.iteh.ai)

[ISO/IEC 17025:1999](https://standards.iteh.ai/catalog/standards/sist/58601841-23af-4024-a75f-f27267850d5c/iso-iec-17025-1999)

<https://standards.iteh.ai/catalog/standards/sist/58601841-23af-4024-a75f-f27267850d5c/iso-iec-17025-1999>

PDF disclaimer

This PDF file may contain embedded typefaces. In accordance with Adobe's licensing policy, this file may be printed or viewed but shall not be edited unless the typefaces which are embedded are licensed to and installed on the computer performing the editing. In downloading this file, parties accept therein the responsibility of not infringing Adobe's licensing policy. The ISO Central Secretariat accepts no liability in this area.

Adobe is a trademark of Adobe Systems Incorporated.

Details of the software products used to create this PDF file can be found in the General Info relative to the file; the PDF-creation parameters were optimized for printing. Every care has been taken to ensure that the file is suitable for use by ISO member bodies. In the unlikely event that a problem relating to it is found, please inform the Central Secretariat at the address given below.

iTeh STANDARD PREVIEW
(standards.iteh.ai)

[ISO/IEC 17025:1999](#)

<https://standards.iteh.ai/catalog/standards/sist/58601841-23af-4024-a75f-f27267850d5c/iso-iec-17025-1999>

© ISO/IEC 1999

All rights reserved. Unless otherwise specified, no part of this publication may be reproduced or utilized in any form or by any means, electronic or mechanical, including photocopying and microfilm, without permission in writing from either ISO at the address below or ISO's member body in the country of the requester.

ISO copyright office
Case postale 56 • CH-1211 Geneva 20
Tel. + 41 22 749 01 11
Fax + 41 22 734 10 79
E-mail copyright@iso.ch
Web www.iso.ch

Printed in Switzerland

Contents

	Page
Foreword.....	iv
Introduction.....	v
1 Scope	1
2 Normative references	1
3 Terms and definitions	2
4 Management requirements	2
4.1 Organization.....	2
4.2 Quality system	3
4.3 Document control	4
4.4 Review of requests, tenders and contracts	5
4.5 Subcontracting of tests and calibrations.....	5
4.6 Purchasing services and supplies.....	6
4.7 Service to the client.....	6
4.8 Complaints	6
4.9 Control of nonconforming testing and/or calibration work	7
4.10 Corrective action.....	7
4.11 Preventive action	8
4.12 Control of records.....	8
4.13 Internal audits	9
4.14 Management reviews.....	9
5 Technical requirements	10
5.1 General.....	10
5.2 Personnel.....	10
5.3 Accommodation and environmental conditions	11
5.4 Test and calibration methods and method validation	12
5.5 Equipment	15
5.6 Measurement traceability.....	16
5.7 Sampling.....	18
5.8 Handling of test and calibration items.....	18
5.9 Assuring the quality of test and calibration results.....	19
5.10 Reporting the results.....	19
Annex A (informative) Nominal cross-references to ISO 9001:1994 and ISO 9002:1994.....	23
Annex B (informative) Guidelines for establishing applications for specific fields.....	24
Bibliography	25

Foreword

ISO (the International Organization for Standardization) and IEC (the International Electrotechnical Commission) form the specialized system for worldwide standardization. National bodies that are members of ISO or IEC participate in the development of International Standards through technical committees established by the respective organization to deal with particular fields of technical activity. ISO and IEC technical committees collaborate in fields of mutual interest. Other international organizations, governmental and non-governmental, in liaison with ISO and IEC, also take part in the work.

International Standards are drafted in accordance with the rules given in the ISO/IEC Directives, Part 3.

Draft International Standards adopted by the technical committees are circulated to member bodies for voting. Publication as an International Standard requires approval by at least 75 % of the member bodies casting a vote.

Attention is drawn to the possibility that some of the elements of this International Standard may be the subject of patent rights. ISO and IEC shall not be held responsible for identifying any or all such patent rights.

International Standard ISO/IEC 17025 was prepared by ISO/CASCO, *Committee on Conformity Assessment*.

This first edition of ISO/IEC 17025 cancels and replaces ISO/IEC Guide 25:1990.

Annexes A and B of this International Standard are for information only.

iTeh STANDARD PREVIEW
(standards.iteh.ai)

[ISO/IEC 17025:1999](https://standards.iteh.ai/catalog/standards/sist/58601841-23af-4024-a75f-f27267850d5c/iso-iec-17025-1999)

<https://standards.iteh.ai/catalog/standards/sist/58601841-23af-4024-a75f-f27267850d5c/iso-iec-17025-1999>

Introduction

This International Standard has been produced as the result of extensive experience in the implementation of ISO/IEC Guide 25 and EN 45001, both of which it now replaces. It contains all of the requirements that testing and calibration laboratories have to meet if they wish to demonstrate that they operate a quality system, are technically competent, and are able to generate technically valid results.

Accreditation bodies that recognize the competence of testing and calibration laboratories should use this International Standard as the basis for their accreditation. Clause 4 specifies the requirements for sound management. Clause 5 specifies the requirements for technical competence for the type of tests and/or calibrations the laboratory undertakes.

The growth in use of quality systems generally has increased the need to ensure that laboratories which form part of larger organizations or offer other services can operate to a quality system that is seen as compliant with ISO 9001 or ISO 9002 as well as with this International Standard. Care has been taken, therefore, to incorporate all those requirements of ISO 9001 and ISO 9002 that are relevant to the scope of testing and calibration services that are covered by the laboratory's quality system.

Testing and calibration laboratories that comply with this International Standard will therefore also operate in accordance with ISO 9001 or ISO 9002.

Certification against ISO 9001 and ISO 9002 does not of itself demonstrate the competence of the laboratory to produce technically valid data and results.

The acceptance of testing and calibration results between countries should be facilitated if laboratories comply with this International Standard and if they obtain accreditation from bodies which have entered into mutual recognition agreements with equivalent bodies in other countries using this International Standard.

The use of this International Standard will facilitate cooperation between laboratories and other bodies, and assist in the exchange of information and experience, and in the harmonization of standards and procedures.

iTeh STANDARD PREVIEW
(standards.iteh.ai)

ISO/IEC 17025:1999

<https://standards.iteh.ai/catalog/standards/sist/58601841-23af-4024-a75f-f27267850d5c/iso-iec-17025-1999>

General requirements for the competence of testing and calibration laboratories

1 Scope

1.1 This International Standard specifies the general requirements for the competence to carry out tests and/or calibrations, including sampling. It covers testing and calibration performed using standard methods, non-standard methods, and laboratory-developed methods.

1.2 This International Standard is applicable to all organizations performing tests and/or calibrations. These include, for example, first-, second- and third-party laboratories, and laboratories where testing and/or calibration forms part of inspection and product certification.

This International Standard is applicable to all laboratories regardless of the number of personnel or the extent of the scope of testing and/or calibration activities. When a laboratory does not undertake one or more of the activities covered by this International Standard, such as sampling and the design/development of new methods, the requirements of those clauses do not apply.

1.3 The notes given provide clarification of the text, examples and guidance. They do not contain requirements and do not form an integral part of this International Standard.

1.4 This International Standard is for use by laboratories in developing their quality, administrative and technical systems that govern their operations. Laboratory clients, regulatory authorities and accreditation bodies may also use it in confirming or recognizing the competence of laboratories.

1.5 Compliance with regulatory and safety requirements on the operation of laboratories is not covered by this International Standard.

1.6 If testing and calibration laboratories comply with the requirements of this International Standard they will operate a quality system for their testing and calibration activities that also meets the requirements of ISO 9001 when they engage in the design/development of new methods, and/or develop test programmes combining standard and non-standard test and calibration methods, and ISO 9002 when they only use standard methods. Annex A provides nominal cross-references between this International Standard and ISO 9001 and ISO 9002. ISO/IEC 17025 covers several technical competence requirements that are not covered by ISO 9001 and ISO 9002.

NOTE 1 It might be necessary to explain or interpret certain requirements in this International Standard to ensure that the requirements are applied in a consistent manner. Guidance for establishing applications for specific fields, especially for accreditation bodies (see ISO/IEC Guide 58:1993, 4.1.3) is given in annex B.

NOTE 2 If a laboratory wishes accreditation for part or all of its testing and calibration activities, it should select an accreditation body that operates in accordance with ISO/IEC Guide 58.

2 Normative references

The following normative documents contain provisions which, through reference in this text, constitute provisions of this International Standard. For dated references, subsequent amendments to, or revisions of, any of these publications do not apply. However, parties to agreements based on this International Standard are encouraged to investigate the possibility of applying the most recent editions of the normative documents indicated below. For

ISO/IEC 17025:1999(E)

undated references, the latest edition of the normative document referred to applies. Members of ISO and IEC maintain registers of currently valid International Standards.

ISO 9001:1994, *Quality systems — Model for quality assurance in design, development, production, installation and servicing*.

ISO 9002:1994, *Quality systems — Model for quality assurance in production, installation and servicing*.

ISO/IEC Guide 2, *General terms and their definitions concerning standardization and related activities*.

VIM, *International vocabulary of basic and general terms in metrology*, issued by BIPM, IEC, IFCC, ISO, IUPAC, IUPAP and OIML.

NOTE 1 Further related standards, guides, etc. on subjects included in this International Standard are given in the bibliography.

NOTE 2 It should be noted that when this International Standard was being developed, the revisions of ISO 9001 and ISO 9002 were anticipated to be published in late 2000 as a merged ISO 9001:2000. This is no longer the case.

3 Terms and definitions

For the purposes of this International Standard, the relevant terms and definitions given in ISO/IEC Guide 2 and VIM apply.

NOTE General definitions related to quality are given in ISO 8402, whereas ISO/IEC Guide 2 gives definitions specifically related to standardization, certification and laboratory accreditation. Where different definitions are given in ISO 8402, the definitions in ISO/IEC Guide 2 and VIM are preferred.

4 Management requirements

[ISO/IEC 17025:1999](https://standards.iteh.ai/catalog/standards/sist/58601841-23af-4024-a75f-f27267850d5c/iso-iec-17025-1999)

<https://standards.iteh.ai/catalog/standards/sist/58601841-23af-4024-a75f-f27267850d5c/iso-iec-17025-1999>

4.1 Organization

4.1.1 The laboratory or the organization of which it is part shall be an entity that can be held legally responsible.

4.1.2 It is the responsibility of the laboratory to carry out its testing and calibration activities in such a way as to meet the requirements of this International Standard and to satisfy the needs of the client, the regulatory authorities or organizations providing recognition.

4.1.3 The laboratory management system shall cover work carried out in the laboratory's permanent facilities, at sites away from its permanent facilities, or in associated temporary or mobile facilities.

4.1.4 If the laboratory is part of an organization performing activities other than testing and/or calibration, the responsibilities of key personnel in the organization that have an involvement or influence on the testing and/or calibration activities of the laboratory shall be defined in order to identify potential conflicts of interest.

NOTE 1 Where a laboratory is part of a larger organization, the organizational arrangements should be such that departments having conflicting interests, such as production, commercial marketing or financing do not adversely influence the laboratory's compliance with the requirements of this International Standard.

NOTE 2 If the laboratory wishes to be recognized as a third-party laboratory, it should be able to demonstrate that it is impartial and that it and its personnel are free from any undue commercial, financial and other pressures which might influence their technical judgement. The third-party testing or calibration laboratory should not engage in any activities that may endanger the trust in its independence of judgement and integrity in relation to its testing or calibration activities.

4.1.5 The laboratory shall

- a) have managerial and technical personnel with the authority and resources needed to carry out their duties and to identify the occurrence of departures from the quality system or from the procedures for performing tests and/or calibrations, and to initiate actions to prevent or minimize such departures (see also 5.2);
- b) have arrangements to ensure that its management and personnel are free from any undue internal and external commercial, financial and other pressures and influences that may adversely affect the quality of their work;
- c) have policies and procedures to ensure the protection of its clients' confidential information and proprietary rights, including procedures for protecting the electronic storage and transmission of results;
- d) have policies and procedures to avoid involvement in any activities that would diminish confidence in its competence, impartiality, judgement or operational integrity;
- e) define the organization and management structure of the laboratory, its place in any parent organization, and the relationships between quality management, technical operations and support services;
- f) specify the responsibility, authority and interrelationships of all personnel who manage, perform or verify work affecting the quality of the tests and/or calibrations;
- g) provide adequate supervision of testing and calibration staff, including trainees, by persons familiar with methods and procedures, purpose of each test and/or calibration, and with the assessment of the test or calibration results;
- h) have technical management which has overall responsibility for the technical operations and the provision of the resources needed to ensure the required quality of laboratory operations;
- i) appoint a member of staff as quality manager (however named) who, irrespective of other duties and responsibilities, shall have defined responsibility and authority for ensuring that the quality system is implemented and followed at all times; the quality manager shall have direct access to the highest level of management at which decisions are made on laboratory policy or resources;
- j) appoint deputies for key managerial personnel (see note).

NOTE Individuals may have more than one function and it may be impractical to appoint deputies for every function.

4.2 Quality system

4.2.1 The laboratory shall establish, implement and maintain a quality system appropriate to the scope of its activities. The laboratory shall document its policies, systems, programmes, procedures and instructions to the extent necessary to assure the quality of the test and/or calibration results. The system's documentation shall be communicated to, understood by, available to, and implemented by the appropriate personnel.

4.2.2 The laboratory's quality system policies and objectives shall be defined in a quality manual (however named). The overall objectives shall be documented in a quality policy statement. The quality policy statement shall be issued under the authority of the chief executive. It shall include at least the following:

- a) the laboratory management's commitment to good professional practice and to the quality of its testing and calibration in servicing its clients;
- b) the management's statement of the laboratory's standard of service;
- c) the objectives of the quality system;
- d) a requirement that all personnel concerned with testing and calibration activities within the laboratory familiarize themselves with the quality documentation and implement the policies and procedures in their work; and

e) the laboratory management's commitment to compliance with this International Standard.

NOTE The quality policy statement should be concise and may include the requirement that tests and/or calibrations shall always be carried out in accordance with stated methods and clients' requirements. When the test and/or calibration laboratory is part of a larger organization, some quality policy elements may be in other documents.

4.2.3 The quality manual shall include or make reference to the supporting procedures including technical procedures. It shall outline the structure of the documentation used in the quality system.

4.2.4 The roles and responsibilities of technical management and the quality manager, including their responsibility for ensuring compliance with this International Standard, shall be defined in the quality manual.

4.3 Document control

4.3.1 General

The laboratory shall establish and maintain procedures to control all documents that form part of its quality system (internally generated or from external sources), such as regulations, standards, other normative documents, test and/or calibration methods, as well as drawings, software, specifications, instructions and manuals.

NOTE 1 In this context "document" could be policy statements, procedures, specifications, calibration tables, charts, text books, posters, notices, memoranda, software, drawings, plans, etc. These may be on various media, whether hard copy or electronic, and they may be digital, analog, photographic or written.

NOTE 2 The control of data related to testing and calibration is covered in 5.4.7. The control of records is covered in 4.12.

4.3.2 Document approval and issue

(standards.iteh.ai)

4.3.2.1 All documents issued to personnel in the laboratory as part of the quality system shall be reviewed and approved for use by authorized personnel prior to issue. A master list or an equivalent document control procedure identifying the current revision status and distribution of documents in the quality system shall be established and be readily available to preclude the use of invalid and/or obsolete documents.

4.3.2.2 The procedure(s) adopted shall ensure that:

- a) authorized editions of appropriate documents are available at all locations where operations essential to the effective functioning of the laboratory are performed;
- b) documents are periodically reviewed and, where necessary, revised to ensure continuing suitability and compliance with applicable requirements;
- c) invalid or obsolete documents are promptly removed from all points of issue or use, or otherwise assured against unintended use;
- d) obsolete documents retained for either legal or knowledge preservation purposes are suitably marked.

4.3.2.3 Quality system documents generated by the laboratory shall be uniquely identified. Such identification shall include the date of issue and/or revision identification, page numbering, the total number of pages or a mark to signify the end of the document, and the issuing authority(ies).

4.3.3 Document changes

4.3.3.1 Changes to documents shall be reviewed and approved by the same function that performed the original review unless specifically designated otherwise. The designated personnel shall have access to pertinent background information upon which to base their review and approval.

4.3.3.2 Where practicable, the altered or new text shall be identified in the document or the appropriate attachments.

4.3.3.3 If the laboratory's documentation control system allows for the amendment of documents by hand pending the re-issue of the documents, the procedures and authorities for such amendments shall be defined. Amendments shall be clearly marked, initialled and dated. A revised document shall be formally re-issued as soon as practicable.

4.3.3.4 Procedures shall be established to describe how changes in documents maintained in computerized systems are made and controlled.

4.4 Review of requests, tenders and contracts

4.4.1 The laboratory shall establish and maintain procedures for the review of requests, tenders and contracts. The policies and procedures for these reviews leading to a contract for testing and/or calibration shall ensure that:

- a) the requirements, including the methods to be used, are adequately defined, documented and understood (see 5.4.2);
- b) the laboratory has the capability and resources to meet the requirements;
- c) the appropriate test and/or calibration method is selected and capable of meeting the clients' requirements (see 5.4.2).

Any differences between the request or tender and the contract shall be resolved before any work commences. Each contract shall be acceptable both to the laboratory and the client.

NOTE 1 The request, tender and contract review should be conducted in a practical and efficient manner, and the effect of financial, legal and time schedule aspects should be taken into account. For internal clients, reviews of requests, tenders and contracts can be performed in a simplified way.

NOTE 2 The review of capability should establish that the laboratory possesses the necessary physical, personnel and information resources, and that the laboratory's personnel have the skills and expertise necessary for the performance of the tests and/or calibrations in question. The review may also encompass results of earlier participation in interlaboratory comparisons or proficiency testing and/or the running of trial test or calibration programmes using samples or items of known value in order to determine uncertainties of measurement, limits of detection, confidence limits, etc.

NOTE 3 A contract may be any written or oral agreement to provide a client with testing and/or calibration services.

4.4.2 Records of reviews, including any significant changes, shall be maintained. Records shall also be maintained of pertinent discussions with a client relating to the client's requirements or the results of the work during the period of execution of the contract.

NOTE For review of routine and other simple tasks, the date and the identification (e.g. the initials) of the person in the laboratory responsible for carrying out the contracted work are considered adequate. For repetitive routine tasks, the review need be made only at the initial enquiry stage or on granting of the contract for on-going routine work performed under a general agreement with the client, provided that the client's requirements remain unchanged. For new, complex or advanced testing and/or calibration tasks, a more comprehensive record should be maintained.

4.4.3 The review shall also cover any work that is subcontracted by the laboratory.

4.4.4 The client shall be informed of any deviation from the contract.

4.4.5 If a contract needs to be amended after work has commenced, the same contract review process shall be repeated and any amendments shall be communicated to all affected personnel.

4.5 Subcontracting of tests and calibrations

4.5.1 When a laboratory subcontracts work whether because of unforeseen reasons (e.g. workload, need for further expertise or temporary incapacity) or on a continuing basis (e.g. through permanent subcontracting, agency or franchising arrangements), this work shall be placed with a competent subcontractor. A competent subcontractor is one that, for example, complies with this International Standard for the work in question.

4.5.2 The laboratory shall advise the client of the arrangement in writing and, when appropriate, gain the approval of the client, preferably in writing.

4.5.3 The laboratory is responsible to the client for the subcontractor's work, except in the case where the client or a regulatory authority specifies which subcontractor is to be used.

4.5.4 The laboratory shall maintain a register of all subcontractors that it uses for tests and/or calibrations and a record of the evidence of compliance with this International Standard for the work in question.

4.6 Purchasing services and supplies

4.6.1 The laboratory shall have a policy and procedure(s) for the selection and purchasing of services and supplies it uses that affect the quality of the tests and/or calibrations. Procedures shall exist for the purchase, reception and storage of reagents and laboratory consumable materials relevant for the tests and calibrations.

4.6.2 The laboratory shall ensure that purchased supplies and reagents and consumable materials that affect the quality of tests and/or calibrations are not used until they have been inspected or otherwise verified as complying with standard specifications or requirements defined in the methods for the tests and/or calibrations concerned. These services and supplies used shall comply with specified requirements. Records of actions taken to check compliance shall be maintained.

4.6.3 Purchasing documents for items affecting the quality of laboratory output shall contain data describing the services and supplies ordered. These purchasing documents shall be reviewed and approved for technical content prior to release.

NOTE The description may include type, class, grade, precise identification, specifications, drawings, inspection instructions, other technical data including approval of test results, the quality required and the quality system standard under which they were made.

4.6.4 The laboratory shall evaluate suppliers of critical consumables, supplies and services which affect the quality of testing and calibration, and shall maintain records of these evaluations and list those approved.

4.7 Service to the client

The laboratory shall afford clients or their representatives cooperation to clarify the client's request and to monitor the laboratory's performance in relation to the work performed, provided that the laboratory ensures confidentiality to other clients.

NOTE 1 Such cooperation may include:

- a) providing the client or the client's representative reasonable access to relevant areas of the laboratory for the witnessing of tests and/or calibrations performed for the client;
- b) preparation, packaging, and dispatch of test and/or calibration items needed by the client for verification purposes.

NOTE 2 Clients value the maintenance of good communication, advice and guidance in technical matters, and opinions and interpretations based on results. Communication with the client, especially in large assignments, should be maintained throughout the work. The laboratory should inform the client of any delays or major deviations in the performance of the tests and/or calibrations.

NOTE 3 Laboratories are encouraged to obtain other feedback, both positive and negative, from their clients (e.g. client surveys). The feedback should be used to improve the quality system, testing and calibration activities and client service.

4.8 Complaints

The laboratory shall have a policy and procedure for the resolution of complaints received from clients or other parties. Records shall be maintained of all complaints and of the investigations and corrective actions taken by the laboratory (see also 4.10).