

---

---

**Microbeam analysis — Scanning electron  
microscopy — Guidelines for calibrating  
image magnification**

*Analyse par microfaisceaux — Microscopie électronique à balayage —  
Lignes directrices pour l'étalonnage du grossissement d'image*

**iTeh STANDARD PREVIEW**  
**(standards.iteh.ai)**

[ISO 16700:2004](#)

<https://standards.iteh.ai/catalog/standards/sist/e16cb0b9-f931-4aa4-9add-402832c4a8e2/iso-16700-2004>



**PDF disclaimer**

This PDF file may contain embedded typefaces. In accordance with Adobe's licensing policy, this file may be printed or viewed but shall not be edited unless the typefaces which are embedded are licensed to and installed on the computer performing the editing. In downloading this file, parties accept therein the responsibility of not infringing Adobe's licensing policy. The ISO Central Secretariat accepts no liability in this area.

Adobe is a trademark of Adobe Systems Incorporated.

Details of the software products used to create this PDF file can be found in the General Info relative to the file; the PDF-creation parameters were optimized for printing. Every care has been taken to ensure that the file is suitable for use by ISO member bodies. In the unlikely event that a problem relating to it is found, please inform the Central Secretariat at the address given below.

**iTeh STANDARD PREVIEW**  
**(standards.iteh.ai)**

ISO 16700:2004

<https://standards.iteh.ai/catalog/standards/sist/e16cb0b9-f931-4aa4-9add-402832c4a8e2/iso-16700-2004>

© ISO 2004

All rights reserved. Unless otherwise specified, no part of this publication may be reproduced or utilized in any form or by any means, electronic or mechanical, including photocopying and microfilm, without permission in writing from either ISO at the address below or ISO's member body in the country of the requester.

ISO copyright office  
Case postale 56 • CH-1211 Geneva 20  
Tel. + 41 22 749 01 11  
Fax + 41 22 749 09 47  
E-mail [copyright@iso.org](mailto:copyright@iso.org)  
Web [www.iso.org](http://www.iso.org)

Published in Switzerland

## Contents

Page

Foreword .....	iv
Introduction .....	v
<b>1 Scope</b> .....	<b>1</b>
<b>2 Normative references</b> .....	<b>1</b>
<b>3 Terms and definitions</b> .....	<b>1</b>
<b>4 Image magnification</b> .....	<b>3</b>
4.1 Scale marker .....	3
4.2 Expressing magnification .....	3
<b>5 Reference material</b> .....	<b>4</b>
5.1 General .....	4
5.2 Requirements for CRM .....	4
5.3 Pitch patterns on CRM .....	4
5.4 Storage and handling .....	4
<b>6 Calibration procedures</b> .....	<b>5</b>
6.1 General .....	5
6.2 Mounting CRM .....	5
6.3 Setting SEM operation conditions for calibration .....	5
6.4 Image recording .....	6
6.5 Measurement of image .....	6
6.6 Calibration of magnification and scale marker .....	7
<b>7 Accuracy of image magnification and scale marker</b> .....	<b>8</b>
<b>8 Calibration report</b> .....	<b>9</b>
8.1 General .....	9
8.2 Contents of calibration report .....	9
<b>Annex A (informative) Reference materials for magnification</b> .....	<b>10</b>
<b>Annex B (informative) Parameters that influence the resultant magnification of an SEM</b> .....	<b>12</b>
<b>Annex C (informative) Uncertainties in magnification measurements</b> .....	<b>14</b>
<b>Annex D (informative) Example of a test report</b> .....	<b>15</b>

## Foreword

ISO (the International Organization for Standardization) is a worldwide federation of national standards bodies (ISO member bodies). The work of preparing International Standards is normally carried out through ISO technical committees. Each member body interested in a subject for which a technical committee has been established has the right to be represented on that committee. International organizations, governmental and non-governmental, in liaison with ISO, also take part in the work. ISO collaborates closely with the International Electrotechnical Commission (IEC) on all matters of electrotechnical standardization.

International Standards are drafted in accordance with the rules given in the ISO/IEC Directives, Part 2.

The main task of technical committees is to prepare International Standards. Draft International Standards adopted by the technical committees are circulated to the member bodies for voting. Publication as an International Standard requires approval by at least 75 % of the member bodies casting a vote.

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. ISO shall not be held responsible for identifying any or all such patent rights.

ISO 16700 was prepared by Technical Committee ISO/TC 202, *Microbeam analysis*, Subcommittee SC 4, *Scanning electron microscopy*.

## iTeh STANDARD PREVIEW (standards.iteh.ai)

[ISO 16700:2004](https://standards.iteh.ai/catalog/standards/sist/e16cb0b9-f931-4aa4-9add-402832c4a8e2/iso-16700-2004)

<https://standards.iteh.ai/catalog/standards/sist/e16cb0b9-f931-4aa4-9add-402832c4a8e2/iso-16700-2004>

## Introduction

The scanning electron microscope is widely used to investigate the surface structure of a range of important materials such as semiconductors, metals, polymers, glass, food and biological materials and this International Standard is relevant to the need for magnification calibration of the images. It describes the requirements for calibration of the image magnification in the scanning electron microscope using a reference material or a certified reference material.

## iTeh STANDARD PREVIEW (standards.iteh.ai)

[ISO 16700:2004](https://standards.iteh.ai/catalog/standards/sist/e16cb0b9-f931-4aa4-9add-402832c4a8e2/iso-16700-2004)

<https://standards.iteh.ai/catalog/standards/sist/e16cb0b9-f931-4aa4-9add-402832c4a8e2/iso-16700-2004>

**iTeh STANDARD PREVIEW**  
**(standards.iteh.ai)**

ISO 16700:2004

<https://standards.iteh.ai/catalog/standards/sist/e16cb0b9-f931-4aa4-9add-402832c4a8e2/iso-16700-2004>

# Microbeam analysis — Scanning electron microscopy — Guidelines for calibrating image magnification

## 1 Scope

This International Standard specifies a method for calibrating the magnification of images generated by a scanning electron microscope (SEM) using an appropriate reference material. This method is limited to magnifications determined by the available size range of structures in the calibrating reference material. This International Standard does not apply to the dedicated critical dimension measurement SEM.

## 2 Normative references

The following referenced documents are indispensable for the application of this document. For dated references, only the edition cited applies. For undated references, the latest edition of the referenced document (including any amendments) applies.

ISO Guide 30:1992, *Terms and definitions used in connection with reference materials*

ISO Guide 34:1996, *Quality system guidelines for the production of reference materials*

ISO Guide 35:1989, *Natural rubber latex concentrate — Determination of mechanical stability*

ISO 5725-1:1994, *Accuracy (trueness and precision) of measurement methods and results — Part 1: General principles and definitions*

ISO/IEC 17025:1999, *General requirements for the competence of testing and calibration laboratories*

## 3 Terms and definitions

For the purposes of this document, the following terms and definitions apply.

### 3.1

#### **scanning electron microscope (SEM)**

instrument that produces magnified images of a specimen by scanning its surface with an electron beam

### 3.2

#### **image**

two-dimensional representation of the specimen surface generated by SEM

NOTE A photograph of a specimen taken using an SEM is a good example of an image.

### 3.3

#### **image magnification**

ratio of the linear dimension of the scan display to the corresponding linear dimension of the specimen scan field

**3.4**  
**scale marker**

line on the image representing a designated actual length in the specimen

**3.5**  
**reference material (RM)**

material or substance one or more of whose property values are sufficiently homogeneous and well established to be used for the calibration of an apparatus, the assessment of a measurement method, or for assigning values to materials

**3.6**  
**certified reference material (CRM)**

reference material, accompanied by a certificate, one or more of whose property values are certified by a procedure which establishes its traceability to an accurate realization of the unit in which the property values are expressed, and for which each certified value is accompanied by an uncertainty at a stated level of confidence

NOTE For the purposes of this document, an RM/CRM possesses pitch pattern(s) with the desired range of pitch size(s) and accuracy, to be used for the calibration of the image magnification.

**3.7**  
**calibration**

set of operations which establish, under specified conditions, the relationship between the magnification indicated by the SEM and the corresponding magnification determined by examination of an RM or a CRM


**3.8**  
**tilt angle**

angle of the inclined specimen surface from the plane perpendicular to the electron beam axis

iTeh STANDARD PREVIEW  
(standards.iteh.ai)

See Figure 1.

ISO 16700:2004  
<https://standards.iteh.ai/catalog/standards/sist/e16cb0b9-f931-4aa4-9add-402832c4a8e2/iso-16700-2004>



**Key**

- 1 tilted specimen
- 2 electron beam
- 3 specimen
- 4 tilt angle

Figure 1 — Tilt angle



**3.9****display**

analog or digital device used for visualization of images

NOTE Examples of display are a cathode ray tube, plasma display panel, liquid crystal display, etc.

**3.10****working distance**

distance between the specimen surface and the bottom plane of the objective lens of the SEM

**3.11****pitch**

closest separation of two similar features on a specimen which are equivalent points on a repeat pattern

**3.12****accuracy**

the closeness of agreement between a test result and the accepted reference value

[ISO 5725-1:1994]

NOTE 1 A “test result” constitutes the observed values of a pitch of a CRM obtained by the procedure outlined in this International Standard.

NOTE 2 The term “accepted reference value” is a value certified by a national or an international calibrating laboratory. There will be an uncertainty associated with this value which should also appear on the certificate.

NOTE 3 Accuracy and precision are different. Precision is defined as the closeness of agreement between independent test results obtained under stipulated conditions. See ISO 5725-1.

(standards.iteh.ai)


**4 Image magnification**

ISO 16700:2004

**4.1 Scale marker**

<https://standards.iteh.ai/catalog/standards/sist/e16cb0b9-f931-4aa4-9add-402832c4a8e2/iso-16700-2004>

To indicate magnification, superimpose on the image a scale marker and the corresponding length, in SI units, that it actually represents on the specimen. An example is shown in Figure 2.



**Figure 2 — Scale marker and its length**

NOTE The value of the scale marker generated by the hardware and/or the software of the SEM may not correspond to the calibrated value.

**4.2 Expressing magnification**

Magnification of an image is given by a number representing the number of times the object has been magnified and it is accompanied by the symbol “x” (e.g. 100 x, 10 000 x, 10k x or x 100, x 10 000, x 10 k where 100, 10 000 and 10 k are magnitude numbers). See Annex A.

NOTE 1 It is not always necessary to show the magnification when the scale marker is shown on the image.

NOTE 2 The magnification shown on the image corresponds to a chosen output device, which can be a display monitor or a printer or a photographing device. The scale marker shown on the image is independent from the output device chosen by the operator of the SEM. The magnification shown corresponds to the scale marker only when the image is displayed or printed on the operator-chosen output device.

## 5 Reference material

### 5.1 General

See ISO Guide 30.

For calibrating the magnification of an image, wherever possible, choose a CRM that is produced in accordance with ISO Guide 34 and certified in accordance with ISO Guide 35.

When a suitable CRM is not available, an RM produced in accordance with ISO Guide 34 may be used.

### 5.2 Requirements for CRM

Ensure that the chosen CRM:

- is stable with respect to vacuum and repeated electron beam exposure;
- provides good contrast in the SEM image;
- is electrically conductive;
- can be cleaned to remove contamination occurring during normal use without causing mechanical damage or distortion;
- has an associated valid calibration certificate.

### 5.3 Pitch patterns on CRM

Pitch patterns on the CRM may be in any one or more of the following forms:

- an orthogonal cross grid;
- a line array;
- a dot array;
- an orthogonal dot array.

Ensure that the chosen CRM contains pitch patterns that allow for calibration in at least one direction, and that the uncertainty in the pitch patterns is consistent with the targeted accuracy.

NOTE 1 The CRM may contain pitch patterns both in X and Y directions so that the measurements can be performed in orthogonal directions without the necessity of mechanically rotating the CRM. The CRM may additionally contain other structures for testing image distortion and/or resolution.

NOTE 2 The chosen CRM may have different sized pitch patterns to cover the whole range of magnifications for which calibration is needed. It may also be necessary to have more than one CRM to cover the desired range of magnifications.

### 5.4 Storage and handling

Store the CRM in a desiccating cabinet or in a vacuum container.

NOTE To ensure minimal handling of the actual CRM, it may be permanently mounted on a stub.

Handle the CRM using fingerstalls, clean room gloves or tweezers.

Visually inspect the CRM surface for contamination and deterioration, as this may affect calibration. Do not use the CRM if it is damaged or grossly contaminated.

Remove any dust, loose debris or other contamination from the CRM using clean dry air or nitrogen gas, taking care not to damage the CRM.

Check the calibration of the CRM at intervals by comparison with other CRMs; record the results. The frequency of verification may depend on the nature and usage of the CRM.

Use the CRM for calibration purposes only.

## 6 Calibration procedures

### 6.1 General

Parameters that influence the resultant magnification of an SEM may cause systematic errors. These are listed in Annex B.

The stability of the SEM will be a major factor in determining the calibration interval. Initially it will be necessary to perform calibration at frequent intervals in order to verify that the SEM is stable.

The results obtained will provide an estimate of the reproducibility within the laboratory and the bias inherent in both the display and the data automatically superimposed on any output.

The selection of the CRM depends on the magnification being used and accuracy required. For the purposes of this International Standard, ensure that the accuracy of calibration is better than 10 %.

### 6.2 Mounting CRM

At the time of mounting the specimen, ensure that handling of the CRM is carried out in accordance with 5.4.

Mount the CRM in accordance with the SEM and the CRM manufacturer's instructions.

Ascertain that there is a good electrical contact between the CRM and the specimen stage of the SEM.

Check that the CRM is securely fixed on the specimen stage so that it does not move from its mounting. This enables one to minimize any image degradation caused by vibration.

### 6.3 Setting SEM operation conditions for calibration

Evacuate the specimen chamber to the working vacuum in accordance with the SEM manufacturer's instructions.

Optimize the electron beam brightness and alignment in accordance with the SEM manufacturer's instructions.

Set tilt angle to 0° following the SEM manufacturer's instructions so that CRM surface is perpendicular to the electron beam axis during operation.

Check the tilt of the CRM by the following procedures.

- a) Turn off the tilt angle correction, the scan rotation and the zoom control of the magnification.
- b) Select the imaging mode (secondary electron and/or back scattered electron).
- c) Bring the image into focus without visible stigmatic distortions in the image.