
**Technical systems and aids for disabled
or handicapped persons — Wheelchair
tiedown and occupant-restraint
systems —**

Part 5:

Systems for specific wheelchairs

iTeh STANDARD PREVIEW

(standards.iteh.ai)
*Assistances et aides techniques pour les personnes invalides ou
handicapées — Systèmes d'attache du fauteuil roulant et de retenue de
l'occupant —*

[ISO 10542-5:2004](https://standards.iteh.ai/catalog/standards/sist/73b45b2a3e73/iso-10542-5-2004)

[Partie 5: Systèmes pour fauteuils roulants particuliers](https://standards.iteh.ai/catalog/standards/sist/73b45b2a3e73/iso-10542-5-2004)
73b45b2a3e73/iso-10542-5-2004



PDF disclaimer

This PDF file may contain embedded typefaces. In accordance with Adobe's licensing policy, this file may be printed or viewed but shall not be edited unless the typefaces which are embedded are licensed to and installed on the computer performing the editing. In downloading this file, parties accept therein the responsibility of not infringing Adobe's licensing policy. The ISO Central Secretariat accepts no liability in this area.

Adobe is a trademark of Adobe Systems Incorporated.

Details of the software products used to create this PDF file can be found in the General Info relative to the file; the PDF-creation parameters were optimized for printing. Every care has been taken to ensure that the file is suitable for use by ISO member bodies. In the unlikely event that a problem relating to it is found, please inform the Central Secretariat at the address given below.

iTeh STANDARD PREVIEW
(standards.iteh.ai)

[ISO 10542-5:2004](https://standards.iteh.ai/catalog/standards/sist/0f7362b4-a3b4-4d21-84f8-73b45b2a3e73/iso-10542-5-2004)

<https://standards.iteh.ai/catalog/standards/sist/0f7362b4-a3b4-4d21-84f8-73b45b2a3e73/iso-10542-5-2004>

© ISO 2004

All rights reserved. Unless otherwise specified, no part of this publication may be reproduced or utilized in any form or by any means, electronic or mechanical, including photocopying and microfilm, without permission in writing from either ISO at the address below or ISO's member body in the country of the requester.

ISO copyright office
Case postale 56 • CH-1211 Geneva 20
Tel. + 41 22 749 01 11
Fax + 41 22 749 09 47
E-mail copyright@iso.org
Web www.iso.org

Published in Switzerland

Foreword

ISO (the International Organization for Standardization) is a worldwide federation of national standards bodies (ISO member bodies). The work of preparing International Standards is normally carried out through ISO technical committees. Each member body interested in a subject for which a technical committee has been established has the right to be represented on that committee. International organizations, governmental and non-governmental, in liaison with ISO, also take part in the work. ISO collaborates closely with the International Electrotechnical Commission (IEC) on all matters of electrotechnical standardization.

International Standards are drafted in accordance with the rules given in the ISO/IEC Directives, Part 2.

The main task of technical committees is to prepare International Standards. Draft International Standards adopted by the technical committees are circulated to the member bodies for voting. Publication as an International Standard requires approval by at least 75 % of the member bodies casting a vote.

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. ISO shall not be held responsible for identifying any or all such patent rights.

ISO 10542-5 was prepared by Technical Committee ISO/TC 173, *Technical systems and aids for disabled or handicapped persons*, Subcommittee SC 1, *Wheelchairs*.

ISO 10542 consists of the following parts, under the general title *Technical systems and aids for disabled or handicapped persons — Wheelchair tiedown and occupant restraint systems*:

- Part 1: Requirements and test methods for all systems
- Part 2: Four-point strap-type tiedown systems
- Part 3: Docking-type tiedown systems
- Part 4: Clamping tiedown systems
- Part 5: Systems for specific wheelchairs

Introduction

Providing effective crash protection for wheelchair-seated occupants of motor vehicles generally requires the installation and use of equipment to tie down the wheelchair and restrain the occupant of the wheelchair. ISO 10542-1 establishes general requirements and test methods for wheelchair tiedown and occupant restraint systems (WTORS) that are designed for use with a wide range of wheelchairs and adult occupants. This part of ISO 10542 establishes additional or alternative requirements for WTORS that are intended to be used with specific makes and models of wheelchairs. The belt-type occupant restraints may attach to the wheelchair such that occupant-restraint loads will be transferred through the wheelchair. This part of ISO 10542 applies to WTORS intended for use with children whose mass is greater than 22 kg, in addition to WTORS intended for use by adults. Children whose mass is less than 22 kg should be transferred from their wheelchairs into appropriate child restraint systems intended for use in vehicles.

The provisions of ISO 10542-1 apply except as amended and supplemented by this part of ISO 10542. ISO 10542-1 and this part of ISO 10542 place particular emphasis on design requirements, performance requirements, and associated test methods with regard to the performance of WTORS for forward-facing wheelchairs and occupants in frontal impacts. Because this part of ISO 10542 addresses WTORS intended for use with specific makes and models of wheelchairs, the 48-km/h frontal impact test is conducted with a specific make or model of wheelchair, rather than with a surrogate wheelchair specified in ISO 10542-1. As such, the performance of both the WTORS and wheelchair are evaluated as a total system.

Transportation-related requirements for wheelchairs that are suitable for adult occupant seating during motor vehicle transportation are specified in ISO 7176-19.

iTeh STANDARD PREVIEW
(standards.iteh.ai)
<https://standards.iteh.ai/catalog/standards/sist/0f7362b4-a3b4-4d21-84f8-73b45b2a3e73/iso-10542-5-2004>

Technical systems and aids for disabled or handicapped persons — Wheelchair tiedown and occupant-restraint systems —

Part 5: Systems for specific wheelchairs

1 Scope

This part of ISO 10542 specifies test methods as well as requirements for design and performance, for instructions and warnings to installers and users, and for product marking and labelling of wheelchair tiedown and occupant restraint systems (WTORS).

It applies only to WTORS that are intended to be used with particular makes and models of wheelchairs when used as a forward-facing seat by a passenger or a driver of a motor vehicle.

This part of ISO 10542 applies to WTORS intended for use with manual or powered wheelchairs, including scooters with three or more wheels, intended for use by children or adults of mass equal to or greater than 22 kg.

This part of ISO 10542 applies primarily to complete WTORS, but portions of this International Standard can also be applied to components and subassemblies sold separately and for replacement parts.

2 Normative references

The following referenced documents are indispensable for the application of this document. For dated references, only the edition cited applies. For undated references, the latest edition of the referenced document (including any amendments) applies.

ISO 7176-22, *Wheelchairs — Part 22: Set-up procedures*

ISO 10542-1:2001, *Technical systems and aids for disabled or handicapped persons — Wheelchair tiedown and occupant-restraint systems — Part 1: Requirements and test methods for all systems*

3 Terms and definitions

For the purposes of this document, the terms and definitions given in ISO 10542-1:2001, with the exception of 3.5, and the following apply.

3.1

anthropomorphic test device

ATD

articulated physical analogue used to represent a wheelchair occupant in a test

3.2

child

person having a mass equal to or greater than 22 kg and less than 36 kg

[ECE R44]

3.3

H Points

points located on the left and right sides of the pelvic region of an anthropomorphic test device (ATD) that represent the approximate locations of the human hip joint centres in the side views, as specified by the ATD manufacturer

4 Design requirements

The requirements of ISO 10542-1:2001 Clause 4 apply with the exception of the following:

- a) Delete 4.1 b), 4.3.1 b) and 4.3.1 c) NOTE.
- b) Add the following to 4.1 WTORS.

WTORS shall include a belt-type occupant restraint either by specifying use of the belt-type occupant restraint and anchorages provided with the vehicle, or as a part of a complete WTORS from one manufacturer or as a part of a WTORS compiled from different manufacturers' equipment. The belt-type occupant restraint of the WTORS may be fixed to the vehicle, the wheelchair or a part (or parts) of the tiedown system.

- c) Add the following to 4.3 Belt-type occupant restraints.

Occupant restraints shall be designed to either:

- function independently of the wheelchair, so that restraint belts anchor to either the vehicle or the wheelchair tiedown components, or
- anchor to the wheelchair so that occupant restraint loads are transmitted through the wheelchair frame.

5 Identification, labelling, instruction and warning requirements

The requirements of ISO 10542-1:2001 Clause 5 apply with the exception of the following.

- a) Delete 5.1.1 c), 5.2.2 c) and 5.3.2 a).
- b) Add the following to 5.1 Identification and labelling:
 - 1) WTORS and replacement parts shall be permanently and legibly marked showing that they conform to this part of ISO 10542.
 - 2) WTORS components and subassemblies shall be accompanied by information that includes the make and model of the wheelchairs and the maximum wheelchair plus occupant mass which can be used with the WTORS.
- c) Add the following to 5.2 Instructions for installers:

Manufacturers of WTORS components and subassemblies shall provide written information that includes the following:

 - 1) the make and model of wheelchairs and the maximum wheelchair plus occupant mass which can be used with the components and or subassemblies or WTORS;
 - 2) a statement that the component or subassemblies or complete WTORS conforms to the requirement of this part of ISO 10542;
 - 3) diagrams and/or drawings which describe the method of installation if the occupant restraint is to be attached to the wheelchair.
- d) Add the following to 5.3 User and maintenance instructions:
 - 1) a statement that the WTORS conforms to this part of ISO 10542;
 - 2) where applicable, descriptions of the features required of a wheelchair or tiedown system for effective attachment of occupant restraint end fittings;
 - 3) warnings that the WTORS, or if applicable a component or subassemblies, are only for use with specific makes and models of wheelchairs and the stated maximum wheelchair plus occupant mass.

6 Performance requirements

The performance requirements of ISO 10542-1:2001 Clause 6 apply with the exception of the following:

- a) Substitute the term “SWC” with the term “test wheelchair” throughout 6.2.1.
- b) Delete 6.2.2 a) and Table 3.
- c) Add the following:
 - 1) When tested in accordance with Annex A, the horizontal excursions of the ATD and the wheelchair with respect to the impact sled shall not exceed the values given in Table 1.
 - 2) During the test, batteries of powered wheelchairs or their surrogate replacement parts shall
 - i) not move completely outside the plan view of the wheelchair,
 - ii) not move into the wheelchair user's space (e.g. shall not contact the back of the ATD's legs).
 - 3) When tested as specified in Annex A, the following requirements shall be met at the end of the test.
 - i) The wheelchair shall remain in an upright position on the test platform and the ATD shall be retained in the wheelchair in a seated posture, as determined by the ATD torso being oriented at not more than 45° to the vertical when viewed from any direction.
 - ii) The wheelchair securement points shall not show visible signs of failure.
 - iii) Release of wheelchair from the tiedown system shall not require the use of tools.
 - iv) For wheelchairs designed for securement by manually operated tiedowns, securement points shall not show deformation or distortion that prevents manual disengagement and removal of the end fittings of the tiedown system.
 - v) Components, fragments or accessories of the wheelchair with a mass in excess of 100 g shall not have completely separated from the wheelchair.
 - vi) Wheelchair components that may contact the occupant shall not fragment or separate in a manner that produces sharp edges with a radius of less than 2 mm.
 - vii) Locking mechanisms of tilt-in-space seat adjusters shall not show signs of failure.
 - viii) Removal of the ATD from the wheelchair shall not require the use of tools other than a hoist.
 - ix) The post test height of the average of left and right ATD H-points, relative to the wheelchair ground plane, shall not decrease by more than 20 % from the pre-test height.

Table 1 — Horizontal excursion limits (mm)

Measurement point	Excursion variable	ATD		
		6-year old child	Small adult female	Midsized adult and large adult male
Wheelchair point P	d_{wc}^a	150	200	200
ATD knee centre	d_{knee}^b	300	375	375
ATD front of head	d_{headF}^c	450	550	650
ATD back of head	d_{headR}^d	-350	-400	-450

^a d_{wc} = the horizontal distance relative to the sled platform between the contrast marker placed at or near point P on the test wheelchair at time t_o , to the Point P marker at the time of peak wheelchair excursion.

^b d_{knee} = the horizontal distance relative to the sled platform between the dummy knee joint marker at time t_o , to the knee joint marker at the time of peak knee excursion.

^c d_{headF} = the horizontal distance relative to the sled platform between the most forward point on the dummy's head above the nose at time t_o , to the most forward point on the dummy's head at the time of peak forward head excursion.

^d d_{headR} = the horizontal distance relative to the sled platform between the most rearward point on the dummy's head at time t_o , to the most rearward point on the dummy's head at the time of peak rearward head excursion.

7 Test report

The test report requirements of ISO 10542-1:2001 Clause 7 apply with the exception of the following:

- a) Delete 7.2 d) and 7.2 e).
- b) Add the following:
 - 1) a description and total mass of the ATD used in the test;
 - 2) the model, designation and serial number of the wheelchair used in the test;
 - 3) the manufacturer of the wheelchair used in the test;
 - 4) the angle of the seat and backrest from A.2 f);
 - 5) a statement that the wheelchair meets or does not meet the requirements of all applicable parts of Clause 6.

iTeh STANDARD PREVIEW
(standards.iteh.ai)

[ISO 10542-5:2004](https://standards.iteh.ai/catalog/standards/sist/0f7362b4-a3b4-4d21-84f8-73b45b2a3e73/iso-10542-5-2004)

<https://standards.iteh.ai/catalog/standards/sist/0f7362b4-a3b4-4d21-84f8-73b45b2a3e73/iso-10542-5-2004>

Annex A (normative)

Test method for frontal impact

A.1 Principle

The wheelchair specified for the test is mounted in a forward-facing configuration on the impact sled of an impact simulator and is loaded with an anthropomorphic test device (ATD). The WTORS is installed to secure the wheelchair and restrain the ATD. The sled platform is subjected to a defined acceleration/deceleration-time pulse, in order to achieve a defined horizontal velocity change (Δv). Observations and measurements are made to determine if the WTORS and the wheelchair conform to the requirements of this part of ISO 10542.

A.2 Test requirements

A.2, A.3, A.4 and A.5 of Annex A of ISO 10542-1:2001 apply with the exception of the following:

- a) Substitute the term "SWC or surrogate wheelchair" with the term "wheelchair specified for the test" throughout the text.
- b) Delete A.3.1 e), A.3.1 f) and A.4.5.
- c) The ATD shall be selected from Table A.1 below and represent the wheelchair manufacturer's intended maximum user mass.
- d) For wheelchairs intended for use by adults, either the midsize adult male or large adult male ATD shall be used.
- e) The midsize male ATD shall be a Hybrid II or Hybrid III type. The other sizes of ATDs shall be Hybrid II, Hybrid III or VIP types.

NOTE See Annex G of ISO 10542-1:2001 for information on sources for ATDs.

- f) The wheelchair specified for the test shall be in addition to the test sample A.2 of ISO 10542-1:2001.
- g) A.4.1, A.4.4, A.4.9 and A.4.15 shall be replaced by the following:
 - 1) A.4.1 The procedures for setting up the test may be undertaken in any order.

EXAMPLE It may be more convenient to position the ATD in the wheelchair prior to placing the assembly on the impact sled.

- 2) A.4.4 Set up the test apparatus in accordance with A.3 and set up the test wheelchair as below.
 - i) Set up the wheelchair in accordance with ISO 7176-22.
 - ii) Prepare the wheelchair for use in a motor vehicle as specified by the wheelchair manufacturer's user instructions. If a range is specified for any adjustments, the midpoint of the range should be used.
 - iii) Equip the wheelchair with any required add-on components.
 - iv) If a pelvic belt, intended for use as an occupant restraint, is provided as a component of the wheelchair, install it on the wheelchair according to the manufacturer's instructions.
 - v) If the wheelchair is equipped with liquid electrolyte batteries, they should be replaced by the nearest equivalent gel, sealed or surrogate battery. Supplementary weights, if used, shall provide equivalent mass distribution to the original batteries.