



SLOVENSKI STANDARD

SIST EN 15479:2009

01-junij-2009

Nadomeglj a:

SIST-TS CEN/TS 15479:2006

Gnojila - Spektrofotometrijsko določevanje biureta v seji nini

Fertilizers - Spectrophotometric determination of biuret in urea

Düngemittel - Spektrometrische Bestimmung von Biuret in Harnstoff

Engrais - Détermination photométrique du biuret dans l'urée

iTeh STANDARD PREVIEW
(standards.iteh.ai)

Ta slovenski standard je istoveten z: EN 15479:2009

<https://standards.iteh.ai/catalog/standards/sist/d96a926d-6941-4ba6-bb7d-89d1106ceefc/sist-en-15479-2009>

ICS:

65.080

Gnojila

Fertilizers

SIST EN 15479:2009

en,fr,de

iTeh STANDARD PREVIEW
(standards.iteh.ai)

SIST EN 15479:2009

<https://standards.iteh.ai/catalog/standards/sist/d96a926d-6941-4ba6-bb7d-89d1106ecef8/sist-en-15479-2009>

EUROPEAN STANDARD
NORME EUROPÉENNE
EUROPÄISCHE NORM

EN 15479

January 2009

ICS 65.080

Supersedes CEN/TS 15479:2006

English Version

Fertilizers - Spectrophotometric determination of biuret in urea

Engrais - Détermination photométrique du biuret dans l'urée

Düngemittel - Spektrometrische Bestimmung von Biuret in Harnstoff

This European Standard was approved by CEN on 30 November 2008.

CEN members are bound to comply with the CEN/CENELEC Internal Regulations which stipulate the conditions for giving this European Standard the status of a national standard without any alteration. Up-to-date lists and bibliographical references concerning such national standards may be obtained on application to the CEN Management Centre or to any CEN member.

This European Standard exists in three official versions (English, French, German). A version in any other language made by translation under the responsibility of a CEN member into its own language and notified to the CEN Management Centre has the same status as the official versions.

CEN members are the national standards bodies of Austria, Belgium, Bulgaria, Cyprus, Czech Republic, Denmark, Estonia, Finland, France, Germany, Greece, Hungary, Iceland, Ireland, Italy, Latvia, Lithuania, Luxembourg, Malta, Netherlands, Norway, Poland, Portugal, Romania, Slovakia, Slovenia, Spain, Sweden, Switzerland and United Kingdom.

ITh STANDARD PREVIEW
(standards.iteh.ai)
<https://standards.iteh.ai/catalog/standards/sist/d96a926d-6941-4ba6-bb7d-89d1106ecef/sist-en-15479-2009>



EUROPEAN COMMITTEE FOR STANDARDIZATION
COMITÉ EUROPÉEN DE NORMALISATION
EUROPÄISCHES KOMITEE FÜR NORMUNG

Management Centre: rue de Stassart, 36 B-1050 Brussels

Contents

Page

Foreword.....	3
1 Scope	4
2 Normative references	4
3 Terms and definitions	4
4 Principle.....	4
5 Reagents.....	4
6 Apparatus	5
7 Sampling and sample preparation	6
8 Procedure	6
9 Calculation and expression of the result	7
10 Precision	7
11 Test report	8
Annex A (informative) Results of the inter-laboratory tests	9
Bibliography	10

[SIST EN 15479:2009](https://standards.itch.ai/catalog/standards/sist/d96a926d-6941-4ba6-bb7d-89d1106ecef/sist-en-15479-2009)
<https://standards.itch.ai/catalog/standards/sist/d96a926d-6941-4ba6-bb7d-89d1106ecef/sist-en-15479-2009>

Foreword

This document (EN 15479:2009) has been prepared by Technical Committee CEN/TC 260 “Fertilizers and liming materials”, the secretariat of which is held by DIN.

This European Standard shall be given the status of a national standard, either by publication of an identical text or by endorsement, at the latest by July 2009, and conflicting national standards shall be withdrawn at the latest by July 2009.

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. CEN [and/or CENELEC] shall not be held responsible for identifying any or all such patent rights.

This document supersedes CEN/TS 15479:2006.

This document has been prepared under a mandate given to CEN by the European Commission and the European Free Trade Association.

According to the CEN/CENELEC Internal Regulations, the national standards organizations of the following countries are bound to implement this European Standard: Austria, Belgium, Bulgaria, Cyprus, Czech Republic, Denmark, Estonia, Finland, France, Germany, Greece, Hungary, Iceland, Ireland, Italy, Latvia, Lithuania, Luxembourg, Malta, Netherlands, Norway, Poland, Portugal, Romania, Slovakia, Slovenia, Spain, Sweden, Switzerland and the United Kingdom.

(standards.iteh.ai)

SIST EN 15479:2009

<https://standards.iteh.ai/catalog/standards/sist/d96a926d-6941-4ba6-bb7d-89d1106ecef/sist-en-15479-2009>

EN 15479:2009 (E)

1 Scope

This European Standard specifies a method for the determination of biuret in urea. The method is applicable to urea and urea-based fertilizers.

2 Normative references

The following referenced documents are indispensable for the application of this document. For dated references, only the edition cited applies. For undated references, the latest edition of the referenced document (including any amendments) applies.

EN 1482-2, *Fertilizers and liming materials — Sampling and sample preparation — Part 2: Sample preparation*

EN 12944-1:1999, *Fertilizers and liming materials and soil improvers — Vocabulary — Part 1: General terms*

EN 12944-2:1999, *Fertilizers and liming materials and soil improvers — Vocabulary — Part 2: Terms relating to fertilizers*

EN ISO 3696:1995, *Water for analytical laboratory use — Specification and test methods (ISO 3696:1987)*

3 Terms and definitions

For the purposes of this document, the terms and definitions given in EN 12944-1:1999 and EN 12944-2:1999 apply.

SIST EN 15479:2009
<https://standards.iteh.ai/catalog/standards/sist/d96a926d-6941-4ba6-bb7d-89d1106ecef/sist-en-15479-2009>

4 Principle

In an alkaline medium in the presence of potassium sodium tartrate, biuret and bivalent copper form a violet cupric compound. The absorbance of the solution is measured at a wavelength of about 546 nm.

5 Reagents

5.1 General

Use only reagents of recognized analytical grade and distilled or demineralized water, free from carbon dioxide and ammonia (grade 3 according to EN ISO 3696:1995). The quality of the water is particularly important in this determination.

5.2 Methanol

5.3 Sulfuric acid solution, c = approximately 0,05 mol/l.

5.4 Sodium hydroxide solution, c = approximately 0,1 mol/l.

5.5 Alkaline solution of potassium sodium tartrate

In a graduated one-litre flask, dissolve 40 g of sodium hydroxide in 500 ml of water and leave to cool. Add 50 g of potassium sodium tartrate ($\text{NaKC}_4\text{H}_4\text{O}_6 \cdot 4\text{H}_2\text{O}$). Make up to the mark and leave standing for 24 h before use.

5.6 Solution of copper sulfate

In a graduated one-litre flask dissolve 15 g of copper sulfate ($\text{CuSO}_4 \cdot 5\text{H}_2\text{O}$) in 500 ml of water. Make up to the mark.

5.7 Freshly prepared biuret standard solution

In a 250 ml graduated flask, dissolve 0,250 g of pure biuret¹ in water. Make up to 250 ml. 1 ml of this solution contains 0,001 g of biuret.

5.8 Indicator solution

In a graduated 100 ml flask, dissolve 0,1 g of methyl red in 50 ml of 95 % ethanol, make up to 100 ml with water. Filter if any insoluble remain. This indicator is red in acid solution and yellow in alkaline solution.

6 Apparatus

6.1 Spectrometer or photometer

With filters with a sensitivity and precision to permit measures of less than 0,5 % factor of transmission to be produced. The factor of transmission, T , is given by the following equation:

$$T = \frac{J}{J_0} \quad \text{iTeh STANDARD PREVIEW} \quad (1)$$

(standards.iteh.ai)

The opacity, O , is given by the following equation:

$$O = \frac{J_0}{J} \quad \text{SIST EN 15479:2009} \quad (2)$$

<https://standards.iteh.ai/catalog/standards/sist/d96a926d-6941-4ba6-bb7d-89d1106ecef/sist-en-15479-2009>

The absorbance, E , is given by the following equation:

$$E = \log O \quad (3)$$

The absorbance per unit of optical run, k , is given by the following equation:

$$k = \frac{E}{s} \quad (4)$$

The coefficient of specific absorbance, K , is given by the following equation:

$$K = \frac{E}{\rho \times s} \quad (5)$$

where

J_0 is the intensity of the beam of monochromatic rays (of a determinate wavelength) before it passes through a transparent body;

J is the intensity of this beam after the passage;

¹ Biuret can be purified beforehand by washing with an ammoniacal solution (10 %), then with acetone and dried in a vacuum.

EN 15479:2009 (E)

- s is the thickness of the layer in centimetres;
- ρ is the mass concentration in milligrams per litre;
- k is the specific factor for each substance in the Lambert-Beer law.

6.2 Graduated flasks, capacity of 100 ml, 250 ml and 1 000 ml.

6.3 Graduated pipettes, capacity of 2 ml, 5 ml, 10 ml, 20 ml, 25 ml and 50 ml,

or a **25 ml burette**, graduated to 0,05 ml.

6.4 Beaker, capacity 250 ml.

6.5 Glass column for cation exchange resin

Height 20 cm, diameter 25 mm, with a bottom plate of nylon gauze with a mesh width of 0,1 mm and ending in a narrow outlet with tap.

The column is filled with cation exchange resin (strong cation: e. g. Duolite C20) to a height of about 120 mm. Before using the column, the resin has first to be regenerated with 100 ml hydrochloric acid (hydrochloric acid, approximately 4 mol/l solution) and subsequently rinsed with water until the eluent is free from acid. After each determination the resin has to be regenerated. The usable capacity of the cation resin has to be great enough to be sure all the ammonia is caught. The total volume of the resin is about 60 cm³; the usable capacity is about 60 milli-equivalents.

ITEH STANDARD PREVIEW
(standards.iteh.ai)

7 Sampling and sample preparation SIST EN 15479:2009

Sampling is not part of the method specified in this document. A recommended sampling method is given in EN 1482-1.

Sample preparation shall be carried out in accordance with EN 1482-2.

8 Procedure

8.1 Preparation of the calibration curve

Transfer 0 ml, 2 ml, 5 ml, 10 ml, 20 ml, 25 ml and 50 ml aliquots of biuret standard solution (5.7) into a series of seven graduated 100 ml flasks. Make up the volumes to about 50 ml with water, add one drop of indicator (5.8) and neutralize, if necessary, with sulfuric acid 0,05 mol/l (5.3). Stir in 20 ml of the alkaline tartrate solution (5.5) then 20 ml of the copper sulfate solution (5.6).

These solutions shall be measured with two precision burettes or better still with pipettes.

Make up to 100 ml with water, mix and leave standing for 15 min at $(30 \pm 2) ^\circ\text{C}$.

With the "0" biuret standard solution as a reference, measure the absorbance of each solution at a wavelength of about 546 nm using cells of a suitable path length.

Plot the calibration curve using the absorbance rate as the ordinate and the corresponding quantities of biuret, in milligrams, as the abscissa.

8.2 Preparation of the solution to be analysed

Weigh, to the nearest 0,001 g, 10 g of the prepared sample; dissolve in about 150 ml of water in a 250 ml graduated flask, and make up to the mark. Filter if necessary.

Weigh to the nearest 0,001 g about 25 g of the ammoniumsulfate containing sample. Transfer into a beaker glass of 250 ml with about 100 ml of water at about 70 °C and dissolve. Filter the solution over a Büchner funnel with a glass fibre filter and wash the filter with approximately 50 ml of water. Transfer the filtrate in a volumetric flask of 250 ml.

Bring the temperature to 20 °C, dilute to the mark with water and mix. Pipette 100 ml of the solution and pass it through the cation resin with a speed of approximately 150 ml/h. Collect the eluate in a 250 ml volumetric flask. Wash the resin with water until a total volume of approximately 220 ml is obtained in the volumetric flask. Dilute to the mark with water and mix. Then proceed as indicated in 8.3.

Elimination of the opalescence: if any colloid substance is present, difficulties might arise during filtering. The solution intended for analysis is in that case prepared as follows:

- dissolve the sample for analysis in 150 ml of water, add 2 ml of 1 mol/l hydrochloric acid and filter the solution through two flat very fine filters into a graduated 250 ml flask. Wash the filters with water and make up to volume. Continue the process according to the method as described in 8.3.

8.3 Determination

According to the presumed biuret content, transfer 25 ml or 50 ml from the solution mentioned in 8.2 with a pipette, place this quantity in a 100 ml graduated flask and neutralize if necessary with sulfuric acid solution or sodium hydroxide solution (5.3 or 5.4) as required, using methyl red as an indicator and add, with the same accuracy as that used when drawing up a calibration curve, 20 ml of the alkaline solution of potassium sodium tartrate (5.5) and 20 ml of the copper solution (5.6). Make up to volume, mix thoroughly and leave standing for 15 min at (30 ± 2) °C.

<https://standards.iteh.ai/catalog/standards/sist/d96a926d-6941-4ba6-bb7d-89d1106ace6/sist-en-15479-2009>

Then carry out the photometric measurements and calculate the quantity of biuret present in the urea.

9 Calculation and expression of the result

Calculate the mass concentration, ρ_B , of biuret in percent of the sample according to the following equation:

$$\rho_B = \frac{m \times 2,5}{V} \quad (6)$$

where

m is the mass, in milligrams, of biuret, read from the calibration graph;

V is the volume of the aliquot in millilitre.

10 Precision

10.1 Inter-laboratory test

An inter-laboratory test was carried out in 2004 with 13 participating laboratories and two different samples. This test yielded the data given in Annex A. Repeatability and reproducibility were calculated according to ISO 5725-1.

The values derived from this inter-laboratory test might not be applicable to concentration ranges and matrices other than those given in Annex A.