



INTERNATIONAL STANDARD ISO 10942:1998
TECHNICAL CORRIGENDUM 1

Published 1998-09-01

INTERNATIONAL ORGANIZATION FOR STANDARDIZATION • МЕЖДУНАРОДНАЯ ОРГАНИЗАЦИЯ ПО СТАНДАРТИЗАЦИИ • ORGANISATION INTERNATIONALE DE NORMALISATION

Ophthalmic instruments — Direct ophthalmoscopes

TECHNICAL CORRIGENDUM 1

Instruments ophtalmiques — Ophtalmoscopes directs

RECTIFICATIF TECHNIQUE 1

Technical Corrigendum 1 to International Standard ISO 10942:1998 was prepared by Technical Committee ISO/TC 172, *Optics and optical instruments*, Subcommittee SC 7, *Ophthalmic optics and instruments*.

(standards.iteh.ai)

[ISO 10942:1998/Cor 1:1998](https://standards.iteh.ai/catalog/standards/sist/888d6748-0737-4e18-ada2-92d5a0e5c377/iso-10942-1998-cor-1-1998)

<https://standards.iteh.ai/catalog/standards/sist/888d6748-0737-4e18-ada2-92d5a0e5c377/iso-10942-1998-cor-1-1998>

Page 3

In Table 1, Criterion Angle of field of view φ , in column 2 for Group A replace $\geq 6^\circ$ with $\geq 3^\circ$, and in column 3 for Group B replace $\geq 5^\circ$ with $\geq 2,5^\circ$.