



INTERNATIONAL STANDARD ISO 5360:1993
TECHNICAL CORRIGENDUM 1

Published 1998-12-01

INTERNATIONAL ORGANIZATION FOR STANDARDIZATION • МЕЖДУНАРОДНАЯ ОРГАНИЗАЦИЯ ПО СТАНДАРТИЗАЦИИ • ORGANISATION INTERNATIONALE DE NORMALISATION

Anaesthetic vaporizers — Agent-specific filling systems

TECHNICAL CORRIGENDUM 1

Évaporateurs d'anesthésie — Systèmes de remplissage spécifiques à l'agent

RECTIFICATIF TECHNIQUE 1

Technical Corrigendum 1 to International Standard ISO 5360:1993 was prepared by Technical Committee ISO/TC 121, *Anaesthetic and respiratory equipment*, Subcommittee SC 1, *Breathing attachments and anaesthetic machines*.

**iTECH STANDARD REVIEW
(standards.iteh.ai)**

[ISO 5360:1993/Cor 1:1998](#)

<https://standards.iteh.ai/catalog/standards/sist/86961f3e-fe7a-480b-bafa-887636e7fc8e/iso-5360-1993-cor-1-1998>

Page 3

In **Table 1 — Dimensions of threaded necks of bottles for anaesthetic agents**

In row 8 (sevoflurane), last column (*d* max.) replace the value "32" with new value "28".