



Standard Test Methods for Iron in Water ¹

This standard is issued under the fixed designation D 1068; the number immediately following the designation indicates the year of original adoption or, in the case of revision, the year of last revision. A number in parentheses indicates the year of last reapproval. A superscript epsilon (ε) indicates an editorial change since the last revision or reapproval.

This standard has been approved for use by agencies of the Department of Defense. Consult the DoD Index of Specifications and Standards for the specific year of issue which has been adopted by the Department of Defense.

1. Scope

1.1 These test methods cover the determination of iron in water. Procedures are given for determining total iron, dissolved iron, and ferrous iron. Undissolved iron may be calculated from the total iron and dissolved iron determinations. The test methods are given as follows:

	Range	Sections
Test Method A—Atomic Absorption, Direct	0.1 to 5.0 mg/L	7 to 15
Test Method C—Atomic Absorption, Graphite Furnace	5 to 100 µg/L	16 to 24
Test Method D—Photometric Bathophenanthrolineµ g/L	40 to 1000 µg/L	25 to 36

1.2 It is the user's responsibility to ensure the validity of these test methods to waters of untested matrices.

1.3 *This standard does not purport to address all of the safety concerns, if any, associated with its use. It is the responsibility of the user of this standard to establish appropriate safety and health practices and determine the applicability of regulatory limitations prior to use.* Specific hazards statements are given in Note 3, Note 5, and Note X1.1.

1.4 Two former photometric test methods were discontinued. See Appendix X2 for historical information.

2. Referenced Documents

2.1 ASTM Standards:

- D 858 Test Methods for Manganese in Water ²
- D 1066 Practice for Sampling Steam ²
- D 1129 Terminology Relating to Water ²
- D 1192 Specification for Equipment for Sampling Water and Steam in Closed Conduits ²
- D 1193 Specification for Reagent Water ²
- D 1687 Test Methods for Chromium in Water ²
- D 1688 Test Methods for Copper in Water ²
- D 1691 Test Methods for Zinc in Water ²
- D 1886 Test Methods for Nickel in Water ²

¹ These test methods are under the jurisdiction of ASTM Committee D-19 on Water and are the direct responsibility of Subcommittee D19.05 on Inorganic Constituents in Water.

Current edition approved Feb. 10, 1996. Published May 1996. Originally published as D 1068 – 49 T. Last previous edition D 1068 – 90.

² Annual Book of ASTM Standards, Vol 11.01.

- D 2777 Practice for Determination of Precision and Bias of Applicable Methods of Committee D-19 on Water ²
- D 3370 Practices for Sampling Water from Closed Conduits ²
- D 3558 Test Methods for Cobalt in Water ²
- D 3559 Test Methods for Lead in Water ²
- D 3919 Practice for Measuring Trace Elements in Water by Graphite Furnace Atomic Absorption Spectrophotometry ²
- D 4841 Practice for Estimation of Holding Time for Water Samples Containing Organic and Inorganic Constituents ²
- E 60 Practice for Photometric and Spectrophotometric Methods for Chemical Analysis of Metals ³
- E 275 Practice for Describing and Measuring Performance of Ultraviolet, Visible, and Near Infrared Spectrophotometers ⁴

3. Terminology

3.1 *Definitions:* —For definitions of terms used in these test methods, refer to Terminology D 1129.

3.2 Definitions of Terms Specific to This Standard:

3.2.1 *total recoverable iron*—an arbitrary analytical term relating to the recoverable forms of iron that are determinable by the digestion method which is included in these test methods.

4. Significance and Use

4.1 Iron is the second most abundant metallic element in the earth's crust and is essential in the metabolism of plants and animals. If presented in excessive amounts, however, it forms oxyhydroxide precipitates that stain laundry and porcelain. As a result, the recommended limit for iron in domestic water supplies is 0.3 mg/L. These test methods are useful for determining iron in many natural waters.

5. Purity of Reagents

5.1 Reagent grade chemicals shall be used in all tests. Unless otherwise indicated, it is intended that all reagents shall

³ Annual Book of ASTM Standards, Vol 03.05.

⁴ Annual Book of ASTM Standards, Vol 03.06.

conform to the specifications of the Committee on Analytical Reagents of the American Chemical Society, where such specifications are available.⁵ Other grades may be used, provided it is first ascertained that the reagent is of sufficiently high purity to permit its use without lessening the accuracy of the determination.

5.2 *Purity of Water*—Unless otherwise indicated, references to water shall be understood to mean reagent water conforming to Specification D 1193, Type I. Other reagent water types may be used, provided it is first ascertained that the water is of sufficiently high purity to permit its use without adversely affecting the bias and precision of the test method. Type II water was specified at the time of round-robin testing of these test methods. In addition, water used in preparing solutions for the determination of ferrous iron shall be freshly boiled and essentially oxygen free.

6. Sampling

6.1 Collect the sample in accordance with Practice D 1066, Specification D 1192, or Practices D 3370, as applicable.

6.2 Samples should be preserved with HNO₃ or HCl (sp gr 1.42) to a pH of 2 or less immediately at the time of collection. If only dissolved iron is to be determined, the sample shall be filtered through a 0.45- μ m membrane filter before acidification. The holding time for samples can be calculated in accordance with Practice D 4841.

6.3 If ferrous iron is to be determined, the sample should be analyzed as soon as possible after collection and contact with atmospheric oxygen should be minimized.

6.4 Additional information on sampling requirements for Test Method D is provided in 32.1.

TEST METHOD A—ATOMIC ABSORPTION, DIRECT

7. Scope

7.1 This test method covers the determination of dissolved and total recoverable iron in most waters and wastewaters.

7.2 This test method is applicable in the range from 0.1 to 5.0 mg/L of iron. The range may be extended to concentrations greater than 5.0 mg/L by dilution of the sample.

7.3 This test method has been used successfully with reagent water; tap, ground, and surface waters; unspecified wastewaters; and a refinery primary treatment water. It is the user's responsibility to ensure the validity of this test method for waters of untested matrices.

8. Summary of Test Method

8.1 Iron is determined by atomic absorption spectrophotometry. Dissolved iron is determined by atomizing the filtered sample directly with no pretreatment. Total recoverable iron is determined by atomizing the sample following hydrochloric-nitric acid digestion and filtration. The same digestion proce-

dures may be used to determine total recoverable nickel (Test Methods D 1886), chromium (Test Methods D 1687), cobalt (Test Methods D 3558), copper (Test Methods D 1688), lead (Test Methods D 3559), manganese (Test Methods D 858), and zinc (Test Methods D 1691).

9. Interferences

9.1 Sodium, potassium, barium, chloride and sulfate (5000 mg/L each), calcium, magnesium, chromium, manganese, cobalt, nickel, copper, zinc, palladium, silver, cadmium, tin, lead, lithium, mercury, selenium, aluminum, antimony, arsenic, vanadium, boron, and molybdenum (100 mg/L) do not interfere.

9.2 Background correction (or chelation-extraction) may be necessary to determine low levels of iron in some waters.

NOTE 1—Instrument manufacturers' instructions for use of the specific correction technique should be followed.

10. Apparatus

10.1 *Atomic Absorption Spectrophotometer*, for use at 248.3 nm.

NOTE 2—The manufacturer's instructions should be followed for all instrumental parameters. A wavelength other than 248.3 nm may be used if it has been determined to be equally suitable.

10.1.1 *Iron Hollow-Cathode Lamp*—Multielement hollow-cathode lamps are available and have also been found satisfactory.

10.2 *Pressure-Reducing Valves*—The supplies of fuel and oxidant shall be maintained at pressures somewhat higher than the controlled operating pressure of the instrument by suitable valves.

11. Reagents and Materials

11.1 *Hydrochloric Acid* (sp gr 1.19)—Concentrated hydrochloric acid (HCl).

NOTE 3—If the reagent blank concentration is greater than the method detection limit, distill the HCl or use a spectrograde acid. **Precaution**—When HCl is distilled an azeotropic mixture is obtained (approximately 6 N HCl). Therefore, when concentrated HCl is specified for the preparation of reagents or in the procedure, use double the volume specified if distilled acid is used.

11.2 *Nitric Acid* (sp gr 1.42)—Concentrated nitric acid (HNO₃).

NOTE 4—If the reagent blank concentration is greater than the method detection limit, distill the HNO₃ or use a spectrograde acid.

11.3 *Nitric Acid* (1 + 499)—Add 1 volume of HNO₃ (sp gr 1.42) to 499 volumes of water.

11.4 *Iron Solution, Stock* (1 mL = 1.0 mg Iron)—Dissolve 1.000 g of pure iron in 100 mL of HCl (1 + 1) with the aid of heat. Cool and dilute to 1 L with water.

11.5 *Iron Solution, Standard* (1 mL = 0.1 mg Iron)—Dilute 100.0 mL of the iron stock solution to 1 L with water.

11.6 *Oxidant*:

11.6.1 *Air*, which has been passed through a suitable filter to remove oil, water, and other foreign substances is the usual oxidant.

11.7 *Fuel*:

⁵ *Reagent Chemicals, American Chemical Society Specifications*, American Chemical Society, Washington, DC. For suggestions on the testing of reagents not listed by the American Chemical Society, see *Analar Standards for Laboratory Chemicals*, BDH Ltd., Poole, Dorset, U.K., and the *United States Pharmacopeia and National Formulary*, U.S. Pharmaceutical Convention, Inc. (USPC), Rockville, MD.

11.7.1 *Acetylene*—Standard, commercially available acetylene is the usual fuel. Acetone, always present in acetylene cylinders can effect analytical results. The cylinder should be replaced at 50 psig (345 kPa).

NOTE 5—**Warning:** “Purified” grade acetylene containing a special proprietary solvent rather than acetone should not be used with poly vinyl chloride tubing as weakening of the tubing walls can cause a potentially hazardous situation.

12. Standardization

12.1 Prepare 100 mL each of a blank and at least four standard solutions to bracket the expected iron concentration range of the samples to be analyzed by diluting the standard iron solution with HNO₃ (1 + 499). Prepare the standards each time the test is to be performed.

12.2 When determining total recoverable iron add 0.5 mL of HNO₃ (sp gr 1.42) and proceed as directed in 13.1 through 13.5. When determining dissolved iron proceed as directed in Note 6, 13.1.

12.3 Aspirate the blank and standards and record the instrument readings. Aspirate HNO₃ (1 + 499) between each standard.

12.4 Prepare an analytical curve by plotting the absorbance versus concentration for each standard on linear graph paper. Alternatively read directly in concentration if this capability is provided with the instrument.

13. Procedure

13.1 Measure 100.0 mL of a well-mixed acidified sample into a 125-mL beaker or flask.

NOTE 6—If only dissolved iron is to be determined, start with 13.3.

13.2 Add 5 mL of HCl (sp gr 1.19) to each sample.

13.3 Heat the samples on a steam bath or hotplate in a well-ventilated hood until the volume has been reduced to 15 to 20 mL, making certain that the samples do not boil.

NOTE 7—When analyzing samples of brines or samples containing appreciable amounts of suspended matter or dissolved solids, the amount of reduction in volume is left to the discretion of the analyst.

13.4 Cool and filter the samples through a suitable filter (such as fine-textured, acid-washed, ashless paper), into 100-mL volumetric flasks. Wash the filter paper two or three times with water and adjust a volume.

13.5 Aspirate each filtered and acidified sample and determine its absorbance or concentration at 248.3 nm. Aspirate HNO₃ (1 + 499) between each sample.

14. Calculation

14.1 Calculate the concentration of iron in the sample, in milligrams per litre, referring to 12.4.

15. Precision and Bias ⁶

15.1 The precision of this test method for 10 laboratories, which include 16 operations within its designated range may be expressed as follows:

Reagent Water Type II:

$$S_T = 0.047 X + 0.053$$

$$S_o = 0.030 X + 0.037$$

Water of Choice:

$$S_T = 0.050 X + 0.114$$

$$S_o = 0.024 X + 0.078$$

where:

S_T = overall precision,

S_o = single-operator precision, and

X = determined concentration of iron, mg/L.

15.2 Recoveries of known amounts of iron in a series of prepared standards were as shown in Table 1.

15.3 The collaborative test data were obtained on reagent water; tap, lake, ground and surface water; unspecified wastewater; and a refinery primary treatment water. It is the user’s responsibility to ensure the validity of this test method for waters of untested matrices.

TEST METHOD C—ATOMIC ABSORPTION, GRAPHITE FURNACE

16. Scope

16.1 This test method covers the determination of dissolved and total recoverable iron in most waters and wastewaters.

16.2 This test method is applicable in the range from 5 to 100 µg/L of iron using a 20-µL injection. The range can be increased or decreased by varying the volume of sample injected or the instrumental settings. High concentrations may be diluted but preferably should be analyzed by direct aspiration atomic absorption spectrophotometry (Test Method A).

16.3 This test method has been used successfully with reagent grade water, filtered tap water, well water, demineralized water, boiler blowdown water, and condensate from a medium Btu-coal gasification process. It is the user’s responsibility to ensure validity of this test method to waters of untested matrices.

16.4 The analyst is encouraged to consult Practice D 3919 for a general discussion of interferences and sample analysis procedures for graphite furnace atomic absorption spectrophotometry.

17. Summary of Test Method

17.1 Iron is determined by an atomic absorption spectrophotometer used in conjunction with a graphite furnace.

TABLE 1 Determination of Bias, Atomic Absorption, Direct

Reagent Water Type II:			Bias, %	Statistically Significant (95 % Confidence Level)
Amount Added, mg/L	Amount Found, mg/L	Bias, mg/L		
0.2	0.2	±0.0	0.0	no
2.4	2.4	±0.0	0.0	no
4.4	4.3	-0.1	- 2.3	yes
Natural Water:			Bias, %	Statistically Significant (95 % Confidence Level)
Amount Added, mg/L	Amount Found, mg/L	Bias, mg/L		
0.2	0.2	±0.0	0	no
2.4	2.3	- 0.1	- 4.17	yes
4.4	4.2	- 0.2	- 4.55	yes

⁶ Supporting data are available from ASTM Headquarters. Request RR: D19 - 1035.

A sample is placed in a graphite tube, evaporated to dryness, charred (pyrolyzed or ashed), and atomized. The absorption signal generated during atomization is recorded and compared to standards. A general guide for the application of the graphite furnace is given in Practice D 3919.

17.2 Dissolved iron is determined on a filtered sample with no pretreatment.

17.3 Total recoverable iron is determined following acid digestion and filtration. Because chlorides interfere with furnace procedures for some metals, the use of hydrochloric acid in any digestion or solubilization step is to be avoided. If suspended material is not present, this digestion and filtration may be omitted.

18. Interferences

18.1 For a complete discussion on general interferences with furnace procedures, the analyst is referred to Practice D 3919.

19. Apparatus

19.1 *Atomic Absorption Spectrophotometer*, for use at 248.3 nm with background correction.

NOTE 8—A wavelength other than 248.3 nm may be used if it has been determined to be suitable. Greater linearity may be obtained at high concentrations by using a less sensitive wavelength.

NOTE 9—The manufacturer's instructions should be followed for all instrumental parameters.

19.2 *Iron Hollow-Cathode Lamp*—A single-element lamp is preferred, but multielement lamps may be used.

19.3 *Graphite Furnace*, capable of reaching temperatures sufficient to atomize the element of interest.

19.4 *Graphite Tubes*, compatible with furnace device. Pyrolytically coated graphite tubes are recommended to eliminate the possible formation of carbides.

19.5 *Pipets*, microlitre with disposable tips. Sizes may range from 1 to 100 μL , as required.

19.6 *Data Storage and Reduction Devices, Computer- and Microprocessor-Controlled Devices, or Strip Chart Recorders*, shall be utilized for collection, storage, reduction, and problem recognition (such as drift, incomplete atomization, changes in sensitivity, etc.). Strip chart recorders shall have a full scale deflection time of 0.2 s or less to ensure accuracy.

19.7 Automatic sampling should be used if available.

NOTE 10—Manual injection has been reported to cause widely scattered values even on purified waters due to contamination from pipetting technique.

20. Reagents and Materials

20.1 *Iron Solution, Stock* (1.0 mL = 1000 μg Fe)—See 11.4.

20.2 *Iron Solution, Intermediate* (1.0 mL = 10 μg Fe)—Dilute 10.0 mL of iron solution, stock (20.1) and 1 mL of HNO_3 (sp gr 1.42) to 1 L with water.

20.3 *Iron Solution, Standard* (1.0 mL = 0.2 μg Fe)—Dilute 20.0 mL of iron solution, intermediate (20.2) and 1 mL of HNO_3 (sp gr 1.42) to 1 L water. This standard is used to prepare working standards at the time of the analysis.

20.4 *Nitric Acid* (sp gr 1.42)—Concentrated nitric acid (HNO_3) (see Note 4).

20.5 *Argon, Standard*, welders grade, commercially available. Nitrogen may also be used if recommended by the instrument manufacturer.

21. Standardization

21.1 Initially, set the instrument according to the manufacturer's specifications. Follow the general instructions as provided in Practice D 3919.

22. Procedure

22.1 Clean all glassware to be used for preparation of standard solutions or in the solubilization step, or both, by rinsing first with HNO_3 (1 + 1) and then with water.

22.2 Measure 100.0 mL of each standard and well-mixed sample into 125-mL beakers or flasks. For total recoverable iron add HNO_3 (sp gr 1.42) to each standard and sample at a rate of 5 mL/L and proceed as directed in 22.4 through 22.6.

22.3 If only dissolved iron is to be determined, filter the sample through a 0.45- μm membrane filter prior to acidification and proceed to 22.6.

22.4 Heat the samples at 95°C on a steam bath or hotplate in a well-ventilated fume hood until the volume has been reduced to 15 to 20 mL, making certain that the samples do not boil (see Note 7).

22.5 Cool and filter the sample through a suitable filter (such as fine-textured, acid-washed, ashless paper) into a 100-mL volumetric flask. Wash the filter paper 2 or 3 times with water and bring to volume (Note 11). The acid concentration at this point should be 0.5 % HNO_3 .

NOTE 11—If suspended material is not present, this filtration may be omitted. The sample must be diluted to 100 mL.

22.6 Inject a measured aliquot of sample into the furnace device following the directions as provided by the particular instrument manufacturer. Refer to Practice D 3919.

23. Calculation

23.1 Determine the concentration of iron in each sample by referring to Practice D 3919.

24. Precision and Bias ⁷

24.1 The precision for this test method was developed by 13 laboratories using reagent water and 7 laboratories using tap water, filtered tap water, well water, demineralized water, boiler blowdown water, and condensate from a medium Btu coal gasification process. Although multiple injections may have been made, the report sheets provided allowed only for reporting single values. Thus, no single-operator precision data can be calculated. See Table 2 for bias data and overall precision data.

24.2 These data may not apply to waters of other matrices, therefore, it is the responsibility of the analyst to ensure the validity of this test method in a particular matrix.

⁷ Supporting data are available from ASTM Headquarters. Request RR:D19-1102.